

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q1858 SAS No.: Q1858 SDG NO.: Q1858

EPA Sample No.: SSTDCCC040 Date Analyzed: 04/24/2025

Lab File ID: BP024407.D Time Analyzed: 11:50

Instrument ID: BNA_P GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	264641	7.716	1170080	10.49	779721	14.35
UPPER LIMIT	529282	8.216	2340160	10.987	1559440	14.845
LOWER LIMIT	132321	7.216	585040	9.987	389861	13.845
EPA SAMPLE NO.						
01 PB167711BL	288156	7.72	1116360	10.49	654539	14.34
02 PB167711BS	253522	7.72	1075220	10.49	697166	14.35

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.