

LAB CHRONICLE

OrderID:	Q1858	OrderDate:	4/22/2025 2:54:00 PM
Client:	Kleinfelder	Project:	Henry Lea School
Contact:	Mark Warchol	Location:	L41,VOA Ref. #2 Soil

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q1858-01	COMP-1	SOIL	VOCMS Group1	8260D	04/21/25		04/23/25	04/22/25
Q1858-02	COMP-2	SOIL	VOCMS Group1	8260D	04/21/25		04/23/25	04/22/25
Q1858-03	COMP-3	SOIL	VOCMS Group1	8260D	04/21/25		04/23/25	04/22/25



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Hit Summary Sheet
SW-846

SDG No.: Q1858
Client: Kleinfelder

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
Client ID:				0				
Total Voc :								
Total Concentration:								



SAMPLE DATA

Report of Analysis

Client:	Kleinfelder	Date Collected:	04/21/25
Project:	Henry Lea School	Date Received:	04/22/25
Client Sample ID:	COMP-1	SDG No.:	Q1858
Lab Sample ID:	Q1858-01	Matrix:	SOIL
Analytical Method:	SW8260	% Solid:	83.9
Sample Wt/Vol:	4.23 Units: g	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOCMS Group1
GC Column:	RXI-624 ID : 0.25	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VY021978.D	1		04/23/25 17:47	VY042325

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
156-59-2	cis-1,2-Dichloroethene	1.10	U	1.10	7.00	ug/Kg
71-55-6	1,1,1-Trichloroethane	1.30	U	1.30	7.00	ug/Kg
71-43-2	Benzene	1.10	U	1.10	7.00	ug/Kg
79-01-6	Trichloroethene	1.10	U	1.10	7.00	ug/Kg
108-88-3	Toluene	1.10	U	1.10	7.00	ug/Kg
100-41-4	Ethyl Benzene	0.94	U	0.94	7.00	ug/Kg
1330-20-7	Total Xylenes	2.90	U	2.90	21.1	ug/Kg
98-82-8	Isopropylbenzene	1.10	U	1.10	7.00	ug/Kg
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	55.5		63 - 155	111%	SPK: 50
1868-53-7	Dibromofluoromethane	50.3		70 - 134	101%	SPK: 50
2037-26-5	Toluene-d8	48.6		74 - 123	97%	SPK: 50
460-00-4	4-Bromofluorobenzene	42.0		38 - 136	84%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	320000	7.707			
540-36-3	1,4-Difluorobenzene	605000	8.616			
3114-55-4	Chlorobenzene-d5	542000	11.414			
3855-82-1	1,4-Dichlorobenzene-d4	216000	13.346			

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	Kleinfelder	Date Collected:	04/21/25
Project:	Henry Lea School	Date Received:	04/22/25
Client Sample ID:	COMP-2	SDG No.:	Q1858
Lab Sample ID:	Q1858-02	Matrix:	SOIL
Analytical Method:	SW8260	% Solid:	81.3
Sample Wt/Vol:	4.53 Units: g	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOCMS Group1
GC Column:	RXI-624 ID : 0.25	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VY021979.D	1		04/23/25 18:11	VY042325

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
156-59-2	cis-1,2-Dichloroethene	1.00	U	1.00	6.80	ug/Kg
71-55-6	1,1,1-Trichloroethane	1.30	U	1.30	6.80	ug/Kg
71-43-2	Benzene	1.10	U	1.10	6.80	ug/Kg
79-01-6	Trichloroethene	1.10	U	1.10	6.80	ug/Kg
108-88-3	Toluene	1.10	U	1.10	6.80	ug/Kg
100-41-4	Ethyl Benzene	0.91	U	0.91	6.80	ug/Kg
1330-20-7	Total Xylenes	2.80	U	2.80	20.4	ug/Kg
98-82-8	Isopropylbenzene	1.10	U	1.10	6.80	ug/Kg
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	51.8		63 - 155	104%	SPK: 50
1868-53-7	Dibromofluoromethane	49.8		70 - 134	100%	SPK: 50
2037-26-5	Toluene-d8	48.0		74 - 123	96%	SPK: 50
460-00-4	4-Bromofluorobenzene	39.1		38 - 136	78%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	318000	7.707			
540-36-3	1,4-Difluorobenzene	596000	8.616			
3114-55-4	Chlorobenzene-d5	515000	11.414			
3855-82-1	1,4-Dichlorobenzene-d4	192000	13.346			

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	Kleinfelder	Date Collected:	04/21/25
Project:	Henry Lea School	Date Received:	04/22/25
Client Sample ID:	COMP-3	SDG No.:	Q1858
Lab Sample ID:	Q1858-03	Matrix:	SOIL
Analytical Method:	SW8260	% Solid:	91.1
Sample Wt/Vol:	3.86 Units: g	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOCMS Group1
GC Column:	RXI-624 ID : 0.25	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VY021980.D	1		04/23/25 18:34	VY042325

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
156-59-2	cis-1,2-Dichloroethene	1.10	U	1.10	7.10	ug/Kg
71-55-6	1,1,1-Trichloroethane	1.30	U	1.30	7.10	ug/Kg
71-43-2	Benzene	1.10	U	1.10	7.10	ug/Kg
79-01-6	Trichloroethene	1.20	U	1.20	7.10	ug/Kg
108-88-3	Toluene	1.10	U	1.10	7.10	ug/Kg
100-41-4	Ethyl Benzene	0.95	U	0.95	7.10	ug/Kg
1330-20-7	Total Xylenes	3.00	U	3.00	21.3	ug/Kg
98-82-8	Isopropylbenzene	1.10	U	1.10	7.10	ug/Kg
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	52.9		63 - 155	106%	SPK: 50
1868-53-7	Dibromofluoromethane	50.7		70 - 134	101%	SPK: 50
2037-26-5	Toluene-d8	48.1		74 - 123	96%	SPK: 50
460-00-4	4-Bromofluorobenzene	39.5		38 - 136	79%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	307000	7.707			
540-36-3	1,4-Difluorobenzene	573000	8.616			
3114-55-4	Chlorobenzene-d5	503000	11.414			
3855-82-1	1,4-Dichlorobenzene-d4	190000	13.346			

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products



QC SUMMARY

Surrogate Summary

SDG No.: **Q1858**

Client: **Kleinfelder**

Analytical Method: **SW8260D**

Lab Sample ID	Client ID	Parameter	Spike	Result	RecoveryQual	Limits	
						Low	High
Q1858-01	COMP-1	1,2-Dichloroethane-d4	50	55.5	111	63	155
		Dibromofluoromethane	50	50.3	101	70	134
		Toluene-d8	50	48.6	97	74	123
		4-Bromofluorobenzene	50	42.0	84	38	136
Q1858-02	COMP-2	1,2-Dichloroethane-d4	50	51.8	104	63	155
		Dibromofluoromethane	50	49.8	100	70	134
		Toluene-d8	50	48.0	96	74	123
		4-Bromofluorobenzene	50	39.1	78	38	136
Q1858-03	COMP-3	1,2-Dichloroethane-d4	50	52.9	106	63	155
		Dibromofluoromethane	50	50.7	101	70	134
		Toluene-d8	50	48.1	96	74	123
		4-Bromofluorobenzene	50	39.5	79	38	136
VY0423SBL01	VY0423SBL01	1,2-Dichloroethane-d4	50	51.1	102	63	155
		Dibromofluoromethane	50	50.0	100	70	134
		Toluene-d8	50	47.6	95	74	123
		4-Bromofluorobenzene	50	38.5	77	38	136
VY0423SBS01	VY0423SBS01	1,2-Dichloroethane-d4	50	51.4	103	63	155
		Dibromofluoromethane	50	51.5	103	70	134
		Toluene-d8	50	52.0	104	74	123
		4-Bromofluorobenzene	50	50.6	101	38	136
VY0423SBSD01	VY0423SBSD01	1,2-Dichloroethane-d4	50	57.7	115	63	155
		Dibromofluoromethane	50	55.6	111	70	134
		Toluene-d8	50	56.5	113	74	123
		4-Bromofluorobenzene	50	55.5	111	38	136

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1858

Client: Kleinfelder

Analytical Method: SW8260D

Datafile : VY021964.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits	RPD
									High	
VY0423SBS01	cis-1,2-Dichloroethene	20	20.3	ug/Kg	102			82	123	
	1,1,1-Trichloroethane	20	20.2	ug/Kg	101			80	126	
	Benzene	20	20.7	ug/Kg	104			84	121	
	Trichloroethene	20	20.2	ug/Kg	101			83	122	
	Toluene	20	20.8	ug/Kg	104			83	122	
	Ethyl Benzene	20	19.8	ug/Kg	99			82	124	
	m/p-Xylenes	40	40.3	ug/Kg	101			83	124	
	o-Xylene	20	19.8	ug/Kg	99			83	123	
	Isopropylbenzene	20	19.6	ug/Kg	98			82	124	

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1858

Client: Kleinfelder

Analytical Method: SW8260D

Datafile : VY021965.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits	RPD
									High	
VY0423SBSD01	cis-1,2-Dichloroethene	20	21.4	ug/Kg	107	5		82	123	20
	1,1,1-Trichloroethane	20	21.2	ug/Kg	106	5		80	126	20
	Benzene	20	21.4	ug/Kg	107	3		84	121	20
	Trichloroethene	20	21.0	ug/Kg	105	4		83	122	20
	Toluene	20	21.6	ug/Kg	108	4		83	122	20
	Ethyl Benzene	20	20.3	ug/Kg	102	3		82	124	20
	m/p-Xylenes	40	41.9	ug/Kg	105	4		83	124	20
	o-Xylene	20	20.6	ug/Kg	103	4		83	123	20
	Isopropylbenzene	20	20.2	ug/Kg	101	3		82	124	20

VOLATILE METHOD BLANK SUMMARY

EPA SAMPLE NO.

VY0423SBL01

Lab Name: CHEMTECH

Contract: POWE02

Lab Code: CHEM Case No.: Q1858

SAS No.: Q1858 SDG NO.: Q1858

Lab File ID: VY021963.D

Lab Sample ID: VY0423SBL01

Date Analyzed: 04/23/2025

Time Analyzed: 10:31

GC Column: RXI-624 ID: 0.25 (mm)

Heated Purge: (Y/N) Y

Instrument ID: MSVOA_Y

THIS METHOD BLANK APPLIES TO THE FOLLOWING SAMPLES, MS AND MSD:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED
VY0423SBS01	VY0423SBS01	VY021964.D	04/23/2025
VY0423SBSD01	VY0423SBSD01	VY021965.D	04/23/2025
COMP-1	Q1858-01	VY021978.D	04/23/2025
COMP-2	Q1858-02	VY021979.D	04/23/2025
COMP-3	Q1858-03	VY021980.D	04/23/2025

COMMENTS: _____

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: CHEMTECH Contract: POWE02
Lab Code: CHEM Case No.: Q1858 SAS No.: Q1858 SDG NO.: Q1858
Lab File ID: VY021952.D BFB Injection Date: 04/22/2025
Instrument ID: MSVOA_Y BFB Injection Time: 11:33
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: Y/N Y

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	16.9
75	30.0 - 60.0% of mass 95	49.7
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	6.5
173	Less than 2.0% of mass 174	1.5 (1.7) 1
174	50.0 - 100.0% of mass 95	88.4
175	5.0 - 9.0% of mass 174	7.1 (8) 1
176	95.0 - 101.0% of mass 174	84.9 (96.1) 1
177	5.0 - 9.0% of mass 176	5.5 (6.4) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDIC005	VSTDIC005	VY021953.D	04/22/2025	13:39
VSTDIC010	VSTDIC010	VY021954.D	04/22/2025	14:44
VSTDIC020	VSTDIC020	VY021955.D	04/22/2025	15:07
VSTDIC050	VSTDIC050	VY021956.D	04/22/2025	15:29
VSTDIC100	VSTDIC100	VY021957.D	04/22/2025	15:52
VSTDIC150	VSTDIC150	VY021958.D	04/22/2025	16:15

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: CHEMTECH Contract: POWE02
Lab Code: CHEM Case No.: Q1858 SAS No.: Q1858 SDG NO.: Q1858
Lab File ID: VY021961.D BFB Injection Date: 04/23/2025
Instrument ID: MSVOA_Y BFB Injection Time: 09:27
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: Y/N Y

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	15.6
75	30.0 - 60.0% of mass 95	48.7
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	6.9
173	Less than 2.0% of mass 174	1.4 (1.7) 1
174	50.0 - 100.0% of mass 95	83.9
175	5.0 - 9.0% of mass 174	6.8 (8.1) 1
176	95.0 - 101.0% of mass 174	83 (99) 1
177	5.0 - 9.0% of mass 176	5.3 (6.4) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDCCC050	VSTDCCC050	VY021962.D	04/23/2025	09:57
VY0423SBL01	VY0423SBL01	VY021963.D	04/23/2025	10:31
VY0423SBS01	VY0423SBS01	VY021964.D	04/23/2025	11:07
VY0423SBSD01	VY0423SBSD01	VY021965.D	04/23/2025	11:30
COMP-1	Q1858-01	VY021978.D	04/23/2025	17:47
COMP-2	Q1858-02	VY021979.D	04/23/2025	18:11
COMP-3	Q1858-03	VY021980.D	04/23/2025	18:34

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: POWE02
Lab Code: CHEM Case No.: Q1858 SAS No.: Q1858 SDG NO.: Q1858
Lab File ID: VY021962.D Date Analyzed: 04/23/2025
Instrument ID: MSVOA_Y Time Analyzed: 09:57
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	304636	7.71	465219	8.62	430304	11.42
UPPER LIMIT	609272	8.213	930438	9.116	860608	11.92
LOWER LIMIT	152318	7.213	232610	8.116	215152	10.92
EPA SAMPLE NO.						
COMP-1	320456	7.71	605358	8.62	542247	11.41
COMP-2	317972	7.71	595656	8.62	514825	11.41
COMP-3	306739	7.71	573313	8.62	503025	11.41
VY0423SBL01	319981	7.71	599033	8.62	525205	11.42
VY0423SBS01	295173	7.71	458305	8.62	417679	11.42
VY0423SBSD01	282902	7.71	449051	8.62	413807	11.41

IS1 = Pentafluorobenzene
IS2 = 1,4-Difluorobenzene
IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: POWE02
 Lab Code: CHEM Case No.: Q1858 SAS No.: Q1858 SDG NO.: Q1858
 Lab File ID: VY021962.D Date Analyzed: 04/23/2025
 Instrument ID: MSVOA_Y Time Analyzed: 09:57
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS4 AREA #	RT #				
12 HOUR STD	227723	13.347				
UPPER LIMIT	455446	13.847				
LOWER LIMIT	113862	12.847				
EPA SAMPLE NO.						
COMP-1	216405	13.35				
COMP-2	191838	13.35				
COMP-3	189503	13.35				
VY0423SBL01	194434	13.35				
VY0423SBS01	221603	13.35				
VY0423SBSD01	217306	13.35				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.



QC SAMPLE DATA

Report of Analysis

Client:	Kleinfelder	Date Collected:	
Project:	Henry Lea School	Date Received:	
Client Sample ID:	VY0423SBL01	SDG No.:	Q1858
Lab Sample ID:	VY0423SBL01	Matrix:	SOIL
Analytical Method:	SW8260	% Solid:	100
Sample Wt/Vol:	5 Units: g	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOCMS Group1
GC Column:	RXI-624 ID : 0.25	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VY021963.D	1		04/23/25 10:31	VY042325

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
156-59-2	cis-1,2-Dichloroethene	0.75	U	0.75	5.00	ug/Kg
71-55-6	1,1,1-Trichloroethane	0.93	U	0.93	5.00	ug/Kg
71-43-2	Benzene	0.79	U	0.79	5.00	ug/Kg
79-01-6	Trichloroethene	0.81	U	0.81	5.00	ug/Kg
108-88-3	Toluene	0.78	U	0.78	5.00	ug/Kg
100-41-4	Ethyl Benzene	0.67	U	0.67	5.00	ug/Kg
1330-20-7	Total Xylenes	2.02	U	2.02	15.0	ug/Kg
98-82-8	Isopropylbenzene	0.78	U	0.78	5.00	ug/Kg
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	51.1		63 - 155	102%	SPK: 50
1868-53-7	Dibromofluoromethane	50.0		70 - 134	100%	SPK: 50
2037-26-5	Toluene-d8	47.6		74 - 123	95%	SPK: 50
460-00-4	4-Bromofluorobenzene	38.6		38 - 136	77%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	320000	7.713			
540-36-3	1,4-Difluorobenzene	599000	8.616			
3114-55-4	Chlorobenzene-d5	525000	11.42			
3855-82-1	1,4-Dichlorobenzene-d4	194000	13.347			

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	Kleinfelder	Date Collected:	
Project:	Henry Lea School	Date Received:	
Client Sample ID:	VY0423SBS01	SDG No.:	Q1858
Lab Sample ID:	VY0423SBS01	Matrix:	SOIL
Analytical Method:	SW8260	% Solid:	100
Sample Wt/Vol:	5 Units: g	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOCMS Group1
GC Column:	RXI-624 ID : 0.25	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VY021964.D	1		04/23/25 11:07	VY042325

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
156-59-2	cis-1,2-Dichloroethene	20.3		0.75	5.00	ug/Kg
71-55-6	1,1,1-Trichloroethane	20.2		0.93	5.00	ug/Kg
71-43-2	Benzene	20.7		0.79	5.00	ug/Kg
79-01-6	Trichloroethene	20.2		0.81	5.00	ug/Kg
108-88-3	Toluene	20.8		0.78	5.00	ug/Kg
100-41-4	Ethyl Benzene	19.8		0.67	5.00	ug/Kg
1330-20-7	Total Xylenes	60.1		2.02	15.0	ug/Kg
98-82-8	Isopropylbenzene	19.6		0.78	5.00	ug/Kg
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	51.4		63 - 155	103%	SPK: 50
1868-53-7	Dibromofluoromethane	51.6		70 - 134	103%	SPK: 50
2037-26-5	Toluene-d8	52.0		74 - 123	104%	SPK: 50
460-00-4	4-Bromofluorobenzene	50.6		38 - 136	101%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	295000	7.707			
540-36-3	1,4-Difluorobenzene	458000	8.616			
3114-55-4	Chlorobenzene-d5	418000	11.42			
3855-82-1	1,4-Dichlorobenzene-d4	222000	13.346			

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	Kleinfelder	Date Collected:	
Project:	Henry Lea School	Date Received:	
Client Sample ID:	VY0423SBSD01	SDG No.:	Q1858
Lab Sample ID:	VY0423SBSD01	Matrix:	SOIL
Analytical Method:	SW8260	% Solid:	100
Sample Wt/Vol:	5 Units: g	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOCMS Group1
GC Column:	RXI-624 ID : 0.25	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VY021965.D	1		04/23/25 11:30	VY042325

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
156-59-2	cis-1,2-Dichloroethene	21.4		0.75	5.00	ug/Kg
71-55-6	1,1,1-Trichloroethane	21.2		0.93	5.00	ug/Kg
71-43-2	Benzene	21.4		0.79	5.00	ug/Kg
79-01-6	Trichloroethene	21.0		0.81	5.00	ug/Kg
108-88-3	Toluene	21.6		0.78	5.00	ug/Kg
100-41-4	Ethyl Benzene	20.3		0.67	5.00	ug/Kg
1330-20-7	Total Xylenes	62.5		2.02	15.0	ug/Kg
98-82-8	Isopropylbenzene	20.2		0.78	5.00	ug/Kg
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	57.7		63 - 155	115%	SPK: 50
1868-53-7	Dibromofluoromethane	55.6		70 - 134	111%	SPK: 50
2037-26-5	Toluene-d8	56.5		74 - 123	113%	SPK: 50
460-00-4	4-Bromofluorobenzene	55.5		38 - 136	111%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	283000	7.707			
540-36-3	1,4-Difluorobenzene	449000	8.616			
3114-55-4	Chlorobenzene-d5	414000	11.414			
3855-82-1	1,4-Dichlorobenzene-d4	217000	13.346			

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products



CALIBRATION SUMMARY

VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name:	<u>CHEMTECH</u>	Contract:	<u>POWE02</u>
Lab Code:	<u>CHEM</u>	Case No.:	<u>Q1858</u>
Instrument ID:	<u>MSVOA_Y</u>	SAS No.:	<u>Q1858</u>
Heated Purge: (Y/N)	<u>Y</u>	SDG No.:	<u>Q1858</u>
GC Column:	<u>RXI-624</u>	Calibration Date(s):	<u>04/22/2025</u>
ID:	<u>0.25 (mm)</u>	Calibration Time(s):	<u>13:39</u>
			<u>16:15</u>

LAB FILE ID:		RRF005 = VY021953.D		RRF010 = VY021954.D		RRF020 = VY021955.D		
		RRF050 = VY021956.D		RRF100 = VY021957.D		RRF150 = VY021958.D		
COMPOUND	RRF005	RRF010	RRF020	RRF050	RRF100	RRF150	RRF	% RSD
cis-1,2-Dichloroethene	0.626	0.638	0.604	0.612	0.633	0.621	0.622	2.1
1,1,1-Trichloroethane	0.972	0.944	0.896	0.853	0.865	0.846	0.896	5.8
Benzene	1.423	1.439	1.351	1.337	1.415	1.358	1.387	3.1
Trichloroethene	0.402	0.405	0.369	0.367	0.387	0.375	0.384	4.3
Toluene	0.838	0.915	0.875	0.894	0.952	0.913	0.898	4.4
Ethyl Benzene	1.693	1.840	1.718	1.835	1.990	1.898	1.829	6.1
m/p-Xylenes	0.664	0.720	0.699	0.734	0.797	0.754	0.728	6.3
o-Xylene	0.599	0.639	0.635	0.689	0.747	0.716	0.671	8.3
Isopropylbenzene	3.061	3.316	3.102	3.325	3.639	3.599	3.341	7.2
1,2-Dichloroethane-d4	0.525	0.477	0.459	0.466	0.418	0.427	0.462	8.3
Dibromofluoromethane	0.335	0.341	0.330	0.330	0.319	0.327	0.330	2.3
Toluene-d8	1.220	1.260	1.211	1.276	1.244	1.261	1.245	2
4-Bromofluorobenzene	0.408	0.429	0.401	0.426	0.423	0.428	0.419	2.8

* Compounds with required minimum RRF and maximum %RSD values.
All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: POWE02

Lab Code: CHEM Case No.: Q1858 SAS No.: Q1858 SDG No.: Q1858

Instrument ID: MSVOA_Y Calibration Date/Time: 04/23/2025 09:57

Lab File ID: VY021962.D Init. Calib. Date(s): 04/22/2025 04/22/2025

Heated Purge: (Y/N) Y Init. Calib. Time(s): 13:39 16:15

GC Column: RXI-624 ID: 0.25 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
cis-1,2-Dichloroethene	0.622	0.605		-2.73	20
1,1,1-Trichloroethane	0.896	0.858		-4.24	20
Benzene	1.387	1.381		-0.43	20
Trichloroethene	0.384	0.371		-3.38	20
Toluene	0.898	0.927		3.23	20
Ethyl Benzene	1.829	1.881		2.84	20
m/p-Xylenes	0.728	0.756		3.85	20
o-Xylene	0.671	0.693		3.28	20
Isopropylbenzene	3.341	3.454		3.38	20
1,2-Dichloroethane-d4	0.462	0.438		-5.2	20
Dibromofluoromethane	0.330	0.325		-1.51	20
Toluene-d8	1.245	1.266		1.69	20
4-Bromofluorobenzene	0.419	0.422		0.72	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.