

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: POWE02
 Lab Code: CHEM Case No.: Q1858 SAS No.: Q1858 SDG NO.: Q1858
 Lab File ID: VY021962.D Date Analyzed: 04/23/2025
 Instrument ID: MSVOA_Y Time Analyzed: 09:57
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	304636	7.71	465219	8.62	430304	11.42
UPPER LIMIT	609272	8.213	930438	9.116	860608	11.92
LOWER LIMIT	152318	7.213	232610	8.116	215152	10.92
EPA SAMPLE NO.						
COMP-1	320456	7.71	605358	8.62	542247	11.41
COMP-2	317972	7.71	595656	8.62	514825	11.41
COMP-3	306739	7.71	573313	8.62	503025	11.41
VY0423SBL01	319981	7.71	599033	8.62	525205	11.42
VY0423SBS01	295173	7.71	458305	8.62	417679	11.42
VY0423SBSD01	282902	7.71	449051	8.62	413807	11.41

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.