

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: CHEMTECH Contract: POWE02
Lab Code: CHEM Case No.: Q1858 SAS No.: Q1858 SDG NO.: Q1858
Lab File ID: VY021961.D BFB Injection Date: 04/23/2025
Instrument ID: MSVOA_Y BFB Injection Time: 09:27
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: Y/N Y

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	15.6
75	30.0 - 60.0% of mass 95	48.7
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	6.9
173	Less than 2.0% of mass 174	1.4 (1.7) 1
174	50.0 - 100.0% of mass 95	83.9
175	5.0 - 9.0% of mass 174	6.8 (8.1) 1
176	95.0 - 101.0% of mass 174	83 (99) 1
177	5.0 - 9.0% of mass 176	5.3 (6.4) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDCCC050	VSTDCCC050	VY021962.D	04/23/2025	09:57
VY0423SBL01	VY0423SBL01	VY021963.D	04/23/2025	10:31
VY0423SBS01	VY0423SBS01	VY021964.D	04/23/2025	11:07
VY0423SBSD01	VY0423SBSD01	VY021965.D	04/23/2025	11:30
COMP-1	Q1858-01	VY021978.D	04/23/2025	17:47
COMP-2	Q1858-02	VY021979.D	04/23/2025	18:11
COMP-3	Q1858-03	VY021980.D	04/23/2025	18:34