

Initial and Continuing Calibration Verification

Client: Kleinfelder

SDG No.: Q1859

Project: Lincoln High School

RunNo.: LB135527

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV1						
Bromide	mg/L	10.2	10	102	90-110	04/22/2025
Chloride	mg/L	3.1	3	103	90-110	04/22/2025
Fluoride	mg/L	2	2	100	90-110	04/22/2025
Nitrite	mg/L	3.1	3	103	90-110	04/22/2025
Nitrate	mg/L	2.6	2.5	104	90-110	04/22/2025
Sulfate	mg/L	15.1	15	101	90-110	04/22/2025
Orthophosphate as P	mg/L	5.2	5	104	90-110	04/22/2025
Sample ID: CCV1						
Bromide	mg/L	10.7	10	107	90-110	04/23/2025
Chloride	mg/L	3.2	3	107	90-110	04/23/2025
Fluoride	mg/L	2.1	2	105	90-110	04/23/2025
Nitrite	mg/L	3.2	3	107	90-110	04/23/2025
Nitrate	mg/L	2.7	2.5	108	90-110	04/23/2025
Sulfate	mg/L	15.6	15	104	90-110	04/23/2025
Orthophosphate as P	mg/L	5.4	5	108	90-110	04/23/2025
Sample ID: CCV2						
Bromide	mg/L	10.6	10	106	90-110	04/23/2025
Chloride	mg/L	3.2	3	107	90-110	04/23/2025
Fluoride	mg/L	2.1	2	105	90-110	04/23/2025
Nitrite	mg/L	3.2	3	107	90-110	04/23/2025
Nitrate	mg/L	2.7	2.5	108	90-110	04/23/2025
Sulfate	mg/L	15.5	15	103	90-110	04/23/2025
Orthophosphate as P	mg/L	5.3	5	106	90-110	04/23/2025

Initial and Continuing Calibration Verification

Client: Kleinfelder

SDG No.: Q1859

Project: Lincoln High School

RunNo.: LB135535

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV Hexavalent Chromium	mg/L	0.494	0.5	99	90-110	04/23/2025
Sample ID: CCV1 Hexavalent Chromium	mg/L	0.501	0.5	100	90-110	04/23/2025
Sample ID: CCV2 Hexavalent Chromium	mg/L	0.496	0.5	99	90-110	04/23/2025
Sample ID: CCV3 Hexavalent Chromium	mg/L	0.502	0.5	100	90-110	04/23/2025

Initial and Continuing Calibration Verification

Client: Kleinfelder

SDG No.: Q1859

Project: Lincoln High School

RunNo.: LB135559

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV1 Ammonia as N	mg/L	0.94	1	94	90-110	04/25/2025
Sample ID: CCV1 Ammonia as N	mg/L	0.95	1	95	90-110	04/25/2025
Sample ID: CCV2 Ammonia as N	mg/L	0.99	1	99	90-110	04/25/2025
Sample ID: CCV3 Ammonia as N	mg/L	0.93	1	93	90-110	04/25/2025
Sample ID: CCV4 Ammonia as N	mg/L	0.99	1	99	90-110	04/25/2025

Initial and Continuing Calibration Verification

Client: Kleinfelder

SDG No.: Q1859

Project: Lincoln High School

RunNo.: LB135559

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
---------	-------	--------	------------	---------------	---------------------------	------------------