

Laboratory Control Sample Summary

Client: Kleinfelder

SDG No.: Q1859

Project: Lincoln High School

Run No.: LB135527

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB135527BSS							
Bromide	mg/Kg	200	212		106	1	90-110	04/23/2025
Chloride	mg/Kg	60	64.0		107	1	90-110	04/23/2025
Fluoride	mg/Kg	40	41.8		104	1	90-110	04/23/2025
Nitrite	mg/Kg	60	63.2		105	1	90-110	04/23/2025
Nitrate	mg/Kg	50	53.9		108	1	90-110	04/23/2025
Sulfate	mg/Kg	300	309		103	1	90-110	04/23/2025
Orthophosphate as P	mg/Kg	100	106		106	1	90-110	04/23/2025

Laboratory Control Sample Summary

Client: Kleinfelder

SDG No.: Q1859

Project: Lincoln High School

Run No.: LB135535

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	PB167705BS							
Hexavalent Chromium	mg/Kg	20	20.2		101	1	84-110	04/23/2025

Laboratory Control Sample Summary

Client: Kleinfelder

SDG No.: Q1859

Project: Lincoln High School

Run No.: LB135559

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	PB167731BS							
Ammonia as N	mg/Kg	50	47.1		94	1	90-110	04/25/2025