



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: Kleinfelder SDG No.: Q1859
Contract: POWE02 Lab Code: CHEM Case No.: Q1859 SAS No.: Q1859
Initial Calibration Source: _____
Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRI01	Aluminum	95.1	100	95	40 - 160	P	04/23/2025	15:53	LB135538
	Antimony	51.3	50.0	102	40 - 160	P	04/23/2025	15:53	LB135538
	Arsenic	20.4	20.0	102	40 - 160	P	04/23/2025	15:53	LB135538
	Barium	88.7	100	89	40 - 160	P	04/23/2025	15:53	LB135538
	Beryllium	5.75	6.0	96	40 - 160	P	04/23/2025	15:53	LB135538
	Boron	96.5	100	96	40 - 160	P	04/23/2025	15:53	LB135538
	Cadmium	6.07	6.0	101	40 - 160	P	04/23/2025	15:53	LB135538
	Chromium	9.31	10.0	93	40 - 160	P	04/23/2025	15:53	LB135538
	Cobalt	28.9	30.0	96	40 - 160	P	04/23/2025	15:53	LB135538
	Copper	22.7	20.0	113	40 - 160	P	04/23/2025	15:53	LB135538
	Iron	92.0	100	92	40 - 160	P	04/23/2025	15:53	LB135538
	Lead	10.1	12.0	84	40 - 160	P	04/23/2025	15:53	LB135538
	Manganese	20.6	20.0	103	40 - 160	P	04/23/2025	15:53	LB135538
	Molybdenum	207	200	103	40 - 160	P	04/23/2025	15:53	LB135538
	Nickel	39.4	40.0	98	40 - 160	P	04/23/2025	15:53	LB135538
	Selenium	18.2	20.0	91	40 - 160	P	04/23/2025	15:53	LB135538
	Silver	9.37	10.0	94	40 - 160	P	04/23/2025	15:53	LB135538
	Thallium	41.5	40.0	104	40 - 160	P	04/23/2025	15:53	LB135538
	Vanadium	39.3	40.0	98	40 - 160	P	04/23/2025	15:53	LB135538
	Zinc	42.0	40.0	105	40 - 160	P	04/23/2025	15:53	LB135538
CRA	Mercury	0.24	0.2	118	40 - 160	CV	04/24/2025	08:12	LB135540