



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client:	<u>Kleinfelder</u>				SDG No.:	<u>Q1859</u>				
Contract:	<u>POWE02</u>	Lab Code:	<u>CHEM</u>	Case No.:	<u>Q1859</u>	SAS No.:	<u>Q1859</u>			
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number	
ICB99	Mercury	0.20	+/-0.20	U	0.20	CV	04/24/2025	08:05	LB135540	

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Kleinfelder

SDG No.: Q1859

Contract: POWE02

Lab Code: CHEM

Case No.: Q1859

SAS No.: Q1859

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB17	Mercury	0.20	+/-0.20	U	0.20	CV	04/24/2025	08:10	LB135540
CCB18	Mercury	0.20	+/-0.20	U	0.20	CV	04/24/2025	08:43	LB135540
CCB19	Mercury	0.20	+/-0.20	U	0.20	CV	04/24/2025	09:10	LB135540
CCB20	Mercury	0.20	+/-0.20	U	0.20	CV	04/24/2025	09:22	LB135540

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: <u>Kleinfelder</u>		SDG No.: <u>Q1859</u>							
Contract: <u>POWE02</u>	Lab Code: <u>CHEM</u>	Case No.: <u>Q1859</u>	SAS No.: <u>Q1859</u>						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Aluminum	100	+/-100	U	100	P	04/23/2025	15:49	LB135538
	Antimony	50.0	+/-50.0	U	50.0	P	04/23/2025	15:49	LB135538
	Arsenic	20.0	+/-20.0	U	20.0	P	04/23/2025	15:49	LB135538
	Barium	100	+/-100	U	100	P	04/23/2025	15:49	LB135538
	Beryllium	6.00	+/-6.00	U	6.00	P	04/23/2025	15:49	LB135538
	Boron	100	+/-100	U	100	P	04/23/2025	15:49	LB135538
	Cadmium	6.00	+/-6.00	U	6.00	P	04/23/2025	15:49	LB135538
	Chromium	10.0	+/-10.0	U	10.0	P	04/23/2025	15:49	LB135538
	Cobalt	30.0	+/-30.0	U	30.0	P	04/23/2025	15:49	LB135538
	Copper	20.0	+/-20.0	U	20.0	P	04/23/2025	15:49	LB135538
	Iron	100	+/-100	U	100	P	04/23/2025	15:49	LB135538
	Lead	12.0	+/-12.0	U	12.0	P	04/23/2025	15:49	LB135538
	Manganese	20.0	+/-20.0	U	20.0	P	04/23/2025	15:49	LB135538
	Molybdenum	200	+/-200	U	200	P	04/23/2025	15:49	LB135538
	Nickel	40.0	+/-40.0	U	40.0	P	04/23/2025	15:49	LB135538
	Selenium	20.0	+/-20.0	U	20.0	P	04/23/2025	15:49	LB135538
	Silver	10.0	+/-10.0	U	10.0	P	04/23/2025	15:49	LB135538
	Thallium	40.0	+/-40.0	U	40.0	P	04/23/2025	15:49	LB135538
	Vanadium	40.0	+/-40.0	U	40.0	P	04/23/2025	15:49	LB135538
	Zinc	40.0	+/-40.0	U	40.0	P	04/23/2025	15:49	LB135538

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Kleinfelder SDG No.: Q1859
Contract: POWE02 Lab Code: CHEM Case No.: Q1859 SAS No.: Q1859

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Aluminum	100	+/-100	U	100	P	04/23/2025	16:19	LB135538
	Antimony	50.0	+/-50.0	U	50.0	P	04/23/2025	16:19	LB135538
	Arsenic	20.0	+/-20.0	U	20.0	P	04/23/2025	16:19	LB135538
	Barium	100	+/-100	U	100	P	04/23/2025	16:19	LB135538
	Beryllium	6.00	+/-6.00	U	6.00	P	04/23/2025	16:19	LB135538
	Boron	100	+/-100	U	100	P	04/23/2025	16:19	LB135538
	Cadmium	6.00	+/-6.00	U	6.00	P	04/23/2025	16:19	LB135538
	Chromium	10.0	+/-10.0	U	10.0	P	04/23/2025	16:19	LB135538
	Cobalt	30.0	+/-30.0	U	30.0	P	04/23/2025	16:19	LB135538
	Copper	20.0	+/-20.0	U	20.0	P	04/23/2025	16:19	LB135538
	Iron	100	+/-100	U	100	P	04/23/2025	16:19	LB135538
	Lead	12.0	+/-12.0	U	12.0	P	04/23/2025	16:19	LB135538
	Manganese	20.0	+/-20.0	U	20.0	P	04/23/2025	16:19	LB135538
	Molybdenum	200	+/-200	U	200	P	04/23/2025	16:19	LB135538
	Nickel	40.0	+/-40.0	U	40.0	P	04/23/2025	16:19	LB135538
	Selenium	20.0	+/-20.0	U	20.0	P	04/23/2025	16:19	LB135538
	Silver	10.0	+/-10.0	U	10.0	P	04/23/2025	16:19	LB135538
	Thallium	40.0	+/-40.0	U	40.0	P	04/23/2025	16:19	LB135538
	Vanadium	40.0	+/-40.0	U	40.0	P	04/23/2025	16:19	LB135538
	Zinc	40.0	+/-40.0	U	40.0	P	04/23/2025	16:19	LB135538
CCB02	Aluminum	100	+/-100	U	100	P	04/23/2025	17:09	LB135538
	Antimony	50.0	+/-50.0	U	50.0	P	04/23/2025	17:09	LB135538
	Arsenic	20.0	+/-20.0	U	20.0	P	04/23/2025	17:09	LB135538
	Barium	100	+/-100	U	100	P	04/23/2025	17:09	LB135538
	Beryllium	6.00	+/-6.00	U	6.00	P	04/23/2025	17:09	LB135538
	Boron	100	+/-100	U	100	P	04/23/2025	17:09	LB135538
	Cadmium	6.00	+/-6.00	U	6.00	P	04/23/2025	17:09	LB135538
	Chromium	10.0	+/-10.0	U	10.0	P	04/23/2025	17:09	LB135538
	Cobalt	30.0	+/-30.0	U	30.0	P	04/23/2025	17:09	LB135538
	Copper	20.0	+/-20.0	U	20.0	P	04/23/2025	17:09	LB135538
	Iron	100	+/-100	U	100	P	04/23/2025	17:09	LB135538
	Lead	12.0	+/-12.0	U	12.0	P	04/23/2025	17:09	LB135538
	Manganese	20.0	+/-20.0	U	20.0	P	04/23/2025	17:09	LB135538
	Molybdenum	200	+/-200	U	200	P	04/23/2025	17:09	LB135538
	Nickel	40.0	+/-40.0	U	40.0	P	04/23/2025	17:09	LB135538
	Selenium	20.0	+/-20.0	U	20.0	P	04/23/2025	17:09	LB135538
	Silver	10.0	+/-10.0	U	10.0	P	04/23/2025	17:09	LB135538
	Thallium	40.0	+/-40.0	U	40.0	P	04/23/2025	17:09	LB135538
	Vanadium	40.0	+/-40.0	U	40.0	P	04/23/2025	17:09	LB135538

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: <u>Kleinfelder</u>		SDG No.: <u>Q1859</u>							
Contract: <u>POWE02</u>	Lab Code: <u>CHEM</u>	Case No.: <u>Q1859</u>	SAS No.: <u>Q1859</u>						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB02	Zinc	40.0	+/-40.0	U	40.0	P	04/23/2025	17:09	LB135538
CCB03	Aluminum	100	+/-100	U	100	P	04/23/2025	17:59	LB135538
	Antimony	50.0	+/-50.0	U	50.0	P	04/23/2025	17:59	LB135538
	Arsenic	20.0	+/-20.0	U	20.0	P	04/23/2025	17:59	LB135538
	Barium	100	+/-100	U	100	P	04/23/2025	17:59	LB135538
	Beryllium	6.00	+/-6.00	U	6.00	P	04/23/2025	17:59	LB135538
	Boron	100	+/-100	U	100	P	04/23/2025	17:59	LB135538
	Cadmium	6.00	+/-6.00	U	6.00	P	04/23/2025	17:59	LB135538
	Chromium	10.0	+/-10.0	U	10.0	P	04/23/2025	17:59	LB135538
	Cobalt	30.0	+/-30.0	U	30.0	P	04/23/2025	17:59	LB135538
	Copper	20.0	+/-20.0	U	20.0	P	04/23/2025	17:59	LB135538
	Iron	100	+/-100	U	100	P	04/23/2025	17:59	LB135538
	Lead	12.0	+/-12.0	U	12.0	P	04/23/2025	17:59	LB135538
	Manganese	20.0	+/-20.0	U	20.0	P	04/23/2025	17:59	LB135538
	Molybdenum	200	+/-200	U	200	P	04/23/2025	17:59	LB135538
	Nickel	40.0	+/-40.0	U	40.0	P	04/23/2025	17:59	LB135538
	Selenium	20.0	+/-20.0	U	20.0	P	04/23/2025	17:59	LB135538
	Silver	10.0	+/-10.0	U	10.0	P	04/23/2025	17:59	LB135538
	Thallium	40.0	+/-40.0	U	40.0	P	04/23/2025	17:59	LB135538
	Vanadium	40.0	+/-40.0	U	40.0	P	04/23/2025	17:59	LB135538
	Zinc	40.0	+/-40.0	U	40.0	P	04/23/2025	17:59	LB135538
CCB04	Aluminum	100	+/-100	U	100	P	04/23/2025	18:42	LB135538
	Antimony	50.0	+/-50.0	U	50.0	P	04/23/2025	18:42	LB135538
	Arsenic	20.0	+/-20.0	U	20.0	P	04/23/2025	18:42	LB135538
	Barium	100	+/-100	U	100	P	04/23/2025	18:42	LB135538
	Beryllium	6.00	+/-6.00	U	6.00	P	04/23/2025	18:42	LB135538
	Boron	100	+/-100	U	100	P	04/23/2025	18:42	LB135538
	Cadmium	6.00	+/-6.00	U	6.00	P	04/23/2025	18:42	LB135538
	Chromium	10.0	+/-10.0	U	10.0	P	04/23/2025	18:42	LB135538
	Cobalt	30.0	+/-30.0	U	30.0	P	04/23/2025	18:42	LB135538
	Copper	20.0	+/-20.0	U	20.0	P	04/23/2025	18:42	LB135538
	Iron	100	+/-100	U	100	P	04/23/2025	18:42	LB135538
	Lead	12.0	+/-12.0	U	12.0	P	04/23/2025	18:42	LB135538
	Manganese	20.0	+/-20.0	U	20.0	P	04/23/2025	18:42	LB135538
	Molybdenum	200	+/-200	U	200	P	04/23/2025	18:42	LB135538
	Nickel	40.0	+/-40.0	U	40.0	P	04/23/2025	18:42	LB135538
	Selenium	20.0	+/-20.0	U	20.0	P	04/23/2025	18:42	LB135538
	Silver	10.0	+/-10.0	U	10.0	P	04/23/2025	18:42	LB135538
	Thallium	40.0	+/-40.0	U	40.0	P	04/23/2025	18:42	LB135538

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Kleinfelder **SDG No.:** Q1859
Contract: POWE02 **Lab Code:** CHEM **Case No.:** Q1859 **SAS No.:** Q1859

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB04	Vanadium	40.0	+/-40.0	U	40.0	P	04/23/2025	18:42	LB135538
	Zinc	40.0	+/-40.0	U	40.0	P	04/23/2025	18:42	LB135538