

**CALIBRATION VERIFICATION SUMMARY**

**Contract:** POWE02

**Lab Code:** CHEM      **Case No.:** Q1859      **SAS No.:** Q1859      **SDG NO.:** Q1859

**Continuing Calib Date:** 04/24/2025      **Initial Calibration Date(s):** 04/10/2025      04/10/2025

**Continuing Calib Time:** 00:03      **Initial Calibration Time(s):** 09:36      17:52

**GC Column:** ZB-MR1      **ID:** 0.32      (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Tetrachloro-m-xylene	3.69	3.69	3.59	3.79	0.00
Decachlorobiphenyl	8.73	8.73	8.63	8.83	0.00

**CALIBRATION VERIFICATION SUMMARY**

**Contract:** POWE02

**Lab Code:** CHEM      **Case No.:** Q1859      **SAS No.:** Q1859      **SDG NO.:** Q1859

**Continuing Calib Date:** 04/24/2025      **Initial Calibration Date(s):** 04/10/2025      04/10/2025

**Continuing Calib Time:** 00:03      **Initial Calibration Time(s):** 09:36      17:52

**GC Column:** ZB-MR2      **ID:** 0.32      (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Tetrachloro-m-xylene	3.68	3.69	3.59	3.79	0.01
Decachlorobiphenyl	8.68	8.69	8.59	8.79	0.01