

Report of Analysis

Client:	Kleinfelder					Date Collected:	04/18/25	
Project:	Lincoln Hig	gh School				Date Received:	04/22/25	
Client Sample ID:	COMP-2					SDG No.:	Q1859	
Lab Sample ID:	Q1859-02					Matrix:	SOIL	
Analytical Method:	SW8082A					% Solid:	83.2 D	ecanted:
Sample Wt/Vol:	30.02	Units: g				Final Vol:	10000	uL
Soil Aliquot Vol:		uL				Test:	PCB Group1	
Extraction Type:						Injection Volume :		
GPC Factor :	1.0	PH :						
Prep Method :	SW3541B							
File ID/Qc Batch:	Dilution:		Prep	o Date		Date Analyzed	Prep Bat	tch ID
	Dilution: 1		-	o Date 23/25 08:35		Date Analyzed 04/23/25 20:14	Prep Bat PB1677	
File ID/Qc Batch:			-				-	08
File ID/Qc Batch: PO110663.D CAS Number	1		04/2	23/25 08:35			PB1677	08
File ID/Qc Batch: PO110663.D	1		04/2	23/25 08:35			PB1677	08 L Units(Dry Weight)
File ID/Qc Batch: PO110663.D CAS Number TARGETS	1 Parameter		04/2 Conc.	23/25 08:35 Qualifier	MDL		PB1677	08 L Units(Dry Weight) 4 ug/kg
File ID/Qc Batch: PO110663.D CAS Number TARGETS 12674-11-2	1 Parameter Aroclor-1016		04/2 C onc. 4.70	23/25 08:35 Qualifier U	MDL 4.70		PB1677 LOQ / CRQI 20.4	08 L Units(Dry Weight) 4 ug/kg 4 ug/kg
File ID/Qc Batch: PO110663.D CAS Number TARGETS 12674-11-2 11097-69-1	1 Parameter Aroclor-1016 Aroclor-1254		04/2 Conc. 4.70 3.90	23/25 08:35 Qualifier U U	MDL 4.70 3.90		PB16774 LOQ / CRQ1 20.4 20.4	08 L Units(Dry Weight) 4 ug/kg 4 ug/kg
File ID/Qc Batch: PO110663.D CAS Number TARGETS 12674-11-2 11097-69-1 11096-82-5	1 Parameter Aroclor-1016 Aroclor-1254		04/2 Conc. 4.70 3.90	23/25 08:35 Qualifier U U	MDL 4.70 3.90		PB16774 LOQ / CRQ1 20.4 20.4	08 L Units(Dry Weight) 4 ug/kg 4 ug/kg 4 ug/kg

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates > 25% difference for detected

concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

- J = Estimated Value
- B = Analyte Found in Associated Method Blank
- N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

 $\mathbf{S}=\mathbf{Indicates}$ estimated value where valid five-point calibration

was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit