



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1859

Client: Kleinfelder

Analytical Method: **8082A** **Datafile :** PO110645.D

Lab Sample ID	Parameter	Spike	Result	Units	Rec	RPD	Qual	RPD		Limits	
								Qual	Low	High	RPD
PB167708BS	AR1016	166.5	151	ug/kg	91				71	120	
	AR1260	166.5	145	ug/kg	87				65	130	