

CALIBRATION VERIFICATION SUMMARY

Contract: POWE02

Lab Code: CHEM **Case No.:** Q1859 **SAS No.:** Q1859 **SDG NO.:** Q1859

Continuing Calib Date: 04/23/2025 **Initial Calibration Date(s):** 04/18/2025 04/18/2025

Continuing Calib Time: 09:55 **Initial Calibration Time(s):** 13:56 14:51

GC Column: ZB-MR1 **ID:** 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Decachlorobiphenyl	9.09	9.08	8.98	9.18	-0.01
Tetrachloro-m-xylene	3.56	3.55	3.45	3.65	-0.01
Aldrin	5.28	5.27	5.17	5.37	-0.01
Dieldrin	6.36	6.35	6.25	6.45	-0.01
4,4'-DDE	6.21	6.20	6.10	6.30	-0.01
4,4'-DDD	6.72	6.71	6.61	6.81	-0.01
4,4'-DDT	7.03	7.02	6.92	7.12	-0.01

CALIBRATION VERIFICATION SUMMARY

Contract: POWE02

Lab Code: CHEM Case No.: Q1859 SAS No.: Q1859 SDG NO.: Q1859

Continuing Calib Date: 04/23/2025 Initial Calibration Date(s): 04/18/2025 04/18/2025

Continuing Calib Time: 09:55 Initial Calibration Time(s): 13:56 14:51

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
Decachlorobiphenyl	8.08	8.08	7.98	8.18	0.00
Tetrachloro-m-xylene	2.88	2.88	2.78	2.98	0.00
Aldrin	4.37	4.37	4.27	4.47	0.00
Dieldrin	5.52	5.52	5.42	5.62	0.00
4,4'-DDE	5.38	5.38	5.28	5.48	0.00
4,4'-DDD	5.93	5.93	5.83	6.03	0.00
4,4'-DDT	6.19	6.19	6.09	6.29	0.00