

CALIBRATION VERIFICATION SUMMARY

Contract: POWE02

Lab Code: CHEM Case No.: Q1859 SAS No.: Q1859 SDG NO.: Q1859

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 04/18/2025 04/18/2025

Client Sample No.: CCAL03 Date Analyzed: 04/28/2025

Lab Sample No.: PSTDCCC050 Data File : PD088290.D Time Analyzed: 09:46

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	6.711	6.606	6.806	50.200	50.000	0.4
4,4'-DDE	6.203	6.097	6.297	49.670	50.000	-0.7
4,4'-DDT	7.027	6.922	7.122	49.970	50.000	-0.1
Aldrin	5.278	5.173	5.373	49.760	50.000	-0.5
Decachlorobiphenyl	9.079	8.975	9.175	45.380	50.000	-9.2
Dieldrin	6.353	6.249	6.449	50.890	50.000	1.8
Tetrachloro-m-xylene	3.556	3.452	3.652	48.690	50.000	-2.6

CALIBRATION VERIFICATION SUMMARY

Contract: POWE02

Lab Code: CHEM Case No.: Q1859 SAS No.: Q1859 SDG NO.: Q1859

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 04/18/2025 04/18/2025

Client Sample No.: CCAL03 Date Analyzed: 04/28/2025

Lab Sample No.: PSTDCCC050 Data File : PD088290.D Time Analyzed: 09:46

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	5.932	5.834	6.034	44.700	50.000	-10.6
4,4'-DDE	5.378	5.280	5.480	45.050	50.000	-9.9
4,4'-DDT	6.187	6.088	6.288	45.420	50.000	-9.2
Aldrin	4.371	4.273	4.473	45.140	50.000	-9.7
Decachlorobiphenyl	8.076	7.977	8.177	40.250	50.000	-19.5
Dieldrin	5.516	5.417	5.617	45.200	50.000	-9.6
Tetrachloro-m-xylene	2.880	2.783	2.983	44.950	50.000	-10.1