

Surrogate Summary

SDG No.: Q1859

Client: Kleinfelder

Analytical Method: 8081B

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
I.BLK-PD088121.D	PIBLK-PD088121.D	Decachlorobiphenyl	1	20	22.9	115		43	140
		Tetrachloro-m-xylene	1	20	20.1	101		77	126
		Decachlorobiphenyl	2	20	22.7	113		43	140
		Tetrachloro-m-xylene	2	20	20.9	104		77	126
I.BLK-PD088231.D	PIBLK-PD088231.D	Decachlorobiphenyl	1	20	18.4	92		43	140
		Tetrachloro-m-xylene	1	20	19.0	95		77	126
		Decachlorobiphenyl	2	20	17.7	88		43	140
		Tetrachloro-m-xylene	2	20	18.9	94		77	126
PB167709BL	PB167709BL	Decachlorobiphenyl	1	20	18.3	92		20	144
		Tetrachloro-m-xylene	1	20	17.9	89		19	148
		Decachlorobiphenyl	2	20	17.0	85		20	144
		Tetrachloro-m-xylene	2	20	18.5	92		19	148
PB167709BS	PB167709BS	Decachlorobiphenyl	1	20	19.8	99		20	144
		Tetrachloro-m-xylene	1	20	18.8	94		19	148
		Decachlorobiphenyl	2	20	18.8	94		20	144
		Tetrachloro-m-xylene	2	20	19.1	96		19	148
I.BLK-PD088246.D	PIBLK-PD088246.D	Decachlorobiphenyl	1	20	17.9	89		43	140
		Tetrachloro-m-xylene	1	20	18.9	94		77	126
		Decachlorobiphenyl	2	20	17.1	85		43	140
		Tetrachloro-m-xylene	2	20	18.6	93		77	126
I.BLK-PD088288.D	PIBLK-PD088288.D	Decachlorobiphenyl	1	20	18.0	90		43	140
		Tetrachloro-m-xylene	1	20	17.5	88		77	126
		Decachlorobiphenyl	2	20	16.4	82		43	140
		Tetrachloro-m-xylene	2	20	17.2	86		77	126
Q1858-02MS	COMP-2MS	Decachlorobiphenyl	1	20	15.0	75		20	144
		Tetrachloro-m-xylene	1	20	16.8	84		19	148
		Decachlorobiphenyl	2	20	13.0	65		20	144
		Tetrachloro-m-xylene	2	20	16.9	85		19	148
Q1858-02MSD	COMP-2MSD	Decachlorobiphenyl	1	20	15.2	76		20	144
		Tetrachloro-m-xylene	1	20	16.8	84		19	148
		Decachlorobiphenyl	2	20	12.6	63		20	144
		Tetrachloro-m-xylene	2	20	17.0	85		19	148
Q1859-01	COMP-1	Decachlorobiphenyl	1	20	9.17	46		20	144
		Tetrachloro-m-xylene	1	20	17.1	85		19	148
		Decachlorobiphenyl	2	20	6.94	35		20	144
		Tetrachloro-m-xylene	2	20	16.6	83		19	148
Q1859-02	COMP-2	Decachlorobiphenyl	1	20	12.3	62		20	144
		Tetrachloro-m-xylene	1	20	18.1	90		19	148
		Decachlorobiphenyl	2	20	10.6	53		20	144
		Tetrachloro-m-xylene	2	20	17.9	89		19	148
Q1859-03	COMP-3	Decachlorobiphenyl	1	20	15.3	76		20	144

Surrogate Summary

SDG No.: Q1859

Client: Kleinfelder

Analytical Method: 8081B

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
Q1859-03	COMP-3	Tetrachloro-m-xylene	1	20	17.5	87		19	148
		Decachlorobiphenyl	2	20	11.1	55		20	144
		Tetrachloro-m-xylene	2	20	17.1	86		19	148
I.BLK-PD088302.D	PIBLK-PD088302.D	Decachlorobiphenyl	1	20	17.6	88		43	140
		Tetrachloro-m-xylene	1	20	18.1	91		77	126
		Decachlorobiphenyl	2	20	12.7	63		43	140
		Tetrachloro-m-xylene	2	20	17.7	89		77	126