



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1859

Client: Kleinfelder

Analytical Method: **8081B** **Datafile :** PD088239.D

Lab Sample ID	Parameter	Spike	Result	Units	Rec	RPD	Qual	RPD		Limits	
								Qual	Low	High	RPD
PB167709BS	Aldrin	16.66	16.4	ug/kg	98				82	124	
	Dieldrin	16.66	16.6	ug/kg	100				85	121	
	4,4'-DDE	16.66	16.1	ug/kg	97				81	123	
	4,4'-DDD	16.66	16.7	ug/kg	100				80	131	
	4,4'-DDT	16.66	15.3	ug/kg	92				70	129	