

CALIBRATION VERIFICATION SUMMARY

Contract: POWE02

Lab Code: CHEM Case No.: Q1859 SAS No.: Q1859 SDG NO.: Q1859

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 04/18/2025 04/18/2025

Client Sample No.: CCAL03 Date Analyzed: 04/23/2025

Lab Sample No.: PSTDCCC050 Data File : PD088257.D Time Analyzed: 21:58

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	6.706	6.606	6.806	59.130	50.000	18.3
4,4'-DDE	6.195	6.097	6.297	56.460	50.000	12.9
4,4'-DDT	7.022	6.922	7.122	43.360	50.000	-13.3
Aldrin	5.273	5.173	5.373	58.610	50.000	17.2
Decachlorobiphenyl	9.074	8.975	9.175	48.850	50.000	-2.3
Dieldrin	6.348	6.249	6.449	58.100	50.000	16.2
Tetrachloro-m-xylene	3.552	3.452	3.652	56.540	50.000	13.1

CALIBRATION VERIFICATION SUMMARY

Contract: POWE02

Lab Code: CHEM Case No.: Q1859 SAS No.: Q1859 SDG NO.: Q1859

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 04/18/2025 04/18/2025

Client Sample No.: CCAL03 Date Analyzed: 04/23/2025

Lab Sample No.: PSTDCCC050 Data File : PD088257.D Time Analyzed: 21:58

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	5.932	5.834	6.034	52.870	50.000	5.7
4,4'-DDE	5.377	5.280	5.480	50.860	50.000	1.7
4,4'-DDT	6.187	6.088	6.288	40.850	50.000	-18.3
Aldrin	4.373	4.273	4.473	52.190	50.000	4.4
Decachlorobiphenyl	8.076	7.977	8.177	43.690	50.000	-12.6
Dieldrin	5.516	5.417	5.617	51.090	50.000	2.2
Tetrachloro-m-xylene	2.884	2.783	2.983	51.730	50.000	3.5