

Instrument ID: BNA\_M

**Daily Analysis Runlog For Sequence/QC Batch ID # BM040825**

|              |          |                      |                      |
|--------------|----------|----------------------|----------------------|
| Review By    | anahy    | Review On            | 4/9/2025 10:38:14 AM |
| Supervise By | Jagrut   | Supervise On         | 4/9/2025 4:50:06 PM  |
| SubDirectory | BM040825 | HP Acquire Method    | BNA_M                |
|              |          | HP Processing Method | bm040825             |

| STD. NAME                | STD REF.#                                               |
|--------------------------|---------------------------------------------------------|
| Tune/Reschk              | SP6757                                                  |
| Initial Calibration Stds | SP6722,SP6723,SP6724,SP6725,SP6726,SP6727,SP6728,SP6729 |
| CCC                      | SP6725                                                  |
| Internal Standard/PEM    | S12657,10ul/1000ul sample                               |
| ICV/I.BLK                | SP6686                                                  |
| Surrogate Standard       |                                                         |
| MS/MSD Standard          |                                                         |
| LCS Standard             |                                                         |

| Sr# | Sampleld    | ClientID    | Data File Name | Date-Time         | Comment                                                                        | Operator | Status |
|-----|-------------|-------------|----------------|-------------------|--------------------------------------------------------------------------------|----------|--------|
| 1   | DFTPP       | DFTPP       | BM049847.D     | 08 Apr 2025 12:55 |                                                                                | RC/JU    | Ok     |
| 2   | SSTDICC2.5  | SSTDICC2.5  | BM049848.D     | 08 Apr 2025 13:35 |                                                                                | RC/JU    | Ok     |
| 3   | SSTDICC005  | SSTDICC005  | BM049849.D     | 08 Apr 2025 14:14 | Compound #32,54,65 removed from 5ppm                                           | RC/JU    | Ok,M   |
| 4   | SSTDICC010  | SSTDICC010  | BM049850.D     | 08 Apr 2025 14:53 | Compound #32,54 kept on LR                                                     | RC/JU    | Ok,M   |
| 5   | SSTDICC020  | SSTDICC020  | BM049851.D     | 08 Apr 2025 15:32 | This calibration is fail for Com#77                                            | RC/JU    | Ok     |
| 6   | SSTDICCC040 | SSTDICCC040 | BM049852.D     | 08 Apr 2025 16:12 | This calibration is good for 8270E DOD except Com#77 and good for 625.1 method | RC/JU    | Ok     |
| 7   | SSTDICC050  | SSTDICC050  | BM049853.D     | 08 Apr 2025 17:30 |                                                                                | RC/JU    | Ok     |
| 8   | SSTDICC060  | SSTDICC060  | BM049854.D     | 08 Apr 2025 19:28 | Compound #69 Removed from 60PPM                                                | RC/JU    | Ok,M   |
| 9   | SSTDICC080  | SSTDICC080  | BM049855.D     | 08 Apr 2025 20:07 | Compound #69 Removed from 80PPM                                                | RC/JU    | Ok,M   |
| 10  | SSTDICV040  | ICVBM040825 | BM049856.D     | 08 Apr 2025 20:47 |                                                                                | RC/JU    | Ok     |
| 11  | PB167474BL  | PB167474BL  | BM049857.D     | 08 Apr 2025 21:26 | Use for Q1730 only                                                             | RC/JU    | Not Ok |

M : Manual Integration

Instrument ID: BNA\_M

**Daily Analysis Runlog For Sequence/QC Batch ID # BM042325**

|              |          |                      |                      |
|--------------|----------|----------------------|----------------------|
| Review By    | Rahul    | Review On            | 4/24/2025 9:48:24 AM |
| Supervise By | Jagrut   | Supervise On         | 4/24/2025 2:07:47 PM |
| SubDirectory | BM042325 | HP Acquire Method    | BNA_M                |
|              |          | HP Processing Method | BM040825             |

| STD. NAME                | STD REF.#                                               |
|--------------------------|---------------------------------------------------------|
| Tune/Reschk              | SP6757                                                  |
| Initial Calibration Stds | SP6722,SP6723,SP6724,SP6725,SP6726,SP6727,SP6728,SP6729 |
| CCC                      | SP6725                                                  |
| Internal Standard/PEM    | S12660,10ul/1000ul sample                               |
| ICV/I.BLK                | SP6686                                                  |
| Surrogate Standard       |                                                         |
| MS/MSD Standard          |                                                         |
| LCS Standard             |                                                         |

| Sr# | SampleID    | ClientID       | Data File Name | Date-Time         | Comment | Operator | Status |
|-----|-------------|----------------|----------------|-------------------|---------|----------|--------|
| 1   | DFTPP       | DFTPP          | BM050006.D     | 23 Apr 2025 09:18 |         | RC/JU    | Ok     |
| 2   | SSTDCCC040  | SSTDCCC040     | BM050007.D     | 23 Apr 2025 10:38 |         | RC/JU    | Ok     |
| 3   | PB167694BL  | PB167694BL     | BM050008.D     | 23 Apr 2025 11:17 |         | RC/JU    | Ok     |
| 4   | PB167694BS  | PB167694BS     | BM050009.D     | 23 Apr 2025 11:57 |         | RC/JU    | Ok,M   |
| 5   | PB167694BSD | PB167694BSD    | BM050010.D     | 23 Apr 2025 12:36 |         | RC/JU    | Ok,M   |
| 6   | Q1848-03DL  | ETGI-275DL     | BM050011.D     | 23 Apr 2025 13:19 |         | RC/JU    | Ok,M   |
| 7   | Q1852-07    | 72-12013       | BM050012.D     | 23 Apr 2025 13:58 |         | RC/JU    | Ok,M   |
| 8   | Q1852-07MS  | 72-12013MS     | BM050013.D     | 23 Apr 2025 14:37 |         | RC/JU    | Ok,M   |
| 9   | Q1852-07MSD | 72-12013MSD    | BM050014.D     | 23 Apr 2025 15:16 |         | RC/JU    | Ok,M   |
| 10  | Q1858-01    | COMP-1         | BM050015.D     | 23 Apr 2025 15:55 |         | RC/JU    | Ok     |
| 11  | Q1858-02    | COMP-2         | BM050016.D     | 23 Apr 2025 16:35 |         | RC/JU    | Ok,M   |
| 12  | Q1858-03    | COMP-3         | BM050017.D     | 23 Apr 2025 17:14 |         | RC/JU    | Ok     |
| 13  | Q1859-02    | COMP-2         | BM050018.D     | 23 Apr 2025 17:53 |         | RC/JU    | Ok     |
| 14  | Q1859-03    | COMP-3         | BM050019.D     | 23 Apr 2025 18:32 |         | RC/JU    | Ok     |
| 15  | Q1859-01    | COMP-1         | BM050020.D     | 23 Apr 2025 19:11 |         | RC/JU    | Ok     |
| 16  | Q1853-01    | EO-02-04222025 | BM050021.D     | 23 Apr 2025 19:50 |         | RC/JU    | Ok,M   |
| 17  | Q1854-01    | NB-08-04222025 | BM050022.D     | 23 Apr 2025 20:29 |         | RC/JU    | Ok,M   |

M : Manual Integration

Instrument ID: BNA\_P

**Daily Analysis Runlog For Sequence/QC Batch ID # BP041425**

|              |          |                      |                       |
|--------------|----------|----------------------|-----------------------|
| Review By    | Rahul    | Review On            | 4/15/2025 9:55:40 AM  |
| Supervise By | Jagrut   | Supervise On         | 4/15/2025 10:53:04 AM |
| SubDirectory | BP041425 | HP Acquire Method    | BNA_P                 |
|              |          | HP Processing Method | bp041425              |

| STD. NAME                | STD REF.#                                               |
|--------------------------|---------------------------------------------------------|
| Tune/Reschk              | SP6757                                                  |
| Initial Calibration Stds | SP6722,SP6723,SP6724,SP6725,SP6726,SP6727,SP6728,SP6729 |
| CCC                      | SP6725                                                  |
| Internal Standard/PEM    | S12659,10ul/1000ul sample                               |
| ICV/I.BLK                | SP6770                                                  |
| Surrogate Standard       |                                                         |
| MS/MSD Standard          |                                                         |
| LCS Standard             |                                                         |

| Sr# | SampleID    | ClientID    | Data File Name | Date-Time         | Comment                                                                           | Operator | Status |
|-----|-------------|-------------|----------------|-------------------|-----------------------------------------------------------------------------------|----------|--------|
| 1   | DFTPP       | DFTPP       | BP024274.D     | 14 Apr 2025 10:25 |                                                                                   | RC/JU    | Ok     |
| 2   | SSTDICC2.5  | SSTDICC2.5  | BP024275.D     | 14 Apr 2025 11:06 |                                                                                   | RC/JU    | Ok     |
| 3   | SSTDICC005  | SSTDICC005  | BP024276.D     | 14 Apr 2025 11:47 | Compound #32,54,65 removed from 5ppm                                              | RC/JU    | Ok,M   |
| 4   | SSTDICC010  | SSTDICC010  | BP024277.D     | 14 Apr 2025 12:27 |                                                                                   | RC/JU    | Ok,M   |
| 5   | SSTDICC020  | SSTDICC020  | BP024278.D     | 14 Apr 2025 13:08 |                                                                                   | RC/JU    | Ok,M   |
| 6   | SSTDICCC040 | SSTDICCC040 | BP024279.D     | 14 Apr 2025 13:49 | This calibration is fail for Com#77                                               | RC/JU    | Ok     |
| 7   | SSTDICC050  | SSTDICC050  | BP024280.D     | 14 Apr 2025 15:10 | This calibration is good for 8270E, 8270 DOD and 625.1 methods except Compound#77 | RC/JU    | Ok,M   |
| 8   | SSTDICC060  | SSTDICC060  | BP024281.D     | 14 Apr 2025 16:32 | Compound #69 Removed from 60PPM                                                   | RC/JU    | Ok,M   |
| 9   | SSTDICC080  | SSTDICC080  | BP024282.D     | 14 Apr 2025 17:13 | Compound #69 Removed from 80PPM                                                   | RC/JU    | Ok,M   |
| 10  | SSTDICV040  | ICVBP041425 | BP024283.D     | 14 Apr 2025 18:35 |                                                                                   | RC/JU    | Ok,M   |
| 11  | PB167488TB  | PB167488TB  | BP024284.D     | 14 Apr 2025 19:16 |                                                                                   | RC/JU    | Ok     |

M : Manual Integration

Instrument ID: BNA\_P

**Daily Analysis Runlog For Sequence/QC Batch ID # BP042425**

|              |          |                      |                       |
|--------------|----------|----------------------|-----------------------|
| Review By    | Jagrut   | Review On            | 4/25/2025 10:24:06 AM |
| Supervise By | Rahul    | Supervise On         | 4/25/2025 10:24:26 AM |
| SubDirectory | BP042425 | HP Acquire Method    | BNA_P                 |
|              |          | HP Processing Method | bp041425              |

| STD. NAME                | STD REF.#                                               |
|--------------------------|---------------------------------------------------------|
| Tune/Reschk              | SP6757                                                  |
| Initial Calibration Stds | SP6722,SP6723,SP6724,SP6725,SP6726,SP6727,SP6728,SP6729 |
| CCC                      | SP6725                                                  |
| Internal Standard/PEM    | S12660,10ul/1000ul sample                               |
| ICV/I.BLK                | SP6770                                                  |
| Surrogate Standard       |                                                         |
| MS/MSD Standard          |                                                         |
| LCS Standard             |                                                         |

| Sr# | SampleID    | ClientID    | Data File Name | Date-Time         | Comment                | Operator | Status |
|-----|-------------|-------------|----------------|-------------------|------------------------|----------|--------|
| 1   | DFTPP       | DFTPP       | BP024406.D     | 24 Apr 2025 11:10 |                        | RC/JU    | Ok     |
| 2   | SSTDCCC040  | SSTDCCC040  | BP024407.D     | 24 Apr 2025 11:50 |                        | RC/JU    | Ok,M   |
| 3   | PB167711BL  | PB167711BL  | BP024408.D     | 24 Apr 2025 12:31 |                        | RC/JU    | Ok     |
| 4   | PB167711BS  | PB167711BS  | BP024409.D     | 24 Apr 2025 13:12 |                        | RC/JU    | Ok,M   |
| 5   | Q1852-02    | ETGI-354    | BP024410.D     | 24 Apr 2025 16:35 | Internal Standard Fail | RC/JU    | ReRun  |
| 6   | Q1852-02MS  | ETGI-354MS  | BP024411.D     | 24 Apr 2025 17:16 |                        | RC/JU    | Ok,M   |
| 7   | Q1852-02MSD | ETGI-354MSD | BP024412.D     | 24 Apr 2025 17:56 |                        | RC/JU    | Ok,M   |
| 8   | Q1852-04    | 72-11977    | BP024413.D     | 24 Apr 2025 18:37 |                        | RC/JU    | Ok     |
| 9   | Q1852-06    | ETGI-278    | BP024414.D     | 24 Apr 2025 19:18 | Internal Standard Fail | RC/JU    | ReRun  |
| 10  | Q1852-08    | 72-12013    | BP024415.D     | 24 Apr 2025 19:58 | Internal Standard Fail | RC/JU    | ReRun  |

M : Manual Integration