

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECH

Contract: POWE02

Lab Code: CHEM

SAS No.: Q1859 SDG NO.: Q1859

Lab File ID: BM050006.D

DFTPP Injection Date: 04/23/2025

Instrument ID: BNA_M

DFTPP Injection Time: 09:18

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	21
68	Less than 2.0% of mass 69	0.4 (1.6) 1
69	Mass 69 relative abundance	25.4
70	Less than 2.0% of mass 69	0.1 (0.6) 1
127	10.0 - 80.0% of mass 198	32.7
197	Less than 2.0% of mass 198	0.4
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.9
275	10.0 - 60.0% of mass 198	26.3
365	Greater than 1% of mass 198	3.4
441	Present, but less than mass 443	11.7
442	Greater than 50% of mass 198	75.4
443	15.0 - 24.0% of mass 442	14.7 (19.5) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDCCC040	SSTDCCC040	BM050007.D	04/23/2025	10:38
72-12013MS	Q1852-07MS	BM050013.D	04/23/2025	14:37
72-12013MSD	Q1852-07MSD	BM050014.D	04/23/2025	15:16
COMP-2	Q1859-02	BM050018.D	04/23/2025	17:53
COMP-3	Q1859-03	BM050019.D	04/23/2025	18:32
COMP-1	Q1859-01	BM050020.D	04/23/2025	19:11