

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK  
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECH

Contract: POWE02

Lab Code: CHEM

SAS No.: Q1859

SDG NO.: Q1859

Lab File ID: BP024406.D

DFTPP Injection Date: 04/24/2025

Instrument ID: BNA\_P

DFTPP Injection Time: 11:10

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	26.2
68	Less than 2.0% of mass 69	0.5 ( 1.7 ) 1
69	Mass 69 relative abundance	30.1
70	Less than 2.0% of mass 69	0.1 ( 0.5 ) 1
127	10.0 - 80.0% of mass 198	42.2
197	Less than 2.0% of mass 198	0.0
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6
275	10.0 - 60.0% of mass 198	27.3
365	Greater than 1% of mass 198	4.1
441	Present, but less than mass 443	15.3
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	19.3 ( 19.3 ) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDCCC040	SSTDCCC040	BP024407.D	04/24/2025	11:50
PB167711BL	PB167711BL	BP024408.D	04/24/2025	12:31
PB167711BS	PB167711BS	BP024409.D	04/24/2025	13:12