

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: POWE02

Lab Code: CHEM Case No.: Q1859 SAS No.: Q1859 SDG No.: Q1859

Instrument ID: MSVOA_Y Calibration Date/Time: 04/23/2025 09:57

Lab File ID: VY021962.D Init. Calib. Date(s): 04/22/2025 04/22/2025

Heated Purge: (Y/N) Y Init. Calib. Time(s): 13:39 16:15

GC Column: RXI-624 ID: 0.25 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
cis-1,2-Dichloroethene	0.622	0.605		-2.73	20
1,1,1-Trichloroethane	0.896	0.858		-4.24	20
Benzene	1.387	1.381		-0.43	20
Trichloroethene	0.384	0.371		-3.38	20
Toluene	0.898	0.927		3.23	20
Ethyl Benzene	1.829	1.881		2.84	20
m/p-Xylenes	0.728	0.756		3.85	20
o-Xylene	0.671	0.693		3.28	20
Isopropylbenzene	3.341	3.454		3.38	20
1,2-Dichloroethane-d4	0.462	0.438		-5.2	20
Dibromofluoromethane	0.330	0.325		-1.51	20
Toluene-d8	1.245	1.266		1.69	20
4-Bromofluorobenzene	0.419	0.422		0.72	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.