

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1859

Client: Kleinfelder

Analytical Method: SW8260D

Datafile : VY021964.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VY0423SBS01	cis-1,2-Dichloroethene	20	20.3	ug/Kg	102			82	123	
	1,1,1-Trichloroethane	20	20.2	ug/Kg	101			80	126	
	Benzene	20	20.7	ug/Kg	104			84	121	
	Trichloroethene	20	20.2	ug/Kg	101			83	122	
	Toluene	20	20.8	ug/Kg	104			83	122	
	Ethyl Benzene	20	19.8	ug/Kg	99			82	124	
	m/p-Xylenes	40	40.3	ug/Kg	101			83	124	
	o-Xylene	20	19.8	ug/Kg	99			83	123	
	Isopropylbenzene	20	19.6	ug/Kg	98			82	124	