

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1859

Client: Kleinfelder

Analytical Method: SW8260D

Datafile : VY021965.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VY0423SBSD01	cis-1,2-Dichloroethene	20	21.4	ug/Kg	107	5		82	123	20
	1,1,1-Trichloroethane	20	21.2	ug/Kg	106	5		80	126	20
	Benzene	20	21.4	ug/Kg	107	3		84	121	20
	Trichloroethene	20	21.0	ug/Kg	105	4		83	122	20
	Toluene	20	21.6	ug/Kg	108	4		83	122	20
	Ethyl Benzene	20	20.3	ug/Kg	102	3		82	124	20
	m/p-Xylenes	40	41.9	ug/Kg	105	4		83	124	20
	o-Xylene	20	20.6	ug/Kg	103	4		83	123	20
	Isopropylbenzene	20	20.2	ug/Kg	101	3		82	124	20