

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: POWE02
 Lab Code: CHEM Case No.: Q1859 SAS No.: Q1859 SDG NO.: Q1859
 Lab File ID: VY021962.D Date Analyzed: 04/23/2025
 Instrument ID: MSVOA_Y Time Analyzed: 09:57
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS4 AREA #	RT #				
12 HOUR STD	227723	13.347				
UPPER LIMIT	455446	13.847				
LOWER LIMIT	113862	12.847				
EPA SAMPLE NO.						
COMP-1	198189	13.35				
COMP-2	215851	13.35				
COMP-3	187146	13.35				
VY0423SBL01	194434	13.35				
VY0423SBS01	221603	13.35				
VY0423SBSD01	217306	13.35				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.