

ANALYTICAL RESULTS SUMMARY

GC SEMI-VOLATILES

PROJECT NAME : RFP 911

WESTON SOLUTIONS, INC.
1090 King Georges Post Road
Suite 201
Edison, NJ - 08837-3703
Phone No: 732-585-4410

ORDER ID : Q1860
ATTENTION : Smita Sumbaly



Laboratory Certification ID # 20012



1) Signature Page	3
2) Case Narrative	5
2.1) PCB- Case Narrative	5
3) Qualifier Page	8
4) QA Checklist	9
5) PCB Data	10
6) Shipping Document	88
6.1) CHAIN OF CUSTODY	89
6.2) Lab Certificate	92

1
2
3
4
5
6

Cover Page

Order ID : Q1860

Project ID : RFP 911

Client : Weston Solutions, Inc.

Lab Sample Number

Q1860-01
Q1860-02
Q1860-03
Q1860-04
Q1860-05
Q1860-06
Q1860-07
Q1860-08
Q1860-09
Q1860-10
Q1860-11
Q1860-12
Q1860-13
Q1860-14
Q1860-15
Q1860-16
Q1860-17
Q1860-18
Q1860-19
Q1860-20
Q1860-21
Q1860-22
Q1860-23
Q1860-24
Q1860-25
Q1860-26
Q1860-27
Q1860-28
Q1860-29

Client Sample Number

P001-SS001-01
P001-SS001-01MS
P001-SS001-01MSD
P001-SS002-01
P001-SS002-02
P001-SS003-01
P001-SS004-01
P001-SS005-01
P001-SS006-01
P001-SS007-01
P001-SS008-01
P001-SS009-01
P001-SS010-01
P001-SS011-01
P001-SS012-01
P001-SS013-01
P001-SS014-01
P001-SS015-01
P001-SS016-01
P001-SS016-01MS
P001-SS016-01MSD
P001-SS017-01
P001-SS017-02
P001-SS018-01
P001-SS019-01
P001-SS020-01
P001-SS021-01
P001-SS022-01
P001-SS023-01

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. Release of the data contained in this hard copy data package has been authorized by the laboratory manager or his designee, as verified by the following signature.

Signature :

APPROVED

By Nimisha Pandya, QA/QC Supervisor at 2:36 pm, May 07, 2025

Date: 5/6/2025

NYDOH CERTIFICATION NO - 11376

NJDEP CERTIFICATION NO - 20012

Cover Page

Order ID : Q1860

Project ID : RFP 911

Client : Weston Solutions, Inc.

Lab Sample Number

Q1860-30
Q1860-31
Q1860-32
Q1860-33
Q1860-34
Q1860-35
Q1860-36
Q1860-37
Q1860-38
Q1860-39
Q1860-40
Q1860-41
Q1860-42

Client Sample Number

P001-SS024-01
P001-SS025-01
P001-SS026-01
P001-SS027-01
P001-SS028-01
P001-SS029-01
P001-SS030-01
P001-SS031-01
P001-SS032-01
P001-SS033-01
P001-SS034-01
P001-SS035-01
P001-SS036-01

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. Release of the data contained in this hard copy data package has been authorized by the laboratory manager or his designee, as verified by the following signature.

Signature : _____

Date: 5/6/2025

NYDOH CERTIFICATION NO - 11376

NJDEP CERTIFICATION NO - 20012

CASE NARRATIVE

Weston Solutions, Inc.

Project Name: RFP 911

Project # N/A

Chemtech Project # Q1860

Test Name: PCB

A. Number of Samples and Date of Receipt:

42 Solid samples were received on 04/23/2025.

B. Parameters

According to the Chain of Custody document, the following analyses were requested: PCB. This data package contains results for PCB.

C. Analytical Techniques:

The analyses were performed on instrument GCECD_P. The front column is ZB-MR1 which is 30 meters, 0.32 mm ID, 0.5 um df, Catalogue # 7HM-G016-17. The rear column is ZB-MR2 which is 30 meters, 0.32 mm ID, 0.25 µm; Catalogue # 7HM-G017-11. The analyses were performed on instrument GCECD_O. The front column is ZB-MR1 which is 30 meters, 0.32 mm ID, 0.5 um df, Catalogue # 7HM-G016-17. The rear column is ZB-MR2 which is 30 meters, 0.32 mm ID, 0.25 µm; Catalogue # 7HM-G017-11. The analysis of PCBs was based on method 8082A and extraction was done based on method 3541.

D. QA/ QC Samples:

The Holding Times were met for all analysis.

The Surrogate recoveries met the acceptable criteria except for
P001-SS007-01 [Decachlorobiphenyl(2) - 215%],
P001-SS007-01DL [Decachlorobiphenyl(1) - 178%],
P001-SS013-01DL [Decachlorobiphenyl(2) - 194%],
P001-SS021-01DL [Decachlorobiphenyl(2) - 192%] as per method one surrogate is allowed to failed, therefore no corrective action was taken and
P001-SS009-01 [Decachlorobiphenyl(1) - 248%, Decachlorobiphenyl(2) - 227%],
P001-SS012-01 [Decachlorobiphenyl(1) - 219%, Decachlorobiphenyl(2) - 202%],
P001-SS012-01DL [Decachlorobiphenyl(1) - 219%, Decachlorobiphenyl(2) - 238%],
P001-SS021-01 [Decachlorobiphenyl(1) - 187%, Decachlorobiphenyl(2) - 193%],
P001-SS025-01 [Decachlorobiphenyl(1) - 250%, Decachlorobiphenyl(2) - 222%],
P001-SS026-01 [Decachlorobiphenyl(1) - 373%, Decachlorobiphenyl(2) - 332%],
P001-SS028-01 [Decachlorobiphenyl(1) - 423%, Decachlorobiphenyl(2) - 358%],
P001-SS035-01 [Decachlorobiphenyl(1) - 179%, Decachlorobiphenyl(2) - 181%],
P001-SS025-01DL [Decachlorobiphenyl(1) - 225%, Decachlorobiphenyl(2) - 240%],
P001-SS026-01DL [Decachlorobiphenyl(1) - 380%, Decachlorobiphenyl(2) - 418%],
P001-SS033-01DL [Decachlorobiphenyl(1) - 181%, Decachlorobiphenyl(2) - 180%],

P001-SS034-01DL [Decachlorobiphenyl(1) - 188%, Decachlorobiphenyl(2) - 254%],
P001-SS035-01DL [Decachlorobiphenyl(1) - 196%, Decachlorobiphenyl(2) - 229%],
P001-SS031-01DL [Decachlorobiphenyl(1) - 247%, Decachlorobiphenyl(2) - 283%],
P001-SS028-01DL [Decachlorobiphenyl(1) - 415%, Decachlorobiphenyl(2) - 434%],
P001-SS008-01 [Decachlorobiphenyl(1) - 204%, Decachlorobiphenyl(2) - 407%,
Tetrachloro-m-xylene(1) - 184%],
P001-SS029-01 [Tetrachloro-m-xylene(1) - 149%, Tetrachloro-m-xylene(2) - 151%],
Samples failed for surrogate on both column due to bad matrix. These samples were
required further dilution as well due to high concentration. Therefore original and
dilution analysis were reported and no further corrective action taken also
P001-SS009-01DL [Decachlorobiphenyl(1) - 0%, Decachlorobiphenyl(2) - 0%,
Tetrachloro-m-xylene(1) - 0%, Tetrachloro-m-xylene(2) - 0%],
P001-SS032-01DL2 [Decachlorobiphenyl(1) - 0%, Decachlorobiphenyl(2) - 0%,
Tetrachloro-m-xylene(1) - 0%, Tetrachloro-m-xylene(2) - 0%],
P001-SS035-01DL2 [Decachlorobiphenyl(1) - 0%, Decachlorobiphenyl(2) - 0%,
Tetrachloro-m-xylene(1) - 0% and Tetrachloro-m-xylene(2) - 0%],
P001-SS031-01DL2 [Decachlorobiphenyl(1) - 0%, Decachlorobiphenyl(2) - 0%,
Tetrachloro-m-xylene(1) - 0%, Tetrachloro-m-xylene(2) - 0%],
P001-SS008-01DL [Decachlorobiphenyl(1) - 0%, Decachlorobiphenyl(2) - 0%,
Tetrachloro-m-xylene(1) - 0%, Tetrachloro-m-xylene(2) - 0%],
Surrogates were diluted out due to the high dilution. No further corrective action was
taken.

The Retention Times were acceptable for all samples.

The MS {Q1860-20MS} with File ID: PO110728.D recoveries met the requirements
for all compounds except for AR1016[649%], AR1260[316%] due to matrix
interference.

The MSD {Q1860-21MSD} with File ID: PO110729.D recoveries met the acceptable
requirements except for AR1016[698%], AR1260[321%] due to matrix interference.

The RPD met criteria .

The Blank Spike met requirements for all samples .

The Blank analysis did not indicate the presence of lab contamination.

The Initial Calibration met the requirements .

The Continuous Calibration File ID PO110750.D met the requirements except for
Aroclor-1016(Peak-02),Tetrachloro-m-xylene is failing in 1st column but passing but
passing in 2nd column therefore no corrective action taken.

The Continuous Calibration File ID PO110783.D met the requirements except for
Tetrachloro-m-xylene is failing in 1st column but passing but passing in 2nd column
therefore no corrective action taken.

Samples P001-SS006-01, P001-SS007-01, P001-SS008-01, P001-SS009-01, P001-SS012-01, P001-SS013-01, P001-SS016-01, P001-SS017-01, P001-SS017-02, P001-SS019-01, P001-SS020-01, P001-SS021-01, P001-SS023-01, P001-SS025-01, P001-SS026-01, P001-SS028-01, P001-SS029-01, P001-SS030-01, P001-SS031-01, P001-SS031-01DL, P001-SS032-01, P001-SS032-01DL, P001-SS033-01, P001-SS034-01, P001-SS035-01 and P001-SS035-01DL were diluted due to high concentrations.

E. Additional Comments:

The soil samples results are based on a dry weight basis.

F. Calculation for Concentration in Soil samples:

$$\text{Concentration ug/Kg (Dry weight basis)} = \frac{(Ax) (Vt) (DF) (GPC)}{(CF) (Vi) (Ws) (D)}$$

Where,

Ax = Response (peak area or height) of the compound to be measured.

CF = Mean Calibration Factor from the initial calibration (area/ng).

Vt = Volume of the concentrated extract in uL

Vi = Volume of extract injected (uL). (If a single injection is made onto two columns, use ½ the volume in the syringe as the volume injected onto each column).

Ws = Weight of sample extracted (g).

D = % dry weight or $\frac{100 - \% \text{Moisture}}{100}$

GPC = $\frac{V_{in}}{V_{out}}$ = GPC factor (If no GPC is performed, GPC=1)

DF = Dilution Factor

G. Manual Integration Comments:

Please refer to the Manual integration Report included with the Run Logs for information on the manual integrations performed.

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature_____

APPROVED

By Nimisha Pandya, QA/QC Supervisor at 2:36 pm, May 07, 2025

DATA REPORTING QUALIFIERS- ORGANIC

For reporting results, the following “ Results Qualifiers” are used:

Value	If the result is a value greater than or equal to the detection limit, report the value
U	Indicates the compound was analyzed for but was not detected. Report the minimum detection limit for the sample with the U, i.e. “10 U”. This is not necessarily the instrument detection limit attainable for this particular sample based on any concentration or dilution that may have been required.
ND	Indicates the analyte was analyzed for, but not detected
J	Indicates an estimated value. This flag is used: (1) When estimating a concentration for a tentatively identified compound (library search hits, where a 1:1 response is assumed.) (2) When the mass spectral data indicated the identification, however the result was less than the specified detection limit greater than zero. If the detection limit was 10ug/L and a concentration of 3 ug/L was calculated report as 3 J. This flag is used when similar situation arise on any organic parameter i.e. Pest, PCB and others.
B	Indicates the analyte was found in the blank as well as the sample report as “12 B”.
E	Indicates the analyte ‘s concentration exceeds the calibrated range of the instrument for that specific analysis.
D	This flag identifies all compounds identified in an analysis at a secondary dilution factor.
P	This flag is used for Pesticide/PCB target analyte when there is >25% difference for detected concentrations between the two GC columns. The lower of the two values is reported on Form 1 and flagged with a “P”.
N	This flag indicates presumptive evidence of a compound. This is only used for tentatively identified compounds (TICs), where the identification is based on a mass spectral library search. It applies to all TIC results. For generic characterization of a TIC, such as chlorinated hydrocarbon, the flag is not used.
A	This flag indicates that a Tentatively Identified Compound is a suspected aldol-condensation product.
Q	Indicates the LCS did not meet the control limits requirements

APPENDIX A

QA REVIEW GENERAL DOCUMENTATION

Project #: Q1860

Completed

For thorough review, the report must have the following:

GENERAL:

Are all original paperwork present (chain of custody, record of communication,airbill, sample management lab chronicle, login page)

✓

Check chain-of-custody for proper relinquish/return of samples

✓

Is the chain of custody signed and complete

✓

Check internal chain-of-custody for proper relinquish/return of samples /sample extracts

✓

Collect information for each project id from server. Were all requirements followed

✓

COVER PAGE:

Do numbers of samples correspond to the number of samples in the Chain of Custody on login page

✓

Do lab numbers and client Ids on cover page agree with the Chain of Custody

✓

CHAIN OF CUSTODY:

Do requested analyses on Chain of Custody agree with form I results

✓

Do requested analyses on Chain of Custody agree with the log-in page

✓

Were the correct method log-in for analysis according to the Analytical Request and Chain of Custody

✓

Were the samples received within hold time

✓

Were any problems found with the samples at arrival recorded in the Sample Management Laboratory Chronicle

✓

ANALYTICAL:

Was method requirement followed?

✓

Was client requirement followed?

✓

Does the case narrative summarize all QC failure?

✓

All runlogs and manual integration are reviewed for requirements

✓

All manual calculations and /or hand notations verified

✓

QA Review Signature: SOHIL JODHANI

Date: 05/06/2025

Hit Summary Sheet SW-846

SDG No.: Q1860
Client: Weston Solutions, Inc.

Order ID: Q1860
Project ID: Q1908

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
Client ID : P001-SS001-01								
Q1860-01	P001-SS001-01	SOIL	Aroclor-1260	58.6		3.70	19.6	ug/kg
Total Concentration:				58.600				
Client ID : P001-SS002-01								
Q1860-04	P001-SS002-01	SOIL	Aroclor-1254	200		3.90	20.8	ug/kg
Total Concentration:				200.000				
Client ID : P001-SS002-02								
Q1860-05	P001-SS002-02	SOIL	Aroclor-1254	302		3.80	20.2	ug/kg
Total Concentration:				302.000				
Client ID : P001-SS003-01								
Q1860-06	P001-SS003-01	SOIL	Aroclor-1254	154		3.70	19.4	ug/kg
Q1860-06	P001-SS003-01	SOIL	Aroclor-1260	126		3.70	19.4	ug/kg
Total Concentration:				280.000				
Client ID : P001-SS004-01								
Q1860-07	P001-SS004-01	SOIL	Aroclor-1254	179		3.70	19.6	ug/kg
Q1860-07	P001-SS004-01	SOIL	Aroclor-1260	198		3.70	19.6	ug/kg
Total Concentration:				377.000				
Client ID : P001-SS005-01								
Q1860-08	P001-SS005-01	SOIL	Aroclor-1260	299		3.70	19.7	ug/kg
Total Concentration:				299.000				
Client ID : P001-SS006-01								
Q1860-09	P001-SS006-01	SOIL	Aroclor-1260	439 E		3.60	18.9	ug/kg
Total Concentration:				439.000				
Client ID : P001-SS006-01DL								
Q1860-09DL	P001-SS006-01DL	SOIL	Aroclor-1260	375 D		18.0	94.7	ug/kg
Total Concentration:				375.000				
Client ID : P001-SS007-01								
Q1860-10	P001-SS007-01	SOIL	Aroclor-1248	4200 E		6.80	19.4	ug/kg
Q1860-10	P001-SS007-01	SOIL	Aroclor-1254	3900 E		3.70	19.4	ug/kg
Total Concentration:				8,100.000				

Hit Summary Sheet SW-846

SDG No.: Q1860

Order ID: Q1860

Client: Weston Solutions, Inc.

Project ID: Q1908

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
Client ID : P001-SS007-01DL								
Q1860-10DL	P001-SS007-01DL	SOIL	Aroclor-1248	4000	D	169	485	ug/kg
Q1860-10DL	P001-SS007-01DL	SOIL	Aroclor-1254	3400	D	91.6	485	ug/kg
Total Concentration:				7,400.000				
Client ID : P001-SS008-01								
Q1860-11	P001-SS008-01	SOIL	Aroclor-1248	7200	E	6.70	19.1	ug/kg
Q1860-11	P001-SS008-01	SOIL	Aroclor-1254	7000	E	3.60	19.1	ug/kg
Total Concentration:				14,200.000				
Client ID : P001-SS008-01DL								
Q1860-11DL	P001-SS008-01DL	SOIL	Aroclor-1248	9900	D	666	1900	ug/kg
Q1860-11DL	P001-SS008-01DL	SOIL	Aroclor-1254	7900	D	361	1900	ug/kg
Total Concentration:				17,800.000				
Client ID : P001-SS009-01								
Q1860-12	P001-SS009-01	SOIL	Aroclor-1248	7300	E	6.80	19.6	ug/kg
Q1860-12	P001-SS009-01	SOIL	Aroclor-1254	5700	E	3.70	19.6	ug/kg
Total Concentration:				13,000.000				
Client ID : P001-SS009-01DL								
Q1860-12DL	P001-SS009-01DL	SOIL	Aroclor-1248	9400	D	684	2000	ug/kg
Q1860-12DL	P001-SS009-01DL	SOIL	Aroclor-1254	6700	D	371	2000	ug/kg
Total Concentration:				16,100.000				
Client ID : P001-SS010-01								
Q1860-13	P001-SS010-01	SOIL	Aroclor-1254	57.4		4.00	21.0	ug/kg
Q1860-13	P001-SS010-01	SOIL	Aroclor-1260	54.6		4.00	21.0	ug/kg
Total Concentration:				112.000				
Client ID : P001-SS011-01								
Q1860-14	P001-SS011-01	SOIL	Aroclor-1248	325		7.10	20.5	ug/kg
Q1860-14	P001-SS011-01	SOIL	Aroclor-1254	282		3.90	20.5	ug/kg
Total Concentration:				607.000				
Client ID : P001-SS012-01								
Q1860-15	P001-SS012-01	SOIL	Aroclor-1248	6500	E	7.30	20.9	ug/kg
Q1860-15	P001-SS012-01	SOIL	Aroclor-1254	3900	E	3.90	20.9	ug/kg
Total Concentration:				10,400.000				

Hit Summary Sheet

SW-846

SDG No.: Q1860

Order ID: Q1860

Client: Weston Solutions, Inc.

Project ID: Q1908

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
Client ID : P001-SS012-01DL								
Q1860-15DL	P001-SS012-01DL	SOIL	Aroclor-1248	7000	D	182	522	ug/kg
Q1860-15DL	P001-SS012-01DL	SOIL	Aroclor-1254	3900	D	98.6	522	ug/kg
Total Concentration:				10,900.000				
Client ID : P001-SS013-01								
Q1860-16	P001-SS013-01	SOIL	Aroclor-1248	4300	E	6.90	19.9	ug/kg
Q1860-16	P001-SS013-01	SOIL	Aroclor-1254	3200	E	3.80	19.9	ug/kg
Total Concentration:				7,500.000				
Client ID : P001-SS013-01DL								
Q1860-16DL	P001-SS013-01DL	SOIL	Aroclor-1248	5100	D	173	498	ug/kg
Q1860-16DL	P001-SS013-01DL	SOIL	Aroclor-1254	3500	D	94.0	498	ug/kg
Total Concentration:				8,600.000				
Client ID : P001-SS014-01								
Q1860-17	P001-SS014-01	SOIL	Aroclor-1248	398		7.40	21.3	ug/kg
Q1860-17	P001-SS014-01	SOIL	Aroclor-1254	412		4.00	21.3	ug/kg
Total Concentration:				810.000				
Client ID : P001-SS015-01								
Q1860-18	P001-SS015-01	SOIL	Aroclor-1248	71.7		7.90	22.7	ug/kg
Q1860-18	P001-SS015-01	SOIL	Aroclor-1254	102		4.30	22.7	ug/kg
Total Concentration:				173.700				
Client ID : P001-SS016-01								
Q1860-19	P001-SS016-01	SOIL	Aroclor-1248	932	E	7.10	20.4	ug/kg
Q1860-19	P001-SS016-01	SOIL	Aroclor-1254	734	E	3.90	20.4	ug/kg
Total Concentration:				1,666.000				
Client ID : P001-SS016-01DL								
Q1860-19DL	P001-SS016-01DL	SOIL	Aroclor-1248	986	D	71.1	204	ug/kg
Q1860-19DL	P001-SS016-01DL	SOIL	Aroclor-1254	725	D	38.6	204	ug/kg
Total Concentration:				1,711.000				
Client ID : P001-SS017-01								
Q1860-22	P001-SS017-01	SOIL	Aroclor-1248	770	E	7.00	20.1	ug/kg
Q1860-22	P001-SS017-01	SOIL	Aroclor-1254	759	E	3.80	20.1	ug/kg

Hit Summary Sheet SW-846

SDG No.: Q1860
Client: Weston Solutions, Inc.

Order ID: Q1860
Project ID: Q1908

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
Total Concentration:				1,529.000				
Client ID : P001-SS017-01DL								
Q1860-22DL	P001-SS017-01DL	SOIL	Aroclor-1248	749	D	34.9	100	ug/kg
Q1860-22DL	P001-SS017-01DL	SOIL	Aroclor-1254	663	D	18.9	100	ug/kg
Total Concentration:				1,412.000				
Client ID : P001-SS017-02								
Q1860-23	P001-SS017-02	SOIL	Aroclor-1248	925	E	7.20	20.6	ug/kg
Q1860-23	P001-SS017-02	SOIL	Aroclor-1254	631	E	3.90	20.6	ug/kg
Total Concentration:				1,556.000				
Client ID : P001-SS017-02DL								
Q1860-23DL	P001-SS017-02DL	SOIL	Aroclor-1248	878	D	35.8	103	ug/kg
Q1860-23DL	P001-SS017-02DL	SOIL	Aroclor-1254	525	D	19.4	103	ug/kg
Total Concentration:				1,403.000				
Client ID : P001-SS018-01								
Q1860-24	P001-SS018-01	SOIL	Aroclor-1254	76.3		3.60	18.9	ug/kg
Total Concentration:				76.300				
Client ID : P001-SS019-01								
Q1860-25	P001-SS019-01	SOIL	Aroclor-1254	435	E	3.60	19.2	ug/kg
Q1860-25	P001-SS019-01	SOIL	Aroclor-1260	297		3.70	19.2	ug/kg
Total Concentration:				732.000				
Client ID : P001-SS019-01DL								
Q1860-25DL	P001-SS019-01DL	SOIL	Aroclor-1254	435	D	18.1	96.1	ug/kg
Q1860-25DL	P001-SS019-01DL	SOIL	Aroclor-1260	298	D	18.3	96.1	ug/kg
Total Concentration:				733.000				
Client ID : P001-SS020-01								
Q1860-26	P001-SS020-01	SOIL	Aroclor-1254	5900	EP	3.80	20.0	ug/kg
Total Concentration:				5,900.000				
Client ID : P001-SS020-01DL								
Q1860-26DL	P001-SS020-01DL	SOIL	Aroclor-1254	4500	D	94.3	500	ug/kg
Total Concentration:				4,500.000				

Hit Summary Sheet SW-846

SDG No.: Q1860
Client: Weston Solutions, Inc.

Order ID: Q1860
Project ID: Q1908

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
Client ID : P001-SS021-01								
Q1860-27	P001-SS021-01	SOIL	Aroclor-1254	381	E	3.60	19.2	ug/kg
Total Concentration:				381.000				
Client ID : P001-SS021-01DL								
Q1860-27DL	P001-SS021-01DL	SOIL	Aroclor-1254	400	D	18.1	95.9	ug/kg
Total Concentration:				400.000				
Client ID : P001-SS022-01								
Q1860-28	P001-SS022-01	SOIL	Aroclor-1254	203		3.80	20.3	ug/kg
Total Concentration:				203.000				
Client ID : P001-SS023-01								
Q1860-29	P001-SS023-01	SOIL	Aroclor-1254	410	E	3.80	20.2	ug/kg
Total Concentration:				410.000				
Client ID : P001-SS023-01DL								
Q1860-29DL	P001-SS023-01DL	SOIL	Aroclor-1254	361	D	19.1	101	ug/kg
Total Concentration:				361.000				
Client ID : P001-SS024-01								
Q1860-30	P001-SS024-01	SOIL	Aroclor-1254	347		3.80	20.1	ug/kg
Total Concentration:				347.000				
Client ID : P001-SS025-01								
Q1860-31	P001-SS025-01	SOIL	Aroclor-1254	1200	E	3.70	19.5	ug/kg
Q1860-31	P001-SS025-01	SOIL	Aroclor-1260	1200	E	3.70	19.5	ug/kg
Total Concentration:				2,400.000				
Client ID : P001-SS025-01DL								
Q1860-31DL	P001-SS025-01DL	SOIL	Aroclor-1254	1200	D	18.5	97.7	ug/kg
Q1860-31DL	P001-SS025-01DL	SOIL	Aroclor-1260	1300	D	18.6	97.7	ug/kg
Total Concentration:				2,500.000				
Client ID : P001-SS026-01								
Q1860-32	P001-SS026-01	SOIL	Aroclor-1254	3800	E	4.10	21.5	ug/kg
Total Concentration:				3,800.000				
Client ID : P001-SS026-01DL								

Hit Summary Sheet SW-846

SDG No.: Q1860

Order ID: Q1860

Client: Weston Solutions, Inc.

Project ID: Q1908

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
Q1860-32DL	P001-SS026-01DL	SOIL	Aroclor-1254	4800	D	81.0	429	ug/kg
Total Concentration:				4,800.000				
Client ID : P001-SS027-01								
Q1860-33	P001-SS027-01	SOIL	Aroclor-1248	114		6.80	19.6	ug/kg
Q1860-33	P001-SS027-01	SOIL	Aroclor-1254	225		3.70	19.6	ug/kg
Q1860-33	P001-SS027-01	SOIL	Aroclor-1260	176		3.70	19.6	ug/kg
Total Concentration:				515.000				
Client ID : P001-SS028-01								
Q1860-34	P001-SS028-01	SOIL	Aroclor-1248	1800	E	7.60	21.8	ug/kg
Q1860-34	P001-SS028-01	SOIL	Aroclor-1254	3000	E	4.10	21.8	ug/kg
Q1860-34	P001-SS028-01	SOIL	Aroclor-1260	2100	E	4.20	21.8	ug/kg
Total Concentration:				6,900.000				
Client ID : P001-SS028-01DL								
Q1860-34DL	P001-SS028-01DL	SOIL	Aroclor-1248	2100	D	152	437	ug/kg
Q1860-34DL	P001-SS028-01DL	SOIL	Aroclor-1254	2900	D	82.5	437	ug/kg
Q1860-34DL	P001-SS028-01DL	SOIL	Aroclor-1260	2300	D	83.0	437	ug/kg
Total Concentration:				7,300.000				
Client ID : P001-SS029-01								
Q1860-35	P001-SS029-01	SOIL	Aroclor-1248	317		7.30	21.1	ug/kg
Q1860-35	P001-SS029-01	SOIL	Aroclor-1254	781	E	4.00	21.1	ug/kg
Q1860-35	P001-SS029-01	SOIL	Aroclor-1260	532	E	4.00	21.1	ug/kg
Total Concentration:				1,630.000				
Client ID : P001-SS029-01DL								
Q1860-35DL	P001-SS029-01DL	SOIL	Aroclor-1248	294	D	36.7	105	ug/kg
Q1860-35DL	P001-SS029-01DL	SOIL	Aroclor-1254	665	D	19.9	105	ug/kg
Q1860-35DL	P001-SS029-01DL	SOIL	Aroclor-1260	504	D	20.0	105	ug/kg
Total Concentration:				1,463.000				
Client ID : P001-SS030-01								
Q1860-36	P001-SS030-01	SOIL	Aroclor-1248	2100	E	7.10	20.3	ug/kg
Q1860-36	P001-SS030-01	SOIL	Aroclor-1254	3300	E	3.80	20.3	ug/kg
Q1860-36	P001-SS030-01	SOIL	Aroclor-1260	3500	E	3.90	20.3	ug/kg
Total Concentration:				8,900.000				

Hit Summary Sheet SW-846

SDG No.: Q1860

Order ID: Q1860

Client: Weston Solutions, Inc.

Project ID: Q1908

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
Client ID : P001-SS030-01DL								
Q1860-36DL	P001-SS030-01DL	SOIL	Aroclor-1248	2800	D	141	405	ug/kg
Q1860-36DL	P001-SS030-01DL	SOIL	Aroclor-1254	4100	D	76.5	405	ug/kg
Q1860-36DL	P001-SS030-01DL	SOIL	Aroclor-1260	3900	D	77.0	405	ug/kg
Total Concentration:				10,800.000				
Client ID : P001-SS031-01								
Q1860-37	P001-SS031-01	SOIL	Aroclor-1254	7700	E	3.60	19.1	ug/kg
Total Concentration:				7,700.000				
Client ID : P001-SS031-01DL								
Q1860-37DL	P001-SS031-01DL	SOIL	Aroclor-1254	9100	ED	72.0	381	ug/kg
Total Concentration:				9,100.000				
Client ID : P001-SS031-01DL2								
Q1860-37DL2	P001-SS031-01DL2	SOIL	Aroclor-1254	9100	D	360	1900	ug/kg
Total Concentration:				9,100.000				
Client ID : P001-SS032-01								
Q1860-38	P001-SS032-01	SOIL	Aroclor-1248	2100	E	6.80	19.5	ug/kg
Q1860-38	P001-SS032-01	SOIL	Aroclor-1254	5300	E	3.70	19.5	ug/kg
Total Concentration:				7,400.000				
Client ID : P001-SS032-01DL								
Q1860-38DL	P001-SS032-01DL	SOIL	Aroclor-1248	1900	D	68.0	195	ug/kg
Q1860-38DL	P001-SS032-01DL	SOIL	Aroclor-1254	4000	ED	36.9	195	ug/kg
Total Concentration:				5,900.000				
Client ID : P001-SS032-01DL2								
Q1860-38DL2	P001-SS032-01DL2	SOIL	Aroclor-1248	2400	D	340	976	ug/kg
Q1860-38DL2	P001-SS032-01DL2	SOIL	Aroclor-1254	4200	D	184	976	ug/kg
Total Concentration:				6,600.000				
Client ID : P001-SS033-01								
Q1860-39	P001-SS033-01	SOIL	Aroclor-1248	3900	E	6.70	19.4	ug/kg
Q1860-39	P001-SS033-01	SOIL	Aroclor-1254	3600	E	3.70	19.4	ug/kg
Total Concentration:				7,500.000				
Client ID : P001-SS033-01DL								

Hit Summary Sheet SW-846

SDG No.: Q1860

Order ID: Q1860

Client: Weston Solutions, Inc.

Project ID: Q1908

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
Q1860-39DL	P001-SS033-01DL	SOIL	Aroclor-1248	4900	D	135	387	ug/kg
Q1860-39DL	P001-SS033-01DL	SOIL	Aroclor-1254	4000	D	73.1	387	ug/kg
Total Concentration:				8,900.000				
Client ID : P001-SS034-01								
Q1860-40	P001-SS034-01	SOIL	Aroclor-1248	1700	E	6.70	19.3	ug/kg
Q1860-40	P001-SS034-01	SOIL	Aroclor-1254	2700	E	3.60	19.3	ug/kg
Total Concentration:				4,400.000				
Client ID : P001-SS034-01DL								
Q1860-40DL	P001-SS034-01DL	SOIL	Aroclor-1248	2600	D	135	386	ug/kg
Q1860-40DL	P001-SS034-01DL	SOIL	Aroclor-1254	3600	D	72.9	386	ug/kg
Total Concentration:				6,200.000				
Client ID : P001-SS035-01								
Q1860-41	P001-SS035-01	SOIL	Aroclor-1248	2500	E	6.30	18.1	ug/kg
Q1860-41	P001-SS035-01	SOIL	Aroclor-1254	3900	E	3.40	18.1	ug/kg
Total Concentration:				6,400.000				
Client ID : P001-SS035-01DL								
Q1860-41DL	P001-SS035-01DL	SOIL	Aroclor-1248	2900	D	63.0	181	ug/kg
Q1860-41DL	P001-SS035-01DL	SOIL	Aroclor-1254	4100	ED	34.1	181	ug/kg
Total Concentration:				7,000.000				
Client ID : P001-SS035-01DL2								
Q1860-41DL2	P001-SS035-01DL2	SOIL	Aroclor-1248	3200	D	315	904	ug/kg
Q1860-41DL2	P001-SS035-01DL2	SOIL	Aroclor-1254	4200	DP	171	904	ug/kg
Total Concentration:				7,400.000				
Client ID : P001-SS036-01								
Q1860-42	P001-SS036-01	SOIL	Aroclor-1254	19.6	P	3.40	18.0	ug/kg
Total Concentration:				19.600				



SAMPLE DATA

Report of Analysis

Client:	Weston Solutions, Inc.		Date Collected:	04/16/25	
Project:	RFP 911		Date Received:	04/23/25	
Client Sample ID:	P001-SS001-01		SDG No.:	Q1860	
Lab Sample ID:	Q1860-01		Matrix:	SOIL	
Analytical Method:	SW8082A		% Solid:	86.7	Decanted:
Sample Wt/Vol:	30.07	Units: g	Final Vol:	10000	uL
Soil Aliquot Vol:		uL	Test:	PCB	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	SW3541B				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PP071430.D	1	04/24/25 08:32	04/24/25 13:30	PB167721

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	4.50	U	4.50	19.6	ug/kg
11104-28-2	Aroclor-1221	4.60	U	4.60	19.6	ug/kg
11141-16-5	Aroclor-1232	4.30	U	4.30	19.6	ug/kg
53469-21-9	Aroclor-1242	4.60	U	4.60	19.6	ug/kg
12672-29-6	Aroclor-1248	6.80	U	6.80	19.6	ug/kg
11097-69-1	Aroclor-1254	3.70	U	3.70	19.6	ug/kg
37324-23-5	Aroclor-1262	5.80	U	5.80	19.6	ug/kg
11100-14-4	Aroclor-1268	4.10	U	4.10	19.6	ug/kg
11096-82-5	Aroclor-1260	58.6		3.70	19.6	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	18.6		32 - 144	93%	SPK: 20
2051-24-3	Decachlorobiphenyl	16.5		32 - 175	82%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	Weston Solutions, Inc.		Date Collected:	04/16/25	
Project:	RFP 911		Date Received:	04/23/25	
Client Sample ID:	P001-SS002-01		SDG No.:	Q1860	
Lab Sample ID:	Q1860-04		Matrix:	SOIL	
Analytical Method:	SW8082A		% Solid:	81.8	Decanted:
Sample Wt/Vol:	30.02	Units: g	Final Vol:	10000	uL
Soil Aliquot Vol:		uL	Test:	PCB	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	SW3541B				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PP071433.D	1	04/24/25 08:32	04/24/25 14:19	PB167721

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	4.80	U	4.80	20.8	ug/kg
11104-28-2	Aroclor-1221	4.90	U	4.90	20.8	ug/kg
11141-16-5	Aroclor-1232	4.50	U	4.50	20.8	ug/kg
53469-21-9	Aroclor-1242	4.90	U	4.90	20.8	ug/kg
12672-29-6	Aroclor-1248	7.20	U	7.20	20.8	ug/kg
11097-69-1	Aroclor-1254	200		3.90	20.8	ug/kg
37324-23-5	Aroclor-1262	6.10	U	6.10	20.8	ug/kg
11100-14-4	Aroclor-1268	4.40	U	4.40	20.8	ug/kg
11096-82-5	Aroclor-1260	3.90	U	3.90	20.8	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	15.9		32 - 144	80%	SPK: 20
2051-24-3	Decachlorobiphenyl	31.6		32 - 175	158%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	Weston Solutions, Inc.		Date Collected:	04/16/25	
Project:	RFP 911		Date Received:	04/23/25	
Client Sample ID:	P001-SS002-02		SDG No.:	Q1860	
Lab Sample ID:	Q1860-05		Matrix:	SOIL	
Analytical Method:	SW8082A		% Solid:	84.1	Decanted:
Sample Wt/Vol:	30.01	Units: g	Final Vol:	10000	uL
Soil Aliquot Vol:		uL	Test:	PCB	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	SW3541B				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PO110678.D	1	04/24/25 08:32	04/24/25 13:01	PB167721

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	4.70	U	4.70	20.2	ug/kg
11104-28-2	Aroclor-1221	4.80	U	4.80	20.2	ug/kg
11141-16-5	Aroclor-1232	4.40	U	4.40	20.2	ug/kg
53469-21-9	Aroclor-1242	4.80	U	4.80	20.2	ug/kg
12672-29-6	Aroclor-1248	7.00	U	7.00	20.2	ug/kg
11097-69-1	Aroclor-1254	302		3.80	20.2	ug/kg
37324-23-5	Aroclor-1262	6.00	U	6.00	20.2	ug/kg
11100-14-4	Aroclor-1268	4.30	U	4.30	20.2	ug/kg
11096-82-5	Aroclor-1260	3.80	U	3.80	20.2	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	18.2		32 - 144	91%	SPK: 20
2051-24-3	Decachlorobiphenyl	21.8		32 - 175	109%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	Weston Solutions, Inc.		Date Collected:	04/16/25	
Project:	RFP 911		Date Received:	04/23/25	
Client Sample ID:	P001-SS003-01		SDG No.:	Q1860	
Lab Sample ID:	Q1860-06		Matrix:	SOIL	
Analytical Method:	SW8082A		% Solid:	87.6	Decanted:
Sample Wt/Vol:	30.06	Units: g	Final Vol:	10000	uL
Soil Aliquot Vol:		uL	Test:	PCB	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	SW3541B				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PO110679.D	1	04/24/25 08:32	04/24/25 13:19	PB167721

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	4.50	U	4.50	19.4	ug/kg
11104-28-2	Aroclor-1221	4.60	U	4.60	19.4	ug/kg
11141-16-5	Aroclor-1232	4.20	U	4.20	19.4	ug/kg
53469-21-9	Aroclor-1242	4.60	U	4.60	19.4	ug/kg
12672-29-6	Aroclor-1248	6.70	U	6.70	19.4	ug/kg
11097-69-1	Aroclor-1254	154		3.70	19.4	ug/kg
37324-23-5	Aroclor-1262	5.70	U	5.70	19.4	ug/kg
11100-14-4	Aroclor-1268	4.10	U	4.10	19.4	ug/kg
11096-82-5	Aroclor-1260	126		3.70	19.4	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	20.4		32 - 144	102%	SPK: 20
2051-24-3	Decachlorobiphenyl	15.0		32 - 175	75%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	Weston Solutions, Inc.		Date Collected:	04/16/25	
Project:	RFP 911		Date Received:	04/23/25	
Client Sample ID:	P001-SS004-01		SDG No.:	Q1860	
Lab Sample ID:	Q1860-07		Matrix:	SOIL	
Analytical Method:	SW8082A		% Solid:	86.5	Decanted:
Sample Wt/Vol:	30.03	Units: g	Final Vol:	10000	uL
Soil Aliquot Vol:		uL	Test:	PCB	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	SW3541B				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PO110680.D	1	04/24/25 08:32	04/24/25 13:37	PB167721

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	4.60	U	4.60	19.6	ug/kg
11104-28-2	Aroclor-1221	4.70	U	4.70	19.6	ug/kg
11141-16-5	Aroclor-1232	4.30	U	4.30	19.6	ug/kg
53469-21-9	Aroclor-1242	4.60	U	4.60	19.6	ug/kg
12672-29-6	Aroclor-1248	6.80	U	6.80	19.6	ug/kg
11097-69-1	Aroclor-1254	179		3.70	19.6	ug/kg
37324-23-5	Aroclor-1262	5.80	U	5.80	19.6	ug/kg
11100-14-4	Aroclor-1268	4.20	U	4.20	19.6	ug/kg
11096-82-5	Aroclor-1260	198		3.70	19.6	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	17.3		32 - 144	86%	SPK: 20
2051-24-3	Decachlorobiphenyl	12.6		32 - 175	63%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	Weston Solutions, Inc.		Date Collected:	04/16/25	
Project:	RFP 911		Date Received:	04/23/25	
Client Sample ID:	P001-SS005-01		SDG No.:	Q1860	
Lab Sample ID:	Q1860-08		Matrix:	SOIL	
Analytical Method:	SW8082A		% Solid:	86.2	Decanted:
Sample Wt/Vol:	30.07	Units: g	Final Vol:	10000	uL
Soil Aliquot Vol:		uL	Test:	PCB	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	SW3541B				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PP071434.D	1	04/24/25 08:32	04/24/25 14:35	PB167721

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	4.60	U	4.60	19.7	ug/kg
11104-28-2	Aroclor-1221	4.70	U	4.70	19.7	ug/kg
11141-16-5	Aroclor-1232	4.30	U	4.30	19.7	ug/kg
53469-21-9	Aroclor-1242	4.60	U	4.60	19.7	ug/kg
12672-29-6	Aroclor-1248	6.90	U	6.90	19.7	ug/kg
11097-69-1	Aroclor-1254	3.70	U	3.70	19.7	ug/kg
37324-23-5	Aroclor-1262	5.80	U	5.80	19.7	ug/kg
11100-14-4	Aroclor-1268	4.20	U	4.20	19.7	ug/kg
11096-82-5	Aroclor-1260	299		3.70	19.7	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	17.8		32 - 144	89%	SPK: 20
2051-24-3	Decachlorobiphenyl	15.2		32 - 175	76%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	Weston Solutions, Inc.		Date Collected:	04/16/25	
Project:	RFP 911		Date Received:	04/23/25	
Client Sample ID:	P001-SS006-01		SDG No.:	Q1860	
Lab Sample ID:	Q1860-09		Matrix:	SOIL	
Analytical Method:	SW8082A		% Solid:	89.5	Decanted:
Sample Wt/Vol:	30.08	Units: g	Final Vol:	10000	uL
Soil Aliquot Vol:		uL	Test:	PCB	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	SW3541B				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PP071645.D	1	04/24/25 08:32	04/30/25 13:44	PB167721

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	4.40	U	4.40	18.9	ug/kg
11104-28-2	Aroclor-1221	4.50	U	4.50	18.9	ug/kg
11141-16-5	Aroclor-1232	4.10	U	4.10	18.9	ug/kg
53469-21-9	Aroclor-1242	4.50	U	4.50	18.9	ug/kg
12672-29-6	Aroclor-1248	6.60	U	6.60	18.9	ug/kg
11097-69-1	Aroclor-1254	3.60	U	3.60	18.9	ug/kg
37324-23-5	Aroclor-1262	5.60	U	5.60	18.9	ug/kg
11100-14-4	Aroclor-1268	4.00	U	4.00	18.9	ug/kg
11096-82-5	Aroclor-1260	439	E	3.60	18.9	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	21.8		32 - 144	109%	SPK: 20
2051-24-3	Decachlorobiphenyl	20.4		32 - 175	102%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	Weston Solutions, Inc.		Date Collected:	04/16/25	
Project:	RFP 911		Date Received:	04/23/25	
Client Sample ID:	P001-SS006-01DL		SDG No.:	Q1860	
Lab Sample ID:	Q1860-09DL		Matrix:	SOIL	
Analytical Method:	SW8082A		% Solid:	89.5	Decanted:
Sample Wt/Vol:	30.08	Units: g	Final Vol:	10000	uL
Soil Aliquot Vol:		uL	Test:	PCB	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	SW3541B				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PP071646.D	5	04/24/25 08:32	04/30/25 14:01	PB167721

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	22.0	UD	22.0	94.7	ug/kg
11104-28-2	Aroclor-1221	22.5	UD	22.5	94.7	ug/kg
11141-16-5	Aroclor-1232	20.7	UD	20.7	94.7	ug/kg
53469-21-9	Aroclor-1242	22.3	UD	22.3	94.7	ug/kg
12672-29-6	Aroclor-1248	33.0	UD	33.0	94.7	ug/kg
11097-69-1	Aroclor-1254	17.9	UD	17.9	94.7	ug/kg
37324-23-5	Aroclor-1262	28.0	UD	28.0	94.7	ug/kg
11100-14-4	Aroclor-1268	20.1	UD	20.1	94.7	ug/kg
11096-82-5	Aroclor-1260	375	D	18.0	94.7	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	18.6		32 - 144	93%	SPK: 20
2051-24-3	Decachlorobiphenyl	16.4		32 - 175	82%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	Weston Solutions, Inc.		Date Collected:	04/16/25	
Project:	RFP 911		Date Received:	04/23/25	
Client Sample ID:	P001-SS007-01		SDG No.:	Q1860	
Lab Sample ID:	Q1860-10		Matrix:	SOIL	
Analytical Method:	SW8082A		% Solid:	87.4	Decanted:
Sample Wt/Vol:	30.06	Units: g	Final Vol:	10000	uL
Soil Aliquot Vol:		uL	Test:	PCB	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	SW3541B				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PP071436.D	1	04/24/25 08:32	04/24/25 15:08	PB167721

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	4.50	U	4.50	19.4	ug/kg
11104-28-2	Aroclor-1221	4.60	U	4.60	19.4	ug/kg
11141-16-5	Aroclor-1232	4.20	U	4.20	19.4	ug/kg
53469-21-9	Aroclor-1242	4.60	U	4.60	19.4	ug/kg
12672-29-6	Aroclor-1248	4200	E	6.80	19.4	ug/kg
11097-69-1	Aroclor-1254	3900	E	3.70	19.4	ug/kg
37324-23-5	Aroclor-1262	5.70	U	5.70	19.4	ug/kg
11100-14-4	Aroclor-1268	4.10	U	4.10	19.4	ug/kg
11096-82-5	Aroclor-1260	3.70	U	3.70	19.4	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	27.3		32 - 144	136%	SPK: 20
2051-24-3	Decachlorobiphenyl	42.9	*	32 - 175	215%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	Weston Solutions, Inc.		Date Collected:	04/16/25	
Project:	RFP 911		Date Received:	04/23/25	
Client Sample ID:	P001-SS007-01DL		SDG No.:	Q1860	
Lab Sample ID:	Q1860-10DL		Matrix:	SOIL	
Analytical Method:	SW8082A		% Solid:	87.4	Decanted:
Sample Wt/Vol:	30.06	Units: g	Final Vol:	10000	uL
Soil Aliquot Vol:		uL	Test:	PCB	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	SW3541B				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PP071447.D	25	04/24/25 08:32	04/24/25 19:45	PB167721

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	113	UD	113	485	ug/kg
11104-28-2	Aroclor-1221	115	UD	115	485	ug/kg
11141-16-5	Aroclor-1232	106	UD	106	485	ug/kg
53469-21-9	Aroclor-1242	114	UD	114	485	ug/kg
12672-29-6	Aroclor-1248	4000	D	169	485	ug/kg
11097-69-1	Aroclor-1254	3400	D	91.6	485	ug/kg
37324-23-5	Aroclor-1262	143	UD	143	485	ug/kg
11100-14-4	Aroclor-1268	103	UD	103	485	ug/kg
11096-82-5	Aroclor-1260	92.2	UD	92.2	485	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	21.8		32 - 144	109%	SPK: 20
2051-24-3	Decachlorobiphenyl	35.5	*	32 - 175	178%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	Weston Solutions, Inc.		Date Collected:	04/16/25	
Project:	RFP 911		Date Received:	04/23/25	
Client Sample ID:	P001-SS008-01		SDG No.:	Q1860	
Lab Sample ID:	Q1860-11		Matrix:	SOIL	
Analytical Method:	SW8082A		% Solid:	88.8	Decanted:
Sample Wt/Vol:	30.03	Units: g	Final Vol:	10000	uL
Soil Aliquot Vol:		uL	Test:	PCB	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	SW3541B				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PP071437.D	1	04/24/25 08:32	04/24/25 15:25	PB167721

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	4.40	U	4.40	19.1	ug/kg
11104-28-2	Aroclor-1221	4.50	U	4.50	19.1	ug/kg
11141-16-5	Aroclor-1232	4.20	U	4.20	19.1	ug/kg
53469-21-9	Aroclor-1242	4.50	U	4.50	19.1	ug/kg
12672-29-6	Aroclor-1248	7200	E	6.70	19.1	ug/kg
11097-69-1	Aroclor-1254	7000	E	3.60	19.1	ug/kg
37324-23-5	Aroclor-1262	5.60	U	5.60	19.1	ug/kg
11100-14-4	Aroclor-1268	4.10	U	4.10	19.1	ug/kg
11096-82-5	Aroclor-1260	3.60	U	3.60	19.1	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	36.9	*	32 - 144	184%	SPK: 20
2051-24-3	Decachlorobiphenyl	81.3	*	32 - 175	407%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	Weston Solutions, Inc.		Date Collected:	04/16/25	
Project:	RFP 911		Date Received:	04/23/25	
Client Sample ID:	P001-SS008-01DL		SDG No.:	Q1860	
Lab Sample ID:	Q1860-11DL		Matrix:	SOIL	
Analytical Method:	SW8082A		% Solid:	88.8	Decanted:
Sample Wt/Vol:	30.03	Units: g	Final Vol:	10000	uL
Soil Aliquot Vol:		uL	Test:	PCB	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	SW3541B				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PP071449.D	100	04/24/25 08:32	04/24/25 20:18	PB167721

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	444	UD	444	1900	ug/kg
11104-28-2	Aroclor-1221	453	UD	453	1900	ug/kg
11141-16-5	Aroclor-1232	419	UD	419	1900	ug/kg
53469-21-9	Aroclor-1242	451	UD	451	1900	ug/kg
12672-29-6	Aroclor-1248	9900	D	666	1900	ug/kg
11097-69-1	Aroclor-1254	7900	D	361	1900	ug/kg
37324-23-5	Aroclor-1262	565	UD	565	1900	ug/kg
11100-14-4	Aroclor-1268	405	UD	405	1900	ug/kg
11096-82-5	Aroclor-1260	363	UD	363	1900	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	0	*	32 - 144	0%	SPK: 20
2051-24-3	Decachlorobiphenyl	0	*	32 - 175	0%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	Weston Solutions, Inc.		Date Collected:	04/17/25	
Project:	RFP 911		Date Received:	04/23/25	
Client Sample ID:	P001-SS009-01		SDG No.:	Q1860	
Lab Sample ID:	Q1860-12		Matrix:	SOIL	
Analytical Method:	SW8082A		% Solid:	86.5	Decanted:
Sample Wt/Vol:	30.01	Units: g	Final Vol:	10000	uL
Soil Aliquot Vol:		uL	Test:	PCB	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	SW3541B				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PO110681.D	1	04/24/25 08:32	04/24/25 13:56	PB167721

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	4.60	U	4.60	19.6	ug/kg
11104-28-2	Aroclor-1221	4.70	U	4.70	19.6	ug/kg
11141-16-5	Aroclor-1232	4.30	U	4.30	19.6	ug/kg
53469-21-9	Aroclor-1242	4.60	U	4.60	19.6	ug/kg
12672-29-6	Aroclor-1248	7300	E	6.80	19.6	ug/kg
11097-69-1	Aroclor-1254	5700	E	3.70	19.6	ug/kg
37324-23-5	Aroclor-1262	5.80	U	5.80	19.6	ug/kg
11100-14-4	Aroclor-1268	4.20	U	4.20	19.6	ug/kg
11096-82-5	Aroclor-1260	3.70	U	3.70	19.6	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	23.1		32 - 144	116%	SPK: 20
2051-24-3	Decachlorobiphenyl	49.7	*	32 - 175	248%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	Weston Solutions, Inc.		Date Collected:	04/17/25	
Project:	RFP 911		Date Received:	04/23/25	
Client Sample ID:	P001-SS009-01DL		SDG No.:	Q1860	
Lab Sample ID:	Q1860-12DL		Matrix:	SOIL	
Analytical Method:	SW8082A		% Solid:	86.5	Decanted:
Sample Wt/Vol:	30.01	Units: g	Final Vol:	10000	uL
Soil Aliquot Vol:		uL	Test:	PCB	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	SW3541B				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PO110709.D	100	04/24/25 08:32	04/25/25 02:02	PB167721

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	456	UD	456	2000	ug/kg
11104-28-2	Aroclor-1221	466	UD	466	2000	ug/kg
11141-16-5	Aroclor-1232	430	UD	430	2000	ug/kg
53469-21-9	Aroclor-1242	463	UD	463	2000	ug/kg
12672-29-6	Aroclor-1248	9400	D	684	2000	ug/kg
11097-69-1	Aroclor-1254	6700	D	371	2000	ug/kg
37324-23-5	Aroclor-1262	580	UD	580	2000	ug/kg
11100-14-4	Aroclor-1268	416	UD	416	2000	ug/kg
11096-82-5	Aroclor-1260	373	UD	373	2000	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	0	*	32 - 144	0%	SPK: 20
2051-24-3	Decachlorobiphenyl	0	*	32 - 175	0%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	Weston Solutions, Inc.		Date Collected:	04/17/25	
Project:	RFP 911		Date Received:	04/23/25	
Client Sample ID:	P001-SS010-01		SDG No.:	Q1860	
Lab Sample ID:	Q1860-13		Matrix:	SOIL	
Analytical Method:	SW8082A		% Solid:	80.7	Decanted:
Sample Wt/Vol:	30.04	Units: g	Final Vol:	10000	uL
Soil Aliquot Vol:		uL	Test:	PCB	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	SW3541B				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PO110708.D	1	04/24/25 08:32	04/25/25 01:43	PB167721

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	4.90	U	4.90	21.0	ug/kg
11104-28-2	Aroclor-1221	5.00	U	5.00	21.0	ug/kg
11141-16-5	Aroclor-1232	4.60	U	4.60	21.0	ug/kg
53469-21-9	Aroclor-1242	5.00	U	5.00	21.0	ug/kg
12672-29-6	Aroclor-1248	7.30	U	7.30	21.0	ug/kg
11097-69-1	Aroclor-1254	57.4		4.00	21.0	ug/kg
37324-23-5	Aroclor-1262	6.20	U	6.20	21.0	ug/kg
11100-14-4	Aroclor-1268	4.50	U	4.50	21.0	ug/kg
11096-82-5	Aroclor-1260	54.6		4.00	21.0	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	16.1		32 - 144	81%	SPK: 20
2051-24-3	Decachlorobiphenyl	11.4		32 - 175	57%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	Weston Solutions, Inc.		Date Collected:	04/17/25	
Project:	RFP 911		Date Received:	04/23/25	
Client Sample ID:	P001-SS011-01		SDG No.:	Q1860	
Lab Sample ID:	Q1860-14		Matrix:	SOIL	
Analytical Method:	SW8082A		% Solid:	82.8	Decanted:
Sample Wt/Vol:	30.06	Units: g	Final Vol:	10000	uL
Soil Aliquot Vol:		uL	Test:	PCB	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	SW3541B				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PO110683.D	1	04/24/25 08:32	04/24/25 14:32	PB167721

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	4.80	U	4.80	20.5	ug/kg
11104-28-2	Aroclor-1221	4.90	U	4.90	20.5	ug/kg
11141-16-5	Aroclor-1232	4.50	U	4.50	20.5	ug/kg
53469-21-9	Aroclor-1242	4.80	U	4.80	20.5	ug/kg
12672-29-6	Aroclor-1248	325		7.10	20.5	ug/kg
11097-69-1	Aroclor-1254	282		3.90	20.5	ug/kg
37324-23-5	Aroclor-1262	6.10	U	6.10	20.5	ug/kg
11100-14-4	Aroclor-1268	4.30	U	4.30	20.5	ug/kg
11096-82-5	Aroclor-1260	3.90	U	3.90	20.5	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	16.5		32 - 144	83%	SPK: 20
2051-24-3	Decachlorobiphenyl	14.8		32 - 175	74%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	Weston Solutions, Inc.		Date Collected:	04/17/25	
Project:	RFP 911		Date Received:	04/23/25	
Client Sample ID:	P001-SS012-01		SDG No.:	Q1860	
Lab Sample ID:	Q1860-15		Matrix:	SOIL	
Analytical Method:	SW8082A		% Solid:	81.3	Decanted:
Sample Wt/Vol:	30.03	Units: g	Final Vol:	10000	uL
Soil Aliquot Vol:		uL	Test:	PCB	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	SW3541B				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PO110684.D	1	04/24/25 08:32	04/24/25 14:51	PB167721

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	4.90	U	4.90	20.9	ug/kg
11104-28-2	Aroclor-1221	5.00	U	5.00	20.9	ug/kg
11141-16-5	Aroclor-1232	4.60	U	4.60	20.9	ug/kg
53469-21-9	Aroclor-1242	4.90	U	4.90	20.9	ug/kg
12672-29-6	Aroclor-1248	6500	E	7.30	20.9	ug/kg
11097-69-1	Aroclor-1254	3900	E	3.90	20.9	ug/kg
37324-23-5	Aroclor-1262	6.20	U	6.20	20.9	ug/kg
11100-14-4	Aroclor-1268	4.40	U	4.40	20.9	ug/kg
11096-82-5	Aroclor-1260	4.00	U	4.00	20.9	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	14.8		32 - 144	74%	SPK: 20
2051-24-3	Decachlorobiphenyl	43.8	*	32 - 175	219%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	Weston Solutions, Inc.		Date Collected:	04/17/25	
Project:	RFP 911		Date Received:	04/23/25	
Client Sample ID:	P001-SS012-01DL		SDG No.:	Q1860	
Lab Sample ID:	Q1860-15DL		Matrix:	SOIL	
Analytical Method:	SW8082A		% Solid:	81.3	Decanted:
Sample Wt/Vol:	30.03	Units: g	Final Vol:	10000	uL
Soil Aliquot Vol:		uL	Test:	PCB	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	SW3541B				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PO110710.D	25	04/24/25 08:32	04/25/25 02:20	PB167721

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	121	UD	121	522	ug/kg
11104-28-2	Aroclor-1221	124	UD	124	522	ug/kg
11141-16-5	Aroclor-1232	114	UD	114	522	ug/kg
53469-21-9	Aroclor-1242	123	UD	123	522	ug/kg
12672-29-6	Aroclor-1248	7000	D	182	522	ug/kg
11097-69-1	Aroclor-1254	3900	D	98.6	522	ug/kg
37324-23-5	Aroclor-1262	154	UD	154	522	ug/kg
11100-14-4	Aroclor-1268	111	UD	111	522	ug/kg
11096-82-5	Aroclor-1260	99.2	UD	99.2	522	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	15.3		32 - 144	76%	SPK: 20
2051-24-3	Decachlorobiphenyl	47.5	*	32 - 175	238%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	Weston Solutions, Inc.		Date Collected:	04/17/25	
Project:	RFP 911		Date Received:	04/23/25	
Client Sample ID:	P001-SS013-01		SDG No.:	Q1860	
Lab Sample ID:	Q1860-16		Matrix:	SOIL	
Analytical Method:	SW8082A		% Solid:	85.2	Decanted:
Sample Wt/Vol:	30.05	Units: g	Final Vol:	10000	uL
Soil Aliquot Vol:		uL	Test:	PCB	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	SW3541B				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PO110685.D	1	04/24/25 08:32	04/24/25 15:09	PB167721

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	4.60	U	4.60	19.9	ug/kg
11104-28-2	Aroclor-1221	4.70	U	4.70	19.9	ug/kg
11141-16-5	Aroclor-1232	4.40	U	4.40	19.9	ug/kg
53469-21-9	Aroclor-1242	4.70	U	4.70	19.9	ug/kg
12672-29-6	Aroclor-1248	4300	E	6.90	19.9	ug/kg
11097-69-1	Aroclor-1254	3200	E	3.80	19.9	ug/kg
37324-23-5	Aroclor-1262	5.90	U	5.90	19.9	ug/kg
11100-14-4	Aroclor-1268	4.20	U	4.20	19.9	ug/kg
11096-82-5	Aroclor-1260	3.80	U	3.80	19.9	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	18.3		32 - 144	92%	SPK: 20
2051-24-3	Decachlorobiphenyl	27.1		32 - 175	136%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	Weston Solutions, Inc.		Date Collected:	04/17/25	
Project:	RFP 911		Date Received:	04/23/25	
Client Sample ID:	P001-SS013-01DL		SDG No.:	Q1860	
Lab Sample ID:	Q1860-16DL		Matrix:	SOIL	
Analytical Method:	SW8082A		% Solid:	85.2	Decanted:
Sample Wt/Vol:	30.05	Units: g	Final Vol:	10000	uL
Soil Aliquot Vol:		uL	Test:	PCB	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	SW3541B				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PO110712.D	25	04/24/25 08:32	04/25/25 02:57	PB167721

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	116	UD	116	498	ug/kg
11104-28-2	Aroclor-1221	118	UD	118	498	ug/kg
11141-16-5	Aroclor-1232	109	UD	109	498	ug/kg
53469-21-9	Aroclor-1242	117	UD	117	498	ug/kg
12672-29-6	Aroclor-1248	5100	D	173	498	ug/kg
11097-69-1	Aroclor-1254	3500	D	94.0	498	ug/kg
37324-23-5	Aroclor-1262	147	UD	147	498	ug/kg
11100-14-4	Aroclor-1268	105	UD	105	498	ug/kg
11096-82-5	Aroclor-1260	94.6	UD	94.6	498	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	20.8		32 - 144	104%	SPK: 20
2051-24-3	Decachlorobiphenyl	38.8	*	32 - 175	194%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	Weston Solutions, Inc.		Date Collected:	04/17/25	
Project:	RFP 911		Date Received:	04/23/25	
Client Sample ID:	P001-SS014-01		SDG No.:	Q1860	
Lab Sample ID:	Q1860-17		Matrix:	SOIL	
Analytical Method:	SW8082A		% Solid:	79.8	Decanted:
Sample Wt/Vol:	30.01	Units: g	Final Vol:	10000	uL
Soil Aliquot Vol:		uL	Test:	PCB	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	SW3541B				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PO110686.D	1	04/24/25 08:32	04/24/25 15:27	PB167721

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	4.90	U	4.90	21.3	ug/kg
11104-28-2	Aroclor-1221	5.00	U	5.00	21.3	ug/kg
11141-16-5	Aroclor-1232	4.70	U	4.70	21.3	ug/kg
53469-21-9	Aroclor-1242	5.00	U	5.00	21.3	ug/kg
12672-29-6	Aroclor-1248	398		7.40	21.3	ug/kg
11097-69-1	Aroclor-1254	412		4.00	21.3	ug/kg
37324-23-5	Aroclor-1262	6.30	U	6.30	21.3	ug/kg
11100-14-4	Aroclor-1268	4.50	U	4.50	21.3	ug/kg
11096-82-5	Aroclor-1260	4.00	U	4.00	21.3	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	18.4		32 - 144	92%	SPK: 20
2051-24-3	Decachlorobiphenyl	16.8		32 - 175	84%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	Weston Solutions, Inc.		Date Collected:	04/17/25	
Project:	RFP 911		Date Received:	04/23/25	
Client Sample ID:	P001-SS015-01		SDG No.:	Q1860	
Lab Sample ID:	Q1860-18		Matrix:	SOIL	
Analytical Method:	SW8082A		% Solid:	74.9	Decanted:
Sample Wt/Vol:	30.02	Units: g	Final Vol:	10000	uL
Soil Aliquot Vol:		uL	Test:	PCB	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	SW3541B				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PO110687.D	1	04/24/25 08:32	04/24/25 15:46	PB167721

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	5.30	U	5.30	22.7	ug/kg
11104-28-2	Aroclor-1221	5.40	U	5.40	22.7	ug/kg
11141-16-5	Aroclor-1232	5.00	U	5.00	22.7	ug/kg
53469-21-9	Aroclor-1242	5.40	U	5.40	22.7	ug/kg
12672-29-6	Aroclor-1248	71.7		7.90	22.7	ug/kg
11097-69-1	Aroclor-1254	102		4.30	22.7	ug/kg
37324-23-5	Aroclor-1262	6.70	U	6.70	22.7	ug/kg
11100-14-4	Aroclor-1268	4.80	U	4.80	22.7	ug/kg
11096-82-5	Aroclor-1260	4.30	U	4.30	22.7	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	18.9		32 - 144	95%	SPK: 20
2051-24-3	Decachlorobiphenyl	13.0		32 - 175	65%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	Weston Solutions, Inc.		Date Collected:	04/17/25	
Project:	RFP 911		Date Received:	04/23/25	
Client Sample ID:	P001-SS016-01		SDG No.:	Q1860	
Lab Sample ID:	Q1860-19		Matrix:	SOIL	
Analytical Method:	SW8082A		% Solid:	83.1	Decanted:
Sample Wt/Vol:	30.05	Units: g	Final Vol:	10000	uL
Soil Aliquot Vol:		uL	Test:	PCB	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	SW3541B				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PO110727.D	1	04/25/25 08:20	04/25/25 12:31	PB167737

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	4.70	U	4.70	20.4	ug/kg
11104-28-2	Aroclor-1221	4.80	U	4.80	20.4	ug/kg
11141-16-5	Aroclor-1232	4.50	U	4.50	20.4	ug/kg
53469-21-9	Aroclor-1242	4.80	U	4.80	20.4	ug/kg
12672-29-6	Aroclor-1248	932	E	7.10	20.4	ug/kg
11097-69-1	Aroclor-1254	734	E	3.90	20.4	ug/kg
37324-23-5	Aroclor-1262	6.00	U	6.00	20.4	ug/kg
11100-14-4	Aroclor-1268	4.30	U	4.30	20.4	ug/kg
11096-82-5	Aroclor-1260	3.90	U	3.90	20.4	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	16.5		32 - 144	83%	SPK: 20
2051-24-3	Decachlorobiphenyl	17.1		32 - 175	86%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	Weston Solutions, Inc.		Date Collected:	04/17/25	
Project:	RFP 911		Date Received:	04/23/25	
Client Sample ID:	P001-SS016-01DL		SDG No.:	Q1860	
Lab Sample ID:	Q1860-19DL		Matrix:	SOIL	
Analytical Method:	SW8082A		% Solid:	83.1	Decanted:
Sample Wt/Vol:	30.05	Units: g	Final Vol:	10000	uL
Soil Aliquot Vol:		uL	Test:	PCB	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	SW3541B				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PO110788.D	10	04/25/25 08:20	04/28/25 10:39	PB167737

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	47.5	UD	47.5	204	ug/kg
11104-28-2	Aroclor-1221	48.4	UD	48.4	204	ug/kg
11141-16-5	Aroclor-1232	44.7	UD	44.7	204	ug/kg
53469-21-9	Aroclor-1242	48.2	UD	48.2	204	ug/kg
12672-29-6	Aroclor-1248	986	D	71.1	204	ug/kg
11097-69-1	Aroclor-1254	725	D	38.6	204	ug/kg
37324-23-5	Aroclor-1262	60.3	UD	60.3	204	ug/kg
11100-14-4	Aroclor-1268	43.2	UD	43.2	204	ug/kg
11096-82-5	Aroclor-1260	38.8	UD	38.8	204	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	15.4		32 - 144	77%	SPK: 20
2051-24-3	Decachlorobiphenyl	16.0		32 - 175	80%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	Weston Solutions, Inc.		Date Collected:	04/17/25	
Project:	RFP 911		Date Received:	04/23/25	
Client Sample ID:	P001-SS017-01		SDG No.:	Q1860	
Lab Sample ID:	Q1860-22		Matrix:	SOIL	
Analytical Method:	SW8082A		% Solid:	84.5	Decanted:
Sample Wt/Vol:	30.08	Units: g	Final Vol:	10000	uL
Soil Aliquot Vol:		uL	Test:	PCB	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	SW3541B				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PO110821.D	1	04/24/25 08:32	04/28/25 22:42	PB167721

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	4.70	U	4.70	20.1	ug/kg
11104-28-2	Aroclor-1221	4.80	U	4.80	20.1	ug/kg
11141-16-5	Aroclor-1232	4.40	U	4.40	20.1	ug/kg
53469-21-9	Aroclor-1242	4.70	U	4.70	20.1	ug/kg
12672-29-6	Aroclor-1248	770	E	7.00	20.1	ug/kg
11097-69-1	Aroclor-1254	759	E	3.80	20.1	ug/kg
37324-23-5	Aroclor-1262	5.90	U	5.90	20.1	ug/kg
11100-14-4	Aroclor-1268	4.20	U	4.20	20.1	ug/kg
11096-82-5	Aroclor-1260	3.80	U	3.80	20.1	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	17.3		32 - 144	86%	SPK: 20
2051-24-3	Decachlorobiphenyl	18.4		32 - 175	92%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	Weston Solutions, Inc.		Date Collected:	04/17/25	
Project:	RFP 911		Date Received:	04/23/25	
Client Sample ID:	P001-SS017-01DL		SDG No.:	Q1860	
Lab Sample ID:	Q1860-22DL		Matrix:	SOIL	
Analytical Method:	SW8082A		% Solid:	84.5	Decanted:
Sample Wt/Vol:	30.08	Units: g	Final Vol:	10000	uL
Soil Aliquot Vol:		uL	Test:	PCB	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	SW3541B				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PO110826.D	5	04/24/25 08:32	04/29/25 00:12	PB167721

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	23.3	UD	23.3	100	ug/kg
11104-28-2	Aroclor-1221	23.8	UD	23.8	100	ug/kg
11141-16-5	Aroclor-1232	22.0	UD	22.0	100	ug/kg
53469-21-9	Aroclor-1242	23.7	UD	23.7	100	ug/kg
12672-29-6	Aroclor-1248	749	D	34.9	100	ug/kg
11097-69-1	Aroclor-1254	663	D	18.9	100	ug/kg
37324-23-5	Aroclor-1262	29.6	UD	29.6	100	ug/kg
11100-14-4	Aroclor-1268	21.2	UD	21.2	100	ug/kg
11096-82-5	Aroclor-1260	19.1	UD	19.1	100	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	15.6		32 - 144	78%	SPK: 20
2051-24-3	Decachlorobiphenyl	15.4		32 - 175	77%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	Weston Solutions, Inc.		Date Collected:	04/17/25	
Project:	RFP 911		Date Received:	04/23/25	
Client Sample ID:	P001-SS017-02		SDG No.:	Q1860	
Lab Sample ID:	Q1860-23		Matrix:	SOIL	
Analytical Method:	SW8082A		% Solid:	82.6	Decanted:
Sample Wt/Vol:	30.04	Units: g	Final Vol:	10000	uL
Soil Aliquot Vol:		uL	Test:	PCB	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	SW3541B				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PO110822.D	1	04/24/25 08:32	04/28/25 23:00	PB167721

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	4.80	U	4.80	20.6	ug/kg
11104-28-2	Aroclor-1221	4.90	U	4.90	20.6	ug/kg
11141-16-5	Aroclor-1232	4.50	U	4.50	20.6	ug/kg
53469-21-9	Aroclor-1242	4.80	U	4.80	20.6	ug/kg
12672-29-6	Aroclor-1248	925	E	7.20	20.6	ug/kg
11097-69-1	Aroclor-1254	631	E	3.90	20.6	ug/kg
37324-23-5	Aroclor-1262	6.10	U	6.10	20.6	ug/kg
11100-14-4	Aroclor-1268	4.40	U	4.40	20.6	ug/kg
11096-82-5	Aroclor-1260	3.90	U	3.90	20.6	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	20.2		32 - 144	101%	SPK: 20
2051-24-3	Decachlorobiphenyl	18.5		32 - 175	93%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	Weston Solutions, Inc.		Date Collected:	04/17/25	
Project:	RFP 911		Date Received:	04/23/25	
Client Sample ID:	P001-SS017-02DL		SDG No.:	Q1860	
Lab Sample ID:	Q1860-23DL		Matrix:	SOIL	
Analytical Method:	SW8082A		% Solid:	82.6	Decanted:
Sample Wt/Vol:	30.04	Units: g	Final Vol:	10000	uL
Soil Aliquot Vol:		uL	Test:	PCB	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	SW3541B				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PO110827.D	5	04/24/25 08:32	04/29/25 00:31	PB167721

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	23.9	UD	23.9	103	ug/kg
11104-28-2	Aroclor-1221	24.4	UD	24.4	103	ug/kg
11141-16-5	Aroclor-1232	22.5	UD	22.5	103	ug/kg
53469-21-9	Aroclor-1242	24.2	UD	24.2	103	ug/kg
12672-29-6	Aroclor-1248	878	D	35.8	103	ug/kg
11097-69-1	Aroclor-1254	525	D	19.4	103	ug/kg
37324-23-5	Aroclor-1262	30.3	UD	30.3	103	ug/kg
11100-14-4	Aroclor-1268	21.8	UD	21.8	103	ug/kg
11096-82-5	Aroclor-1260	19.5	UD	19.5	103	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	17.9		32 - 144	90%	SPK: 20
2051-24-3	Decachlorobiphenyl	17.9		32 - 175	90%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	Weston Solutions, Inc.		Date Collected:	04/21/25	
Project:	RFP 911		Date Received:	04/23/25	
Client Sample ID:	P001-SS018-01		SDG No.:	Q1860	
Lab Sample ID:	Q1860-24		Matrix:	SOIL	
Analytical Method:	SW8082A		% Solid:	89.6	Decanted:
Sample Wt/Vol:	30.07	Units: g	Final Vol:	10000	uL
Soil Aliquot Vol:		uL	Test:	PCB	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	SW3541B				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PO110823.D	1	04/24/25 08:32	04/28/25 23:18	PB167721

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	4.40	U	4.40	18.9	ug/kg
11104-28-2	Aroclor-1221	4.50	U	4.50	18.9	ug/kg
11141-16-5	Aroclor-1232	4.10	U	4.10	18.9	ug/kg
53469-21-9	Aroclor-1242	4.50	U	4.50	18.9	ug/kg
12672-29-6	Aroclor-1248	6.60	U	6.60	18.9	ug/kg
11097-69-1	Aroclor-1254	76.3		3.60	18.9	ug/kg
37324-23-5	Aroclor-1262	5.60	U	5.60	18.9	ug/kg
11100-14-4	Aroclor-1268	4.00	U	4.00	18.9	ug/kg
11096-82-5	Aroclor-1260	3.60	U	3.60	18.9	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	20.9		32 - 144	105%	SPK: 20
2051-24-3	Decachlorobiphenyl	16.4		32 - 175	82%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	Weston Solutions, Inc.		Date Collected:	04/21/25	
Project:	RFP 911		Date Received:	04/23/25	
Client Sample ID:	P001-SS019-01		SDG No.:	Q1860	
Lab Sample ID:	Q1860-25		Matrix:	SOIL	
Analytical Method:	SW8082A		% Solid:	88.3	Decanted:
Sample Wt/Vol:	30.06	Units: g	Final Vol:	10000	uL
Soil Aliquot Vol:		uL	Test:	PCB	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	SW3541B				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PO110824.D	1	04/24/25 08:32	04/28/25 23:37	PB167721

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	4.50	U	4.50	19.2	ug/kg
11104-28-2	Aroclor-1221	4.60	U	4.60	19.2	ug/kg
11141-16-5	Aroclor-1232	4.20	U	4.20	19.2	ug/kg
53469-21-9	Aroclor-1242	4.50	U	4.50	19.2	ug/kg
12672-29-6	Aroclor-1248	6.70	U	6.70	19.2	ug/kg
11097-69-1	Aroclor-1254	435	E	3.60	19.2	ug/kg
37324-23-5	Aroclor-1262	5.70	U	5.70	19.2	ug/kg
11100-14-4	Aroclor-1268	4.10	U	4.10	19.2	ug/kg
11096-82-5	Aroclor-1260	297		3.70	19.2	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	21.0		32 - 144	105%	SPK: 20
2051-24-3	Decachlorobiphenyl	18.3		32 - 175	92%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	Weston Solutions, Inc.		Date Collected:	04/21/25	
Project:	RFP 911		Date Received:	04/23/25	
Client Sample ID:	P001-SS019-01DL		SDG No.:	Q1860	
Lab Sample ID:	Q1860-25DL		Matrix:	SOIL	
Analytical Method:	SW8082A		% Solid:	88.3	Decanted:
Sample Wt/Vol:	30.06	Units: g	Final Vol:	10000	uL
Soil Aliquot Vol:		uL	Test:	PCB	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	SW3541B				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PO110837.D	5	04/24/25 08:32	04/29/25 04:57	PB167721

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	22.3	UD	22.3	96.1	ug/kg
11104-28-2	Aroclor-1221	22.8	UD	22.8	96.1	ug/kg
11141-16-5	Aroclor-1232	21.0	UD	21.0	96.1	ug/kg
53469-21-9	Aroclor-1242	22.7	UD	22.7	96.1	ug/kg
12672-29-6	Aroclor-1248	33.5	UD	33.5	96.1	ug/kg
11097-69-1	Aroclor-1254	435	D	18.1	96.1	ug/kg
37324-23-5	Aroclor-1262	28.4	UD	28.4	96.1	ug/kg
11100-14-4	Aroclor-1268	20.3	UD	20.3	96.1	ug/kg
11096-82-5	Aroclor-1260	298	D	18.3	96.1	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	18.6		32 - 144	93%	SPK: 20
2051-24-3	Decachlorobiphenyl	18.6		32 - 175	93%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	Weston Solutions, Inc.		Date Collected:	04/21/25	
Project:	RFP 911		Date Received:	04/23/25	
Client Sample ID:	P001-SS020-01		SDG No.:	Q1860	
Lab Sample ID:	Q1860-26		Matrix:	SOIL	
Analytical Method:	SW8082A		% Solid:	85	Decanted:
Sample Wt/Vol:	30.02	Units: g	Final Vol:	10000	uL
Soil Aliquot Vol:		uL	Test:	PCB	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	SW3541B				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PO110730.D	1	04/25/25 08:20	04/25/25 13:26	PB167737

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	4.60	U	4.60	20.0	ug/kg
11104-28-2	Aroclor-1221	4.70	U	4.70	20.0	ug/kg
11141-16-5	Aroclor-1232	4.40	U	4.40	20.0	ug/kg
53469-21-9	Aroclor-1242	4.70	U	4.70	20.0	ug/kg
12672-29-6	Aroclor-1248	7.00	U	7.00	20.0	ug/kg
11097-69-1	Aroclor-1254	5900	EP	3.80	20.0	ug/kg
37324-23-5	Aroclor-1262	5.90	U	5.90	20.0	ug/kg
11100-14-4	Aroclor-1268	4.20	U	4.20	20.0	ug/kg
11096-82-5	Aroclor-1260	3.80	U	3.80	20.0	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	17.8		32 - 144	89%	SPK: 20
2051-24-3	Decachlorobiphenyl	23.8		32 - 175	119%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	Weston Solutions, Inc.		Date Collected:	04/21/25	
Project:	RFP 911		Date Received:	04/23/25	
Client Sample ID:	P001-SS020-01DL		SDG No.:	Q1860	
Lab Sample ID:	Q1860-26DL		Matrix:	SOIL	
Analytical Method:	SW8082A		% Solid:	85	Decanted:
Sample Wt/Vol:	30.02	Units: g	Final Vol:	10000	uL
Soil Aliquot Vol:		uL	Test:	PCB	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	SW3541B				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PO110789.D	25	04/25/25 08:20	04/28/25 10:57	PB167737

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	116	UD	116	500	ug/kg
11104-28-2	Aroclor-1221	118	UD	118	500	ug/kg
11141-16-5	Aroclor-1232	109	UD	109	500	ug/kg
53469-21-9	Aroclor-1242	118	UD	118	500	ug/kg
12672-29-6	Aroclor-1248	174	UD	174	500	ug/kg
11097-69-1	Aroclor-1254	4500	D	94.3	500	ug/kg
37324-23-5	Aroclor-1262	148	UD	148	500	ug/kg
11100-14-4	Aroclor-1268	106	UD	106	500	ug/kg
11096-82-5	Aroclor-1260	94.9	UD	94.9	500	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	15.8		32 - 144	79%	SPK: 20
2051-24-3	Decachlorobiphenyl	24.0		32 - 175	120%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	Weston Solutions, Inc.		Date Collected:	04/21/25	
Project:	RFP 911		Date Received:	04/23/25	
Client Sample ID:	P001-SS021-01		SDG No.:	Q1860	
Lab Sample ID:	Q1860-27		Matrix:	SOIL	
Analytical Method:	SW8082A		% Solid:	88.5	Decanted:
Sample Wt/Vol:	30.04	Units: g	Final Vol:	10000	uL
Soil Aliquot Vol:		uL	Test:	PCB	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	SW3541B				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PO110731.D	1	04/25/25 08:20	04/25/25 13:44	PB167737

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	4.50	U	4.50	19.2	ug/kg
11104-28-2	Aroclor-1221	4.50	U	4.50	19.2	ug/kg
11141-16-5	Aroclor-1232	4.20	U	4.20	19.2	ug/kg
53469-21-9	Aroclor-1242	4.50	U	4.50	19.2	ug/kg
12672-29-6	Aroclor-1248	6.70	U	6.70	19.2	ug/kg
11097-69-1	Aroclor-1254	381	E	3.60	19.2	ug/kg
37324-23-5	Aroclor-1262	5.70	U	5.70	19.2	ug/kg
11100-14-4	Aroclor-1268	4.10	U	4.10	19.2	ug/kg
11096-82-5	Aroclor-1260	3.60	U	3.60	19.2	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	19.9		32 - 144	99%	SPK: 20
2051-24-3	Decachlorobiphenyl	38.6	*	32 - 175	193%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	Weston Solutions, Inc.		Date Collected:	04/21/25	
Project:	RFP 911		Date Received:	04/23/25	
Client Sample ID:	P001-SS021-01DL		SDG No.:	Q1860	
Lab Sample ID:	Q1860-27DL		Matrix:	SOIL	
Analytical Method:	SW8082A		% Solid:	88.5	Decanted:
Sample Wt/Vol:	30.04	Units: g	Final Vol:	10000	uL
Soil Aliquot Vol:		uL	Test:	PCB	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	SW3541B				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PO110790.D	5	04/25/25 08:20	04/28/25 11:16	PB167737

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	22.3	UD	22.3	95.9	ug/kg
11104-28-2	Aroclor-1221	22.7	UD	22.7	95.9	ug/kg
11141-16-5	Aroclor-1232	21.0	UD	21.0	95.9	ug/kg
53469-21-9	Aroclor-1242	22.6	UD	22.6	95.9	ug/kg
12672-29-6	Aroclor-1248	33.4	UD	33.4	95.9	ug/kg
11097-69-1	Aroclor-1254	400	D	18.1	95.9	ug/kg
37324-23-5	Aroclor-1262	28.3	UD	28.3	95.9	ug/kg
11100-14-4	Aroclor-1268	20.3	UD	20.3	95.9	ug/kg
11096-82-5	Aroclor-1260	18.2	UD	18.2	95.9	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	18.8		32 - 144	94%	SPK: 20
2051-24-3	Decachlorobiphenyl	38.4	*	32 - 175	192%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	Weston Solutions, Inc.		Date Collected:	04/21/25	
Project:	RFP 911		Date Received:	04/23/25	
Client Sample ID:	P001-SS022-01		SDG No.:	Q1860	
Lab Sample ID:	Q1860-28		Matrix:	SOIL	
Analytical Method:	SW8082A		% Solid:	83.5	Decanted:
Sample Wt/Vol:	30.06	Units: g	Final Vol:	10000	uL
Soil Aliquot Vol:		uL	Test:	PCB	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	SW3541B				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PO110732.D	1	04/25/25 08:20	04/25/25 14:03	PB167737

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	4.70	U	4.70	20.3	ug/kg
11104-28-2	Aroclor-1221	4.80	U	4.80	20.3	ug/kg
11141-16-5	Aroclor-1232	4.40	U	4.40	20.3	ug/kg
53469-21-9	Aroclor-1242	4.80	U	4.80	20.3	ug/kg
12672-29-6	Aroclor-1248	7.10	U	7.10	20.3	ug/kg
11097-69-1	Aroclor-1254	203		3.80	20.3	ug/kg
37324-23-5	Aroclor-1262	6.00	U	6.00	20.3	ug/kg
11100-14-4	Aroclor-1268	4.30	U	4.30	20.3	ug/kg
11096-82-5	Aroclor-1260	3.90	U	3.90	20.3	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	18.4		32 - 144	92%	SPK: 20
2051-24-3	Decachlorobiphenyl	30.9		32 - 175	155%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	Weston Solutions, Inc.		Date Collected:	04/21/25	
Project:	RFP 911		Date Received:	04/23/25	
Client Sample ID:	P001-SS023-01		SDG No.:	Q1860	
Lab Sample ID:	Q1860-29		Matrix:	SOIL	
Analytical Method:	SW8082A		% Solid:	84	Decanted:
Sample Wt/Vol:	30.09	Units: g	Final Vol:	10000	uL
Soil Aliquot Vol:		uL	Test:	PCB	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	SW3541B				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PO110733.D	1	04/25/25 08:20	04/25/25 14:21	PB167737

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	4.70	U	4.70	20.2	ug/kg
11104-28-2	Aroclor-1221	4.80	U	4.80	20.2	ug/kg
11141-16-5	Aroclor-1232	4.40	U	4.40	20.2	ug/kg
53469-21-9	Aroclor-1242	4.80	U	4.80	20.2	ug/kg
12672-29-6	Aroclor-1248	7.00	U	7.00	20.2	ug/kg
11097-69-1	Aroclor-1254	410	E	3.80	20.2	ug/kg
37324-23-5	Aroclor-1262	6.00	U	6.00	20.2	ug/kg
11100-14-4	Aroclor-1268	4.30	U	4.30	20.2	ug/kg
11096-82-5	Aroclor-1260	3.80	U	3.80	20.2	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	17.1		32 - 144	86%	SPK: 20
2051-24-3	Decachlorobiphenyl	30.4		32 - 175	152%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	Weston Solutions, Inc.		Date Collected:	04/21/25	
Project:	RFP 911		Date Received:	04/23/25	
Client Sample ID:	P001-SS023-01DL		SDG No.:	Q1860	
Lab Sample ID:	Q1860-29DL		Matrix:	SOIL	
Analytical Method:	SW8082A		% Solid:	84	Decanted:
Sample Wt/Vol:	30.09	Units: g	Final Vol:	10000	uL
Soil Aliquot Vol:		uL	Test:	PCB	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	SW3541B				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PO110791.D	5	04/25/25 08:20	04/28/25 11:34	PB167737

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	23.4	UD	23.4	101	ug/kg
11104-28-2	Aroclor-1221	23.9	UD	23.9	101	ug/kg
11141-16-5	Aroclor-1232	22.1	UD	22.1	101	ug/kg
53469-21-9	Aroclor-1242	23.8	UD	23.8	101	ug/kg
12672-29-6	Aroclor-1248	35.1	UD	35.1	101	ug/kg
11097-69-1	Aroclor-1254	361	D	19.1	101	ug/kg
37324-23-5	Aroclor-1262	29.8	UD	29.8	101	ug/kg
11100-14-4	Aroclor-1268	21.4	UD	21.4	101	ug/kg
11096-82-5	Aroclor-1260	19.2	UD	19.2	101	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	13.1		32 - 144	66%	SPK: 20
2051-24-3	Decachlorobiphenyl	27.8		32 - 175	139%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	Weston Solutions, Inc.		Date Collected:	04/21/25	
Project:	RFP 911		Date Received:	04/23/25	
Client Sample ID:	P001-SS024-01		SDG No.:	Q1860	
Lab Sample ID:	Q1860-30		Matrix:	SOIL	
Analytical Method:	SW8082A		% Solid:	84.6	Decanted:
Sample Wt/Vol:	30.05	Units: g	Final Vol:	10000	uL
Soil Aliquot Vol:		uL	Test:	PCB	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	SW3541B				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PO110734.D	1	04/25/25 08:20	04/25/25 14:39	PB167737

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	4.70	U	4.70	20.1	ug/kg
11104-28-2	Aroclor-1221	4.80	U	4.80	20.1	ug/kg
11141-16-5	Aroclor-1232	4.40	U	4.40	20.1	ug/kg
53469-21-9	Aroclor-1242	4.70	U	4.70	20.1	ug/kg
12672-29-6	Aroclor-1248	7.00	U	7.00	20.1	ug/kg
11097-69-1	Aroclor-1254	347		3.80	20.1	ug/kg
37324-23-5	Aroclor-1262	5.90	U	5.90	20.1	ug/kg
11100-14-4	Aroclor-1268	4.20	U	4.20	20.1	ug/kg
11096-82-5	Aroclor-1260	3.80	U	3.80	20.1	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	19.9		32 - 144	100%	SPK: 20
2051-24-3	Decachlorobiphenyl	14.6		32 - 175	73%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	Weston Solutions, Inc.		Date Collected:	04/21/25	
Project:	RFP 911		Date Received:	04/23/25	
Client Sample ID:	P001-SS025-01		SDG No.:	Q1860	
Lab Sample ID:	Q1860-31		Matrix:	SOIL	
Analytical Method:	SW8082A		% Solid:	86.9	Decanted:
Sample Wt/Vol:	30.03	Units: g	Final Vol:	10000	uL
Soil Aliquot Vol:		uL	Test:	PCB	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	SW3541B				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PO110740.D	1	04/25/25 08:20	04/25/25 18:08	PB167737

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	4.50	U	4.50	19.5	ug/kg
11104-28-2	Aroclor-1221	4.60	U	4.60	19.5	ug/kg
11141-16-5	Aroclor-1232	4.30	U	4.30	19.5	ug/kg
53469-21-9	Aroclor-1242	4.60	U	4.60	19.5	ug/kg
12672-29-6	Aroclor-1248	6.80	U	6.80	19.5	ug/kg
11097-69-1	Aroclor-1254	1200	E	3.70	19.5	ug/kg
37324-23-5	Aroclor-1262	5.80	U	5.80	19.5	ug/kg
11100-14-4	Aroclor-1268	4.10	U	4.10	19.5	ug/kg
11096-82-5	Aroclor-1260	1200	E	3.70	19.5	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	19.8		32 - 144	99%	SPK: 20
2051-24-3	Decachlorobiphenyl	50.0	*	32 - 175	250%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	Weston Solutions, Inc.		Date Collected:	04/21/25	
Project:	RFP 911		Date Received:	04/23/25	
Client Sample ID:	P001-SS025-01DL		SDG No.:	Q1860	
Lab Sample ID:	Q1860-31DL		Matrix:	SOIL	
Analytical Method:	SW8082A		% Solid:	86.9	Decanted:
Sample Wt/Vol:	30.03	Units: g	Final Vol:	10000	uL
Soil Aliquot Vol:		uL	Test:	PCB	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	SW3541B				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PO110792.D	5	04/25/25 08:20	04/28/25 11:53	PB167737

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	22.7	UD	22.7	97.7	ug/kg
11104-28-2	Aroclor-1221	23.2	UD	23.2	97.7	ug/kg
11141-16-5	Aroclor-1232	21.4	UD	21.4	97.7	ug/kg
53469-21-9	Aroclor-1242	23.0	UD	23.0	97.7	ug/kg
12672-29-6	Aroclor-1248	34.0	UD	34.0	97.7	ug/kg
11097-69-1	Aroclor-1254	1200	D	18.5	97.7	ug/kg
37324-23-5	Aroclor-1262	28.9	UD	28.9	97.7	ug/kg
11100-14-4	Aroclor-1268	20.7	UD	20.7	97.7	ug/kg
11096-82-5	Aroclor-1260	1300	D	18.6	97.7	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	20.4		32 - 144	102%	SPK: 20
2051-24-3	Decachlorobiphenyl	48.0	*	32 - 175	240%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	Weston Solutions, Inc.		Date Collected:	04/21/25	
Project:	RFP 911		Date Received:	04/23/25	
Client Sample ID:	P001-SS026-01		SDG No.:	Q1860	
Lab Sample ID:	Q1860-32		Matrix:	SOIL	
Analytical Method:	SW8082A		% Solid:	79.2	Decanted:
Sample Wt/Vol:	30.01	Units: g	Final Vol:	10000	uL
Soil Aliquot Vol:		uL	Test:	PCB	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	SW3541B				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PO110741.D	1	04/25/25 08:20	04/25/25 18:26	PB167737

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	5.00	U	5.00	21.5	ug/kg
11104-28-2	Aroclor-1221	5.10	U	5.10	21.5	ug/kg
11141-16-5	Aroclor-1232	4.70	U	4.70	21.5	ug/kg
53469-21-9	Aroclor-1242	5.10	U	5.10	21.5	ug/kg
12672-29-6	Aroclor-1248	7.50	U	7.50	21.5	ug/kg
11097-69-1	Aroclor-1254	3800	E	4.10	21.5	ug/kg
37324-23-5	Aroclor-1262	6.30	U	6.30	21.5	ug/kg
11100-14-4	Aroclor-1268	4.50	U	4.50	21.5	ug/kg
11096-82-5	Aroclor-1260	4.10	U	4.10	21.5	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	19.5		32 - 144	97%	SPK: 20
2051-24-3	Decachlorobiphenyl	74.5	*	32 - 175	373%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	Weston Solutions, Inc.		Date Collected:	04/21/25	
Project:	RFP 911		Date Received:	04/23/25	
Client Sample ID:	P001-SS026-01DL		SDG No.:	Q1860	
Lab Sample ID:	Q1860-32DL		Matrix:	SOIL	
Analytical Method:	SW8082A		% Solid:	79.2	Decanted:
Sample Wt/Vol:	30.01	Units: g	Final Vol:	10000	uL
Soil Aliquot Vol:		uL	Test:	PCB	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	SW3541B				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PO110794.D	20	04/25/25 08:20	04/28/25 12:29	PB167737

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	99.7	UD	99.7	429	ug/kg
11104-28-2	Aroclor-1221	102	UD	102	429	ug/kg
11141-16-5	Aroclor-1232	93.9	UD	93.9	429	ug/kg
53469-21-9	Aroclor-1242	101	UD	101	429	ug/kg
12672-29-6	Aroclor-1248	149	UD	149	429	ug/kg
11097-69-1	Aroclor-1254	4800	D	81.0	429	ug/kg
37324-23-5	Aroclor-1262	127	UD	127	429	ug/kg
11100-14-4	Aroclor-1268	90.9	UD	90.9	429	ug/kg
11096-82-5	Aroclor-1260	81.5	UD	81.5	429	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	26.2		32 - 144	131%	SPK: 20
2051-24-3	Decachlorobiphenyl	83.6	*	32 - 175	418%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	Weston Solutions, Inc.		Date Collected:	04/21/25	
Project:	RFP 911		Date Received:	04/23/25	
Client Sample ID:	P001-SS027-01		SDG No.:	Q1860	
Lab Sample ID:	Q1860-33		Matrix:	SOIL	
Analytical Method:	SW8082A		% Solid:	86.5	Decanted:
Sample Wt/Vol:	30.04	Units: g	Final Vol:	10000	uL
Soil Aliquot Vol:		uL	Test:	PCB	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	SW3541B				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PP071536.D	1	04/25/25 08:20	04/28/25 10:43	PB167737

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	4.60	U	4.60	19.6	ug/kg
11104-28-2	Aroclor-1221	4.70	U	4.70	19.6	ug/kg
11141-16-5	Aroclor-1232	4.30	U	4.30	19.6	ug/kg
53469-21-9	Aroclor-1242	4.60	U	4.60	19.6	ug/kg
12672-29-6	Aroclor-1248	114		6.80	19.6	ug/kg
11097-69-1	Aroclor-1254	225		3.70	19.6	ug/kg
37324-23-5	Aroclor-1262	5.80	U	5.80	19.6	ug/kg
11100-14-4	Aroclor-1268	4.20	U	4.20	19.6	ug/kg
11096-82-5	Aroclor-1260	176		3.70	19.6	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	17.3		32 - 144	87%	SPK: 20
2051-24-3	Decachlorobiphenyl	15.7		32 - 175	78%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	Weston Solutions, Inc.		Date Collected:	04/21/25	
Project:	RFP 911		Date Received:	04/23/25	
Client Sample ID:	P001-SS028-01		SDG No.:	Q1860	
Lab Sample ID:	Q1860-34		Matrix:	SOIL	
Analytical Method:	SW8082A		% Solid:	77.7	Decanted:
Sample Wt/Vol:	30.05	Units: g	Final Vol:	10000	uL
Soil Aliquot Vol:		uL	Test:	PCB	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	SW3541B				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PO110743.D	1	04/25/25 08:20	04/25/25 19:03	PB167737

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	5.10	U	5.10	21.8	ug/kg
11104-28-2	Aroclor-1221	5.20	U	5.20	21.8	ug/kg
11141-16-5	Aroclor-1232	4.80	U	4.80	21.8	ug/kg
53469-21-9	Aroclor-1242	5.20	U	5.20	21.8	ug/kg
12672-29-6	Aroclor-1248	1800	E	7.60	21.8	ug/kg
11097-69-1	Aroclor-1254	3000	E	4.10	21.8	ug/kg
37324-23-5	Aroclor-1262	6.50	U	6.50	21.8	ug/kg
11100-14-4	Aroclor-1268	4.60	U	4.60	21.8	ug/kg
11096-82-5	Aroclor-1260	2100	E	4.20	21.8	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	26.6		32 - 144	133%	SPK: 20
2051-24-3	Decachlorobiphenyl	84.5	*	32 - 175	423%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	Weston Solutions, Inc.		Date Collected:	04/21/25	
Project:	RFP 911		Date Received:	04/23/25	
Client Sample ID:	P001-SS028-01DL		SDG No.:	Q1860	
Lab Sample ID:	Q1860-34DL		Matrix:	SOIL	
Analytical Method:	SW8082A		% Solid:	77.7	Decanted:
Sample Wt/Vol:	30.05	Units: g	Final Vol:	10000	uL
Soil Aliquot Vol:		uL	Test:	PCB	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	SW3541B				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PO110855.D	20	04/25/25 08:20	04/29/25 12:11	PB167737

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	102	UD	102	437	ug/kg
11104-28-2	Aroclor-1221	104	UD	104	437	ug/kg
11141-16-5	Aroclor-1232	95.6	UD	95.6	437	ug/kg
53469-21-9	Aroclor-1242	103	UD	103	437	ug/kg
12672-29-6	Aroclor-1248	2100	D	152	437	ug/kg
11097-69-1	Aroclor-1254	2900	D	82.5	437	ug/kg
37324-23-5	Aroclor-1262	129	UD	129	437	ug/kg
11100-14-4	Aroclor-1268	92.5	UD	92.5	437	ug/kg
11096-82-5	Aroclor-1260	2300	D	83.0	437	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	23.8		32 - 144	119%	SPK: 20
2051-24-3	Decachlorobiphenyl	86.8	*	32 - 175	434%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	Weston Solutions, Inc.		Date Collected:	04/21/25	
Project:	RFP 911		Date Received:	04/23/25	
Client Sample ID:	P001-SS029-01		SDG No.:	Q1860	
Lab Sample ID:	Q1860-35		Matrix:	SOIL	
Analytical Method:	SW8082A		% Solid:	80.6	Decanted:
Sample Wt/Vol:	30.02	Units: g	Final Vol:	10000	uL
Soil Aliquot Vol:		uL	Test:	PCB	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	SW3541B				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PP071537.D	1	04/25/25 08:20	04/28/25 10:59	PB167737

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	4.90	U	4.90	21.1	ug/kg
11104-28-2	Aroclor-1221	5.00	U	5.00	21.1	ug/kg
11141-16-5	Aroclor-1232	4.60	U	4.60	21.1	ug/kg
53469-21-9	Aroclor-1242	5.00	U	5.00	21.1	ug/kg
12672-29-6	Aroclor-1248	317		7.30	21.1	ug/kg
11097-69-1	Aroclor-1254	781	E	4.00	21.1	ug/kg
37324-23-5	Aroclor-1262	6.20	U	6.20	21.1	ug/kg
11100-14-4	Aroclor-1268	4.50	U	4.50	21.1	ug/kg
11096-82-5	Aroclor-1260	532	E	4.00	21.1	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	30.2	*	32 - 144	151%	SPK: 20
2051-24-3	Decachlorobiphenyl	32.8		32 - 175	164%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	Weston Solutions, Inc.		Date Collected:	04/21/25	
Project:	RFP 911		Date Received:	04/23/25	
Client Sample ID:	P001-SS029-01DL		SDG No.:	Q1860	
Lab Sample ID:	Q1860-35DL		Matrix:	SOIL	
Analytical Method:	SW8082A		% Solid:	80.6	Decanted:
Sample Wt/Vol:	30.02	Units: g	Final Vol:	10000	uL
Soil Aliquot Vol:		uL	Test:	PCB	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	SW3541B				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PP071604.D	5	04/25/25 08:20	04/29/25 11:56	PB167737

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	24.5	UD	24.5	105	ug/kg
11104-28-2	Aroclor-1221	25.0	UD	25.0	105	ug/kg
11141-16-5	Aroclor-1232	23.1	UD	23.1	105	ug/kg
53469-21-9	Aroclor-1242	24.9	UD	24.9	105	ug/kg
12672-29-6	Aroclor-1248	294	D	36.7	105	ug/kg
11097-69-1	Aroclor-1254	665	D	19.9	105	ug/kg
37324-23-5	Aroclor-1262	31.1	UD	31.1	105	ug/kg
11100-14-4	Aroclor-1268	22.3	UD	22.3	105	ug/kg
11096-82-5	Aroclor-1260	504	D	20.0	105	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	26.7		32 - 144	134%	SPK: 20
2051-24-3	Decachlorobiphenyl	33.0		32 - 175	165%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	Weston Solutions, Inc.		Date Collected:	04/21/25	
Project:	RFP 911		Date Received:	04/23/25	
Client Sample ID:	P001-SS030-01		SDG No.:	Q1860	
Lab Sample ID:	Q1860-36		Matrix:	SOIL	
Analytical Method:	SW8082A		% Solid:	83.7	Decanted:
Sample Wt/Vol:	30.06	Units: g	Final Vol:	10000	uL
Soil Aliquot Vol:		uL	Test:	PCB	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	SW3541B				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PO110745.D	1	04/25/25 08:20	04/25/25 19:39	PB167737

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	4.70	U	4.70	20.3	ug/kg
11104-28-2	Aroclor-1221	4.80	U	4.80	20.3	ug/kg
11141-16-5	Aroclor-1232	4.40	U	4.40	20.3	ug/kg
53469-21-9	Aroclor-1242	4.80	U	4.80	20.3	ug/kg
12672-29-6	Aroclor-1248	2100	E	7.10	20.3	ug/kg
11097-69-1	Aroclor-1254	3300	E	3.80	20.3	ug/kg
37324-23-5	Aroclor-1262	6.00	U	6.00	20.3	ug/kg
11100-14-4	Aroclor-1268	4.30	U	4.30	20.3	ug/kg
11096-82-5	Aroclor-1260	3500	E	3.90	20.3	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	14.7		32 - 144	74%	SPK: 20
2051-24-3	Decachlorobiphenyl	24.3		32 - 175	121%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	Weston Solutions, Inc.		Date Collected:	04/21/25	
Project:	RFP 911		Date Received:	04/23/25	
Client Sample ID:	P001-SS030-01DL		SDG No.:	Q1860	
Lab Sample ID:	Q1860-36DL		Matrix:	SOIL	
Analytical Method:	SW8082A		% Solid:	83.7	Decanted:
Sample Wt/Vol:	30.06	Units: g	Final Vol:	10000	uL
Soil Aliquot Vol:		uL	Test:	PCB	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	SW3541B				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PO110797.D	20	04/25/25 08:20	04/28/25 13:24	PB167737

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	94.2	UD	94.2	405	ug/kg
11104-28-2	Aroclor-1221	96.1	UD	96.1	405	ug/kg
11141-16-5	Aroclor-1232	88.7	UD	88.7	405	ug/kg
53469-21-9	Aroclor-1242	95.6	UD	95.6	405	ug/kg
12672-29-6	Aroclor-1248	2800	D	141	405	ug/kg
11097-69-1	Aroclor-1254	4100	D	76.5	405	ug/kg
37324-23-5	Aroclor-1262	120	UD	120	405	ug/kg
11100-14-4	Aroclor-1268	85.8	UD	85.8	405	ug/kg
11096-82-5	Aroclor-1260	3900	D	77.0	405	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	16.8		32 - 144	84%	SPK: 20
2051-24-3	Decachlorobiphenyl	24.4		32 - 175	122%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	Weston Solutions, Inc.		Date Collected:	04/21/25	
Project:	RFP 911		Date Received:	04/23/25	
Client Sample ID:	P001-SS031-01		SDG No.:	Q1860	
Lab Sample ID:	Q1860-37		Matrix:	SOIL	
Analytical Method:	SW8082A		% Solid:	88.9	Decanted:
Sample Wt/Vol:	30.09	Units: g	Final Vol:	10000	uL
Soil Aliquot Vol:		uL	Test:	PCB	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	SW3541B				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PO110746.D	1	04/25/25 08:20	04/25/25 19:58	PB167737

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	4.40	U	4.40	19.1	ug/kg
11104-28-2	Aroclor-1221	4.50	U	4.50	19.1	ug/kg
11141-16-5	Aroclor-1232	4.20	U	4.20	19.1	ug/kg
53469-21-9	Aroclor-1242	4.50	U	4.50	19.1	ug/kg
12672-29-6	Aroclor-1248	6.60	U	6.60	19.1	ug/kg
11097-69-1	Aroclor-1254	7700	E	3.60	19.1	ug/kg
37324-23-5	Aroclor-1262	5.60	U	5.60	19.1	ug/kg
11100-14-4	Aroclor-1268	4.00	U	4.00	19.1	ug/kg
11096-82-5	Aroclor-1260	3.60	U	3.60	19.1	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	18.9		32 - 144	94%	SPK: 20
2051-24-3	Decachlorobiphenyl	27.3		32 - 175	137%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	Weston Solutions, Inc.		Date Collected:	04/21/25	
Project:	RFP 911		Date Received:	04/23/25	
Client Sample ID:	P001-SS031-01DL		SDG No.:	Q1860	
Lab Sample ID:	Q1860-37DL		Matrix:	SOIL	
Analytical Method:	SW8082A		% Solid:	88.9	Decanted:
Sample Wt/Vol:	30.09	Units: g	Final Vol:	10000	uL
Soil Aliquot Vol:		uL	Test:	PCB	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	SW3541B				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PO110838.D	20	04/25/25 08:20	04/29/25 05:16	PB167737

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	88.6	UD	88.6	381	ug/kg
11104-28-2	Aroclor-1221	90.4	UD	90.4	381	ug/kg
11141-16-5	Aroclor-1232	83.4	UD	83.4	381	ug/kg
53469-21-9	Aroclor-1242	89.9	UD	89.9	381	ug/kg
12672-29-6	Aroclor-1248	133	UD	133	381	ug/kg
11097-69-1	Aroclor-1254	9100	ED	72.0	381	ug/kg
37324-23-5	Aroclor-1262	113	UD	113	381	ug/kg
11100-14-4	Aroclor-1268	80.7	UD	80.7	381	ug/kg
11096-82-5	Aroclor-1260	72.4	UD	72.4	381	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	24.8		32 - 144	124%	SPK: 20
2051-24-3	Decachlorobiphenyl	56.6	*	32 - 175	283%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	Weston Solutions, Inc.		Date Collected:	04/21/25	
Project:	RFP 911		Date Received:	04/23/25	
Client Sample ID:	P001-SS031-01DL2		SDG No.:	Q1860	
Lab Sample ID:	Q1860-37DL2		Matrix:	SOIL	
Analytical Method:	SW8082A		% Solid:	88.9	Decanted:
Sample Wt/Vol:	30.09	Units: g	Final Vol:	10000	uL
Soil Aliquot Vol:		uL	Test:	PCB	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	SW3541B				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PO110839.D	100	04/25/25 08:20	04/29/25 05:34	PB167737

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	443	UD	443	1900	ug/kg
11104-28-2	Aroclor-1221	452	UD	452	1900	ug/kg
11141-16-5	Aroclor-1232	417	UD	417	1900	ug/kg
53469-21-9	Aroclor-1242	450	UD	450	1900	ug/kg
12672-29-6	Aroclor-1248	664	UD	664	1900	ug/kg
11097-69-1	Aroclor-1254	9100	D	360	1900	ug/kg
37324-23-5	Aroclor-1262	563	UD	563	1900	ug/kg
11100-14-4	Aroclor-1268	404	UD	404	1900	ug/kg
11096-82-5	Aroclor-1260	362	UD	362	1900	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	0	*	32 - 144	0%	SPK: 20
2051-24-3	Decachlorobiphenyl	0	*	32 - 175	0%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	Weston Solutions, Inc.		Date Collected:	04/21/25	
Project:	RFP 911		Date Received:	04/23/25	
Client Sample ID:	P001-SS032-01		SDG No.:	Q1860	
Lab Sample ID:	Q1860-38		Matrix:	SOIL	
Analytical Method:	SW8082A		% Solid:	87	Decanted:
Sample Wt/Vol:	30.02	Units: g	Final Vol:	10000	uL
Soil Aliquot Vol:		uL	Test:	PCB	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	SW3541B				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PO110747.D	1	04/25/25 08:20	04/25/25 20:16	PB167737

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	4.50	U	4.50	19.5	ug/kg
11104-28-2	Aroclor-1221	4.60	U	4.60	19.5	ug/kg
11141-16-5	Aroclor-1232	4.30	U	4.30	19.5	ug/kg
53469-21-9	Aroclor-1242	4.60	U	4.60	19.5	ug/kg
12672-29-6	Aroclor-1248	2100	E	6.80	19.5	ug/kg
11097-69-1	Aroclor-1254	5300	E	3.70	19.5	ug/kg
37324-23-5	Aroclor-1262	5.80	U	5.80	19.5	ug/kg
11100-14-4	Aroclor-1268	4.10	U	4.10	19.5	ug/kg
11096-82-5	Aroclor-1260	3.70	U	3.70	19.5	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	20.9		32 - 144	104%	SPK: 20
2051-24-3	Decachlorobiphenyl	28.0		32 - 175	140%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	Weston Solutions, Inc.		Date Collected:	04/21/25	
Project:	RFP 911		Date Received:	04/23/25	
Client Sample ID:	P001-SS032-01DL		SDG No.:	Q1860	
Lab Sample ID:	Q1860-38DL		Matrix:	SOIL	
Analytical Method:	SW8082A		% Solid:	87	Decanted:
Sample Wt/Vol:	30.02	Units: g	Final Vol:	10000	uL
Soil Aliquot Vol:		uL	Test:	PCB	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	SW3541B				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PO110805.D	10	04/25/25 08:20	04/28/25 16:41	PB167737

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	45.4	UD	45.4	195	ug/kg
11104-28-2	Aroclor-1221	46.3	UD	46.3	195	ug/kg
11141-16-5	Aroclor-1232	42.7	UD	42.7	195	ug/kg
53469-21-9	Aroclor-1242	46.1	UD	46.1	195	ug/kg
12672-29-6	Aroclor-1248	1900	D	68.0	195	ug/kg
11097-69-1	Aroclor-1254	4000	ED	36.9	195	ug/kg
37324-23-5	Aroclor-1262	57.7	UD	57.7	195	ug/kg
11100-14-4	Aroclor-1268	41.4	UD	41.4	195	ug/kg
11096-82-5	Aroclor-1260	37.1	UD	37.1	195	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	17.8		32 - 144	89%	SPK: 20
2051-24-3	Decachlorobiphenyl	32.1		32 - 175	160%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	Weston Solutions, Inc.		Date Collected:	04/21/25	
Project:	RFP 911		Date Received:	04/23/25	
Client Sample ID:	P001-SS032-01DL2		SDG No.:	Q1860	
Lab Sample ID:	Q1860-38DL2		Matrix:	SOIL	
Analytical Method:	SW8082A		% Solid:	87	Decanted:
Sample Wt/Vol:	30.02	Units: g	Final Vol:	10000	uL
Soil Aliquot Vol:		uL	Test:	PCB	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	SW3541B				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PO110806.D	50	04/25/25 08:20	04/28/25 16:59	PB167737

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	227	UD	227	976	ug/kg
11104-28-2	Aroclor-1221	231	UD	231	976	ug/kg
11141-16-5	Aroclor-1232	214	UD	214	976	ug/kg
53469-21-9	Aroclor-1242	230	UD	230	976	ug/kg
12672-29-6	Aroclor-1248	2400	D	340	976	ug/kg
11097-69-1	Aroclor-1254	4200	D	184	976	ug/kg
37324-23-5	Aroclor-1262	288	UD	288	976	ug/kg
11100-14-4	Aroclor-1268	207	UD	207	976	ug/kg
11096-82-5	Aroclor-1260	186	UD	186	976	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	0	*	32 - 144	0%	SPK: 20
2051-24-3	Decachlorobiphenyl	0	*	32 - 175	0%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	Weston Solutions, Inc.		Date Collected:	04/21/25	
Project:	RFP 911		Date Received:	04/23/25	
Client Sample ID:	P001-SS033-01		SDG No.:	Q1860	
Lab Sample ID:	Q1860-39		Matrix:	SOIL	
Analytical Method:	SW8082A		% Solid:	87.7	Decanted:
Sample Wt/Vol:	30.05	Units: g	Final Vol:	10000	uL
Soil Aliquot Vol:		uL	Test:	PCB	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	SW3541B				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PO110748.D	1	04/25/25 08:20	04/25/25 20:34	PB167737

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	4.50	U	4.50	19.4	ug/kg
11104-28-2	Aroclor-1221	4.60	U	4.60	19.4	ug/kg
11141-16-5	Aroclor-1232	4.20	U	4.20	19.4	ug/kg
53469-21-9	Aroclor-1242	4.60	U	4.60	19.4	ug/kg
12672-29-6	Aroclor-1248	3900	E	6.70	19.4	ug/kg
11097-69-1	Aroclor-1254	3600	E	3.70	19.4	ug/kg
37324-23-5	Aroclor-1262	5.70	U	5.70	19.4	ug/kg
11100-14-4	Aroclor-1268	4.10	U	4.10	19.4	ug/kg
11096-82-5	Aroclor-1260	3.70	U	3.70	19.4	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	19.1		32 - 144	95%	SPK: 20
2051-24-3	Decachlorobiphenyl	30.5		32 - 175	153%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	Weston Solutions, Inc.		Date Collected:	04/21/25	
Project:	RFP 911		Date Received:	04/23/25	
Client Sample ID:	P001-SS033-01DL		SDG No.:	Q1860	
Lab Sample ID:	Q1860-39DL		Matrix:	SOIL	
Analytical Method:	SW8082A		% Solid:	87.7	Decanted:
Sample Wt/Vol:	30.05	Units: g	Final Vol:	10000	uL
Soil Aliquot Vol:		uL	Test:	PCB	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	SW3541B				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PO110807.D	20	04/25/25 08:20	04/28/25 17:17	PB167737

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	89.9	UD	89.9	387	ug/kg
11104-28-2	Aroclor-1221	91.8	UD	91.8	387	ug/kg
11141-16-5	Aroclor-1232	84.7	UD	84.7	387	ug/kg
53469-21-9	Aroclor-1242	91.3	UD	91.3	387	ug/kg
12672-29-6	Aroclor-1248	4900	D	135	387	ug/kg
11097-69-1	Aroclor-1254	4000	D	73.1	387	ug/kg
37324-23-5	Aroclor-1262	114	UD	114	387	ug/kg
11100-14-4	Aroclor-1268	82.0	UD	82.0	387	ug/kg
11096-82-5	Aroclor-1260	73.5	UD	73.5	387	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	22.0		32 - 144	110%	SPK: 20
2051-24-3	Decachlorobiphenyl	36.2	*	32 - 175	181%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	Weston Solutions, Inc.		Date Collected:	04/21/25	
Project:	RFP 911		Date Received:	04/23/25	
Client Sample ID:	P001-SS034-01		SDG No.:	Q1860	
Lab Sample ID:	Q1860-40		Matrix:	SOIL	
Analytical Method:	SW8082A		% Solid:	87.9	Decanted:
Sample Wt/Vol:	30.04	Units: g	Final Vol:	10000	uL
Soil Aliquot Vol:		uL	Test:	PCB	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	SW3541B				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PO110749.D	1	04/25/25 08:20	04/25/25 20:53	PB167737

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	4.50	U	4.50	19.3	ug/kg
11104-28-2	Aroclor-1221	4.60	U	4.60	19.3	ug/kg
11141-16-5	Aroclor-1232	4.20	U	4.20	19.3	ug/kg
53469-21-9	Aroclor-1242	4.60	U	4.60	19.3	ug/kg
12672-29-6	Aroclor-1248	1700	E	6.70	19.3	ug/kg
11097-69-1	Aroclor-1254	2700	E	3.60	19.3	ug/kg
37324-23-5	Aroclor-1262	5.70	U	5.70	19.3	ug/kg
11100-14-4	Aroclor-1268	4.10	U	4.10	19.3	ug/kg
11096-82-5	Aroclor-1260	3.70	U	3.70	19.3	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	13.7		32 - 144	69%	SPK: 20
2051-24-3	Decachlorobiphenyl	26.8		32 - 175	134%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	Weston Solutions, Inc.		Date Collected:	04/21/25	
Project:	RFP 911		Date Received:	04/23/25	
Client Sample ID:	P001-SS034-01DL		SDG No.:	Q1860	
Lab Sample ID:	Q1860-40DL		Matrix:	SOIL	
Analytical Method:	SW8082A		% Solid:	87.9	Decanted:
Sample Wt/Vol:	30.04	Units: g	Final Vol:	10000	uL
Soil Aliquot Vol:		uL	Test:	PCB	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	SW3541B				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PO110808.D	20	04/25/25 08:20	04/28/25 17:36	PB167737

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	89.8	UD	89.8	386	ug/kg
11104-28-2	Aroclor-1221	91.6	UD	91.6	386	ug/kg
11141-16-5	Aroclor-1232	84.5	UD	84.5	386	ug/kg
53469-21-9	Aroclor-1242	91.1	UD	91.1	386	ug/kg
12672-29-6	Aroclor-1248	2600	D	135	386	ug/kg
11097-69-1	Aroclor-1254	3600	D	72.9	386	ug/kg
37324-23-5	Aroclor-1262	114	UD	114	386	ug/kg
11100-14-4	Aroclor-1268	81.8	UD	81.8	386	ug/kg
11096-82-5	Aroclor-1260	73.4	UD	73.4	386	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	21.6		32 - 144	108%	SPK: 20
2051-24-3	Decachlorobiphenyl	50.8	*	32 - 175	254%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	Weston Solutions, Inc.		Date Collected:	04/21/25	
Project:	RFP 911		Date Received:	04/23/25	
Client Sample ID:	P001-SS035-01		SDG No.:	Q1860	
Lab Sample ID:	Q1860-41		Matrix:	SOIL	
Analytical Method:	SW8082A		% Solid:	94	Decanted:
Sample Wt/Vol:	30.01	Units: g	Final Vol:	10000	uL
Soil Aliquot Vol:		uL	Test:	PCB	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	SW3541B				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PO110758.D	1	04/25/25 08:20	04/26/25 01:34	PB167737

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	4.20	U	4.20	18.1	ug/kg
11104-28-2	Aroclor-1221	4.30	U	4.30	18.1	ug/kg
11141-16-5	Aroclor-1232	4.00	U	4.00	18.1	ug/kg
53469-21-9	Aroclor-1242	4.30	U	4.30	18.1	ug/kg
12672-29-6	Aroclor-1248	2500	E	6.30	18.1	ug/kg
11097-69-1	Aroclor-1254	3900	E	3.40	18.1	ug/kg
37324-23-5	Aroclor-1262	5.30	U	5.30	18.1	ug/kg
11100-14-4	Aroclor-1268	3.80	U	3.80	18.1	ug/kg
11096-82-5	Aroclor-1260	3.40	U	3.40	18.1	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	21.9		32 - 144	109%	SPK: 20
2051-24-3	Decachlorobiphenyl	36.2	*	32 - 175	181%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	Weston Solutions, Inc.		Date Collected:	04/21/25	
Project:	RFP 911		Date Received:	04/23/25	
Client Sample ID:	P001-SS035-01DL		SDG No.:	Q1860	
Lab Sample ID:	Q1860-41DL		Matrix:	SOIL	
Analytical Method:	SW8082A		% Solid:	94	Decanted:
Sample Wt/Vol:	30.01	Units: g	Final Vol:	10000	uL
Soil Aliquot Vol:		uL	Test:	PCB	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	SW3541B				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PO110809.D	10	04/25/25 08:20	04/28/25 17:53	PB167737

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	42.0	UD	42.0	181	ug/kg
11104-28-2	Aroclor-1221	42.9	UD	42.9	181	ug/kg
11141-16-5	Aroclor-1232	39.6	UD	39.6	181	ug/kg
53469-21-9	Aroclor-1242	42.6	UD	42.6	181	ug/kg
12672-29-6	Aroclor-1248	2900	D	63.0	181	ug/kg
11097-69-1	Aroclor-1254	4100	ED	34.1	181	ug/kg
37324-23-5	Aroclor-1262	53.4	UD	53.4	181	ug/kg
11100-14-4	Aroclor-1268	38.3	UD	38.3	181	ug/kg
11096-82-5	Aroclor-1260	34.4	UD	34.4	181	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	23.2		32 - 144	116%	SPK: 20
2051-24-3	Decachlorobiphenyl	45.8	*	32 - 175	229%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	Weston Solutions, Inc.		Date Collected:	04/21/25	
Project:	RFP 911		Date Received:	04/23/25	
Client Sample ID:	P001-SS035-01DL2		SDG No.:	Q1860	
Lab Sample ID:	Q1860-41DL2		Matrix:	SOIL	
Analytical Method:	SW8082A		% Solid:	94	Decanted:
Sample Wt/Vol:	30.01	Units: g	Final Vol:	10000	uL
Soil Aliquot Vol:		uL	Test:	PCB	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	SW3541B				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PO110810.D	50	04/25/25 08:20	04/28/25 18:12	PB167737

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	210	UD	210	904	ug/kg
11104-28-2	Aroclor-1221	214	UD	214	904	ug/kg
11141-16-5	Aroclor-1232	198	UD	198	904	ug/kg
53469-21-9	Aroclor-1242	213	UD	213	904	ug/kg
12672-29-6	Aroclor-1248	3200	D	315	904	ug/kg
11097-69-1	Aroclor-1254	4200	DP	171	904	ug/kg
37324-23-5	Aroclor-1262	267	UD	267	904	ug/kg
11100-14-4	Aroclor-1268	191	UD	191	904	ug/kg
11096-82-5	Aroclor-1260	172	UD	172	904	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	0	*	32 - 144	0%	SPK: 20
2051-24-3	Decachlorobiphenyl	0	*	32 - 175	0%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	Weston Solutions, Inc.		Date Collected:	04/21/25	
Project:	RFP 911		Date Received:	04/23/25	
Client Sample ID:	P001-SS036-01		SDG No.:	Q1860	
Lab Sample ID:	Q1860-42		Matrix:	SOIL	
Analytical Method:	SW8082A		% Solid:	94.4	Decanted:
Sample Wt/Vol:	30.07	Units: g	Final Vol:	10000	uL
Soil Aliquot Vol:		uL	Test:	PCB	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	SW3541B				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PP071539.D	1	04/25/25 08:20	04/28/25 11:32	PB167737

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	4.20	U	4.20	18.0	ug/kg
11104-28-2	Aroclor-1221	4.30	U	4.30	18.0	ug/kg
11141-16-5	Aroclor-1232	3.90	U	3.90	18.0	ug/kg
53469-21-9	Aroclor-1242	4.20	U	4.20	18.0	ug/kg
12672-29-6	Aroclor-1248	6.30	U	6.30	18.0	ug/kg
11097-69-1	Aroclor-1254	19.6	P	3.40	18.0	ug/kg
37324-23-5	Aroclor-1262	5.30	U	5.30	18.0	ug/kg
11100-14-4	Aroclor-1268	3.80	U	3.80	18.0	ug/kg
11096-82-5	Aroclor-1260	3.40	U	3.40	18.0	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	18.2		32 - 144	91%	SPK: 20
2051-24-3	Decachlorobiphenyl	16.4		32 - 175	82%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

LAB CHRONICLE

OrderID:	Q1860	OrderDate:	4/23/2025 10:15:04 AM
Client:	Weston Solutions, Inc.	Project:	RFP 911
Contact:	Smita Sumbaly	Location:	L51

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q1860-01	P001-SS001-01	SOIL	PCB	8082A	04/16/25	04/24/25	04/24/25	04/23/25
Q1860-04	P001-SS002-01	SOIL	PCB	8082A	04/16/25	04/24/25	04/24/25	04/23/25
Q1860-05	P001-SS002-02	SOIL	PCB	8082A	04/16/25	04/24/25	04/24/25	04/23/25
Q1860-06	P001-SS003-01	SOIL	PCB	8082A	04/16/25	04/24/25	04/24/25	04/23/25
Q1860-07	P001-SS004-01	SOIL	PCB	8082A	04/16/25	04/24/25	04/24/25	04/23/25
Q1860-08	P001-SS005-01	SOIL	PCB	8082A	04/16/25	04/24/25	04/24/25	04/23/25
Q1860-09	P001-SS006-01	SOIL	PCB	8082A	04/16/25	04/24/25	04/30/25	04/23/25
Q1860-09DL	P001-SS006-01DL	SOIL	PCB	8082A	04/16/25	04/24/25	04/30/25	04/23/25
Q1860-10	P001-SS007-01	SOIL	PCB	8082A	04/16/25	04/24/25	04/24/25	04/23/25
Q1860-10DL	P001-SS007-01DL	SOIL	PCB	8082A	04/16/25	04/24/25	04/24/25	04/23/25
Q1860-11	P001-SS008-01	SOIL	PCB	8082A	04/16/25	04/24/25	04/24/25	04/23/25
Q1860-11DL	P001-SS008-01DL	SOIL			04/16/25			04/23/25

LAB CHRONICLE

			PCB	8082A		04/24/25	04/24/25	
Q1860-12	P001-SS009-01	SOIL			04/17/25			04/23/25
			PCB	8082A		04/24/25	04/24/25	
Q1860-12DL	P001-SS009-01DL	SOIL			04/17/25			04/23/25
			PCB	8082A		04/24/25	04/25/25	
Q1860-13	P001-SS010-01	SOIL			04/17/25			04/23/25
			PCB	8082A		04/24/25	04/25/25	
Q1860-14	P001-SS011-01	SOIL			04/17/25			04/23/25
			PCB	8082A		04/24/25	04/24/25	
Q1860-15	P001-SS012-01	SOIL			04/17/25			04/23/25
			PCB	8082A		04/24/25	04/24/25	
Q1860-15DL	P001-SS012-01DL	SOIL			04/17/25			04/23/25
			PCB	8082A		04/24/25	04/25/25	
Q1860-16	P001-SS013-01	SOIL			04/17/25			04/23/25
			PCB	8082A		04/24/25	04/24/25	
Q1860-16DL	P001-SS013-01DL	SOIL			04/17/25			04/23/25
			PCB	8082A		04/24/25	04/25/25	
Q1860-17	P001-SS014-01	SOIL			04/17/25			04/23/25
			PCB	8082A		04/24/25	04/24/25	
Q1860-18	P001-SS015-01	SOIL			04/17/25			04/23/25
			PCB	8082A		04/24/25	04/24/25	
Q1860-19	P001-SS016-01	SOIL			04/17/25			04/23/25
			PCB	8082A		04/25/25	04/25/25	
Q1860-19DL	P001-SS016-01DL	SOIL			04/17/25			04/23/25
			PCB	8082A		04/25/25	04/28/25	
Q1860-22	P001-SS017-01	SOIL			04/17/25			04/23/25
			PCB	8082A		04/24/25	04/28/25	
Q1860-22DL	P001-SS017-01DL	SOIL			04/17/25			04/23/25
			PCB	8082A		04/24/25	04/29/25	

LAB CHRONICLE

Q1860-23	P001-SS017-02	SOIL			04/17/25			04/23/25
			PCB	8082A		04/24/25	04/28/25	
Q1860-23DL	P001-SS017-02DL	SOIL			04/17/25			04/23/25
			PCB	8082A		04/24/25	04/29/25	
Q1860-24	P001-SS018-01	SOIL			04/21/25			04/23/25
			PCB	8082A		04/24/25	04/28/25	
Q1860-25	P001-SS019-01	SOIL			04/21/25			04/23/25
			PCB	8082A		04/24/25	04/28/25	
Q1860-25DL	P001-SS019-01DL	SOIL			04/21/25			04/23/25
			PCB	8082A		04/24/25	04/29/25	
Q1860-26	P001-SS020-01	SOIL			04/21/25			04/23/25
			PCB	8082A		04/25/25	04/25/25	
Q1860-26DL	P001-SS020-01DL	SOIL			04/21/25			04/23/25
			PCB	8082A		04/25/25	04/28/25	
Q1860-27	P001-SS021-01	SOIL			04/21/25			04/23/25
			PCB	8082A		04/25/25	04/25/25	
Q1860-27DL	P001-SS021-01DL	SOIL			04/21/25			04/23/25
			PCB	8082A		04/25/25	04/28/25	
Q1860-28	P001-SS022-01	SOIL			04/21/25			04/23/25
			PCB	8082A		04/25/25	04/25/25	
Q1860-29	P001-SS023-01	SOIL			04/21/25			04/23/25
			PCB	8082A		04/25/25	04/25/25	
Q1860-29DL	P001-SS023-01DL	SOIL			04/21/25			04/23/25
			PCB	8082A		04/25/25	04/28/25	
Q1860-30	P001-SS024-01	SOIL			04/21/25			04/23/25
			PCB	8082A		04/25/25	04/25/25	
Q1860-31	P001-SS025-01	SOIL			04/21/25			04/23/25
			PCB	8082A		04/25/25	04/25/25	
Q1860-31DL	P001-SS025-01DL	SOIL			04/21/25			04/23/25

LAB CHRONICLE

PCB 8082A 04/25/25 04/28/25

Q1860-32	P001-SS026-01	SOIL			04/21/25		04/23/25
			PCB	8082A		04/25/25	04/25/25
Q1860-32DL	P001-SS026-01DL	SOIL			04/21/25		04/23/25
			PCB	8082A		04/25/25	04/28/25
Q1860-33	P001-SS027-01	SOIL			04/21/25		04/23/25
			PCB	8082A		04/25/25	04/28/25
Q1860-34	P001-SS028-01	SOIL			04/21/25		04/23/25
			PCB	8082A		04/25/25	04/25/25
Q1860-34DL	P001-SS028-01DL	SOIL			04/21/25		04/23/25
			PCB	8082A		04/25/25	04/29/25
Q1860-35	P001-SS029-01	SOIL			04/21/25		04/23/25
			PCB	8082A		04/25/25	04/28/25
Q1860-35DL	P001-SS029-01DL	SOIL			04/21/25		04/23/25
			PCB	8082A		04/25/25	04/29/25
Q1860-36	P001-SS030-01	SOIL			04/21/25		04/23/25
			PCB	8082A		04/25/25	04/25/25
Q1860-36DL	P001-SS030-01DL	SOIL			04/21/25		04/23/25
			PCB	8082A		04/25/25	04/28/25
Q1860-37	P001-SS031-01	SOIL			04/21/25		04/23/25
			PCB	8082A		04/25/25	04/25/25
Q1860-37DL	P001-SS031-01DL	SOIL			04/21/25		04/23/25
			PCB	8082A		04/25/25	04/29/25
Q1860-37DL 2	P001-SS031-01DL2	SOIL			04/21/25		04/23/25
			PCB	8082A		04/25/25	04/29/25
Q1860-38	P001-SS032-01	SOIL			04/21/25		04/23/25
			PCB	8082A		04/25/25	04/25/25
Q1860-38DL	P001-SS032-01DL	SOIL			04/21/25		04/23/25

LAB CHRONICLE

			PCB	8082A	04/25/25	04/28/25	
Q1860-38DL	P001-SS032-01DL2	SOIL			04/21/25		04/23/25
2							
			PCB	8082A	04/25/25	04/28/25	
Q1860-39	P001-SS033-01	SOIL			04/21/25		04/23/25
			PCB	8082A	04/25/25	04/25/25	
Q1860-39DL	P001-SS033-01DL	SOIL			04/21/25		04/23/25
			PCB	8082A	04/25/25	04/28/25	
Q1860-40	P001-SS034-01	SOIL			04/21/25		04/23/25
			PCB	8082A	04/25/25	04/25/25	
Q1860-40DL	P001-SS034-01DL	SOIL			04/21/25		04/23/25
			PCB	8082A	04/25/25	04/28/25	
Q1860-41	P001-SS035-01	SOIL			04/21/25		04/23/25
			PCB	8082A	04/25/25	04/26/25	
Q1860-41DL	P001-SS035-01DL	SOIL			04/21/25		04/23/25
			PCB	8082A	04/25/25	04/28/25	
Q1860-41DL	P001-SS035-01DL2	SOIL			04/21/25		04/23/25
2							
			PCB	8082A	04/25/25	04/28/25	
Q1860-42	P001-SS036-01	SOIL			04/21/25		04/23/25
			PCB	8082A	04/25/25	04/28/25	



SHIPPING DOCUMENTS

USEPA

DateShipped: 4/22/2025

CarrierName: FedEx

AirbillNo: 880713748116

CHAIN OF CUSTODY RECORD

Site #: 02FP

Contact Name Josh Frizzell

(470) 277-4600

Q1860

No: 2-042225-0004-0037-01

TAT 5 Days

Lab: Alliance Technical Group, LLC - Non
CLP


Lab Phone: 908-728-3144

Lab #	Sample #	Location	Analyses	Matrix	Sample Date	Sample Time	Numb Cont	Container	Preservative	Lab QC
	P001-SS001-01	FHG West	TAL PCBs (TAT 5 Days)	Soil	4/16/2025	11:15	2	8 oz glass	4 C	Y
	P001-SS002-01	FHG West	TAL PCBs (TAT 5 Days)	Soil	4/16/2025	11:30	1	8 oz glass	4 C	N
	P001-SS002-02	FHG West	TAL PCBs (TAT 5 Days)	Soil	4/16/2025	11:30	1	8 oz glass	4 C	N
	P001-SS003-01	FHG West	TAL PCBs (TAT 5 Days)	Soil	4/16/2025	11:45	1	8 oz glass	4 C	N
	P001-SS004-01	FHG West	TAL PCBs (TAT 5 Days)	Soil	4/16/2025	11:50	1	8 oz glass	4 C	N
	P001-SS005-01	FHG West	TAL PCBs (TAT 5 Days)	Soil	4/16/2025	12:00	1	8 oz glass	4 C	N
	P001-SS006-01	FHG West	TAL PCBs (TAT 5 Days)	Soil	4/16/2025	16:10	1	8 oz glass	4 C	N
	P001-SS007-01	FHG West	TAL PCBs (TAT 5 Days)	Soil	4/16/2025	16:15	1	8 oz glass	4 C	N
	P001-SS008-01	FHG West	TAL PCBs (TAT 5 Days)	Soil	4/16/2025	16:30	1	8 oz glass	4 C	N
	P001-SS009-01	FHG West	TAL PCBs (TAT 5 Days)	Soil	4/17/2025	12:45	1	8 oz glass	4 C	N
	P001-SS010-01	FHG West	TAL PCBs (TAT 5 Days)	Soil	4/17/2025	12:50	1	8 oz glass	4 C	N
	P001-SS011-01	FHG West	TAL PCBs (TAT 5 Days)	Soil	4/17/2025	15:35	1	8 oz glass	4 C	N
	P001-SS012-01	FHG West	TAL PCBs (TAT 5 Days)	Soil	4/17/2025	15:40	1	8 oz glass	4 C	N
	P001-SS013-01	FHG West	TAL PCBs (TAT 5 Days)	Soil	4/17/2025	15:45	1	8 oz glass	4 C	N
	P001-SS014-01	FHG West	TAL PCBs (TAT 5 Days)	Soil	4/17/2025	15:50	1	8 oz glass	4 C	N
	P001-SS015-01	FHG West	TAL PCBs (TAT 5 Days)	Soil	4/17/2025	15:55	1	8 oz glass	4 C	N
	P001-SS016-01	FHG West	TAL PCBs (TAT 5 Days)	Soil	4/17/2025	16:05	2	8 oz glass	4 C	Y
	P001-SS017-01	FHG West	TAL PCBs (TAT 5 Days)	Soil	4/17/2025	16:15	1	8 oz glass	4 C	N

Special Instructions: Please email results to S.Sumbaly@WestonSolutions.com and Josh.Frizzell@WestonSolutions.com. TAT 5 days.

RFP #911

SAMPLES TRANSFERRED FROM
CHAIN OF CUSTODY #

Items/Reason	Relinquished by (Signature and Organization)	Date/Time	Received by (Signature and Organization)	Date/Time	Sample Condition Upon Receipt
All Samples All Analyses	 Weston	4-22-25 10900		0935 4-23-25	IR Gun #1 Temp 3.2°C Custody Seal intact Temp blank present

USEPA

DateShipped: 4/22/2025

CarrierName: FedEx

AirbillNo: 880713748116

CHAIN OF CUSTODY RECORD

Site #: 02FP

Contact Name Josh Frizzell

(470) 277-4600

No: 2-042225-0004-0037-01

TAT 5 Days

Lab: Alliance Technical Group, LLC - Non
CLP



Lab Phone: 908-728-3144

Lab #	Sample #	Location	Analyses	Matrix	Sample Date	Sample Time	Numb Cont	Container	Preservative	Lab QC
	P001-SS017-02	FHG West	TAL PCBs (TAT 5 Days)	Soil	4/17/2025	16:15	1	8 oz glass	4 C	N
	P001-SS018-01	FHG West	TAL PCBs (TAT 5 Days)	Soil	4/21/2025	08:00	1	8 oz glass	4 C	N.
	P001-SS019-01	FHG West	TAL PCBs (TAT 5 Days)	Soil	4/21/2025	08:15	1	8 oz glass	4 C	N
	P001-SS020-01	FHG West	TAL PCBs (TAT 5 Days)	Soil	4/21/2025	08:17	1	8 oz glass	4 C	N
	P001-SS021-01	FHG West	TAL PCBs (TAT 5 Days)	Soil	4/21/2025	08:20	1	8 oz glass	4 C	N
	P001-SS022-01	FHG West	TAL PCBs (TAT 5 Days)	Soil	4/21/2025	08:25	1	8 oz glass	4 C	N.
	P001-SS023-01	FHG West	TAL PCBs (TAT 5 Days)	Soil	4/21/2025	08:30	1	8 oz glass	4 C	N.
	P001-SS024-01	FHG West	TAL PCBs (TAT 5 Days)	Soil	4/21/2025	08:40	1	8 oz glass	4 C	N
	P001-SS025-01	FHG West	TAL PCBs (TAT 5 Days)	Soil	4/21/2025	08:45	1	8 oz glass	4 C	N.
	P001-SS026-01	FHG West	TAL PCBs (TAT 5 Days)	Soil	4/21/2025	08:50	1	8 oz glass	4 C	N
	P001-SS027-01	FHG West	TAL PCBs (TAT 5 Days)	Soil	4/21/2025	09:00	1	8 oz glass	4 C	N
	P001-SS028-01	FHG West	TAL PCBs (TAT 5 Days)	Soil	4/21/2025	09:08	1	8 oz glass	4 C	N
	P001-SS029-01	FHG West	TAL PCBs (TAT 5 Days)	Soil	4/21/2025	09:06	1	8 oz glass	4 C	N
	P001-SS030-01	FHG West	TAL PCBs (TAT 5 Days)	Soil	4/21/2025	09:30	1	8 oz glass	4 C	N
	P001-SS031-01	FHG West	TAL PCBs (TAT 5 Days)	Soil	4/21/2025	09:50	1	8 oz glass	4 C	N.
	P001-SS032-01	FHG West	TAL PCBs (TAT 5 Days)	Soil	4/21/2025	10:00	1	8 oz glass	4 C	N
	P001-SS033-01	FHG West	TAL PCBs (TAT 5 Days)	Soil	4/21/2025	10:05	1	8 oz glass	4 C	N
	P001-SS034-01	FHG West	TAL PCBs (TAT 5 Days)	Soil	4/21/2025	10:10	1	8 oz glass	4 C	N

Special Instructions: Please email results to S.Sumbaly@WestonSolutions.com and Josh.Frizzell@WestonSolutions.com. TAT 5 days.

RFP # 911

SAMPLES TRANSFERRED FROM
CHAIN OF CUSTODY #

Items/Reason	Relinquished by (Signature and Organization)	Date/Time	Received by (Signature and Organization)	Date/Time	Sample Condition Upon Receipt
Act samples Ail Analy ses	 Weston	4-22-25 / 0900		0935 4-23-25	IR Gun #1 Temp 3.2C Custody Seal intact Temp blank present

No: 2-042225-0004-0037-01

TAT 5 Days


Lab: Alliance Technical Group, LLC - Non-CLP

Lab Phone: 908-728-3144

~~JOE~~
~~4-22-2025~~

SAMPLES TRANSFERRED FROM
CHAIN OF CUSTODY #

RFP # 911

Items/Reason	Relinquished by (Signature and Organization)	Date/Time	Received by (Signature and Organization)	Date/Time	Sample Condition Upon Receipt
All Samples All Analyses	 Weston	4-22-25 0900		0935 4-23-25	IR Gun #1 Temp 3.2C Custody seal intact Temp blank present

Laboratory Certification

Certified By	License No.
CAS EPA CLP Contract	68HERH20D0011
Connecticut	PH-0830
DOD ELAP (ANAB)	L2219
Maine	2024021
Maryland	296
New Hampshire	255424 Rev 1
New Jersey	20012
New York	11376
Pennsylvania	68-00548
Soil Permit	525-24-234-08441
Texas	T104704488