

CALIBRATION VERIFICATION SUMMARY

Contract: ROYF02

Lab Code: CHEM Case No.: Q1860 SAS No.: Q1860 SDG NO.: Q1860

Continuing Calib Date: 04/28/2025 Initial Calibration Date(s): 04/10/2025 04/10/2025

Continuing Calib Time: 08:50 Initial Calibration Time(s): 09:36 17:52

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	4.78	4.78	4.68	4.88	0.00
Aroclor-1016-2	(2)	4.80	4.80	4.70	4.90	0.00
Aroclor-1016-3	(3)	4.85	4.86	4.76	4.96	0.01
Aroclor-1016-4	(4)	4.97	4.98	4.88	5.08	0.01
Aroclor-1016-5	(5)	5.23	5.23	5.13	5.33	0.00
Aroclor-1260-1	(1)	6.27	6.27	6.17	6.37	0.00
Aroclor-1260-2	(2)	6.46	6.46	6.36	6.56	0.00
Aroclor-1260-3	(3)	6.82	6.83	6.73	6.93	0.01
Aroclor-1260-4	(4)	7.08	7.09	6.99	7.19	0.01
Aroclor-1260-5	(5)	7.33	7.33	7.23	7.43	0.00
Tetrachloro-m-xylene		3.69	3.69	3.59	3.79	0.00
Decachlorobiphenyl		8.73	8.73	8.63	8.83	0.00

CALIBRATION VERIFICATION SUMMARY

Contract: ROYF02

Lab Code: CHEM Case No.: Q1860 SAS No.: Q1860 SDG NO.: Q1860

Continuing Calib Date: 04/28/2025 Initial Calibration Date(s): 04/10/2025 04/10/2025

Continuing Calib Time: 08:50 Initial Calibration Time(s): 09:36 17:52

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
Aroclor-1016-1	(1)	4.76	4.77	4.67	4.87	0.01
Aroclor-1016-2	(2)	4.78	4.78	4.68	4.88	0.00
Aroclor-1016-3	(3)	4.96	4.96	4.86	5.06	0.00
Aroclor-1016-4	(4)	5.00	5.00	4.90	5.10	0.00
Aroclor-1016-5	(5)	5.21	5.22	5.12	5.32	0.01
Aroclor-1260-1	(1)	6.24	6.25	6.15	6.35	0.01
Aroclor-1260-2	(2)	6.43	6.43	6.33	6.53	0.00
Aroclor-1260-3	(3)	6.58	6.59	6.49	6.69	0.01
Aroclor-1260-4	(4)	7.05	7.06	6.96	7.16	0.01
Aroclor-1260-5	(5)	7.29	7.30	7.20	7.40	0.01
Tetrachloro-m-xylene		3.68	3.69	3.59	3.79	0.01
Decachlorobiphenyl		8.68	8.68	8.58	8.78	0.00