

### CALIBRATION VERIFICATION SUMMARY

Contract: ROYF02

Lab Code: CHEM Case No.: Q1860 SAS No.: Q1860 SDG NO.: Q1860

Continuing Calib Date: 04/29/2025 Initial Calibration Date(s): 04/10/2025 04/10/2025

Continuing Calib Time: 08:11 Initial Calibration Time(s): 09:36 17:52

GC Column: ZB-MR1 ID: 0.32 (mm)

| COMPOUND             |     | CCAL<br>RT | AVG<br>RT | RT WINDOW |      | DIFF<br>RT |
|----------------------|-----|------------|-----------|-----------|------|------------|
|                      |     |            |           | FROM      | TO   |            |
| Aroclor-1254-1       | (1) | 5.58       | 5.59      | 5.49      | 5.69 | 0.01       |
| Aroclor-1254-2       | (2) | 5.73       | 5.74      | 5.64      | 5.84 | 0.01       |
| Aroclor-1254-3       | (3) | 6.14       | 6.14      | 6.04      | 6.24 | 0.00       |
| Aroclor-1254-4       | (4) | 6.37       | 6.37      | 6.27      | 6.47 | 0.00       |
| Aroclor-1254-5       | (5) | 6.79       | 6.79      | 6.69      | 6.89 | 0.00       |
| Tetrachloro-m-xylene |     | 3.69       | 3.69      | 3.59      | 3.79 | 0.00       |
| Decachlorobiphenyl   |     | 8.73       | 8.73      | 8.63      | 8.83 | 0.00       |

### CALIBRATION VERIFICATION SUMMARY

Contract: ROYF02

Lab Code: CHEM Case No.: Q1860 SAS No.: Q1860 SDG NO.: Q1860

Continuing Calib Date: 04/29/2025 Initial Calibration Date(s): 04/10/2025 04/10/2025

Continuing Calib Time: 08:11 Initial Calibration Time(s): 09:36 17:52

GC Column: ZB-MR2 ID: 0.32 (mm)

| COMPOUND             |     | CCAL<br>RT | AVG<br>RT | RT WINDOW |      | DIFF<br>RT |
|----------------------|-----|------------|-----------|-----------|------|------------|
|                      |     |            |           | FROM      | TO   |            |
| Aroclor-1254-1       | (1) | 5.56       | 5.57      | 5.47      | 5.67 | 0.01       |
| Aroclor-1254-2       | (2) | 5.71       | 5.71      | 5.61      | 5.81 | 0.00       |
| Aroclor-1254-3       | (3) | 6.11       | 6.12      | 6.02      | 6.22 | 0.01       |
| Aroclor-1254-4       | (4) | 6.34       | 6.34      | 6.24      | 6.44 | 0.00       |
| Aroclor-1254-5       | (5) | 6.76       | 6.76      | 6.66      | 6.86 | 0.00       |
| Tetrachloro-m-xylene |     | 3.68       | 3.69      | 3.59      | 3.79 | 0.01       |
| Decachlorobiphenyl   |     | 8.68       | 8.68      | 8.58      | 8.78 | 0.00       |