

### CALIBRATION VERIFICATION SUMMARY

Contract: ROYF02

Lab Code: CHEM Case No.: Q1860 SAS No.: Q1860 SDG NO.: Q1860

Continuing Calib Date: 04/30/2025 Initial Calibration Date(s): 04/10/2025 04/10/2025

Continuing Calib Time: 02:48 Initial Calibration Time(s): 09:36 17:52

GC Column: ZB-MR1 ID: 0.32 (mm)

| COMPOUND             | CCAL<br>RT | AVG<br>RT | RT WINDOW |      | DIFF<br>RT |
|----------------------|------------|-----------|-----------|------|------------|
|                      |            |           | FROM      | TO   |            |
| Aroclor-1242-1 (1)   | 4.78       | 4.78      | 4.68      | 4.88 | 0.00       |
| Aroclor-1242-2 (2)   | 4.80       | 4.80      | 4.70      | 4.90 | 0.00       |
| Aroclor-1242-3 (3)   | 4.85       | 4.86      | 4.76      | 4.96 | 0.01       |
| Aroclor-1242-4 (4)   | 4.97       | 4.98      | 4.88      | 5.08 | 0.01       |
| Aroclor-1242-5 (5)   | 5.62       | 5.63      | 5.53      | 5.73 | 0.01       |
| Tetrachloro-m-xylene | 3.69       | 3.69      | 3.59      | 3.79 | 0.00       |
| Decachlorobiphenyl   | 8.73       | 8.73      | 8.63      | 8.83 | 0.00       |

### CALIBRATION VERIFICATION SUMMARY

Contract: ROYF02

Lab Code: CHEM Case No.: Q1860 SAS No.: Q1860 SDG NO.: Q1860

Continuing Calib Date: 04/30/2025 Initial Calibration Date(s): 04/10/2025 04/10/2025

Continuing Calib Time: 02:48 Initial Calibration Time(s): 09:36 17:52

GC Column: ZB-MR2 ID: 0.32 (mm)

| COMPOUND             | CCAL<br>RT | AVG<br>RT | RT WINDOW |      | DIFF<br>RT |
|----------------------|------------|-----------|-----------|------|------------|
|                      |            |           | FROM      | TO   |            |
| Aroclor-1242-1 (1)   | 4.76       | 4.77      | 4.67      | 4.87 | 0.01       |
| Aroclor-1242-2 (2)   | 4.78       | 4.78      | 4.68      | 4.88 | 0.00       |
| Aroclor-1242-3 (3)   | 4.96       | 4.96      | 4.86      | 5.06 | 0.00       |
| Aroclor-1242-4 (4)   | 5.04       | 5.04      | 4.94      | 5.14 | 0.00       |
| Aroclor-1242-5 (5)   | 5.56       | 5.56      | 5.46      | 5.66 | 0.00       |
| Tetrachloro-m-xylene | 3.68       | 3.68      | 3.58      | 3.78 | 0.00       |
| Decachlorobiphenyl   | 8.68       | 8.68      | 8.58      | 8.78 | 0.00       |