

### CALIBRATION VERIFICATION SUMMARY

Contract: ROYF02

Lab Code: CHEM Case No.: Q1860 SAS No.: Q1860 SDG NO.: Q1860

Continuing Calib Date: 04/28/2025 Initial Calibration Date(s): 04/22/2025 04/22/2025

Continuing Calib Time: 09:20 Initial Calibration Time(s): 10:29 17:49

GC Column: ZB-MR1 ID: 0.32 (mm)

| COMPOUND             |     | CCAL<br>RT | AVG<br>RT | RT WINDOW |       | DIFF<br>RT |
|----------------------|-----|------------|-----------|-----------|-------|------------|
|                      |     |            |           | FROM      | TO    |            |
| Aroclor-1254-1       | (1) | 6.51       | 6.52      | 6.42      | 6.62  | 0.01       |
| Aroclor-1254-2       | (2) | 6.73       | 6.74      | 6.64      | 6.84  | 0.01       |
| Aroclor-1254-3       | (3) | 7.09       | 7.10      | 7.00      | 7.20  | 0.01       |
| Aroclor-1254-4       | (4) | 7.37       | 7.38      | 7.28      | 7.48  | 0.01       |
| Aroclor-1254-5       | (5) | 7.79       | 7.80      | 7.70      | 7.90  | 0.01       |
| Tetrachloro-m-xylene |     | 4.51       | 4.52      | 4.42      | 4.62  | 0.01       |
| Decachlorobiphenyl   |     | 10.23      | 10.24     | 10.14     | 10.34 | 0.01       |

### CALIBRATION VERIFICATION SUMMARY

Contract: ROYF02

Lab Code: CHEM Case No.: Q1860 SAS No.: Q1860 SDG NO.: Q1860

Continuing Calib Date: 04/28/2025 Initial Calibration Date(s): 04/22/2025 04/22/2025

Continuing Calib Time: 09:20 Initial Calibration Time(s): 10:29 17:49

GC Column: ZB-MR2 ID: 0.32 (mm)

| COMPOUND             | CCAL<br>RT | AVG<br>RT | RT WINDOW |      | DIFF<br>RT |
|----------------------|------------|-----------|-----------|------|------------|
|                      |            |           | FROM      | TO   |            |
| Aroclor-1254-1 (1)   | 5.70       | 5.70      | 5.60      | 5.80 | 0.00       |
| Aroclor-1254-2 (2)   | 5.85       | 5.85      | 5.75      | 5.95 | 0.01       |
| Aroclor-1254-3 (3)   | 6.25       | 6.25      | 6.15      | 6.35 | 0.00       |
| Aroclor-1254-4 (4)   | 6.48       | 6.48      | 6.38      | 6.58 | 0.00       |
| Aroclor-1254-5 (5)   | 6.90       | 6.90      | 6.80      | 7.00 | 0.01       |
| Tetrachloro-m-xylene | 3.81       | 3.81      | 3.71      | 3.91 | 0.00       |
| Decachlorobiphenyl   | 8.84       | 8.85      | 8.75      | 8.95 | 0.01       |