

CALIBRATION VERIFICATION SUMMARY

Contract: ROYF02

Lab Code: CHEM Case No.: Q1860 SAS No.: Q1860 SDG NO.: Q1860

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 04/10/2025 04/10/2025

Client Sample No.: CCAL02 Date Analyzed: 04/24/2025

Lab Sample No.: AR1660CCC500 Data File : PO110688.D Time Analyzed: 17:22

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.776	4.679	4.879	527.140	500.000	5.4
Aroclor-1016-2	4.795	4.698	4.898	534.060	500.000	6.8
Aroclor-1016-3	4.852	4.755	4.955	515.500	500.000	3.1
Aroclor-1016-4	4.972	4.875	5.075	528.050	500.000	5.6
Aroclor-1016-5	5.229	5.132	5.332	557.750	500.000	11.6
Aroclor-1260-1	6.268	6.172	6.372	505.840	500.000	1.2
Aroclor-1260-2	6.457	6.360	6.560	484.770	500.000	-3.0
Aroclor-1260-3	6.825	6.729	6.929	490.280	500.000	-1.9
Aroclor-1260-4	7.086	6.988	7.188	483.790	500.000	-3.2
Aroclor-1260-5	7.327	7.230	7.430	485.640	500.000	-2.9
Decachlorobiphenyl	8.728	8.632	8.832	42.720	50.000	-14.6
Tetrachloro-m-xylene	3.687	3.588	3.788	52.990	50.000	6.0

CALIBRATION VERIFICATION SUMMARY

Contract: ROYF02

Lab Code: CHEM Case No.: Q1860 SAS No.: Q1860 SDG NO.: Q1860

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 04/10/2025 04/10/2025

Client Sample No.: CCAL02 Date Analyzed: 04/24/2025

Lab Sample No.: AR1660CCC500 Data File : PO110688.D Time Analyzed: 17:22

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.763	4.666	4.866	496.730	500.000	-0.7
Aroclor-1016-2	4.782	4.684	4.884	497.570	500.000	-0.5
Aroclor-1016-3	4.957	4.860	5.060	487.620	500.000	-2.5
Aroclor-1016-4	5.000	4.902	5.102	457.790	500.000	-8.4
Aroclor-1016-5	5.212	5.115	5.315	503.600	500.000	0.7
Aroclor-1260-1	6.244	6.146	6.346	476.820	500.000	-4.6
Aroclor-1260-2	6.431	6.334	6.534	465.690	500.000	-6.9
Aroclor-1260-3	6.583	6.487	6.687	458.890	500.000	-8.2
Aroclor-1260-4	7.054	6.957	7.157	449.330	500.000	-10.1
Aroclor-1260-5	7.296	7.197	7.397	452.820	500.000	-9.4
Decachlorobiphenyl	8.681	8.584	8.784	43.660	50.000	-12.7
Tetrachloro-m-xylene	3.684	3.586	3.786	50.850	50.000	1.7