

CALIBRATION VERIFICATION SUMMARY

Contract: ROYF02

Lab Code: CHEM Case No.: Q1860 SAS No.: Q1860 SDG NO.: Q1860

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 04/10/2025 04/10/2025

Client Sample No.: CCAL03 Date Analyzed: 04/24/2025

Lab Sample No.: AR1660CCC500 Data File : PO110703.D Time Analyzed: 23:35

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.777	4.679	4.879	533.100	500.000	6.6
Aroclor-1016-2	4.796	4.698	4.898	537.560	500.000	7.5
Aroclor-1016-3	4.852	4.755	4.955	524.530	500.000	4.9
Aroclor-1016-4	4.972	4.875	5.075	533.240	500.000	6.6
Aroclor-1016-5	5.230	5.132	5.332	570.520	500.000	14.1
Aroclor-1260-1	6.269	6.172	6.372	537.990	500.000	7.6
Aroclor-1260-2	6.458	6.360	6.560	516.960	500.000	3.4
Aroclor-1260-3	6.825	6.729	6.929	512.420	500.000	2.5
Aroclor-1260-4	7.085	6.988	7.188	501.680	500.000	0.3
Aroclor-1260-5	7.327	7.230	7.430	500.140	500.000	0.0
Decachlorobiphenyl	8.728	8.632	8.832	45.370	50.000	-9.3
Tetrachloro-m-xylene	3.687	3.588	3.788	53.550	50.000	7.1

CALIBRATION VERIFICATION SUMMARY

Contract: ROYF02

Lab Code: CHEM Case No.: Q1860 SAS No.: Q1860 SDG NO.: Q1860

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 04/10/2025 04/10/2025

Client Sample No.: CCAL03 Date Analyzed: 04/24/2025

Lab Sample No.: AR1660CCC500 Data File : PO110703.D Time Analyzed: 23:35

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.763	4.666	4.866	501.920	500.000	0.4
Aroclor-1016-2	4.782	4.684	4.884	499.310	500.000	-0.1
Aroclor-1016-3	4.957	4.860	5.060	492.420	500.000	-1.5
Aroclor-1016-4	4.999	4.902	5.102	452.970	500.000	-9.4
Aroclor-1016-5	5.212	5.115	5.315	513.740	500.000	2.7
Aroclor-1260-1	6.243	6.146	6.346	490.050	500.000	-2.0
Aroclor-1260-2	6.430	6.334	6.534	479.640	500.000	-4.1
Aroclor-1260-3	6.583	6.487	6.687	475.250	500.000	-5.0
Aroclor-1260-4	7.054	6.957	7.157	472.900	500.000	-5.4
Aroclor-1260-5	7.295	7.197	7.397	476.250	500.000	-4.8
Decachlorobiphenyl	8.680	8.584	8.784	46.300	50.000	-7.4
Tetrachloro-m-xylene	3.683	3.586	3.786	51.560	50.000	3.1