

CALIBRATION VERIFICATION SUMMARY

Contract: ROYF02

Lab Code: CHEM Case No.: Q1860 SAS No.: Q1860 SDG NO.: Q1860

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 04/10/2025 04/10/2025

Client Sample No.: CCAL04 Date Analyzed: 04/25/2025

Lab Sample No.: AR1660CCC500 Data File : PO110714.D Time Analyzed: 04:38

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.776	4.679	4.879	533.230	500.000	6.6
Aroclor-1016-2	4.795	4.698	4.898	539.670	500.000	7.9
Aroclor-1016-3	4.852	4.755	4.955	524.650	500.000	4.9
Aroclor-1016-4	4.972	4.875	5.075	534.760	500.000	7.0
Aroclor-1016-5	5.228	5.132	5.332	533.320	500.000	6.7
Aroclor-1260-1	6.269	6.172	6.372	552.460	500.000	10.5
Aroclor-1260-2	6.457	6.360	6.560	528.810	500.000	5.8
Aroclor-1260-3	6.825	6.729	6.929	522.060	500.000	4.4
Aroclor-1260-4	7.085	6.988	7.188	512.740	500.000	2.5
Aroclor-1260-5	7.327	7.230	7.430	517.050	500.000	3.4
Decachlorobiphenyl	8.727	8.632	8.832	46.830	50.000	-6.3
Tetrachloro-m-xylene	3.686	3.588	3.788	53.670	50.000	7.3

CALIBRATION VERIFICATION SUMMARY

Contract: ROYF02

Lab Code: CHEM Case No.: Q1860 SAS No.: Q1860 SDG NO.: Q1860

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 04/10/2025 04/10/2025

Client Sample No.: CCAL04 Date Analyzed: 04/25/2025

Lab Sample No.: AR1660CCC500 Data File : PO110714.D Time Analyzed: 04:38

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.763	4.666	4.866	499.230	500.000	-0.2
Aroclor-1016-2	4.782	4.684	4.884	500.630	500.000	0.1
Aroclor-1016-3	4.957	4.860	5.060	489.660	500.000	-2.1
Aroclor-1016-4	4.999	4.902	5.102	426.090	500.000	-14.8
Aroclor-1016-5	5.212	5.115	5.315	538.280	500.000	7.7
Aroclor-1260-1	6.243	6.146	6.346	498.320	500.000	-0.3
Aroclor-1260-2	6.431	6.334	6.534	486.000	500.000	-2.8
Aroclor-1260-3	6.583	6.487	6.687	477.710	500.000	-4.5
Aroclor-1260-4	7.054	6.957	7.157	475.560	500.000	-4.9
Aroclor-1260-5	7.295	7.197	7.397	479.150	500.000	-4.2
Decachlorobiphenyl	8.681	8.584	8.784	46.100	50.000	-7.8
Tetrachloro-m-xylene	3.683	3.586	3.786	51.400	50.000	2.8