

CALIBRATION VERIFICATION SUMMARY

Contract: ROYF02

Lab Code: CHEM Case No.: Q1860 SAS No.: Q1860 SDG NO.: Q1860

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 04/10/2025 04/10/2025

Client Sample No.: CCAL05 Date Analyzed: 04/25/2025

Lab Sample No.: AR1660CCC500 Data File : PO110720.D Time Analyzed: 09:24

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.776	4.679	4.879	519.800	500.000	4.0
Aroclor-1016-2	4.795	4.698	4.898	529.150	500.000	5.8
Aroclor-1016-3	4.852	4.755	4.955	515.150	500.000	3.0
Aroclor-1016-4	4.972	4.875	5.075	525.650	500.000	5.1
Aroclor-1016-5	5.230	5.132	5.332	522.090	500.000	4.4
Aroclor-1260-1	6.269	6.172	6.372	511.280	500.000	2.3
Aroclor-1260-2	6.458	6.360	6.560	502.580	500.000	0.5
Aroclor-1260-3	6.825	6.729	6.929	501.760	500.000	0.4
Aroclor-1260-4	7.085	6.988	7.188	497.560	500.000	-0.5
Aroclor-1260-5	7.327	7.230	7.430	495.450	500.000	-0.9
Decachlorobiphenyl	8.728	8.632	8.832	45.580	50.000	-8.8
Tetrachloro-m-xylene	3.686	3.588	3.788	53.220	50.000	6.4

CALIBRATION VERIFICATION SUMMARY

Contract: ROYF02

Lab Code: CHEM Case No.: Q1860 SAS No.: Q1860 SDG NO.: Q1860

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 04/10/2025 04/10/2025

Client Sample No.: CCAL05 Date Analyzed: 04/25/2025

Lab Sample No.: AR1660CCC500 Data File : PO110720.D Time Analyzed: 09:24

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.762	4.666	4.866	486.210	500.000	-2.8
Aroclor-1016-2	4.781	4.684	4.884	484.370	500.000	-3.1
Aroclor-1016-3	4.956	4.860	5.060	461.780	500.000	-7.6
Aroclor-1016-4	4.998	4.902	5.102	460.000	500.000	-8.0
Aroclor-1016-5	5.211	5.115	5.315	471.920	500.000	-5.6
Aroclor-1260-1	6.242	6.146	6.346	470.310	500.000	-5.9
Aroclor-1260-2	6.429	6.334	6.534	462.170	500.000	-7.6
Aroclor-1260-3	6.582	6.487	6.687	458.670	500.000	-8.3
Aroclor-1260-4	7.053	6.957	7.157	459.830	500.000	-8.0
Aroclor-1260-5	7.294	7.197	7.397	456.120	500.000	-8.8
Decachlorobiphenyl	8.679	8.584	8.784	43.970	50.000	-12.1
Tetrachloro-m-xylene	3.683	3.586	3.786	50.330	50.000	0.7