

CALIBRATION VERIFICATION SUMMARY

Contract: ROYF02

Lab Code: CHEM Case No.: Q1860 SAS No.: Q1860 SDG NO.: Q1860

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 04/10/2025 04/10/2025

Client Sample No.: CCAL07 Date Analyzed: 04/25/2025

Lab Sample No.: AR1660CCC500 Data File : PO110750.D Time Analyzed: 23:07

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.776	4.679	4.879	592.280	500.000	18.5
Aroclor-1016-2	4.795	4.698	4.898	601.570	500.000	20.3
Aroclor-1016-3	4.852	4.755	4.955	579.250	500.000	15.9
Aroclor-1016-4	4.972	4.875	5.075	591.810	500.000	18.4
Aroclor-1016-5	5.229	5.132	5.332	562.360	500.000	12.5
Aroclor-1260-1	6.269	6.172	6.372	535.110	500.000	7.0
Aroclor-1260-2	6.459	6.360	6.560	502.530	500.000	0.5
Aroclor-1260-3	6.826	6.729	6.929	515.560	500.000	3.1
Aroclor-1260-4	7.086	6.988	7.188	521.290	500.000	4.3
Aroclor-1260-5	7.329	7.230	7.430	497.400	500.000	-0.5
Decachlorobiphenyl	8.729	8.632	8.832	43.050	50.000	-13.9
Tetrachloro-m-xylene	3.687	3.588	3.788	60.850	50.000	21.7

CALIBRATION VERIFICATION SUMMARY

Contract: ROYF02

Lab Code: CHEM Case No.: Q1860 SAS No.: Q1860 SDG NO.: Q1860

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 04/10/2025 04/10/2025

Client Sample No.: CCAL07 Date Analyzed: 04/25/2025

Lab Sample No.: AR1660CCC500 Data File : PO110750.D Time Analyzed: 23:07

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.763	4.666	4.866	566.480	500.000	13.3
Aroclor-1016-2	4.782	4.684	4.884	564.460	500.000	12.9
Aroclor-1016-3	4.957	4.860	5.060	553.640	500.000	10.7
Aroclor-1016-4	5.000	4.902	5.102	463.690	500.000	-7.3
Aroclor-1016-5	5.212	5.115	5.315	541.520	500.000	8.3
Aroclor-1260-1	6.243	6.146	6.346	537.220	500.000	7.4
Aroclor-1260-2	6.431	6.334	6.534	516.220	500.000	3.2
Aroclor-1260-3	6.584	6.487	6.687	497.970	500.000	-0.4
Aroclor-1260-4	7.055	6.957	7.157	487.730	500.000	-2.5
Aroclor-1260-5	7.296	7.197	7.397	482.310	500.000	-3.5
Decachlorobiphenyl	8.681	8.584	8.784	46.320	50.000	-7.4
Tetrachloro-m-xylene	3.684	3.586	3.786	58.410	50.000	16.8