

CALIBRATION VERIFICATION SUMMARY

Contract: ROYF02

Lab Code: CHEM Case No.: Q1860 SAS No.: Q1860 SDG NO.: Q1860

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 04/10/2025 04/10/2025

Client Sample No.: CCAL08 Date Analyzed: 04/26/2025

Lab Sample No.: AR1660CCC500 Data File : PO110765.D Time Analyzed: 05:02

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.777	4.679	4.879	547.010	500.000	9.4
Aroclor-1016-2	4.796	4.698	4.898	548.410	500.000	9.7
Aroclor-1016-3	4.853	4.755	4.955	532.670	500.000	6.5
Aroclor-1016-4	4.973	4.875	5.075	545.810	500.000	9.2
Aroclor-1016-5	5.230	5.132	5.332	572.870	500.000	14.6
Aroclor-1260-1	6.270	6.172	6.372	532.990	500.000	6.6
Aroclor-1260-2	6.459	6.360	6.560	513.020	500.000	2.6
Aroclor-1260-3	6.826	6.729	6.929	509.410	500.000	1.9
Aroclor-1260-4	7.086	6.988	7.188	492.120	500.000	-1.6
Aroclor-1260-5	7.328	7.230	7.430	475.800	500.000	-4.8
Decachlorobiphenyl	8.729	8.632	8.832	42.700	50.000	-14.6
Tetrachloro-m-xylene	3.687	3.588	3.788	54.780	50.000	9.6

CALIBRATION VERIFICATION SUMMARY

Contract: ROYF02

Lab Code: CHEM Case No.: Q1860 SAS No.: Q1860 SDG NO.: Q1860

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 04/10/2025 04/10/2025

Client Sample No.: CCAL08 Date Analyzed: 04/26/2025

Lab Sample No.: AR1660CCC500 Data File : PO110765.D Time Analyzed: 05:02

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.764	4.666	4.866	508.470	500.000	1.7
Aroclor-1016-2	4.783	4.684	4.884	511.570	500.000	2.3
Aroclor-1016-3	4.958	4.860	5.060	498.960	500.000	-0.2
Aroclor-1016-4	5.000	4.902	5.102	472.650	500.000	-5.5
Aroclor-1016-5	5.213	5.115	5.315	521.480	500.000	4.3
Aroclor-1260-1	6.244	6.146	6.346	492.430	500.000	-1.5
Aroclor-1260-2	6.432	6.334	6.534	480.940	500.000	-3.8
Aroclor-1260-3	6.585	6.487	6.687	477.240	500.000	-4.6
Aroclor-1260-4	7.056	6.957	7.157	469.790	500.000	-6.0
Aroclor-1260-5	7.296	7.197	7.397	471.610	500.000	-5.7
Decachlorobiphenyl	8.682	8.584	8.784	43.510	50.000	-13.0
Tetrachloro-m-xylene	3.685	3.586	3.786	52.590	50.000	5.2