

CALIBRATION VERIFICATION SUMMARY

Contract: ROYF02

Lab Code: CHEM Case No.: Q1860 SAS No.: Q1860 SDG NO.: Q1860

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 04/10/2025 04/10/2025

Client Sample No.: CCAL09 Date Analyzed: 04/28/2025

Lab Sample No.: AR1660CCC500 Data File : PO110783.D Time Analyzed: 08:50

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.776	4.679	4.879	585.950	500.000	17.2
Aroclor-1016-2	4.795	4.698	4.898	597.980	500.000	19.6
Aroclor-1016-3	4.851	4.755	4.955	577.510	500.000	15.5
Aroclor-1016-4	4.971	4.875	5.075	590.620	500.000	18.1
Aroclor-1016-5	5.228	5.132	5.332	580.930	500.000	16.2
Aroclor-1260-1	6.268	6.172	6.372	563.280	500.000	12.7
Aroclor-1260-2	6.457	6.360	6.560	544.360	500.000	8.9
Aroclor-1260-3	6.824	6.729	6.929	541.920	500.000	8.4
Aroclor-1260-4	7.083	6.988	7.188	530.580	500.000	6.1
Aroclor-1260-5	7.326	7.230	7.430	539.380	500.000	7.9
Decachlorobiphenyl	8.727	8.632	8.832	47.350	50.000	-5.3
Tetrachloro-m-xylene	3.686	3.588	3.788	60.140	50.000	20.3

CALIBRATION VERIFICATION SUMMARY

Contract: ROYF02

Lab Code: CHEM Case No.: Q1860 SAS No.: Q1860 SDG NO.: Q1860

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 04/10/2025 04/10/2025

Client Sample No.: CCAL09 Date Analyzed: 04/28/2025

Lab Sample No.: AR1660CCC500 Data File : PO110783.D Time Analyzed: 08:50

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.762	4.666	4.866	532.110	500.000	6.4
Aroclor-1016-2	4.781	4.684	4.884	533.020	500.000	6.6
Aroclor-1016-3	4.956	4.860	5.060	520.740	500.000	4.1
Aroclor-1016-4	4.998	4.902	5.102	513.870	500.000	2.8
Aroclor-1016-5	5.211	5.115	5.315	514.310	500.000	2.9
Aroclor-1260-1	6.241	6.146	6.346	511.960	500.000	2.4
Aroclor-1260-2	6.429	6.334	6.534	502.140	500.000	0.4
Aroclor-1260-3	6.582	6.487	6.687	498.290	500.000	-0.3
Aroclor-1260-4	7.053	6.957	7.157	495.070	500.000	-1.0
Aroclor-1260-5	7.293	7.197	7.397	490.000	500.000	-2.0
Decachlorobiphenyl	8.679	8.584	8.784	46.180	50.000	-7.6
Tetrachloro-m-xylene	3.683	3.586	3.786	55.420	50.000	10.8