

CALIBRATION VERIFICATION SUMMARY

Contract: ROYF02

Lab Code: CHEM Case No.: Q1860 SAS No.: Q1860 SDG NO.: Q1860

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 04/10/2025 04/10/2025

Client Sample No.: CCAL10 Date Analyzed: 04/28/2025

Lab Sample No.: AR1660CCC500 Data File : PO110798.D Time Analyzed: 14:33

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.776	4.679	4.879	525.390	500.000	5.1
Aroclor-1016-2	4.795	4.698	4.898	529.570	500.000	5.9
Aroclor-1016-3	4.852	4.755	4.955	516.130	500.000	3.2
Aroclor-1016-4	4.972	4.875	5.075	527.400	500.000	5.5
Aroclor-1016-5	5.229	5.132	5.332	573.980	500.000	14.8
Aroclor-1260-1	6.268	6.172	6.372	522.950	500.000	4.6
Aroclor-1260-2	6.457	6.360	6.560	505.960	500.000	1.2
Aroclor-1260-3	6.824	6.729	6.929	500.100	500.000	0.0
Aroclor-1260-4	7.084	6.988	7.188	492.520	500.000	-1.5
Aroclor-1260-5	7.326	7.230	7.430	486.490	500.000	-2.7
Decachlorobiphenyl	8.727	8.632	8.832	43.770	50.000	-12.5
Tetrachloro-m-xylene	3.686	3.588	3.788	53.480	50.000	7.0

CALIBRATION VERIFICATION SUMMARY

Contract: ROYF02

Lab Code: CHEM Case No.: Q1860 SAS No.: Q1860 SDG NO.: Q1860

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 04/10/2025 04/10/2025

Client Sample No.: CCAL10 Date Analyzed: 04/28/2025

Lab Sample No.: AR1660CCC500 Data File : PO110798.D Time Analyzed: 14:33

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.762	4.666	4.866	483.340	500.000	-3.3
Aroclor-1016-2	4.781	4.684	4.884	481.240	500.000	-3.8
Aroclor-1016-3	4.956	4.860	5.060	467.650	500.000	-6.5
Aroclor-1016-4	4.998	4.902	5.102	436.010	500.000	-12.8
Aroclor-1016-5	5.211	5.115	5.315	496.610	500.000	-0.7
Aroclor-1260-1	6.241	6.146	6.346	467.120	500.000	-6.6
Aroclor-1260-2	6.430	6.334	6.534	455.240	500.000	-9.0
Aroclor-1260-3	6.582	6.487	6.687	450.360	500.000	-9.9
Aroclor-1260-4	7.053	6.957	7.157	450.210	500.000	-10.0
Aroclor-1260-5	7.294	7.197	7.397	446.080	500.000	-10.8
Decachlorobiphenyl	8.679	8.584	8.784	42.340	50.000	-15.3
Tetrachloro-m-xylene	3.683	3.586	3.786	49.750	50.000	-0.5