

CALIBRATION VERIFICATION SUMMARY

Contract: ROYF02

Lab Code: CHEM Case No.: Q1860 SAS No.: Q1860 SDG NO.: Q1860

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 04/10/2025 04/10/2025

Client Sample No.: CCAL11 Date Analyzed: 04/28/2025

Lab Sample No.: AR1660CCC500 Data File : PO110813.D Time Analyzed: 19:58

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.775	4.679	4.879	522.610	500.000	4.5
Aroclor-1016-2	4.794	4.698	4.898	523.190	500.000	4.6
Aroclor-1016-3	4.851	4.755	4.955	506.550	500.000	1.3
Aroclor-1016-4	4.971	4.875	5.075	518.260	500.000	3.7
Aroclor-1016-5	5.228	5.132	5.332	568.780	500.000	13.8
Aroclor-1260-1	6.269	6.172	6.372	505.510	500.000	1.1
Aroclor-1260-2	6.458	6.360	6.560	483.880	500.000	-3.2
Aroclor-1260-3	6.825	6.729	6.929	481.120	500.000	-3.8
Aroclor-1260-4	7.085	6.988	7.188	480.470	500.000	-3.9
Aroclor-1260-5	7.327	7.230	7.430	476.280	500.000	-4.7
Decachlorobiphenyl	8.727	8.632	8.832	42.700	50.000	-14.6
Tetrachloro-m-xylene	3.685	3.588	3.788	52.620	50.000	5.2

CALIBRATION VERIFICATION SUMMARY

Contract: ROYF02

Lab Code: CHEM Case No.: Q1860 SAS No.: Q1860 SDG NO.: Q1860

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 04/10/2025 04/10/2025

Client Sample No.: CCAL11 Date Analyzed: 04/28/2025

Lab Sample No.: AR1660CCC500 Data File : PO110813.D Time Analyzed: 19:58

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.762	4.666	4.866	482.910	500.000	-3.4
Aroclor-1016-2	4.781	4.684	4.884	481.530	500.000	-3.7
Aroclor-1016-3	4.956	4.860	5.060	462.860	500.000	-7.4
Aroclor-1016-4	4.998	4.902	5.102	426.070	500.000	-14.8
Aroclor-1016-5	5.211	5.115	5.315	496.540	500.000	-0.7
Aroclor-1260-1	6.242	6.146	6.346	462.680	500.000	-7.5
Aroclor-1260-2	6.430	6.334	6.534	452.580	500.000	-9.5
Aroclor-1260-3	6.583	6.487	6.687	442.880	500.000	-11.4
Aroclor-1260-4	7.053	6.957	7.157	445.230	500.000	-11.0
Aroclor-1260-5	7.294	7.197	7.397	441.430	500.000	-11.7
Decachlorobiphenyl	8.678	8.584	8.784	42.100	50.000	-15.8
Tetrachloro-m-xylene	3.683	3.586	3.786	49.720	50.000	-0.6