

CALIBRATION VERIFICATION SUMMARY

Contract: ROYF02

Lab Code: CHEM Case No.: Q1860 SAS No.: Q1860 SDG NO.: Q1860

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 04/10/2025 04/10/2025

Client Sample No.: CCAL13 Date Analyzed: 04/29/2025

Lab Sample No.: AR1660CCC500 Data File : PO110840.D Time Analyzed: 06:57

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.776	4.679	4.879	535.490	500.000	7.1
Aroclor-1016-2	4.795	4.698	4.898	538.890	500.000	7.8
Aroclor-1016-3	4.851	4.755	4.955	521.500	500.000	4.3
Aroclor-1016-4	4.972	4.875	5.075	531.460	500.000	6.3
Aroclor-1016-5	5.228	5.132	5.332	519.550	500.000	3.9
Aroclor-1260-1	6.269	6.172	6.372	538.450	500.000	7.7
Aroclor-1260-2	6.458	6.360	6.560	520.390	500.000	4.1
Aroclor-1260-3	6.824	6.729	6.929	508.610	500.000	1.7
Aroclor-1260-4	7.085	6.988	7.188	497.730	500.000	-0.5
Aroclor-1260-5	7.327	7.230	7.430	489.760	500.000	-2.0
Decachlorobiphenyl	8.727	8.632	8.832	44.710	50.000	-10.6
Tetrachloro-m-xylene	3.686	3.588	3.788	53.930	50.000	7.9

CALIBRATION VERIFICATION SUMMARY

Contract: ROYF02

Lab Code: CHEM Case No.: Q1860 SAS No.: Q1860 SDG NO.: Q1860

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 04/10/2025 04/10/2025

Client Sample No.: CCAL13 Date Analyzed: 04/29/2025

Lab Sample No.: AR1660CCC500 Data File : PO110840.D Time Analyzed: 06:57

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.762	4.666	4.866	509.860	500.000	2.0
Aroclor-1016-2	4.781	4.684	4.884	514.080	500.000	2.8
Aroclor-1016-3	4.956	4.860	5.060	490.500	500.000	-1.9
Aroclor-1016-4	4.999	4.902	5.102	420.220	500.000	-16.0
Aroclor-1016-5	5.211	5.115	5.315	544.220	500.000	8.8
Aroclor-1260-1	6.242	6.146	6.346	496.360	500.000	-0.7
Aroclor-1260-2	6.430	6.334	6.534	486.020	500.000	-2.8
Aroclor-1260-3	6.583	6.487	6.687	471.250	500.000	-5.8
Aroclor-1260-4	7.053	6.957	7.157	470.770	500.000	-5.8
Aroclor-1260-5	7.294	7.197	7.397	480.410	500.000	-3.9
Decachlorobiphenyl	8.679	8.584	8.784	47.270	50.000	-5.5
Tetrachloro-m-xylene	3.683	3.586	3.786	51.790	50.000	3.6