

CALIBRATION VERIFICATION SUMMARY

Contract: ROYF02

Lab Code: CHEM Case No.: Q1860 SAS No.: Q1860 SDG NO.: Q1860

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 04/22/2025 04/22/2025

Client Sample No.: CCAL16 Date Analyzed: 04/24/2025

Lab Sample No.: AR1660CCC500 Data File : PP071423.D Time Analyzed: 09:22

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	5.662	5.571	5.771	521.590	500.000	4.3
Aroclor-1016-2	5.685	5.593	5.793	521.660	500.000	4.3
Aroclor-1016-3	5.746	5.654	5.854	525.790	500.000	5.2
Aroclor-1016-4	5.844	5.752	5.952	525.400	500.000	5.1
Aroclor-1016-5	6.136	6.045	6.245	524.180	500.000	4.8
Aroclor-1260-1	7.255	7.165	7.365	522.560	500.000	4.5
Aroclor-1260-2	7.509	7.418	7.618	527.350	500.000	5.5
Aroclor-1260-3	7.867	7.777	7.977	553.110	500.000	10.6
Aroclor-1260-4	8.091	8.002	8.202	540.030	500.000	8.0
Aroclor-1260-5	8.411	8.322	8.522	547.510	500.000	9.5
Decachlorobiphenyl	10.224	10.137	10.337	54.640	50.000	9.3
Tetrachloro-m-xylene	4.509	4.417	4.617	52.210	50.000	4.4

CALIBRATION VERIFICATION SUMMARY

Contract: ROYF02

Lab Code: CHEM Case No.: Q1860 SAS No.: Q1860 SDG NO.: Q1860

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 04/22/2025 04/22/2025

Client Sample No.: CCAL16 Date Analyzed: 04/24/2025

Lab Sample No.: AR1660CCC500 Data File : PP071423.D Time Analyzed: 09:22

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.896	4.798	4.998	501.210	500.000	0.2
Aroclor-1016-2	4.914	4.816	5.016	503.100	500.000	0.6
Aroclor-1016-3	5.091	4.993	5.193	511.240	500.000	2.2
Aroclor-1016-4	5.133	5.036	5.236	518.420	500.000	3.7
Aroclor-1016-5	5.347	5.250	5.450	528.410	500.000	5.7
Aroclor-1260-1	6.381	6.285	6.485	490.940	500.000	-1.8
Aroclor-1260-2	6.570	6.473	6.673	494.440	500.000	-1.1
Aroclor-1260-3	6.723	6.627	6.827	498.920	500.000	-0.2
Aroclor-1260-4	7.194	7.098	7.298	495.060	500.000	-1.0
Aroclor-1260-5	7.435	7.339	7.539	500.940	500.000	0.2
Decachlorobiphenyl	8.844	8.749	8.949	52.040	50.000	4.1
Tetrachloro-m-xylene	3.811	3.712	3.912	46.690	50.000	-6.6