

CALIBRATION VERIFICATION SUMMARY

Contract: ROYF02

Lab Code: CHEM Case No.: Q1860 SAS No.: Q1860 SDG NO.: Q1860

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 04/22/2025 04/22/2025

Client Sample No.: CCAL18 Date Analyzed: 04/24/2025

Lab Sample No.: AR1660CCC500 Data File : PP071450.D Time Analyzed: 22:12

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	5.667	5.571	5.771	525.270	500.000	5.1
Aroclor-1016-2	5.688	5.593	5.793	529.380	500.000	5.9
Aroclor-1016-3	5.750	5.654	5.854	526.390	500.000	5.3
Aroclor-1016-4	5.848	5.752	5.952	531.600	500.000	6.3
Aroclor-1016-5	6.141	6.045	6.245	533.350	500.000	6.7
Aroclor-1260-1	7.259	7.165	7.365	538.260	500.000	7.7
Aroclor-1260-2	7.513	7.418	7.618	518.160	500.000	3.6
Aroclor-1260-3	7.872	7.777	7.977	533.180	500.000	6.6
Aroclor-1260-4	8.095	8.002	8.202	514.460	500.000	2.9
Aroclor-1260-5	8.416	8.322	8.522	525.980	500.000	5.2
Decachlorobiphenyl	10.229	10.137	10.337	53.720	50.000	7.4
Tetrachloro-m-xylene	4.513	4.417	4.617	52.270	50.000	4.5

CALIBRATION VERIFICATION SUMMARY

Contract: ROYF02

Lab Code: CHEM Case No.: Q1860 SAS No.: Q1860 SDG NO.: Q1860

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 04/22/2025 04/22/2025

Client Sample No.: CCAL18 Date Analyzed: 04/24/2025

Lab Sample No.: AR1660CCC500 Data File : PP071450.D Time Analyzed: 22:12

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.896	4.798	4.998	516.820	500.000	3.4
Aroclor-1016-2	4.915	4.816	5.016	512.710	500.000	2.5
Aroclor-1016-3	5.091	4.993	5.193	521.130	500.000	4.2
Aroclor-1016-4	5.133	5.036	5.236	506.320	500.000	1.3
Aroclor-1016-5	5.348	5.250	5.450	557.360	500.000	11.5
Aroclor-1260-1	6.383	6.285	6.485	513.870	500.000	2.8
Aroclor-1260-2	6.571	6.473	6.673	506.840	500.000	1.4
Aroclor-1260-3	6.723	6.627	6.827	519.300	500.000	3.9
Aroclor-1260-4	7.195	7.098	7.298	530.720	500.000	6.1
Aroclor-1260-5	7.436	7.339	7.539	527.770	500.000	5.6
Decachlorobiphenyl	8.845	8.749	8.949	53.190	50.000	6.4
Tetrachloro-m-xylene	3.811	3.712	3.912	53.430	50.000	6.9