

CALIBRATION VERIFICATION SUMMARY

Contract: ROYF02

Lab Code: CHEM Case No.: Q1860 SAS No.: Q1860 SDG NO.: Q1860

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 04/22/2025 04/22/2025

Client Sample No.: CCAL19 Date Analyzed: 04/28/2025

Lab Sample No.: AR1660CCC500 Data File : PP071528.D Time Analyzed: 08:28

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.667	5.571	5.771	479.640	500.000	-4.1
Aroclor-1016-2	5.688	5.593	5.793	494.560	500.000	-1.1
Aroclor-1016-3	5.750	5.654	5.854	482.000	500.000	-3.6
Aroclor-1016-4	5.848	5.752	5.952	483.910	500.000	-3.2
Aroclor-1016-5	6.140	6.045	6.245	489.670	500.000	-2.1
Aroclor-1260-1	7.259	7.165	7.365	474.390	500.000	-5.1
Aroclor-1260-2	7.513	7.418	7.618	484.300	500.000	-3.1
Aroclor-1260-3	7.871	7.777	7.977	505.470	500.000	1.1
Aroclor-1260-4	8.095	8.002	8.202	504.790	500.000	1.0
Aroclor-1260-5	8.415	8.322	8.522	502.570	500.000	0.5
Decachlorobiphenyl	10.228	10.137	10.337	50.380	50.000	0.8
Tetrachloro-m-xylene	4.514	4.417	4.617	50.770	50.000	1.5

CALIBRATION VERIFICATION SUMMARY

Contract: ROYF02

Lab Code: CHEM Case No.: Q1860 SAS No.: Q1860 SDG NO.: Q1860

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 04/22/2025 04/22/2025

Client Sample No.: CCAL19 Date Analyzed: 04/28/2025

Lab Sample No.: AR1660CCC500 Data File : PP071528.D Time Analyzed: 08:28

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.893	4.798	4.998	503.970	500.000	0.8
Aroclor-1016-2	4.911	4.816	5.016	511.280	500.000	2.3
Aroclor-1016-3	5.089	4.993	5.193	516.340	500.000	3.3
Aroclor-1016-4	5.129	5.036	5.236	523.840	500.000	4.8
Aroclor-1016-5	5.344	5.250	5.450	520.790	500.000	4.2
Aroclor-1260-1	6.379	6.285	6.485	507.560	500.000	1.5
Aroclor-1260-2	6.567	6.473	6.673	512.510	500.000	2.5
Aroclor-1260-3	6.720	6.627	6.827	518.030	500.000	3.6
Aroclor-1260-4	7.191	7.098	7.298	513.530	500.000	2.7
Aroclor-1260-5	7.432	7.339	7.539	516.680	500.000	3.3
Decachlorobiphenyl	8.840	8.749	8.949	48.920	50.000	-2.2
Tetrachloro-m-xylene	3.809	3.712	3.912	48.890	50.000	-2.2