

CALIBRATION VERIFICATION SUMMARY

Contract: ROYF02

Lab Code: CHEM Case No.: Q1860 SAS No.: Q1860 SDG NO.: Q1860

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 04/22/2025 04/22/2025

Client Sample No.: CCAL20 Date Analyzed: 04/28/2025

Lab Sample No.: AR1660CCC500 Data File : PP071542.D Time Analyzed: 12:53

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	5.662	5.571	5.771	472.400	500.000	-5.5
Aroclor-1016-2	5.684	5.593	5.793	469.150	500.000	-6.2
Aroclor-1016-3	5.745	5.654	5.854	465.430	500.000	-6.9
Aroclor-1016-4	5.843	5.752	5.952	472.130	500.000	-5.6
Aroclor-1016-5	6.136	6.045	6.245	465.790	500.000	-6.8
Aroclor-1260-1	7.255	7.165	7.365	453.340	500.000	-9.3
Aroclor-1260-2	7.508	7.418	7.618	459.620	500.000	-8.1
Aroclor-1260-3	7.867	7.777	7.977	480.840	500.000	-3.8
Aroclor-1260-4	8.091	8.002	8.202	461.550	500.000	-7.7
Aroclor-1260-5	8.411	8.322	8.522	479.340	500.000	-4.1
Decachlorobiphenyl	10.224	10.137	10.337	49.080	50.000	-1.8
Tetrachloro-m-xylene	4.509	4.417	4.617	49.030	50.000	-1.9

CALIBRATION VERIFICATION SUMMARY

Contract: ROYF02

Lab Code: CHEM Case No.: Q1860 SAS No.: Q1860 SDG NO.: Q1860

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 04/22/2025 04/22/2025

Client Sample No.: CCAL20 Date Analyzed: 04/28/2025

Lab Sample No.: AR1660CCC500 Data File : PP071542.D Time Analyzed: 12:53

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.894	4.798	4.998	507.840	500.000	1.6
Aroclor-1016-2	4.913	4.816	5.016	509.030	500.000	1.8
Aroclor-1016-3	5.090	4.993	5.193	515.210	500.000	3.0
Aroclor-1016-4	5.132	5.036	5.236	516.570	500.000	3.3
Aroclor-1016-5	5.346	5.250	5.450	515.930	500.000	3.2
Aroclor-1260-1	6.380	6.285	6.485	483.210	500.000	-3.4
Aroclor-1260-2	6.568	6.473	6.673	479.260	500.000	-4.1
Aroclor-1260-3	6.722	6.627	6.827	482.010	500.000	-3.6
Aroclor-1260-4	7.192	7.098	7.298	484.170	500.000	-3.2
Aroclor-1260-5	7.434	7.339	7.539	491.680	500.000	-1.7
Decachlorobiphenyl	8.842	8.749	8.949	49.250	50.000	-1.5
Tetrachloro-m-xylene	3.810	3.712	3.912	51.970	50.000	3.9