



Report of Analysis

Client: Weston Solutions, Inc.

Date Collected: 04/21/25

Project: RFP 911

Date Received: 04/23/25

Client Sample ID: P001-SS032-01DL

SDG No.: Q1860

Lab Sample ID: Q1860-38DL

Matrix: SOIL

Analytical Method: SW8082A

V8082A % Solid: 87 Decanted:

Sample Wt/Vol: 30.02 Units: g

uL

Final Vol: 10000 uL

Extraction Type:

Soil Aliquot Vol:

PO110805.D

Test: PCB

GPC Factor: 1.0 PH:

10

Injection Volume:

Prep Method: SW3541B

File ID/Qc Batch: Dilution:

Prep Date Date Analyzed

Prep Batch ID

04/25/25 08:20 04/28/25 16:41 PB167737

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						_
12674-11-2	Aroclor-1016	45.4	UD	45.4	195	ug/kg
11104-28-2	Aroclor-1221	46.3	UD	46.3	195	ug/kg
11141-16-5	Aroclor-1232	42.7	UD	42.7	195	ug/kg
53469-21-9	Aroclor-1242	46.1	UD	46.1	195	ug/kg
12672-29-6	Aroclor-1248	1900	D	68.0	195	ug/kg
11097-69-1	Aroclor-1254	4000	ED	36.9	195	ug/kg
37324-23-5	Aroclor-1262	57.7	UD	57.7	195	ug/kg
11100-14-4	Aroclor-1268	41.4	UD	41.4	195	ug/kg
11096-82-5	Aroclor-1260	37.1	UD	37.1	195	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	17.8		32 - 144	89%	SPK: 20
2051-24-3	Decachlorobiphenyl	32.1		32 - 175	160%	SPK: 20

Comments:

U = Not Detected

LOO = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit