

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QC Batch ID # PO041025

Review By	yogesh	Review On	4/10/2025 11:27:35 AM
Supervise By	mohammad	Supervise On	4/14/2025 12:40:29 AM
SubDirectory	PO041025	HP Acquire Method	HP Processing Method PO041025

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,P P24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP 24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369
CCC	PP24332,PP24347,PP24352,PP24357
Internal Standard/PEM	
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	Sampleld	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PO110347.D	10 Apr 2025 08:59		YP/AJ	Ok
2	I.BLK	I.BLK	PO110348.D	10 Apr 2025 09:17		YP/AJ	Ok
3	AR1660ICC1000	AR1660ICC1000	PO110349.D	10 Apr 2025 09:36		YP/AJ	Ok
4	AR1660ICC750	AR1660ICC750	PO110350.D	10 Apr 2025 09:54		YP/AJ	Ok
5	AR1660ICC500	AR1660ICC500	PO110351.D	10 Apr 2025 10:13		YP/AJ	Ok
6	AR1660ICC250	AR1660ICC250	PO110352.D	10 Apr 2025 10:31		YP/AJ	Ok
7	AR1660ICC050	AR1660ICC050	PO110353.D	10 Apr 2025 10:49		YP/AJ	Ok
8	AR1221ICC500	AR1221ICC500	PO110354.D	10 Apr 2025 11:08		YP/AJ	Ok
9	AR1232ICC500	AR1232ICC500	PO110355.D	10 Apr 2025 11:26		YP/AJ	Ok
10	AR1242ICC1000	AR1242ICC1000	PO110356.D	10 Apr 2025 11:44		YP/AJ	Ok
11	AR1242ICC750	AR1242ICC750	PO110357.D	10 Apr 2025 12:03		YP/AJ	Ok
12	AR1242ICC500	AR1242ICC500	PO110358.D	10 Apr 2025 12:21		YP/AJ	Ok
13	AR1242ICC250	AR1242ICC250	PO110359.D	10 Apr 2025 12:39		YP/AJ	Ok
14	AR1242ICC050	AR1242ICC050	PO110360.D	10 Apr 2025 12:58		YP/AJ	Ok
15	AR1248ICC1000	AR1248ICC1000	PO110361.D	10 Apr 2025 13:16		YP/AJ	Ok
16	AR1248ICC750	AR1248ICC750	PO110362.D	10 Apr 2025 13:35		YP/AJ	Ok
17	AR1248ICC500	AR1248ICC500	PO110363.D	10 Apr 2025 13:53		YP/AJ	Ok
18	AR1248ICC250	AR1248ICC250	PO110364.D	10 Apr 2025 14:11		YP/AJ	Ok

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QC Batch ID # PO041025

Review By	yogesh	Review On	4/10/2025 11:27:35 AM
Supervise By	mohammad	Supervise On	4/14/2025 12:40:29 AM
SubDirectory	PO041025	HP Acquire Method	HP Processing Method PO041025
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,P P24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP 24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369		
CCC Internal Standard/PEM	PP24332,PP24347,PP24352,PP24357		
ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387		

19	AR1248ICC050	AR1248ICC050	PO110365.D	10 Apr 2025 14:30		YP/AJ	Ok
20	AR1254ICC1000	AR1254ICC1000	PO110366.D	10 Apr 2025 14:48		YP/AJ	Ok
21	AR1254ICC750	AR1254ICC750	PO110367.D	10 Apr 2025 15:06		YP/AJ	Ok
22	AR1254ICC500	AR1254ICC500	PO110368.D	10 Apr 2025 15:25		YP/AJ	Ok
23	AR1254ICC250	AR1254ICC250	PO110369.D	10 Apr 2025 15:43		YP/AJ	Ok
24	AR1254ICC050	AR1254ICC050	PO110370.D	10 Apr 2025 16:02		YP/AJ	Ok
25	AR1262ICC500	AR1262ICC500	PO110371.D	10 Apr 2025 16:20		YP/AJ	Ok
26	AR1268ICC1000	AR1268ICC1000	PO110372.D	10 Apr 2025 16:38		YP/AJ	Ok
27	AR1268ICC750	AR1268ICC750	PO110373.D	10 Apr 2025 16:57		YP/AJ	Ok
28	AR1268ICC500	AR1268ICC500	PO110374.D	10 Apr 2025 17:15		YP/AJ	Ok
29	AR1268ICC250	AR1268ICC250	PO110375.D	10 Apr 2025 17:33		YP/AJ	Ok
30	AR1268ICC050	AR1268ICC050	PO110376.D	10 Apr 2025 17:52		YP/AJ	Ok
31	PO041025ICV500	ICVPO041025	PO110377.D	10 Apr 2025 18:09		YP/AJ	Ok
32	AR1242ICV500	ICVPO041025AR1242	PO110378.D	10 Apr 2025 18:46		YP/AJ	Ok
33	AR1248ICV500	ICVPO041025AR1248	PO110379.D	10 Apr 2025 19:22		YP/AJ	Ok
34	AR1254ICV500	ICVPO041025AR1254	PO110380.D	10 Apr 2025 19:58		YP/AJ	Ok
35	AR1268ICV500	ICVPO041025AR1268	PO110381.D	10 Apr 2025 20:35		YP/AJ	Ok

M : Manual Integration

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QC Batch ID # PO042425

Review By	yogesh	Review On	4/24/2025 12:16:32 PM
Supervise By	mohammad	Supervise On	4/26/2025 2:18:54 AM
SubDirectory	PO042425	HP Acquire Method	HP Processing Method PO041025

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,PP24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369
CCC	PP24332,PP24347,PP24352,PP24357
Internal Standard/PEM	
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	Sampleld	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PO110672.D	24 Apr 2025 09:33		YP/AJ	Ok
2	AR1660CCC500	AR1660CCC500	PO110673.D	24 Apr 2025 10:09		YP/AJ	Ok
3	AR1242CCC500	AR1242CCC500	PO110674.D	24 Apr 2025 10:34		YP/AJ	Ok
4	AR1248CCC500	AR1248CCC500	PO110675.D	24 Apr 2025 10:53		YP/AJ	Ok
5	AR1254CCC500	AR1254CCC500	PO110676.D	24 Apr 2025 11:11		YP/AJ	Ok
6	I.BLK	I.BLK	PO110677.D	24 Apr 2025 11:29		YP/AJ	Ok
7	Q1860-05	P001-SS002-02	PO110678.D	24 Apr 2025 13:01	AR1254 Hit	YP/AJ	Ok,M
8	Q1860-06	P001-SS003-01	PO110679.D	24 Apr 2025 13:19	AR1254 + AR1260 Hit	YP/AJ	Ok
9	Q1860-07	P001-SS004-01	PO110680.D	24 Apr 2025 13:37	AR1254 + AR1260 Hit	YP/AJ	Ok
10	Q1860-12	P001-SS009-01	PO110681.D	24 Apr 2025 13:56	AR1248 + AR1254 Hit, Need 100X Dilution, DCB High in both column,	YP/AJ	Dilution
11	Q1860-13	P001-SS010-01	PO110682.D	24 Apr 2025 14:14	May possible carryover	YP/AJ	Not Ok
12	Q1860-14	P001-SS011-01	PO110683.D	24 Apr 2025 14:32	AR1248 + AR1254 Hit,	YP/AJ	Ok,M
13	Q1860-15	P001-SS012-01	PO110684.D	24 Apr 2025 14:51	AR1248 + AR1254 Hit, Need 25X Dilution, DCB High in both column	YP/AJ	Dilution
14	Q1860-16	P001-SS013-01	PO110685.D	24 Apr 2025 15:09	AR1248 + AR1254 Hit, Need 25X Dilution	YP/AJ	Dilution
15	Q1860-17	P001-SS014-01	PO110686.D	24 Apr 2025 15:27	AR1248 + AR1254 Hit,	YP/AJ	Ok,M
16	Q1860-18	P001-SS015-01	PO110687.D	24 Apr 2025 15:46	AR1248 + AR1254 Hit,	YP/AJ	Ok,M

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QC Batch ID # PO042425

Review By	yogesh	Review On	4/24/2025 12:16:32 PM
Supervise By	mohammad	Supervise On	4/26/2025 2:18:54 AM
SubDirectory	PO042425	HP Acquire Method	HP Processing Method PO041025

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,P P24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP 24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369
CCC Internal Standard/PEM	PP24332,PP24347,PP24352,PP24357
ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387

Run #	Sample Name	Std Name	File Name	Time	Notes	Operator	Status
17	AR1660CCC500	AR1660CCC500	PO110688.D	24 Apr 2025 17:22		YP/AJ	Ok
18	AR1242CCC500	AR1242CCC500	PO110689.D	24 Apr 2025 17:59		YP/AJ	Ok
19	AR1248CCC500	AR1248CCC500	PO110690.D	24 Apr 2025 18:17		YP/AJ	Ok
20	AR1254CCC500	AR1254CCC500	PO110691.D	24 Apr 2025 18:36		YP/AJ	Ok
21	I.BLK	I.BLK	PO110692.D	24 Apr 2025 18:54		YP/AJ	Ok
22	PB167719BL	PB167719BL	PO110693.D	24 Apr 2025 19:13	Contamination	YP/AJ	Not Ok
23	PB167719BS	PB167719BS	PO110694.D	24 Apr 2025 19:31		YP/AJ	Ok
24	Q1860-22	P001-SS017-01	PO110695.D	24 Apr 2025 19:49	AR1248 + AR1254 Hit,	YP/AJ	Not Ok
25	Q1860-23	P001-SS017-02	PO110696.D	24 Apr 2025 20:08	AR1248 + AR1254 Hit,	YP/AJ	Not Ok
26	Q1860-24	P001-SS018-01	PO110697.D	24 Apr 2025 20:26	AR1254 Hit,	YP/AJ	Not Ok
27	Q1860-25	P001-SS019-01	PO110698.D	24 Apr 2025 20:44	AR1254 + AR1260 Hit,	YP/AJ	Not Ok
28	Q1865-02	COMP-2	PO110699.D	24 Apr 2025 21:03		YP/AJ	Ok,M
29	Q1865-03	COMP-3	PO110700.D	24 Apr 2025 21:21		YP/AJ	Ok
30	Q1865-04	COMP-4	PO110701.D	24 Apr 2025 21:39		YP/AJ	Ok
31	Q1865-01	COMP-1	PO110702.D	24 Apr 2025 21:57		YP/AJ	Ok,M
32	AR1660CCC500	AR1660CCC500	PO110703.D	24 Apr 2025 23:35		YP/AJ	Ok
33	AR1242CCC500	AR1242CCC500	PO110704.D	25 Apr 2025 00:30		YP/AJ	Ok
34	AR1248CCC500	AR1248CCC500	PO110705.D	25 Apr 2025 00:48		YP/AJ	Ok
35	AR1254CCC500	AR1254CCC500	PO110706.D	25 Apr 2025 01:07		YP/AJ	Ok

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QC Batch ID # PO042425

Review By	yogesh	Review On	4/24/2025 12:16:32 PM		
Supervise By	mohammad	Supervise On	4/26/2025 2:18:54 AM		
SubDirectory	PO042425	HP Acquire Method	HP Processing Method	PO041025	
STD. NAME	STD REF.#				
Tune/Reschk					
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,P P24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP 24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369				
CCC	PP24332,PP24347,PP24352,PP24357				
Internal Standard/PEM					
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387				
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

36	I.BLK	I.BLK	PO110707.D	25 Apr 2025 01:25		YP/AJ	Ok
37	Q1860-13	P001-SS010-01	PO110708.D	25 Apr 2025 01:43	AR1254 + AR1260 Hit	YP/AJ	Ok,M
38	Q1860-12DL	P001-SS009-01DL	PO110709.D	25 Apr 2025 02:02	AR1248 + AR1254 Hit	YP/AJ	Ok,M
39	Q1860-15DL	P001-SS012-01DL	PO110710.D	25 Apr 2025 02:20	AR1248 + AR1254 Hit, DCB High in both column	YP/AJ	Ok,M
40	Q1860-15DL2	P001-SS012-01DL2	PO110711.D	25 Apr 2025 02:38	AR1248 + AR1254 Hit,	YP/AJ	Not Ok
41	Q1860-16DL	P001-SS013-01DL	PO110712.D	25 Apr 2025 02:57	AR1248 + AR1254 Hit, DCB High in 2nd column	YP/AJ	Ok,M
42	Q1860-16DL2	P001-SS013-01DL2	PO110713.D	25 Apr 2025 03:15	AR1248 + AR1254 Hit,	YP/AJ	Not Ok
43	AR1660CCC500	AR1660CCC500	PO110714.D	25 Apr 2025 04:38		YP/AJ	Ok,M
44	AR1242CCC500	AR1242CCC500	PO110715.D	25 Apr 2025 05:33		YP/AJ	Ok
45	AR1248CCC500	AR1248CCC500	PO110716.D	25 Apr 2025 05:51		YP/AJ	Ok
46	AR1254CCC500	AR1254CCC500	PO110717.D	25 Apr 2025 06:10		YP/AJ	Ok
47	I.BLK	I.BLK	PO110718.D	25 Apr 2025 06:28		YP/AJ	Ok

M : Manual Integration

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QC Batch ID # PO042525

Review By	yogesh	Review On	4/25/2025 12:46:37 PM
Supervise By	mohammad	Supervise On	5/2/2025 2:20:23 AM
SubDirectory	PO042525	HP Acquire Method	HP Processing Method PO041025

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,PP24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369
CCC	PP24332,PP24347,PP24352,PP24357
Internal Standard/PEM	
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PO110719.D	25 Apr 2025 09:05		YP/AJ	Ok
2	AR1660CCC500	AR1660CCC500	PO110720.D	25 Apr 2025 09:24		YP/AJ	Ok
3	AR1242CCC500	AR1242CCC500	PO110721.D	25 Apr 2025 10:19		YP/AJ	Ok
4	AR1248CCC500	AR1248CCC500	PO110722.D	25 Apr 2025 10:37		YP/AJ	Ok
5	AR1254CCC500	AR1254CCC500	PO110723.D	25 Apr 2025 10:56		YP/AJ	Ok
6	I.BLK	I.BLK	PO110724.D	25 Apr 2025 11:14		YP/AJ	Ok
7	PB167737BL	PB167737BL	PO110725.D	25 Apr 2025 11:54	reanalyzed with cleanup	YP/AJ	Not Ok
8	PB167737BS	PB167737BS	PO110726.D	25 Apr 2025 12:13		YP/AJ	Ok
9	Q1860-19	P001-SS016-01	PO110727.D	25 Apr 2025 12:31	AR1248 + AR1254 Hit, Need 4X Dilution	YP/AJ	Dilution
10	Q1860-20MS	P001-SS016-01MS	PO110728.D	25 Apr 2025 12:49	1016 & 1260 Recovery High	YP/AJ	Ok,M
11	Q1860-21MSD	P001-SS016-01MSD	PO110729.D	25 Apr 2025 13:08	1016 & 1260 Recovery High	YP/AJ	Ok,M
12	Q1860-26	P001-SS020-01	PO110730.D	25 Apr 2025 13:26	AR1254 Hit, Need Dilution	YP/AJ	Dilution
13	Q1860-27	P001-SS021-01	PO110731.D	25 Apr 2025 13:44	AR1254 Hit, Need Dilution, DCB High in both column	YP/AJ	Dilution
14	Q1860-28	P001-SS022-01	PO110732.D	25 Apr 2025 14:03	AR1254 Hit,	YP/AJ	Ok
15	Q1860-29	P001-SS023-01	PO110733.D	25 Apr 2025 14:21	AR1254 Hit, Need Dilution	YP/AJ	Dilution
16	Q1860-30	P001-SS024-01	PO110734.D	25 Apr 2025 14:39	AR1254 Hit,	YP/AJ	Ok,M
17	AR1660CCC500	AR1660CCC500	PO110735.D	25 Apr 2025 16:18		YP/AJ	Ok

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QC Batch ID # PO042525

Review By	yogesh	Review On	4/25/2025 12:46:37 PM
Supervise By	mohammad	Supervise On	5/2/2025 2:20:23 AM
SubDirectory	PO042525	HP Acquire Method	HP Processing Method PO041025
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,P P24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP 24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369		
CCC	PP24332,PP24347,PP24352,PP24357		
Internal Standard/PEM			
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

18	AR1242CCC500	AR1242CCC500	PO110736.D	25 Apr 2025 16:54		YP/AJ	Ok
19	AR1248CCC500	AR1248CCC500	PO110737.D	25 Apr 2025 17:13		YP/AJ	Ok
20	AR1254CCC500	AR1254CCC500	PO110738.D	25 Apr 2025 17:31		YP/AJ	Ok
21	I.BLK	I.BLK	PO110739.D	25 Apr 2025 17:49		YP/AJ	Ok
22	Q1860-31	P001-SS025-01	PO110740.D	25 Apr 2025 18:08	AR1254 + AR1260 Hit, Need Dilution, DCB high in both column;Need 5X	YP/AJ	Dilution
23	Q1860-32	P001-SS026-01	PO110741.D	25 Apr 2025 18:26	AR1254 Hit, Need Dilution, DCB high in both column	YP/AJ	Dilution
24	Q1860-33	P001-SS027-01	PO110742.D	25 Apr 2025 18:44	Cleanup Needed	YP/AJ	Not Ok
25	Q1860-34	P001-SS028-01	PO110743.D	25 Apr 2025 19:03	AR1248 + AR1254 + AR1260 Hit, Need 4X Dilution, DCBhigh in both column	YP/AJ	Dilution
26	Q1860-35	P001-SS029-01	PO110744.D	25 Apr 2025 19:21	TCMX high in both column, Check for cleanup	YP/AJ	Not Ok
27	Q1860-36	P001-SS030-01	PO110745.D	25 Apr 2025 19:39	AR1248 + AR1254 + AR1260 Hit, Need dilution	YP/AJ	Dilution
28	Q1860-37	P001-SS031-01	PO110746.D	25 Apr 2025 19:58	AR1254 Hit, Need dilution, DCB High in 2nd column	YP/AJ	Dilution
29	Q1860-38	P001-SS032-01	PO110747.D	25 Apr 2025 20:16	AR1248 + AR1254 Hit, Need Dilution	YP/AJ	Dilution
30	Q1860-39	P001-SS033-01	PO110748.D	25 Apr 2025 20:34	AR1248 + AR1254 Hit, Need Dilution	YP/AJ	Dilution
31	Q1860-40	P001-SS034-01	PO110749.D	25 Apr 2025 20:53	AR1248 + AR1254 Hit, Need Dilution	YP/AJ	Dilution
32	AR1660CCC500	AR1660CCC500	PO110750.D	25 Apr 2025 23:07		YP/AJ	Ok,M
33	AR1242CCC500	AR1242CCC500	PO110751.D	25 Apr 2025 23:25		YP/AJ	Ok

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QC Batch ID # PO042525

Review By	yogesh	Review On	4/25/2025 12:46:37 PM
Supervise By	mohammad	Supervise On	5/2/2025 2:20:23 AM
SubDirectory	PO042525	HP Acquire Method	HP Processing Method PO041025
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,P P24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP 24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369		
CCC	PP24332,PP24347,PP24352,PP24357		
Internal Standard/PEM			
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

34	AR1248CCC500	AR1248CCC500	PO110752.D	25 Apr 2025 23:44		YP/AJ	Ok
35	AR1254CCC500	AR1254CCC500	PO110753.D	26 Apr 2025 00:02		YP/AJ	Ok
36	I.BLK	I.BLK	PO110754.D	26 Apr 2025 00:20		YP/AJ	Ok
37	PB167740BL	PB167740BL	PO110755.D	26 Apr 2025 00:39		YP/AJ	Ok
38	PB167740BS	PB167740BS	PO110756.D	26 Apr 2025 00:57		YP/AJ	Ok
39	PB167740BSD	PB167740BSD	PO110757.D	26 Apr 2025 01:16		YP/AJ	Ok
40	Q1860-41	P001-SS035-01	PO110758.D	26 Apr 2025 01:34	AR1248 + AR1254 Hit, Need Dilution, DCB Hgih in both column	YP/AJ	Dilution
41	Q1860-42	P001-SS036-01	PO110759.D	26 Apr 2025 01:52	Cleanup Needed	YP/AJ	Not Ok
42	Q1855-06	60452	PO110760.D	26 Apr 2025 02:11		YP/AJ	Ok,M
43	Q1861-02	0411-A-0411-B-COMP	PO110761.D	26 Apr 2025 02:29		YP/AJ	Ok
44	Q1861-05	0416A-0416B-0416C-C	PO110762.D	26 Apr 2025 02:47		YP/AJ	Ok
45	Q1875-05	AUD-25-0051	PO110763.D	26 Apr 2025 03:06		YP/AJ	Ok
46	Q1875-06	AUD-25-0052	PO110764.D	26 Apr 2025 03:24		YP/AJ	Ok,M
47	AR1660CCC500	AR1660CCC500	PO110765.D	26 Apr 2025 05:02		YP/AJ	Ok
48	AR1242CCC500	AR1242CCC500	PO110766.D	26 Apr 2025 05:57		YP/AJ	Ok
49	AR1248CCC500	AR1248CCC500	PO110767.D	26 Apr 2025 06:15		YP/AJ	Ok
50	AR1254CCC500	AR1254CCC500	PO110768.D	26 Apr 2025 06:34		YP/AJ	Ok
51	I.BLK	I.BLK	PO110769.D	26 Apr 2025 06:52		YP/AJ	Ok

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QC Batch ID # PO042525

Review By	yogesh	Review On	4/25/2025 12:46:37 PM
Supervise By	mohammad	Supervise On	5/2/2025 2:20:23 AM
SubDirectory	PO042525	HP Acquire Method	HP Processing Method PO041025
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,P P24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP 24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369		
CCC Internal Standard/PEM	PP24332,PP24347,PP24352,PP24357		
ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387		

52	PB167719BL	PB167719BL	PO110770.D	26 Apr 2025 07:28	CCC Missing	YP/AJ	Not Ok
53	Q1875-07	AUD-25-0055	PO110771.D	26 Apr 2025 07:45	CCC Missing	YP/AJ	Not Ok
54	Q1875-08	AUD-25-0056	PO110772.D	26 Apr 2025 08:03	CCC Missing	YP/AJ	Not Ok
55	Q1861-01	50622	PO110773.D	26 Apr 2025 08:22	DCB low in 1st col & TCMX High in 1 col, CCC Missing	YP/AJ	Not Ok
56	Q1860-22	P001-SS017-01	PO110774.D	26 Apr 2025 08:40	AR1248 + AR1254 Hit, Need 20X Dilution, CCC Missing	YP/AJ	Not Ok
57	Q1860-23	P001-SS017-02	PO110775.D	26 Apr 2025 08:58	AR1248 + AR1254 Hit, Need 20X Dilution, CCC Missing	YP/AJ	Not Ok
58	Q1860-24	P001-SS018-01	PO110776.D	26 Apr 2025 09:17	AR1254 + AR1260 Hit, CCC Missing	YP/AJ	Not Ok
59	Q1860-25	P001-SS019-01	PO110777.D	26 Apr 2025 09:35	AR1254 + AR1260 Hit, Need 2X Dilution, CCC Missing	YP/AJ	Not Ok
60	Q1874-05	HT3727	PO110778.D	26 Apr 2025 09:53	Check for cleanup, CCC Missing	YP/AJ	Not Ok
61	Q1875-01	AUD-25-0053	PO110779.D	26 Apr 2025 10:12	CCC Missing	YP/AJ	Not Ok
62	AR1660CCC500	AR1660CCC500	PO110780.D	26 Apr 2025 11:50		YP/AJ	Ok
63	AR1242CCC500	AR1242CCC500	PO110781.D	26 Apr 2025 12:45		YP/AJ	Ok,M

M : Manual Integration

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QC Batch ID # PO042825

Review By	yogesh	Review On	4/29/2025 11:07:05 AM
Supervise By	mohammad	Supervise On	5/2/2025 2:30:39 AM
SubDirectory	PO042825	HP Acquire Method	HP Processing Method PO041025

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,P P24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP 24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369
CCC	PP24332,PP24347,PP24352,PP24357
Internal Standard/PEM	
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	Sampleld	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PO110782.D	28 Apr 2025 08:31		YP/AJ	Ok
2	AR1660CCC500	AR1660CCC500	PO110783.D	28 Apr 2025 08:50		YP/AJ	Ok
3	AR1242CCC500	AR1242CCC500	PO110784.D	28 Apr 2025 09:08		YP/AJ	Ok
4	AR1248CCC500	AR1248CCC500	PO110785.D	28 Apr 2025 09:44		YP/AJ	Ok
5	AR1254CCC500	AR1254CCC500	PO110786.D	28 Apr 2025 10:02		YP/AJ	Ok
6	I.BLK	I.BLK	PO110787.D	28 Apr 2025 10:21		YP/AJ	Ok
7	Q1860-19DL	P001-SS016-01DL	PO110788.D	28 Apr 2025 10:39	AR1248 + AR1254 Hit	YP/AJ	Ok,M
8	Q1860-26DL	P001-SS020-01DL	PO110789.D	28 Apr 2025 10:57	AR1254 Hit	YP/AJ	Ok
9	Q1860-27DL	P001-SS021-01DL	PO110790.D	28 Apr 2025 11:16	AR1254 Hit, DCB High in 2nd column	YP/AJ	Ok,M
10	Q1860-29DL	P001-SS023-01DL	PO110791.D	28 Apr 2025 11:34	AR1254 Hit	YP/AJ	Ok
11	Q1860-31DL	P001-SS025-01DL	PO110792.D	28 Apr 2025 11:53	AR1254 + AR1260 Hit, DCB high in both column	YP/AJ	Ok
12	Q1860-32DL	P001-SS026-01DL	PO110793.D	28 Apr 2025 12:11	AR1254 Hit, Need Dilution, DCB high in both column	YP/AJ	Not Ok
13	Q1860-32DL	P001-SS026-01DL	PO110794.D	28 Apr 2025 12:29	AR1254 hits, DCB high in both column	YP/AJ	Ok,M
14	Q1860-34DL	P001-SS028-01DL	PO110795.D	28 Apr 2025 12:48	AR1248 + AR1254 + AR1260 Hit, Need 2X Dilution, DCB high in both column	YP/AJ	Not Ok
15	Q1860-36DL	P001-SS030-01DL	PO110796.D	28 Apr 2025 13:06	AR1248 + AR1254 + AR1260 Hit, Need Dilution, DCB High in 2nd column	YP/AJ	Not Ok
16	Q1860-36DL	P001-SS030-01DL	PO110797.D	28 Apr 2025 13:24	AR1248 + AR1254 + AR1260 Hit	YP/AJ	Ok,M

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QCBatch ID # PO042825

Review By	yogesh	Review On	4/29/2025 11:07:05 AM
Supervise By	mohammad	Supervise On	5/2/2025 2:30:39 AM
SubDirectory	PO042825	HP Acquire Method	HP Processing Method PO041025
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,P P24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP 24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369		
CCC	PP24332,PP24347,PP24352,PP24357		
Internal Standard/PEM			
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

17	AR1660CCC500	AR1660CCC500	PO110798.D	28 Apr 2025 14:33		YP/AJ	Ok
18	AR1242CCC500	AR1242CCC500	PO110799.D	28 Apr 2025 14:52		YP/AJ	Ok
19	AR1248CCC500	AR1248CCC500	PO110800.D	28 Apr 2025 15:10		YP/AJ	Ok
20	AR1254CCC500	AR1254CCC500	PO110801.D	28 Apr 2025 15:28		YP/AJ	Ok
21	I.BLK	I.BLK	PO110802.D	28 Apr 2025 15:47		YP/AJ	Ok
22	Q1893-01	UGGP-1	PO110803.D	28 Apr 2025 16:05		YP/AJ	Ok,M
23	Q1893-02	UGGP-2	PO110804.D	28 Apr 2025 16:23	AR1248 + AR1254 Hit,	YP/AJ	Ok
24	Q1860-38DL	P001-SS032-01DL	PO110805.D	28 Apr 2025 16:41	AR1248 + AR1254 Hit, Need Further Dilution	YP/AJ	Dilution
25	Q1860-38DL2	P001-SS032-01DL2	PO110806.D	28 Apr 2025 16:59	AR1248 + AR1254 Hit,	YP/AJ	Ok
26	Q1860-39DL	P001-SS033-01DL	PO110807.D	28 Apr 2025 17:17	AR1248 + AR1254 Hit, DCB High in both col	YP/AJ	Ok,M
27	Q1860-40DL	P001-SS034-01DL	PO110808.D	28 Apr 2025 17:36	AR1248 + AR1254 Hit, DCB High in both col	YP/AJ	Ok,M
28	Q1860-41DL	P001-SS035-01DL	PO110809.D	28 Apr 2025 17:53	AR1248 + AR1254 Hit, Need Further Dilution, DCB Hgih in both column	YP/AJ	Dilution
29	Q1860-41DL2	P001-SS035-01DL2	PO110810.D	28 Apr 2025 18:12	AR1248 + AR1254 Hit,	YP/AJ	Ok,M
30	Q1875-07	AUD-25-0055	PO110811.D	28 Apr 2025 18:30		YP/AJ	Ok
31	Q1861-01	50622	PO110812.D	28 Apr 2025 18:48		YP/AJ	Ok
32	AR1660CCC500	AR1660CCC500	PO110813.D	28 Apr 2025 19:58		YP/AJ	Ok
33	AR1242CCC500	AR1242CCC500	PO110814.D	28 Apr 2025 20:34		YP/AJ	Ok
34	AR1248CCC500	AR1248CCC500	PO110815.D	28 Apr 2025 20:53		YP/AJ	Ok

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QC Batch ID # PO042825

Review By	yogesh	Review On	4/29/2025 11:07:05 AM
Supervise By	mohammad	Supervise On	5/2/2025 2:30:39 AM
SubDirectory	PO042825	HP Acquire Method	HP Processing Method PO041025
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,P P24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP 24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369		
CCC	PP24332,PP24347,PP24352,PP24357		
Internal Standard/PEM			
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

35	AR1254CCC500	AR1254CCC500	PO110816.D	28 Apr 2025 21:11		YP/AJ	Ok
36	I.BLK	I.BLK	PO110817.D	28 Apr 2025 21:29		YP/AJ	Ok
37	PB167750BL	PB167750BL	PO110818.D	28 Apr 2025 21:48		YP/AJ	Ok
38	Q1875-01	AUD-25-0053	PO110819.D	28 Apr 2025 22:06		YP/AJ	Ok
39	Q1875-08	AUD-25-0056	PO110820.D	28 Apr 2025 22:24		YP/AJ	Ok
40	Q1860-22	P001-SS017-01	PO110821.D	28 Apr 2025 22:42	AR1248 + AR1254 Hit, Need 5X Dilution	YP/AJ	Dilution
41	Q1860-23	P001-SS017-02	PO110822.D	28 Apr 2025 23:00	AR1248 + AR1254 Hit, Need 5X Dilution	YP/AJ	Dilution
42	Q1860-24	P001-SS018-01	PO110823.D	28 Apr 2025 23:18	AR1254 Hit	YP/AJ	Ok,M
43	Q1860-25	P001-SS019-01	PO110824.D	28 Apr 2025 23:37	AR1254 + AR1260 Hit, Need 5X Dilution	YP/AJ	Dilution
44	Q1874-05	HT3727	PO110825.D	28 Apr 2025 23:54		YP/AJ	Ok
45	Q1860-22DL	P001-SS017-01DL	PO110826.D	29 Apr 2025 00:12	AR1248 + AR1254 Hit,	YP/AJ	Ok
46	Q1860-23DL	P001-SS017-02DL	PO110827.D	29 Apr 2025 00:31	AR1248 + AR1254 Hit,	YP/AJ	Ok,M
47	AR1660CCC500	AR1660CCC500	PO110828.D	29 Apr 2025 01:54		YP/AJ	Ok,M
48	AR1242CCC500	AR1242CCC500	PO110829.D	29 Apr 2025 02:31		YP/AJ	Ok
49	AR1248CCC500	AR1248CCC500	PO110830.D	29 Apr 2025 02:49		YP/AJ	Ok
50	AR1254CCC500	AR1254CCC500	PO110831.D	29 Apr 2025 03:07		YP/AJ	Ok
51	I.BLK	I.BLK	PO110832.D	29 Apr 2025 03:26		YP/AJ	Ok
52	Q1896-01	295-BERGEN-RO	PO110833.D	29 Apr 2025 03:44	AR1254 Hit	YP/AJ	Ok

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QCBatch ID # PO042825

Review By	yogesh	Review On	4/29/2025 11:07:05 AM		
Supervise By	mohammad	Supervise On	5/2/2025 2:30:39 AM		
SubDirectory	PO042825	HP Acquire Method	HP Processing Method	PO041025	
STD. NAME	STD REF.#				
Tune/Reschk					
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,P P24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP 24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369				
CCC	PP24332,PP24347,PP24352,PP24357				
Internal Standard/PEM					
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387				
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

53	Q1900-01	WC-1	PO110834.D	29 Apr 2025 04:02		YP/AJ	Ok,M
54	Q1900-05	WC-2	PO110835.D	29 Apr 2025 04:21		YP/AJ	Ok
55	Q1900-09	WC-3	PO110836.D	29 Apr 2025 04:39		YP/AJ	Ok
56	Q1860-25DL	P001-SS019-01DL	PO110837.D	29 Apr 2025 04:57	AR1254 + AR1260 Hit,	YP/AJ	Ok,M
57	Q1860-37DL	P001-SS031-01DL	PO110838.D	29 Apr 2025 05:16	AR1254 Hit, Need Further dilution, DCB High in both column	YP/AJ	Dilution
58	Q1860-37DL2	P001-SS031-01DL2	PO110839.D	29 Apr 2025 05:34	AR1254 Hit	YP/AJ	Ok
59	AR1660CCC500	AR1660CCC500	PO110840.D	29 Apr 2025 06:57		YP/AJ	Ok,M
60	AR1242CCC500	AR1242CCC500	PO110841.D	29 Apr 2025 07:34		YP/AJ	Ok
61	AR1248CCC500	AR1248CCC500	PO110842.D	29 Apr 2025 07:53		YP/AJ	Ok
62	AR1254CCC500	AR1254CCC500	PO110843.D	29 Apr 2025 08:11		YP/AJ	Ok
63	I.BLK	I.BLK	PO110844.D	29 Apr 2025 08:29		YP/AJ	Ok

M : Manual Integration

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QC Batch ID # PO042925

Review By	yogesh	Review On	4/29/2025 11:07:42 AM
Supervise By	mohammad	Supervise On	5/2/2025 2:51:07 AM
SubDirectory	PO042925	HP Acquire Method	HP Processing Method PO041025

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,P P24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP 24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369
CCC	PP24332,PP24347,PP24352,PP24357
Internal Standard/PEM	
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	Sampleld	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PO110845.D	29 Apr 2025 09:06		YP/AJ	Ok
2	AR1660CCC500	AR1660CCC500	PO110846.D	29 Apr 2025 09:24		YP/AJ	Ok
3	AR1242CCC500	AR1242CCC500	PO110847.D	29 Apr 2025 09:45		YP/AJ	Ok
4	AR1248CCC500	AR1248CCC500	PO110848.D	29 Apr 2025 10:03		YP/AJ	Ok
5	AR1254CCC500	AR1254CCC500	PO110849.D	29 Apr 2025 10:21		YP/AJ	Ok
6	I.BLK	I.BLK	PO110850.D	29 Apr 2025 10:40		YP/AJ	Ok
7	PB167765BL	PB167765BL	PO110851.D	29 Apr 2025 10:58		YP/AJ	Not Ok
8	Q1889-01	COMP-1	PO110852.D	29 Apr 2025 11:16		YP/AJ	Ok
9	Q1889-01MS	COMP-1MS	PO110853.D	29 Apr 2025 11:35		YP/AJ	Ok
10	Q1889-01MSD	COMP-1MSD	PO110854.D	29 Apr 2025 11:53		YP/AJ	Ok
11	Q1860-34DL	P001-SS028-01DL	PO110855.D	29 Apr 2025 12:11	AR1248 + AR1254 + AR1260 Hit, DCB High in both column	YP/AJ	Ok,M
12	Q1904-01	VNJ-210	PO110856.D	29 Apr 2025 13:36		YP/AJ	Not Ok
13	Q1902-02	343	PO110857.D	29 Apr 2025 13:54		YP/AJ	Ok,M
14	Q1903-01	COMP-4	PO110858.D	29 Apr 2025 14:12		YP/AJ	Ok
15	Q1903-02	COMP-5	PO110859.D	29 Apr 2025 14:31		YP/AJ	Ok
16	Q1903-03	COMP-6	PO110860.D	29 Apr 2025 14:49		YP/AJ	Ok
17	AR1660CCC500	AR1660CCC500	PO110861.D	29 Apr 2025 15:58		YP/AJ	Ok

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QC Batch ID # PO042925

Review By	yogesh	Review On	4/29/2025 11:07:42 AM
Supervise By	mohammad	Supervise On	5/2/2025 2:51:07 AM
SubDirectory	PO042925	HP Acquire Method	HP Processing Method PO041025
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,P P24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP 24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369		
CCC	PP24332,PP24347,PP24352,PP24357		
Internal Standard/PEM			
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

18	AR1242CCC500	AR1242CCC500	PO110862.D	29 Apr 2025 16:35		YP/AJ	Ok
19	AR1248CCC500	AR1248CCC500	PO110863.D	29 Apr 2025 16:53		YP/AJ	Ok
20	AR1254CCC500	AR1254CCC500	PO110864.D	29 Apr 2025 17:11		YP/AJ	Ok
21	I.BLK	I.BLK	PO110865.D	29 Apr 2025 17:29		YP/AJ	Ok
22	Q1904-01	VNJ-210	PO110866.D	29 Apr 2025 17:48		YP/AJ	Ok
23	Q1904-01MS	VNJ-210MS	PO110867.D	29 Apr 2025 18:05		YP/AJ	Ok
24	Q1904-01MSD	VNJ-210MSD	PO110868.D	29 Apr 2025 18:23		YP/AJ	Ok
25	Q1906-01	WC-4	PO110869.D	29 Apr 2025 18:42		YP/AJ	Ok
26	Q1906-05	WC-5	PO110870.D	29 Apr 2025 19:00		YP/AJ	Ok
27	Q1906-09	WC-6	PO110871.D	29 Apr 2025 19:19		YP/AJ	Ok,M
28	Q1906-13	WC-7	PO110872.D	29 Apr 2025 19:37		YP/AJ	Ok
29	Q1897-01	40125	PO110873.D	29 Apr 2025 19:55	AR1242 Hit	YP/AJ	Ok,M
30	Q1901-06	FB04262025	PO110874.D	29 Apr 2025 20:14		YP/AJ	Ok
31	Q1897-02	41425	PO110875.D	29 Apr 2025 20:32		YP/AJ	Ok,M
32	AR1660CCC500	AR1660CCC500	PO110876.D	29 Apr 2025 22:00		YP/AJ	Ok
33	AR1242CCC500	AR1242CCC500	PO110877.D	29 Apr 2025 22:36		YP/AJ	Ok
34	AR1248CCC500	AR1248CCC500	PO110878.D	29 Apr 2025 22:55		YP/AJ	Ok
35	AR1254CCC500	AR1254CCC500	PO110879.D	29 Apr 2025 23:13		YP/AJ	Ok
36	I.BLK	I.BLK	PO110880.D	29 Apr 2025 23:31		YP/AJ	Ok

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QC Batch ID # PO042925

Review By	yogesh	Review On	4/29/2025 11:07:42 AM
Supervise By	mohammad	Supervise On	5/2/2025 2:51:07 AM
SubDirectory	PO042925	HP Acquire Method	HP Processing Method PO041025
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,P P24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP 24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369		
CCC	PP24332,PP24347,PP24352,PP24357		
Internal Standard/PEM			
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

37	PB167790BL	PB167790BL	PO110881.D	29 Apr 2025 23:50		YP/AJ	Ok
38	PB167790BS	PB167790BS	PO110882.D	30 Apr 2025 00:08		YP/AJ	Ok
39	Q1909-01	KMH358F-1-1	PO110883.D	30 Apr 2025 00:26	AR1254 Hit	YP/AJ	Ok,M
40	Q1909-02	KMH358F-1-2	PO110884.D	30 Apr 2025 00:45	AR1254 Hit	YP/AJ	Ok,M
41	AR1660CCC500	AR1660CCC500	PO110885.D	30 Apr 2025 02:11		YP/AJ	Ok
42	AR1242CCC500	AR1242CCC500	PO110886.D	30 Apr 2025 02:48		YP/AJ	Ok
43	AR1248CCC500	AR1248CCC500	PO110887.D	30 Apr 2025 03:06		YP/AJ	Ok
44	AR1254CCC500	AR1254CCC500	PO110888.D	30 Apr 2025 03:25		YP/AJ	Ok
45	I.BLK	I.BLK	PO110889.D	30 Apr 2025 03:43		YP/AJ	Ok

M : Manual Integration

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP042225

Review By	yogesh	Review On	4/22/2025 1:34:56 PM
Supervise By	mohammad	Supervise On	4/24/2025 6:33:42 AM
SubDirectory	PP042225	HP Acquire Method	HP Processing Method PP042225

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,PP24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369
CCC	PP24332,PP24347,PP24352,PP24357
Internal Standard/PEM	
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	Sampleld	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PP071387.D	22 Apr 2025 09:56		YPIAJ	Ok
2	I.BLK	I.BLK	PP071388.D	22 Apr 2025 10:13		YPIAJ	Ok
3	AR1660ICC1000	AR1660ICC1000	PP071389.D	22 Apr 2025 10:29		YPIAJ	Ok
4	AR1660ICC750	AR1660ICC750	PP071390.D	22 Apr 2025 10:45		YPIAJ	Ok
5	AR1660ICC500	AR1660ICC500	PP071391.D	22 Apr 2025 11:02		YPIAJ	Ok
6	AR1660ICC250	AR1660ICC250	PP071392.D	22 Apr 2025 11:18		YPIAJ	Ok
7	AR1660ICC050	AR1660ICC050	PP071393.D	22 Apr 2025 11:34		YPIAJ	Ok,M
8	AR1221ICC500	AR1221ICC500	PP071394.D	22 Apr 2025 11:51		YPIAJ	Ok
9	AR1232ICC500	AR1232ICC500	PP071395.D	22 Apr 2025 12:07		YPIAJ	Ok
10	AR1242ICC1000	AR1242ICC1000	PP071396.D	22 Apr 2025 12:23		YPIAJ	Ok
11	AR1242ICC750	AR1242ICC750	PP071397.D	22 Apr 2025 12:39		YPIAJ	Ok
12	AR1242ICC500	AR1242ICC500	PP071398.D	22 Apr 2025 12:56		YPIAJ	Ok
13	AR1242ICC250	AR1242ICC250	PP071399.D	22 Apr 2025 13:12		YPIAJ	Ok
14	AR1242ICC050	AR1242ICC050	PP071400.D	22 Apr 2025 13:28		YPIAJ	Ok
15	AR1248ICC1000	AR1248ICC1000	PP071401.D	22 Apr 2025 13:45		YPIAJ	Ok
16	AR1248ICC750	AR1248ICC750	PP071402.D	22 Apr 2025 14:01		YPIAJ	Ok
17	AR1248ICC500	AR1248ICC500	PP071403.D	22 Apr 2025 14:17		YPIAJ	Ok
18	AR1248ICC250	AR1248ICC250	PP071404.D	22 Apr 2025 14:33		YPIAJ	Ok

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP042225

Review By	yogesh	Review On	4/22/2025 1:34:56 PM
Supervise By	mohammad	Supervise On	4/24/2025 6:33:42 AM
SubDirectory	PP042225	HP Acquire Method	HP Processing Method PP042225
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,P P24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP 24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369		
CCC	PP24332,PP24347,PP24352,PP24357		
Internal Standard/PEM			
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

19	AR1248ICC050	AR1248ICC050	PP071405.D	22 Apr 2025 14:50		YPIAJ	Ok
20	AR1254ICC1000	AR1254ICC1000	PP071406.D	22 Apr 2025 15:06		YPIAJ	Ok
21	AR1254ICC750	AR1254ICC750	PP071407.D	22 Apr 2025 15:22		YPIAJ	Ok,M
22	AR1254ICC500	AR1254ICC500	PP071408.D	22 Apr 2025 15:38		YPIAJ	Ok
23	AR1254ICC250	AR1254ICC250	PP071409.D	22 Apr 2025 15:55		YPIAJ	Ok
24	AR1254ICC050	AR1254ICC050	PP071410.D	22 Apr 2025 16:11		YPIAJ	Ok,M
25	AR1262ICC500	AR1262ICC500	PP071411.D	22 Apr 2025 16:27		YPIAJ	Ok
26	AR1268ICC1000	AR1268ICC1000	PP071412.D	22 Apr 2025 16:44		YPIAJ	Ok,M
27	AR1268ICC750	AR1268ICC750	PP071413.D	22 Apr 2025 17:00		YPIAJ	Ok,M
28	AR1268ICC500	AR1268ICC500	PP071414.D	22 Apr 2025 17:16		YPIAJ	Ok,M
29	AR1268ICC250	AR1268ICC250	PP071415.D	22 Apr 2025 17:33		YPIAJ	Ok,M
30	AR1268ICC050	AR1268ICC050	PP071416.D	22 Apr 2025 17:49		YPIAJ	Ok,M
31	PP042225ICV500	ICVPP042225	PP071417.D	22 Apr 2025 18:05		YPIAJ	Ok
32	AR1242ICV500	ICVPP042225AR1242	PP071418.D	22 Apr 2025 18:38		YPIAJ	Ok
33	AR1248ICV500	ICVPP042225AR1248	PP071419.D	22 Apr 2025 19:27		YPIAJ	Ok
34	AR1254ICV500	ICVPP042225AR1254	PP071420.D	22 Apr 2025 19:43		YPIAJ	Ok
35	AR1268ICV500	ICVPP042225AR1268	PP071421.D	22 Apr 2025 20:16		YPIAJ	Ok,M

M : Manual Integration

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP042425

Review By	yogesh	Review On	4/24/2025 12:15:27 PM
Supervise By	mohammad	Supervise On	4/26/2025 2:18:47 AM
SubDirectory	PP042425	HP Acquire Method	HP Processing Method PP042225

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,PP24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369
CCC	PP24332,PP24347,PP24352,PP24357
Internal Standard/PEM	
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PP071422.D	24 Apr 2025 09:06		YPIAJ	Ok
2	AR1660CCC500	AR1660CCC500	PP071423.D	24 Apr 2025 09:22		YPIAJ	Ok
3	AR1242CCC500	AR1242CCC500	PP071424.D	24 Apr 2025 09:38		YPIAJ	Ok
4	AR1248CCC500	AR1248CCC500	PP071425.D	24 Apr 2025 09:55		YPIAJ	Ok
5	AR1254CCC500	AR1254CCC500	PP071426.D	24 Apr 2025 10:11		YPIAJ	Ok,M
6	I.BLK	I.BLK	PP071427.D	24 Apr 2025 10:27		YPIAJ	Ok
7	PB167721BL	PB167721BL	PP071428.D	24 Apr 2025 12:58		YPIAJ	Ok
8	PB167721BS	PB167721BS	PP071429.D	24 Apr 2025 13:14		YPIAJ	Ok
9	Q1860-01	P001-SS001-01	PP071430.D	24 Apr 2025 13:30	AR1260 Hit	YPIAJ	Ok,M
10	Q1860-02MS	P001-SS001-01MS	PP071431.D	24 Apr 2025 13:47		YPIAJ	Ok,M
11	Q1860-03MSD	P001-SS001-01MSD	PP071432.D	24 Apr 2025 14:03		YPIAJ	Ok,M
12	Q1860-04	P001-SS002-01	PP071433.D	24 Apr 2025 14:19	AR1254 Hit	YPIAJ	Ok,M
13	Q1860-08	P001-SS005-01	PP071434.D	24 Apr 2025 14:35	AR1260 Hit	YPIAJ	Ok,M
14	Q1860-09	P001-SS006-01	PP071435.D	24 Apr 2025 14:52	AR1260 Hit, Need 2X Dilution	YPIAJ	Not Ok
15	Q1860-10	P001-SS007-01	PP071436.D	24 Apr 2025 15:08	AR1248 + AR1254 Hit, Need 25X Dilution, DCB High in 2nd column	YPIAJ	Dilution
16	Q1860-11	P001-SS008-01	PP071437.D	24 Apr 2025 15:25	AR1248 + AR1254 Hit, Need 100X Dilution, DCB High in both column, TCMX High in 1st column	YPIAJ	Dilution

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP042425

Review By	yogesh	Review On	4/24/2025 12:15:27 PM
Supervise By	mohammad	Supervise On	4/26/2025 2:18:47 AM
SubDirectory	PP042425	HP Acquire Method	HP Processing Method PP042225
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,P P24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP 24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369		
CCC	PP24332,PP24347,PP24352,PP24357		
Internal Standard/PEM			
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

17	AR1660CCC500	AR1660CCC500	PP071438.D	24 Apr 2025 17:02		YPIAJ	Ok
18	AR1242CCC500	AR1242CCC500	PP071439.D	24 Apr 2025 17:35		YPIAJ	Ok
19	AR1248CCC500	AR1248CCC500	PP071440.D	24 Apr 2025 17:51		YPIAJ	Ok
20	AR1254CCC500	AR1254CCC500	PP071441.D	24 Apr 2025 18:08		YPIAJ	Ok
21	I.BLK	I.BLK	PP071442.D	24 Apr 2025 18:24		YPIAJ	Ok
22	Q1867-01	RBR200059	PP071443.D	24 Apr 2025 18:40	AR1248 + AR1254 Hit, Need 2X Dilution	YPIAJ	Dilution
23	Q1867-01MS	RBR200059MS	PP071444.D	24 Apr 2025 18:56	Recovery Fail for AR1016 & AR1260	YPIAJ	Ok
24	Q1867-01MSD	RBR200059MSD	PP071445.D	24 Apr 2025 19:13	Recovery Fail for AR1016 & AR1260	YPIAJ	Ok
25	Q1867-03	VNJ211	PP071446.D	24 Apr 2025 19:29	AR1254 Hit	YPIAJ	Ok,M
26	Q1860-10DL	P001-SS007-01DL	PP071447.D	24 Apr 2025 19:45	AR1248 + AR1254 Hit, DCB High in 1st column	YPIAJ	Ok,M
27	Q1860-10DL2	P001-SS007-01DL2	PP071448.D	24 Apr 2025 20:01	AR1248 + AR1254 Hit,	YPIAJ	Not Ok
28	Q1860-11DL	P001-SS008-01DL	PP071449.D	24 Apr 2025 20:18	AR1248 + AR1254 Hit	YPIAJ	Ok,M
29	AR1660CCC500	AR1660CCC500	PP071450.D	24 Apr 2025 22:12		YPIAJ	Ok
30	AR1242CCC500	AR1242CCC500	PP071451.D	24 Apr 2025 22:44		YPIAJ	Ok
31	AR1248CCC500	AR1248CCC500	PP071452.D	24 Apr 2025 23:01		YPIAJ	Ok
32	AR1254CCC500	AR1254CCC500	PP071453.D	24 Apr 2025 23:17		YPIAJ	Ok
33	I.BLK	I.BLK	PP071454.D	24 Apr 2025 23:33		YPIAJ	Ok

M : Manual Integration

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP042825

Review By	yogesh	Review On	4/28/2025 11:52:03 AM
Supervise By	mohammad	Supervise On	5/2/2025 2:36:20 AM
SubDirectory	PP042825	HP Acquire Method	HP Processing Method PP042225

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,PP24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369
CCC	PP24332,PP24347,PP24352,PP24357
Internal Standard/PEM	
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	Sampleld	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PP071527.D	28 Apr 2025 08:12		YPIAJ	Ok
2	AR1660CCC500	AR1660CCC500	PP071528.D	28 Apr 2025 08:28		YPIAJ	Ok
3	AR1242CCC500	AR1242CCC500	PP071529.D	28 Apr 2025 08:47		YPIAJ	Ok
4	AR1248CCC500	AR1248CCC500	PP071530.D	28 Apr 2025 09:03		YPIAJ	Ok
5	AR1254CCC500	AR1254CCC500	PP071531.D	28 Apr 2025 09:20		YPIAJ	Ok
6	I.BLK	I.BLK	PP071532.D	28 Apr 2025 09:36		YPIAJ	Ok
7	PB167737BL	PB167737BL	PP071533.D	28 Apr 2025 09:53		YPIAJ	Ok
8	PB167750BL	PB167750BL	PP071534.D	28 Apr 2025 10:10		YPIAJ	Ok
9	PB167741BS	PB167741BS	PP071535.D	28 Apr 2025 10:26		YPIAJ	Ok
10	Q1860-33	P001-SS027-01	PP071536.D	28 Apr 2025 10:43	AR1248+AR1254+AR1260 Hit	YPIAJ	Ok,M
11	Q1860-35	P001-SS029-01	PP071537.D	28 Apr 2025 10:59	AR1248+AR1254+AR1260 Hit, Need dil, TCMX high in both column	YPIAJ	Dilution
12	Q1875-04	AUD-25-0024	PP071538.D	28 Apr 2025 11:16		YPIAJ	Ok
13	Q1860-42	P001-SS036-01	PP071539.D	28 Apr 2025 11:32	AR1254 Hit	YPIAJ	Ok,M
14	Q1876-04DL	AUD-25-0061DL	PP071540.D	28 Apr 2025 11:48	AR1254 Hit	YPIAJ	Ok,M
15	Q1876-09DL	AUD-25-0066DL	PP071541.D	28 Apr 2025 12:04	AR1254 Hit	YPIAJ	Ok,M
16	AR1660CCC500	AR1660CCC500	PP071542.D	28 Apr 2025 12:53		YPIAJ	Ok
17	AR1242CCC500	AR1242CCC500	PP071543.D	28 Apr 2025 13:09		YPIAJ	Ok

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP042825

Review By	yogesh	Review On	4/28/2025 11:52:03 AM
Supervise By	mohammad	Supervise On	5/2/2025 2:36:20 AM
SubDirectory	PP042825	HP Acquire Method	HP Processing Method PP042225
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,P P24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP 24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369		
CCC	PP24332,PP24347,PP24352,PP24357		
Internal Standard/PEM			
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

18	AR1248CCC500	AR1248CCC500	PP071544.D	28 Apr 2025 13:26		YPIAJ	Ok
19	AR1254CCC500	AR1254CCC500	PP071545.D	28 Apr 2025 13:42		YPIAJ	Ok,M
20	I.BLK	I.BLK	PP071546.D	28 Apr 2025 13:58		YPIAJ	Ok
21	Q1893-03	INTERIOR-1	PP071547.D	28 Apr 2025 14:14	AR1248 Hit, Need 20X	YPIAJ	Dilution
22	Q1893-04	INTERIOR-2	PP071548.D	28 Apr 2025 14:31	AR1248 Hit	YPIAJ	Ok,M
23	Q1893-05	INTERIOR-3	PP071549.D	28 Apr 2025 14:47	AR1248 Hit, Need 50X, TCMX high 2nd column	YPIAJ	Dilution
24	Q1893-06	INTERIOR-4	PP071550.D	28 Apr 2025 15:03	AR1248 Hit	YPIAJ	Ok,M
25	Q1893-07	EXTERIOR-1	PP071551.D	28 Apr 2025 15:19	AR1248 Hit, Need 10X	YPIAJ	Dilution
26	Q1893-03DL	INTERIOR-1DL	PP071552.D	28 Apr 2025 15:35	AR1248 Hit	YPIAJ	Ok
27	Q1893-05DL	INTERIOR-3DL	PP071553.D	28 Apr 2025 15:52	AR1248 Hit	YPIAJ	Ok
28	Q1893-07DL	EXTERIOR-1DL	PP071554.D	28 Apr 2025 16:08	AR1248 Hit	YPIAJ	Ok
29	Q1898-01	41525A	PP071555.D	28 Apr 2025 16:41		YPIAJ	Ok,M
30	Q1898-02	41525B	PP071556.D	28 Apr 2025 16:58		YPIAJ	Ok
31	AR1660CCC500	AR1660CCC500	PP071557.D	28 Apr 2025 18:03		YPIAJ	Ok
32	AR1242CCC500	AR1242CCC500	PP071558.D	28 Apr 2025 18:36		YPIAJ	Ok
33	AR1248CCC500	AR1248CCC500	PP071559.D	28 Apr 2025 18:53		YPIAJ	Ok
34	AR1254CCC500	AR1254CCC500	PP071560.D	28 Apr 2025 19:09		YPIAJ	Ok,M
35	I.BLK	I.BLK	PP071561.D	28 Apr 2025 19:25		YPIAJ	Ok
36	PB167754BL	PB167754BL	PP071562.D	28 Apr 2025 19:42		YPIAJ	Ok

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP042825

Review By	yogesh	Review On	4/28/2025 11:52:03 AM
Supervise By	mohammad	Supervise On	5/2/2025 2:36:20 AM
SubDirectory	PP042825	HP Acquire Method	HP Processing Method PP042225
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,P P24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP 24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369		
CCC	PP24332,PP24347,PP24352,PP24357		
Internal Standard/PEM			
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

37	PB167754BS	PB167754BS	PP071563.D	28 Apr 2025 19:58		YPIAJ	Ok
38	PB167765BL	PB167765BL	PP071564.D	28 Apr 2025 20:15		YPIAJ	Ok
39	PB167765BS	PB167765BS	PP071565.D	28 Apr 2025 20:31		YPIAJ	Ok
40	Q1884-01	P001-SS037-01	PP071566.D	28 Apr 2025 20:47	AR1248 + AR1254 + AR1260 Hit, Need Dilution, DCB High in both col	YPIAJ	Dilution
41	Q1884-02	P001-SS038-01	PP071567.D	28 Apr 2025 21:04	AR1254 + AR1260 Hit, Need dilution, DCB High in both column	YPIAJ	Dilution
42	Q1889-01	COMP-1	PP071568.D	28 Apr 2025 21:20	Check Chromatogram	YPIAJ	Not Ok
43	Q1889-01MS	COMP-1MS	PP071569.D	28 Apr 2025 21:36		YPIAJ	Not Ok
44	Q1889-01MSD	COMP-1MSD	PP071570.D	28 Apr 2025 21:53		YPIAJ	Not Ok
45	Q1889-02	COMP-2	PP071571.D	28 Apr 2025 22:09		YPIAJ	Ok
46	AR1660CCC500	AR1660CCC500	PP071572.D	28 Apr 2025 23:31		YPIAJ	Ok
47	AR1242CCC500	AR1242CCC500	PP071573.D	29 Apr 2025 00:03		YPIAJ	Ok
48	AR1248CCC500	AR1248CCC500	PP071574.D	29 Apr 2025 00:20		YPIAJ	Ok
49	AR1254CCC500	AR1254CCC500	PP071575.D	29 Apr 2025 00:36		YPIAJ	Ok,M
50	I.BLK	I.BLK	PP071576.D	29 Apr 2025 00:52		YPIAJ	Ok
51	Q1889-03	COMP-3	PP071577.D	29 Apr 2025 01:09		YPIAJ	Ok
52	Q1891-01	MH-C	PP071578.D	29 Apr 2025 01:25		YPIAJ	Ok,M
53	Q1891-05	MH-D	PP071579.D	29 Apr 2025 01:42		YPIAJ	Ok,M
54	Q1892-01	MH-G	PP071580.D	29 Apr 2025 01:58		YPIAJ	Ok

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP042825

Review By	yogesh	Review On	4/28/2025 11:52:03 AM
Supervise By	mohammad	Supervise On	5/2/2025 2:36:20 AM
SubDirectory	PP042825	HP Acquire Method	HP Processing Method PP042225
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,P P24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP 24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369		
CCC	PP24332,PP24347,PP24352,PP24357		
Internal Standard/PEM			
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

55	Q1892-05	MH-H	PP071581.D	29 Apr 2025 02:14		YPIAJ	Ok
56	Q1892-09	MH-U2	PP071582.D	29 Apr 2025 02:31		YPIAJ	Ok
57	Q1895-01	COMP-1	PP071583.D	29 Apr 2025 02:47		YPIAJ	Ok
58	Q1895-03	COMP-2	PP071584.D	29 Apr 2025 03:03		YPIAJ	Ok
59	Q1895-05	COMP-3	PP071585.D	29 Apr 2025 03:20	AR1254 Hit	YPIAJ	Ok,M
60	Q1898-04	42525B	PP071586.D	29 Apr 2025 03:36		YPIAJ	Ok
61	AR1660CCC500	AR1660CCC500	PP071587.D	29 Apr 2025 04:58		YPIAJ	Ok
62	AR1242CCC500	AR1242CCC500	PP071588.D	29 Apr 2025 05:31		YPIAJ	Ok
63	AR1248CCC500	AR1248CCC500	PP071589.D	29 Apr 2025 05:47		YPIAJ	Ok
64	AR1254CCC500	AR1254CCC500	PP071590.D	29 Apr 2025 06:03		YPIAJ	Ok
65	I.BLK	I.BLK	PP071591.D	29 Apr 2025 06:20		YPIAJ	Ok
66	Q1898-03	42525A	PP071592.D	29 Apr 2025 06:36		YPIAJ	Ok
67	AR1660CCC500	AR1660CCC500	PP071593.D	29 Apr 2025 07:58		YPIAJ	Ok
68	AR1242CCC500	AR1242CCC500	PP071594.D	29 Apr 2025 08:31		YPIAJ	Ok
69	AR1248CCC500	AR1248CCC500	PP071595.D	29 Apr 2025 08:47		YPIAJ	Ok
70	AR1254CCC500	AR1254CCC500	PP071596.D	29 Apr 2025 09:04		YPIAJ	Ok,M
71	I.BLK	I.BLK	PP071597.D	29 Apr 2025 09:20		YPIAJ	Ok

M : Manual Integration

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP042925

Review By	yogesh	Review On	4/29/2025 11:08:35 AM
Supervise By	mohammad	Supervise On	5/2/2025 2:54:00 AM
SubDirectory	PP042925	HP Acquire Method	HP Processing Method PP042225

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,PP24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369
CCC	PP24332,PP24347,PP24352,PP24357
Internal Standard/PEM	
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PP071598.D	29 Apr 2025 10:16		YPIAJ	Ok
2	AR1660CCC500	AR1660CCC500	PP071599.D	29 Apr 2025 10:33		YPIAJ	Ok
3	AR1242CCC500	AR1242CCC500	PP071600.D	29 Apr 2025 10:51		YPIAJ	Ok
4	AR1248CCC500	AR1248CCC500	PP071601.D	29 Apr 2025 11:07		YPIAJ	Ok
5	AR1254CCC500	AR1254CCC500	PP071602.D	29 Apr 2025 11:24		YPIAJ	Ok
6	I.BLK	I.BLK	PP071603.D	29 Apr 2025 11:40		YPIAJ	Ok
7	Q1860-35DL	P001-SS029-01DL	PP071604.D	29 Apr 2025 11:56	AR1248+AR1254+AR1260 Hit	YPIAJ	Ok,M
8	Q1884-01DL	P001-SS037-01DL	PP071605.D	29 Apr 2025 12:12	AR1248 + AR1254 + AR1260 Hit, DCB High in both col, TCMX High in 2nd col	YPIAJ	Ok,M
9	Q1884-02DL	P001-SS038-01DL	PP071606.D	29 Apr 2025 12:29	AR1254 + AR1260 Hit, DCB High in both col	YPIAJ	Ok,M
10	PB167776BL	PB167776BL	PP071607.D	29 Apr 2025 13:30		YPIAJ	Ok
11	PB167776BS	PB167776BS	PP071608.D	29 Apr 2025 13:46		YPIAJ	Ok
12	Q1883-01	OU4-PCS-TC-27-0423	PP071609.D	29 Apr 2025 14:02		YPIAJ	Ok,M
13	Q1883-03	OU4-PCS-TC-28-0423	PP071610.D	29 Apr 2025 14:19		YPIAJ	Ok
14	Q1883-05	OU4-PCS-TC-29-0423	PP071611.D	29 Apr 2025 14:35		YPIAJ	Ok
15	Q1883-07	OU4-PCS-TC-30-0423	PP071612.D	29 Apr 2025 14:51		YPIAJ	Ok
16	Q1883-09	OU4-PCS-TC-31-0423	PP071613.D	29 Apr 2025 15:08		YPIAJ	Ok
17	AR1660CCC500	AR1660CCC500	PP071614.D	29 Apr 2025 16:29		YPIAJ	Ok

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP042925

Review By	yogesh	Review On	4/29/2025 11:08:35 AM
Supervise By	mohammad	Supervise On	5/2/2025 2:54:00 AM
SubDirectory	PP042925	HP Acquire Method	HP Processing Method PP042225
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,P P24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP 24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369		
CCC	PP24332,PP24347,PP24352,PP24357		
Internal Standard/PEM			
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

18	AR1242CCC500	AR1242CCC500	PP071615.D	29 Apr 2025 17:02		YPIAJ	Ok
19	AR1248CCC500	AR1248CCC500	PP071616.D	29 Apr 2025 17:18		YPIAJ	Ok
20	AR1254CCC500	AR1254CCC500	PP071617.D	29 Apr 2025 17:34		YPIAJ	Ok
21	I.BLK	I.BLK	PP071618.D	29 Apr 2025 17:51		YPIAJ	Ok
22	Q1883-11	OU4-PCS-TC-32-04232	PP071619.D	29 Apr 2025 18:07		YPIAJ	Ok
23	Q1883-13	OU4-VSL-18-042325	PP071620.D	29 Apr 2025 18:23		YPIAJ	Ok
24	Q1883-15	OU4-VSL-19-042325	PP071621.D	29 Apr 2025 18:40		YPIAJ	Ok
25	Q1901-02	B-167-SB01	PP071622.D	29 Apr 2025 18:56		YPIAJ	Ok,M
26	Q1901-04	B-167-SB02	PP071623.D	29 Apr 2025 19:12		YPIAJ	Ok
27	PB167782BL	PB167782BL	PP071624.D	29 Apr 2025 19:29		YPIAJ	Ok
28	PB167782BS	PB167782BS	PP071625.D	29 Apr 2025 19:45		YPIAJ	Ok
29	PB167782BSD	PB167782BSD	PP071626.D	29 Apr 2025 20:01		YPIAJ	Ok
30	Q1890-01	TW-WTS-07	PP071627.D	29 Apr 2025 20:18		YPIAJ	Ok
31	Q1896-03	295-BERGEN-FRAC	PP071628.D	29 Apr 2025 20:34		YPIAJ	Ok,M
32	AR1660CCC500	AR1660CCC500	PP071629.D	29 Apr 2025 21:55		YPIAJ	Ok
33	AR1242CCC500	AR1242CCC500	PP071630.D	29 Apr 2025 22:28		YPIAJ	Ok
34	AR1248CCC500	AR1248CCC500	PP071631.D	29 Apr 2025 22:44		YPIAJ	Ok
35	AR1254CCC500	AR1254CCC500	PP071632.D	29 Apr 2025 23:01		YPIAJ	Ok,M
36	I.BLK	I.BLK	PP071633.D	29 Apr 2025 23:17		YPIAJ	Ok

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP042925

Review By	yogesh	Review On	4/29/2025 11:08:35 AM		
Supervise By	mohammad	Supervise On	5/2/2025 2:54:00 AM		
SubDirectory	PP042925	HP Acquire Method	HP Processing Method	PP042225	
STD. NAME	STD REF.#				
Tune/Reschk					
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,P P24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP 24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369				
CCC	PP24332,PP24347,PP24352,PP24357				
Internal Standard/PEM					
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387				
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

M : Manual Integration

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP043025

Review By	yogesh	Review On	4/30/2025 11:33:57 AM
Supervise By	mohammad	Supervise On	5/2/2025 3:18:29 AM
SubDirectory	PP043025	HP Acquire Method	HP Processing Method PP042225
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,P P24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP 24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369		
CCC	PP24332,PP24347,PP24352,PP24357		
Internal Standard/PEM			
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	Sampleld	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PP071634.D	30 Apr 2025 08:51		YPIAJ	Ok
2	AR1660CCC500	AR1660CCC500	PP071635.D	30 Apr 2025 09:07		YPIAJ	Ok
3	AR1242CCC500	AR1242CCC500	PP071636.D	30 Apr 2025 09:28		YPIAJ	Ok
4	AR1248CCC500	AR1248CCC500	PP071637.D	30 Apr 2025 09:45		YPIAJ	Ok
5	AR1254CCC500	AR1254CCC500	PP071638.D	30 Apr 2025 10:01		YPIAJ	Ok,M
6	I.BLK	I.BLK	PP071639.D	30 Apr 2025 10:17		YPIAJ	Ok
7	Q1867-01DL	RBR200059DL	PP071640.D	30 Apr 2025 11:09	AR1248 + AR1254 Hit	YPIAJ	Ok,M
8	PB167794BL	PB167794BL	PP071641.D	30 Apr 2025 12:39		YPIAJ	Ok
9	PB167794BS	PB167794BS	PP071642.D	30 Apr 2025 12:56		YPIAJ	Ok
10	Q1901-03	B-170-SB01	PP071643.D	30 Apr 2025 13:12		YPIAJ	Ok,M
11	Q1905-01	MH-G	PP071644.D	30 Apr 2025 13:28		YPIAJ	Ok,M
12	Q1860-09	P001-SS006-01	PP071645.D	30 Apr 2025 13:44	AR1260 Hit, Need 5X Dilution	YPIAJ	Dilution
13	Q1860-09DL	P001-SS006-01DL	PP071646.D	30 Apr 2025 14:01	AR1260 Hit	YPIAJ	Ok,M
14	Q1910-01	MOO-25-0123-27	PP071647.D	30 Apr 2025 14:33		YPIAJ	Ok,M
15	Q1910-01MS	MOO-25-0123-27MS	PP071648.D	30 Apr 2025 14:49		YPIAJ	Ok,M
16	Q1910-01MSD	MOO-25-0123-27MSD	PP071649.D	30 Apr 2025 15:05		YPIAJ	Ok,M
17	AR1660CCC500	AR1660CCC500	PP071650.D	30 Apr 2025 16:11		YPIAJ	Ok
18	AR1242CCC500	AR1242CCC500	PP071651.D	30 Apr 2025 16:27		YPIAJ	Ok

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP043025

Review By	yogesh	Review On	4/30/2025 11:33:57 AM
Supervise By	mohammad	Supervise On	5/2/2025 3:18:29 AM
SubDirectory	PP043025	HP Acquire Method	HP Processing Method PP042225
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,P P24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP 24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369		
CCC	PP24332,PP24347,PP24352,PP24357		
Internal Standard/PEM			
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

19	AR1248CCC500	AR1248CCC500	PP071652.D	30 Apr 2025 16:43		YPIAJ	Ok
20	AR1254CCC500	AR1254CCC500	PP071653.D	30 Apr 2025 17:00		YPIAJ	Ok,M
21	I.BLK	I.BLK	PP071654.D	30 Apr 2025 17:16		YPIAJ	Ok
22	Q1911-01	EO-03-04292025	PP071655.D	30 Apr 2025 17:33		YPIAJ	Ok,M
23	Q1912-01	MH-E	PP071656.D	30 Apr 2025 17:49		YPIAJ	Ok,M
24	Q1912-05	MH-F	PP071657.D	30 Apr 2025 18:06		YPIAJ	Ok,M
25	Q1905-05	MH-H	PP071658.D	30 Apr 2025 18:22		YPIAJ	Ok,M
26	Q1907-01	CO-8R-WC	PP071659.D	30 Apr 2025 18:38	AR1254 & AR1260 hits	YPIAJ	Ok,M
27	Q1921-01	BC226751-1-1	PP071660.D	30 Apr 2025 18:55		YPIAJ	Ok,M
28	Q1921-02	BC226751-1-2	PP071661.D	30 Apr 2025 19:11		YPIAJ	Ok,M
29	Q1921-03	BC274768-1-1	PP071662.D	30 Apr 2025 19:27		YPIAJ	Ok,M
30	Q1921-04	BC274768-1-2	PP071663.D	30 Apr 2025 19:44		YPIAJ	Ok,M
31	Q1921-05	BC271327-1-1	PP071664.D	30 Apr 2025 20:00		YPIAJ	Ok,M
32	AR1660CCC500	AR1660CCC500	PP071665.D	30 Apr 2025 21:38		YPIAJ	Ok
33	AR1242CCC500	AR1242CCC500	PP071666.D	30 Apr 2025 22:11		YPIAJ	Ok
34	AR1248CCC500	AR1248CCC500	PP071667.D	30 Apr 2025 22:27		YPIAJ	Ok
35	AR1254CCC500	AR1254CCC500	PP071668.D	30 Apr 2025 22:44		YPIAJ	Ok,M
36	I.BLK	I.BLK	PP071669.D	30 Apr 2025 23:00		YPIAJ	Ok,M
37	PB167809BL	PB167809BL	PP071670.D	30 Apr 2025 23:16		YPIAJ	Ok

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP043025

Review By	yogesh	Review On	4/30/2025 11:33:57 AM
Supervise By	mohammad	Supervise On	5/2/2025 3:18:29 AM
SubDirectory	PP043025	HP Acquire Method	HP Processing Method PP042225
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,P P24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP 24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369		
CCC	PP24332,PP24347,PP24352,PP24357		
Internal Standard/PEM			
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

38	PB167809BS	PB167809BS	PP071671.D	30 Apr 2025 23:33		YPIAJ	Ok
39	Q1921-06	BC271327-1-2	PP071672.D	30 Apr 2025 23:49		YPIAJ	Ok
40	Q1921-07	BC248150-1-1	PP071673.D	01 May 2025 00:05	DCB high in 1st column	YPIAJ	Ok,M
41	Q1921-08	BC248150-1-2	PP071674.D	01 May 2025 00:22	DCB high in 2nd column	YPIAJ	Ok,M
42	AR1660CCC500	AR1660CCC500	PP071675.D	01 May 2025 02:00		YPIAJ	Ok
43	AR1242CCC500	AR1242CCC500	PP071676.D	01 May 2025 02:49		YPIAJ	Ok,M
44	AR1248CCC500	AR1248CCC500	PP071677.D	01 May 2025 03:05		YPIAJ	Ok
45	AR1254CCC500	AR1254CCC500	PP071678.D	01 May 2025 03:21		YPIAJ	Ok,M
46	I.BLK	I.BLK	PP071679.D	01 May 2025 03:38		YPIAJ	Ok,M

M : Manual Integration