

Hit Summary Sheet
SW-846

SDG No.: Q1860

Order ID: Q1860

Client: Weston Solutions, Inc.

Project ID: Q1908

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
Client ID : P001-SS001-01								
Q1860-01	P001-SS001-01	SOIL	Aroclor-1260	58.6		3.70	19.6	ug/kg
			Total Concentration:	58.600				
Client ID : P001-SS002-01								
Q1860-04	P001-SS002-01	SOIL	Aroclor-1254	200		3.90	20.8	ug/kg
			Total Concentration:	200.000				
Client ID : P001-SS002-02								
Q1860-05	P001-SS002-02	SOIL	Aroclor-1254	302		3.80	20.2	ug/kg
			Total Concentration:	302.000				
Client ID : P001-SS003-01								
Q1860-06	P001-SS003-01	SOIL	Aroclor-1254	154		3.70	19.4	ug/kg
Q1860-06	P001-SS003-01	SOIL	Aroclor-1260	126		3.70	19.4	ug/kg
			Total Concentration:	280.000				
Client ID : P001-SS004-01								
Q1860-07	P001-SS004-01	SOIL	Aroclor-1254	179		3.70	19.6	ug/kg
Q1860-07	P001-SS004-01	SOIL	Aroclor-1260	198		3.70	19.6	ug/kg
			Total Concentration:	377.000				
Client ID : P001-SS005-01								
Q1860-08	P001-SS005-01	SOIL	Aroclor-1260	299		3.70	19.7	ug/kg
			Total Concentration:	299.000				
Client ID : P001-SS006-01								
Q1860-09	P001-SS006-01	SOIL	Aroclor-1260	439 E		3.60	18.9	ug/kg
			Total Concentration:	439.000				
Client ID : P001-SS006-01DL								
Q1860-09DL	P001-SS006-01DL	SOIL	Aroclor-1260	375 D		18.0	94.7	ug/kg
			Total Concentration:	375.000				
Client ID : P001-SS007-01								
Q1860-10	P001-SS007-01	SOIL	Aroclor-1248	4200 E		6.80	19.4	ug/kg
Q1860-10	P001-SS007-01	SOIL	Aroclor-1254	3900 E		3.70	19.4	ug/kg
			Total Concentration:	8,100.000				

A

B

C

D

Hit Summary Sheet SW-846

SDG No.: Q1860

Order ID: Q1860

Client: Weston Solutions, Inc.

Project ID: Q1908

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
Client ID : P001-SS007-01DL								
Q1860-10DL	P001-SS007-01DL	SOIL	Aroclor-1248	4000	D	169	485	ug/kg
Q1860-10DL	P001-SS007-01DL	SOIL	Aroclor-1254	3400	D	91.6	485	ug/kg
Total Concentration:				7,400.000				
Client ID : P001-SS008-01								
Q1860-11	P001-SS008-01	SOIL	Aroclor-1248	7200	E	6.70	19.1	ug/kg
Q1860-11	P001-SS008-01	SOIL	Aroclor-1254	7000	E	3.60	19.1	ug/kg
Total Concentration:				14,200.000				
Client ID : P001-SS008-01DL								
Q1860-11DL	P001-SS008-01DL	SOIL	Aroclor-1248	9900	D	666	1900	ug/kg
Q1860-11DL	P001-SS008-01DL	SOIL	Aroclor-1254	7900	D	361	1900	ug/kg
Total Concentration:				17,800.000				
Client ID : P001-SS009-01								
Q1860-12	P001-SS009-01	SOIL	Aroclor-1248	7300	E	6.80	19.6	ug/kg
Q1860-12	P001-SS009-01	SOIL	Aroclor-1254	5700	E	3.70	19.6	ug/kg
Total Concentration:				13,000.000				
Client ID : P001-SS009-01DL								
Q1860-12DL	P001-SS009-01DL	SOIL	Aroclor-1248	9400	D	684	2000	ug/kg
Q1860-12DL	P001-SS009-01DL	SOIL	Aroclor-1254	6700	D	371	2000	ug/kg
Total Concentration:				16,100.000				
Client ID : P001-SS010-01								
Q1860-13	P001-SS010-01	SOIL	Aroclor-1254	57.4		4.00	21.0	ug/kg
Q1860-13	P001-SS010-01	SOIL	Aroclor-1260	54.6		4.00	21.0	ug/kg
Total Concentration:				112.000				
Client ID : P001-SS011-01								
Q1860-14	P001-SS011-01	SOIL	Aroclor-1248	325		7.10	20.5	ug/kg
Q1860-14	P001-SS011-01	SOIL	Aroclor-1254	282		3.90	20.5	ug/kg
Total Concentration:				607.000				
Client ID : P001-SS012-01								
Q1860-15	P001-SS012-01	SOIL	Aroclor-1248	6500	E	7.30	20.9	ug/kg
Q1860-15	P001-SS012-01	SOIL	Aroclor-1254	3900	E	3.90	20.9	ug/kg
Total Concentration:				10,400.000				

Hit Summary Sheet
SW-846

SDG No.: Q1860

Order ID: Q1860

Client: Weston Solutions, Inc.

Project ID: Q1908

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
Client ID : P001-SS012-01DL								
Q1860-15DL	P001-SS012-01DL	SOIL	Aroclor-1248	7000	D	182	522	ug/kg
Q1860-15DL	P001-SS012-01DL	SOIL	Aroclor-1254	3900	D	98.6	522	ug/kg
Total Concentration:				10,900.000				
Client ID : P001-SS013-01								
Q1860-16	P001-SS013-01	SOIL	Aroclor-1248	4300	E	6.90	19.9	ug/kg
Q1860-16	P001-SS013-01	SOIL	Aroclor-1254	3200	E	3.80	19.9	ug/kg
Total Concentration:				7,500.000				
Client ID : P001-SS013-01DL								
Q1860-16DL	P001-SS013-01DL	SOIL	Aroclor-1248	5100	D	173	498	ug/kg
Q1860-16DL	P001-SS013-01DL	SOIL	Aroclor-1254	3500	D	94.0	498	ug/kg
Total Concentration:				8,600.000				
Client ID : P001-SS014-01								
Q1860-17	P001-SS014-01	SOIL	Aroclor-1248	398		7.40	21.3	ug/kg
Q1860-17	P001-SS014-01	SOIL	Aroclor-1254	412		4.00	21.3	ug/kg
Total Concentration:				810.000				
Client ID : P001-SS015-01								
Q1860-18	P001-SS015-01	SOIL	Aroclor-1248	71.7		7.90	22.7	ug/kg
Q1860-18	P001-SS015-01	SOIL	Aroclor-1254	102		4.30	22.7	ug/kg
Total Concentration:				173.700				
Client ID : P001-SS016-01								
Q1860-19	P001-SS016-01	SOIL	Aroclor-1248	932	E	7.10	20.4	ug/kg
Q1860-19	P001-SS016-01	SOIL	Aroclor-1254	734	E	3.90	20.4	ug/kg
Total Concentration:				1,666.000				
Client ID : P001-SS016-01DL								
Q1860-19DL	P001-SS016-01DL	SOIL	Aroclor-1248	986	D	71.1	204	ug/kg
Q1860-19DL	P001-SS016-01DL	SOIL	Aroclor-1254	725	D	38.6	204	ug/kg
Total Concentration:				1,711.000				
Client ID : P001-SS017-01								
Q1860-22	P001-SS017-01	SOIL	Aroclor-1248	770	E	7.00	20.1	ug/kg
Q1860-22	P001-SS017-01	SOIL	Aroclor-1254	759	E	3.80	20.1	ug/kg

Hit Summary Sheet
SW-846

SDG No.: Q1860

Order ID: Q1860

Client: Weston Solutions, Inc.

Project ID: Q1908

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
Total Concentration:				1,529.000				
Client ID :	P001-SS017-01DL							
Q1860-22DL	P001-SS017-01DL	SOIL	Aroclor-1248	749	D	34.9	100	ug/kg
Q1860-22DL	P001-SS017-01DL	SOIL	Aroclor-1254	663	D	18.9	100	ug/kg
Total Concentration:				1,412.000				
Client ID :	P001-SS017-02							
Q1860-23	P001-SS017-02	SOIL	Aroclor-1248	925	E	7.20	20.6	ug/kg
Q1860-23	P001-SS017-02	SOIL	Aroclor-1254	631	E	3.90	20.6	ug/kg
Total Concentration:				1,556.000				
Client ID :	P001-SS017-02DL							
Q1860-23DL	P001-SS017-02DL	SOIL	Aroclor-1248	878	D	35.8	103	ug/kg
Q1860-23DL	P001-SS017-02DL	SOIL	Aroclor-1254	525	D	19.4	103	ug/kg
Total Concentration:				1,403.000				
Client ID :	P001-SS018-01							
Q1860-24	P001-SS018-01	SOIL	Aroclor-1254	76.3		3.60	18.9	ug/kg
Total Concentration:				76.300				
Client ID :	P001-SS019-01							
Q1860-25	P001-SS019-01	SOIL	Aroclor-1254	435	E	3.60	19.2	ug/kg
Q1860-25	P001-SS019-01	SOIL	Aroclor-1260	297		3.70	19.2	ug/kg
Total Concentration:				732.000				
Client ID :	P001-SS019-01DL							
Q1860-25DL	P001-SS019-01DL	SOIL	Aroclor-1254	435	D	18.1	96.1	ug/kg
Q1860-25DL	P001-SS019-01DL	SOIL	Aroclor-1260	298	D	18.3	96.1	ug/kg
Total Concentration:				733.000				
Client ID :	P001-SS020-01							
Q1860-26	P001-SS020-01	SOIL	Aroclor-1254	5900	EP	3.80	20.0	ug/kg
Total Concentration:				5,900.000				
Client ID :	P001-SS020-01DL							
Q1860-26DL	P001-SS020-01DL	SOIL	Aroclor-1254	4500	D	94.3	500	ug/kg
Total Concentration:				4,500.000				

A
B
C
D



A

B

C

D

Order ID: Q1860

Project ID: Q1908

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
Client ID : Q1860-27	P001-SS021-01 P001-SS021-01	SOIL	Aroclor-1254		381 E	3.60	19.2	ug/kg
Total Concentration:				381.000				
Client ID : Q1860-27DL	P001-SS021-01DL P001-SS021-01DL	SOIL	Aroclor-1254		400 D	18.1	95.9	ug/kg
Total Concentration:				400.000				
Client ID : Q1860-28	P001-SS022-01 P001-SS022-01	SOIL	Aroclor-1254		203	3.80	20.3	ug/kg
Total Concentration:				203.000				
Client ID : Q1860-29	P001-SS023-01 P001-SS023-01	SOIL	Aroclor-1254		410 E	3.80	20.2	ug/kg
Total Concentration:				410.000				
Client ID : Q1860-29DL	P001-SS023-01DL P001-SS023-01DL	SOIL	Aroclor-1254		361 D	19.1	101	ug/kg
Total Concentration:				361.000				
Client ID : Q1860-30	P001-SS024-01 P001-SS024-01	SOIL	Aroclor-1254		347	3.80	20.1	ug/kg
Total Concentration:				347.000				
Client ID : Q1860-31	P001-SS025-01 P001-SS025-01	SOIL	Aroclor-1254		1200 E	3.70	19.5	ug/kg
Q1860-31	P001-SS025-01	SOIL	Aroclor-1260		1200 E	3.70	19.5	ug/kg
Total Concentration:				2,400.000				
Client ID : Q1860-31DL	P001-SS025-01DL P001-SS025-01DL	SOIL	Aroclor-1254		1200 D	18.5	97.7	ug/kg
Q1860-31DL	P001-SS025-01DL	SOIL	Aroclor-1260		1300 D	18.6	97.7	ug/kg
Total Concentration:				2,500.000				
Client ID : Q1860-32	P001-SS026-01 P001-SS026-01	SOIL	Aroclor-1254		3800 E	4.10	21.5	ug/kg
Total Concentration:				3,800.000				
Client ID :	P001-SS026-01DL							

Hit Summary Sheet
SW-846

SDG No.: Q1860

Order ID: Q1860

Client: Weston Solutions, Inc.

Project ID: Q1908

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
Q1860-32DL	P001-SS026-01DL	SOIL	Aroclor-1254	4800	D	81.0	429	ug/kg
Total Concentration:				4,800.000				
Client ID : P001-SS027-01								
Q1860-33	P001-SS027-01	SOIL	Aroclor-1248	114		6.80	19.6	ug/kg
Q1860-33	P001-SS027-01	SOIL	Aroclor-1254	225		3.70	19.6	ug/kg
Q1860-33	P001-SS027-01	SOIL	Aroclor-1260	176		3.70	19.6	ug/kg
Total Concentration:				515.000				
Client ID : P001-SS028-01								
Q1860-34	P001-SS028-01	SOIL	Aroclor-1248	1800	E	7.60	21.8	ug/kg
Q1860-34	P001-SS028-01	SOIL	Aroclor-1254	3000	E	4.10	21.8	ug/kg
Q1860-34	P001-SS028-01	SOIL	Aroclor-1260	2100	E	4.20	21.8	ug/kg
Total Concentration:				6,900.000				
Client ID : P001-SS028-01DL								
Q1860-34DL	P001-SS028-01DL	SOIL	Aroclor-1248	2100	D	152	437	ug/kg
Q1860-34DL	P001-SS028-01DL	SOIL	Aroclor-1254	2900	D	82.5	437	ug/kg
Q1860-34DL	P001-SS028-01DL	SOIL	Aroclor-1260	2300	D	83.0	437	ug/kg
Total Concentration:				7,300.000				
Client ID : P001-SS029-01								
Q1860-35	P001-SS029-01	SOIL	Aroclor-1248	317		7.30	21.1	ug/kg
Q1860-35	P001-SS029-01	SOIL	Aroclor-1254	781	E	4.00	21.1	ug/kg
Q1860-35	P001-SS029-01	SOIL	Aroclor-1260	532	E	4.00	21.1	ug/kg
Total Concentration:				1,630.000				
Client ID : P001-SS029-01DL								
Q1860-35DL	P001-SS029-01DL	SOIL	Aroclor-1248	294	D	36.7	105	ug/kg
Q1860-35DL	P001-SS029-01DL	SOIL	Aroclor-1254	665	D	19.9	105	ug/kg
Q1860-35DL	P001-SS029-01DL	SOIL	Aroclor-1260	504	D	20.0	105	ug/kg
Total Concentration:				1,463.000				
Client ID : P001-SS030-01								
Q1860-36	P001-SS030-01	SOIL	Aroclor-1248	2100	E	7.10	20.3	ug/kg
Q1860-36	P001-SS030-01	SOIL	Aroclor-1254	3300	E	3.80	20.3	ug/kg
Q1860-36	P001-SS030-01	SOIL	Aroclor-1260	3500	E	3.90	20.3	ug/kg
Total Concentration:				8,900.000				

Hit Summary Sheet

SW-846

SDG No.: Q1860

Order ID: Q1860

Client: **Weston Solutions, Inc.**

Project ID: Q1908

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
Client ID : P001-SS030-01DL								
Q1860-36DL	P001-SS030-01DL	SOIL	Aroclor-1248	2800	D	141	405	ug/kg
Q1860-36DL	P001-SS030-01DL	SOIL	Aroclor-1254	4100	D	76.5	405	ug/kg
Q1860-36DL	P001-SS030-01DL	SOIL	Aroclor-1260	3900	D	77.0	405	ug/kg
Total Concentration:				10,800.000				
Client ID : P001-SS031-01								
Q1860-37	P001-SS031-01	SOIL	Aroclor-1254	7700	E	3.60	19.1	ug/kg
Total Concentration:				7,700.000				
Client ID : P001-SS031-01DL								
Q1860-37DL	P001-SS031-01DL	SOIL	Aroclor-1254	9100	ED	72.0	381	ug/kg
Total Concentration:				9,100.000				
Client ID : P001-SS031-01DL2								
Q1860-37DL2	P001-SS031-01DL2	SOIL	Aroclor-1254	9100	D	360	1900	ug/kg
Total Concentration:				9,100.000				
Client ID : P001-SS032-01								
Q1860-38	P001-SS032-01	SOIL	Aroclor-1248	2100	E	6.80	19.5	ug/kg
Q1860-38	P001-SS032-01	SOIL	Aroclor-1254	5300	E	3.70	19.5	ug/kg
Total Concentration:				7,400.000				
Client ID : P001-SS032-01DL								
Q1860-38DL	P001-SS032-01DL	SOIL	Aroclor-1248	1900	D	68.0	195	ug/kg
Q1860-38DL	P001-SS032-01DL	SOIL	Aroclor-1254	4000	ED	36.9	195	ug/kg
Total Concentration:				5,900.000				
Client ID : P001-SS032-01DL2								
Q1860-38DL2	P001-SS032-01DL2	SOIL	Aroclor-1248	2400	D	340	976	ug/kg
Q1860-38DL2	P001-SS032-01DL2	SOIL	Aroclor-1254	4200	D	184	976	ug/kg
Total Concentration:				6,600.000				
Client ID : P001-SS033-01								
Q1860-39	P001-SS033-01	SOIL	Aroclor-1248	3900	E	6.70	19.4	ug/kg
Q1860-39	P001-SS033-01	SOIL	Aroclor-1254	3600	E	3.70	19.4	ug/kg
Total Concentration:				7,500.000				
Client ID : P001-SS033-01DL								

Hit Summary Sheet
SW-846

SDG No.: Q1860

Order ID: Q1860

Client: Weston Solutions, Inc.

Project ID: Q1908

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
Q1860-39DL	P001-SS033-01DL	SOIL	Aroclor-1248	4900	D	135	387	ug/kg
Q1860-39DL	P001-SS033-01DL	SOIL	Aroclor-1254	4000	D	73.1	387	ug/kg
Total Concentration:				8,900.000				
Client ID :	P001-SS034-01							
Q1860-40	P001-SS034-01	SOIL	Aroclor-1248	1700	E	6.70	19.3	ug/kg
Q1860-40	P001-SS034-01	SOIL	Aroclor-1254	2700	E	3.60	19.3	ug/kg
Total Concentration:				4,400.000				
Client ID :	P001-SS034-01DL							
Q1860-40DL	P001-SS034-01DL	SOIL	Aroclor-1248	2600	D	135	386	ug/kg
Q1860-40DL	P001-SS034-01DL	SOIL	Aroclor-1254	3600	D	72.9	386	ug/kg
Total Concentration:				6,200.000				
Client ID :	P001-SS035-01							
Q1860-41	P001-SS035-01	SOIL	Aroclor-1248	2500	E	6.30	18.1	ug/kg
Q1860-41	P001-SS035-01	SOIL	Aroclor-1254	3900	E	3.40	18.1	ug/kg
Total Concentration:				6,400.000				
Client ID :	P001-SS035-01DL							
Q1860-41DL	P001-SS035-01DL	SOIL	Aroclor-1248	2900	D	63.0	181	ug/kg
Q1860-41DL	P001-SS035-01DL	SOIL	Aroclor-1254	4100	ED	34.1	181	ug/kg
Total Concentration:				7,000.000				
Client ID :	P001-SS035-01DL2							
Q1860-41DL2	P001-SS035-01DL2	SOIL	Aroclor-1248	3200	D	315	904	ug/kg
Q1860-41DL2	P001-SS035-01DL2	SOIL	Aroclor-1254	4200	DP	171	904	ug/kg
Total Concentration:				7,400.000				
Client ID :	P001-SS036-01							
Q1860-42	P001-SS036-01	SOIL	Aroclor-1254	19.6	P	3.40	18.0	ug/kg
Total Concentration:				19.600				



SAMPLE DATA

A

B

C

D

Report of Analysis

Client:	Weston Solutions, Inc.		Date Collected:	04/16/25	
Project:	RFP 911		Date Received:	04/23/25	
Client Sample ID:	P001-SS001-01		SDG No.:	Q1860	
Lab Sample ID:	Q1860-01		Matrix:	SOIL	
Analytical Method:	SW8082A		% Solid:	86.7	Decanted:
Sample Wt/Vol:	30.07	Units: g	Final Vol:	10000	uL
Soil Aliquot Vol:		uL	Test:	PCB	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	SW3541B				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PP071430.D	1	04/24/25 08:32	04/24/25 13:30	PB167721

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	4.50	U	4.50	19.6	ug/kg
11104-28-2	Aroclor-1221	4.60	U	4.60	19.6	ug/kg
11141-16-5	Aroclor-1232	4.30	U	4.30	19.6	ug/kg
53469-21-9	Aroclor-1242	4.60	U	4.60	19.6	ug/kg
12672-29-6	Aroclor-1248	6.80	U	6.80	19.6	ug/kg
11097-69-1	Aroclor-1254	3.70	U	3.70	19.6	ug/kg
37324-23-5	Aroclor-1262	5.80	U	5.80	19.6	ug/kg
11100-14-4	Aroclor-1268	4.10	U	4.10	19.6	ug/kg
11096-82-5	Aroclor-1260	58.6		3.70	19.6	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	18.6		32 - 144	93%	SPK: 20
2051-24-3	Decachlorobiphenyl	16.5		32 - 175	82%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	Weston Solutions, Inc.		Date Collected:	04/16/25	
Project:	RFP 911		Date Received:	04/23/25	
Client Sample ID:	P001-SS002-01		SDG No.:	Q1860	
Lab Sample ID:	Q1860-04		Matrix:	SOIL	
Analytical Method:	SW8082A		% Solid:	81.8	Decanted:
Sample Wt/Vol:	30.02	Units: g	Final Vol:	10000	uL
Soil Aliquot Vol:		uL	Test:	PCB	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	SW3541B				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PP071433.D	1	04/24/25 08:32	04/24/25 14:19	PB167721

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	4.80	U	4.80	20.8	ug/kg
11104-28-2	Aroclor-1221	4.90	U	4.90	20.8	ug/kg
11141-16-5	Aroclor-1232	4.50	U	4.50	20.8	ug/kg
53469-21-9	Aroclor-1242	4.90	U	4.90	20.8	ug/kg
12672-29-6	Aroclor-1248	7.20	U	7.20	20.8	ug/kg
11097-69-1	Aroclor-1254	200		3.90	20.8	ug/kg
37324-23-5	Aroclor-1262	6.10	U	6.10	20.8	ug/kg
11100-14-4	Aroclor-1268	4.40	U	4.40	20.8	ug/kg
11096-82-5	Aroclor-1260	3.90	U	3.90	20.8	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	15.9		32 - 144	80%	SPK: 20
2051-24-3	Decachlorobiphenyl	31.6		32 - 175	158%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	Weston Solutions, Inc.		Date Collected:	04/16/25	
Project:	RFP 911		Date Received:	04/23/25	
Client Sample ID:	P001-SS002-02		SDG No.:	Q1860	
Lab Sample ID:	Q1860-05		Matrix:	SOIL	
Analytical Method:	SW8082A		% Solid:	84.1	Decanted:
Sample Wt/Vol:	30.01	Units: g	Final Vol:	10000	uL
Soil Aliquot Vol:		uL	Test:	PCB	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	SW3541B				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PO110678.D	1	04/24/25 08:32	04/24/25 13:01	PB167721

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	4.70	U	4.70	20.2	ug/kg
11104-28-2	Aroclor-1221	4.80	U	4.80	20.2	ug/kg
11141-16-5	Aroclor-1232	4.40	U	4.40	20.2	ug/kg
53469-21-9	Aroclor-1242	4.80	U	4.80	20.2	ug/kg
12672-29-6	Aroclor-1248	7.00	U	7.00	20.2	ug/kg
11097-69-1	Aroclor-1254	302		3.80	20.2	ug/kg
37324-23-5	Aroclor-1262	6.00	U	6.00	20.2	ug/kg
11100-14-4	Aroclor-1268	4.30	U	4.30	20.2	ug/kg
11096-82-5	Aroclor-1260	3.80	U	3.80	20.2	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	18.2		32 - 144	91%	SPK: 20
2051-24-3	Decachlorobiphenyl	21.8		32 - 175	109%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	Weston Solutions, Inc.		Date Collected:	04/16/25	
Project:	RFP 911		Date Received:	04/23/25	
Client Sample ID:	P001-SS003-01		SDG No.:	Q1860	
Lab Sample ID:	Q1860-06		Matrix:	SOIL	
Analytical Method:	SW8082A		% Solid:	87.6	Decanted:
Sample Wt/Vol:	30.06	Units: g	Final Vol:	10000	uL
Soil Aliquot Vol:		uL	Test:	PCB	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	SW3541B				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PO110679.D	1	04/24/25 08:32	04/24/25 13:19	PB167721

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	4.50	U	4.50	19.4	ug/kg
11104-28-2	Aroclor-1221	4.60	U	4.60	19.4	ug/kg
11141-16-5	Aroclor-1232	4.20	U	4.20	19.4	ug/kg
53469-21-9	Aroclor-1242	4.60	U	4.60	19.4	ug/kg
12672-29-6	Aroclor-1248	6.70	U	6.70	19.4	ug/kg
11097-69-1	Aroclor-1254	154		3.70	19.4	ug/kg
37324-23-5	Aroclor-1262	5.70	U	5.70	19.4	ug/kg
11100-14-4	Aroclor-1268	4.10	U	4.10	19.4	ug/kg
11096-82-5	Aroclor-1260	126		3.70	19.4	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	20.4		32 - 144	102%	SPK: 20
2051-24-3	Decachlorobiphenyl	15.0		32 - 175	75%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	Weston Solutions, Inc.		Date Collected:	04/16/25	
Project:	RFP 911		Date Received:	04/23/25	
Client Sample ID:	P001-SS004-01		SDG No.:	Q1860	
Lab Sample ID:	Q1860-07		Matrix:	SOIL	
Analytical Method:	SW8082A		% Solid:	86.5	Decanted:
Sample Wt/Vol:	30.03	Units: g	Final Vol:	10000	uL
Soil Aliquot Vol:		uL	Test:	PCB	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	SW3541B				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PO110680.D	1	04/24/25 08:32	04/24/25 13:37	PB167721

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	4.60	U	4.60	19.6	ug/kg
11104-28-2	Aroclor-1221	4.70	U	4.70	19.6	ug/kg
11141-16-5	Aroclor-1232	4.30	U	4.30	19.6	ug/kg
53469-21-9	Aroclor-1242	4.60	U	4.60	19.6	ug/kg
12672-29-6	Aroclor-1248	6.80	U	6.80	19.6	ug/kg
11097-69-1	Aroclor-1254	179		3.70	19.6	ug/kg
37324-23-5	Aroclor-1262	5.80	U	5.80	19.6	ug/kg
11100-14-4	Aroclor-1268	4.20	U	4.20	19.6	ug/kg
11096-82-5	Aroclor-1260	198		3.70	19.6	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	17.3		32 - 144	86%	SPK: 20
2051-24-3	Decachlorobiphenyl	12.6		32 - 175	63%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	Weston Solutions, Inc.		Date Collected:	04/16/25	
Project:	RFP 911		Date Received:	04/23/25	
Client Sample ID:	P001-SS005-01		SDG No.:	Q1860	
Lab Sample ID:	Q1860-08		Matrix:	SOIL	
Analytical Method:	SW8082A		% Solid:	86.2	Decanted:
Sample Wt/Vol:	30.07	Units: g	Final Vol:	10000	uL
Soil Aliquot Vol:		uL	Test:	PCB	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	SW3541B				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PP071434.D	1	04/24/25 08:32	04/24/25 14:35	PB167721

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	4.60	U	4.60	19.7	ug/kg
11104-28-2	Aroclor-1221	4.70	U	4.70	19.7	ug/kg
11141-16-5	Aroclor-1232	4.30	U	4.30	19.7	ug/kg
53469-21-9	Aroclor-1242	4.60	U	4.60	19.7	ug/kg
12672-29-6	Aroclor-1248	6.90	U	6.90	19.7	ug/kg
11097-69-1	Aroclor-1254	3.70	U	3.70	19.7	ug/kg
37324-23-5	Aroclor-1262	5.80	U	5.80	19.7	ug/kg
11100-14-4	Aroclor-1268	4.20	U	4.20	19.7	ug/kg
11096-82-5	Aroclor-1260	299		3.70	19.7	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	17.8		32 - 144	89%	SPK: 20
2051-24-3	Decachlorobiphenyl	15.2		32 - 175	76%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	Weston Solutions, Inc.		Date Collected:	04/16/25	
Project:	RFP 911		Date Received:	04/23/25	
Client Sample ID:	P001-SS006-01		SDG No.:	Q1860	
Lab Sample ID:	Q1860-09		Matrix:	SOIL	
Analytical Method:	SW8082A		% Solid:	89.5	Decanted:
Sample Wt/Vol:	30.08	Units: g	Final Vol:	10000	uL
Soil Aliquot Vol:		uL	Test:	PCB	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	SW3541B				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PP071645.D	1	04/24/25 08:32	04/30/25 13:44	PB167721

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	4.40	U	4.40	18.9	ug/kg
11104-28-2	Aroclor-1221	4.50	U	4.50	18.9	ug/kg
11141-16-5	Aroclor-1232	4.10	U	4.10	18.9	ug/kg
53469-21-9	Aroclor-1242	4.50	U	4.50	18.9	ug/kg
12672-29-6	Aroclor-1248	6.60	U	6.60	18.9	ug/kg
11097-69-1	Aroclor-1254	3.60	U	3.60	18.9	ug/kg
37324-23-5	Aroclor-1262	5.60	U	5.60	18.9	ug/kg
11100-14-4	Aroclor-1268	4.00	U	4.00	18.9	ug/kg
11096-82-5	Aroclor-1260	439	E	3.60	18.9	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	21.8		32 - 144	109%	SPK: 20
2051-24-3	Decachlorobiphenyl	20.4		32 - 175	102%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	Weston Solutions, Inc.		Date Collected:	04/16/25	
Project:	RFP 911		Date Received:	04/23/25	
Client Sample ID:	P001-SS006-01DL		SDG No.:	Q1860	
Lab Sample ID:	Q1860-09DL		Matrix:	SOIL	
Analytical Method:	SW8082A		% Solid:	89.5	Decanted:
Sample Wt/Vol:	30.08	Units: g	Final Vol:	10000	uL
Soil Aliquot Vol:		uL	Test:	PCB	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	SW3541B				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PP071646.D	5	04/24/25 08:32	04/30/25 14:01	PB167721

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	22.0	UD	22.0	94.7	ug/kg
11104-28-2	Aroclor-1221	22.5	UD	22.5	94.7	ug/kg
11141-16-5	Aroclor-1232	20.7	UD	20.7	94.7	ug/kg
53469-21-9	Aroclor-1242	22.3	UD	22.3	94.7	ug/kg
12672-29-6	Aroclor-1248	33.0	UD	33.0	94.7	ug/kg
11097-69-1	Aroclor-1254	17.9	UD	17.9	94.7	ug/kg
37324-23-5	Aroclor-1262	28.0	UD	28.0	94.7	ug/kg
11100-14-4	Aroclor-1268	20.1	UD	20.1	94.7	ug/kg
11096-82-5	Aroclor-1260	375	D	18.0	94.7	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	18.6		32 - 144	93%	SPK: 20
2051-24-3	Decachlorobiphenyl	16.4		32 - 175	82%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	Weston Solutions, Inc.		Date Collected:	04/16/25	
Project:	RFP 911		Date Received:	04/23/25	
Client Sample ID:	P001-SS007-01		SDG No.:	Q1860	
Lab Sample ID:	Q1860-10		Matrix:	SOIL	
Analytical Method:	SW8082A		% Solid:	87.4	Decanted:
Sample Wt/Vol:	30.06	Units: g	Final Vol:	10000	uL
Soil Aliquot Vol:		uL	Test:	PCB	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	SW3541B				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PP071436.D	1	04/24/25 08:32	04/24/25 15:08	PB167721

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	4.50	U	4.50	19.4	ug/kg
11104-28-2	Aroclor-1221	4.60	U	4.60	19.4	ug/kg
11141-16-5	Aroclor-1232	4.20	U	4.20	19.4	ug/kg
53469-21-9	Aroclor-1242	4.60	U	4.60	19.4	ug/kg
12672-29-6	Aroclor-1248	4200	E	6.80	19.4	ug/kg
11097-69-1	Aroclor-1254	3900	E	3.70	19.4	ug/kg
37324-23-5	Aroclor-1262	5.70	U	5.70	19.4	ug/kg
11100-14-4	Aroclor-1268	4.10	U	4.10	19.4	ug/kg
11096-82-5	Aroclor-1260	3.70	U	3.70	19.4	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	27.3		32 - 144	136%	SPK: 20
2051-24-3	Decachlorobiphenyl	42.9	*	32 - 175	215%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	Weston Solutions, Inc.		Date Collected:	04/16/25	
Project:	RFP 911		Date Received:	04/23/25	
Client Sample ID:	P001-SS007-01DL		SDG No.:	Q1860	
Lab Sample ID:	Q1860-10DL		Matrix:	SOIL	
Analytical Method:	SW8082A		% Solid:	87.4	Decanted:
Sample Wt/Vol:	30.06	Units: g	Final Vol:	10000	uL
Soil Aliquot Vol:		uL	Test:	PCB	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	SW3541B				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PP071447.D	25	04/24/25 08:32	04/24/25 19:45	PB167721

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	113	UD	113	485	ug/kg
11104-28-2	Aroclor-1221	115	UD	115	485	ug/kg
11141-16-5	Aroclor-1232	106	UD	106	485	ug/kg
53469-21-9	Aroclor-1242	114	UD	114	485	ug/kg
12672-29-6	Aroclor-1248	4000	D	169	485	ug/kg
11097-69-1	Aroclor-1254	3400	D	91.6	485	ug/kg
37324-23-5	Aroclor-1262	143	UD	143	485	ug/kg
11100-14-4	Aroclor-1268	103	UD	103	485	ug/kg
11096-82-5	Aroclor-1260	92.2	UD	92.2	485	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	21.8		32 - 144	109%	SPK: 20
2051-24-3	Decachlorobiphenyl	35.5	*	32 - 175	178%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	Weston Solutions, Inc.		Date Collected:	04/16/25	
Project:	RFP 911		Date Received:	04/23/25	
Client Sample ID:	P001-SS008-01		SDG No.:	Q1860	
Lab Sample ID:	Q1860-11		Matrix:	SOIL	
Analytical Method:	SW8082A		% Solid:	88.8	Decanted:
Sample Wt/Vol:	30.03	Units: g	Final Vol:	10000	uL
Soil Aliquot Vol:		uL	Test:	PCB	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	SW3541B				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PP071437.D	1	04/24/25 08:32	04/24/25 15:25	PB167721

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	4.40	U	4.40	19.1	ug/kg
11104-28-2	Aroclor-1221	4.50	U	4.50	19.1	ug/kg
11141-16-5	Aroclor-1232	4.20	U	4.20	19.1	ug/kg
53469-21-9	Aroclor-1242	4.50	U	4.50	19.1	ug/kg
12672-29-6	Aroclor-1248	7200	E	6.70	19.1	ug/kg
11097-69-1	Aroclor-1254	7000	E	3.60	19.1	ug/kg
37324-23-5	Aroclor-1262	5.60	U	5.60	19.1	ug/kg
11100-14-4	Aroclor-1268	4.10	U	4.10	19.1	ug/kg
11096-82-5	Aroclor-1260	3.60	U	3.60	19.1	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	36.9	*	32 - 144	184%	SPK: 20
2051-24-3	Decachlorobiphenyl	81.3	*	32 - 175	407%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	Weston Solutions, Inc.		Date Collected:	04/16/25	
Project:	RFP 911		Date Received:	04/23/25	
Client Sample ID:	P001-SS008-01DL		SDG No.:	Q1860	
Lab Sample ID:	Q1860-11DL		Matrix:	SOIL	
Analytical Method:	SW8082A		% Solid:	88.8	Decanted:
Sample Wt/Vol:	30.03	Units: g	Final Vol:	10000	uL
Soil Aliquot Vol:		uL	Test:	PCB	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	SW3541B				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PP071449.D	100	04/24/25 08:32	04/24/25 20:18	PB167721

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	444	UD	444	1900	ug/kg
11104-28-2	Aroclor-1221	453	UD	453	1900	ug/kg
11141-16-5	Aroclor-1232	419	UD	419	1900	ug/kg
53469-21-9	Aroclor-1242	451	UD	451	1900	ug/kg
12672-29-6	Aroclor-1248	9900	D	666	1900	ug/kg
11097-69-1	Aroclor-1254	7900	D	361	1900	ug/kg
37324-23-5	Aroclor-1262	565	UD	565	1900	ug/kg
11100-14-4	Aroclor-1268	405	UD	405	1900	ug/kg
11096-82-5	Aroclor-1260	363	UD	363	1900	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	0	*	32 - 144	0%	SPK: 20
2051-24-3	Decachlorobiphenyl	0	*	32 - 175	0%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	Weston Solutions, Inc.		Date Collected:	04/17/25	
Project:	RFP 911		Date Received:	04/23/25	
Client Sample ID:	P001-SS009-01		SDG No.:	Q1860	
Lab Sample ID:	Q1860-12		Matrix:	SOIL	
Analytical Method:	SW8082A		% Solid:	86.5	Decanted:
Sample Wt/Vol:	30.01	Units: g	Final Vol:	10000	uL
Soil Aliquot Vol:		uL	Test:	PCB	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	SW3541B				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PO110681.D	1	04/24/25 08:32	04/24/25 13:56	PB167721

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	4.60	U	4.60	19.6	ug/kg
11104-28-2	Aroclor-1221	4.70	U	4.70	19.6	ug/kg
11141-16-5	Aroclor-1232	4.30	U	4.30	19.6	ug/kg
53469-21-9	Aroclor-1242	4.60	U	4.60	19.6	ug/kg
12672-29-6	Aroclor-1248	7300	E	6.80	19.6	ug/kg
11097-69-1	Aroclor-1254	5700	E	3.70	19.6	ug/kg
37324-23-5	Aroclor-1262	5.80	U	5.80	19.6	ug/kg
11100-14-4	Aroclor-1268	4.20	U	4.20	19.6	ug/kg
11096-82-5	Aroclor-1260	3.70	U	3.70	19.6	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	23.1		32 - 144	116%	SPK: 20
2051-24-3	Decachlorobiphenyl	49.7	*	32 - 175	248%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	Weston Solutions, Inc.		Date Collected:	04/17/25	
Project:	RFP 911		Date Received:	04/23/25	
Client Sample ID:	P001-SS009-01DL		SDG No.:	Q1860	
Lab Sample ID:	Q1860-12DL		Matrix:	SOIL	
Analytical Method:	SW8082A		% Solid:	86.5	Decanted:
Sample Wt/Vol:	30.01	Units: g	Final Vol:	10000	uL
Soil Aliquot Vol:		uL	Test:	PCB	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	SW3541B				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PO110709.D	100	04/24/25 08:32	04/25/25 02:02	PB167721

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	456	UD	456	2000	ug/kg
11104-28-2	Aroclor-1221	466	UD	466	2000	ug/kg
11141-16-5	Aroclor-1232	430	UD	430	2000	ug/kg
53469-21-9	Aroclor-1242	463	UD	463	2000	ug/kg
12672-29-6	Aroclor-1248	9400	D	684	2000	ug/kg
11097-69-1	Aroclor-1254	6700	D	371	2000	ug/kg
37324-23-5	Aroclor-1262	580	UD	580	2000	ug/kg
11100-14-4	Aroclor-1268	416	UD	416	2000	ug/kg
11096-82-5	Aroclor-1260	373	UD	373	2000	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	0	*	32 - 144	0%	SPK: 20
2051-24-3	Decachlorobiphenyl	0	*	32 - 175	0%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	Weston Solutions, Inc.		Date Collected:	04/17/25	
Project:	RFP 911		Date Received:	04/23/25	
Client Sample ID:	P001-SS010-01		SDG No.:	Q1860	
Lab Sample ID:	Q1860-13		Matrix:	SOIL	
Analytical Method:	SW8082A		% Solid:	80.7	Decanted:
Sample Wt/Vol:	30.04	Units: g	Final Vol:	10000	uL
Soil Aliquot Vol:		uL	Test:	PCB	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	SW3541B				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PO110708.D	1	04/24/25 08:32	04/25/25 01:43	PB167721

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	4.90	U	4.90	21.0	ug/kg
11104-28-2	Aroclor-1221	5.00	U	5.00	21.0	ug/kg
11141-16-5	Aroclor-1232	4.60	U	4.60	21.0	ug/kg
53469-21-9	Aroclor-1242	5.00	U	5.00	21.0	ug/kg
12672-29-6	Aroclor-1248	7.30	U	7.30	21.0	ug/kg
11097-69-1	Aroclor-1254	57.4		4.00	21.0	ug/kg
37324-23-5	Aroclor-1262	6.20	U	6.20	21.0	ug/kg
11100-14-4	Aroclor-1268	4.50	U	4.50	21.0	ug/kg
11096-82-5	Aroclor-1260	54.6		4.00	21.0	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	16.1		32 - 144	81%	SPK: 20
2051-24-3	Decachlorobiphenyl	11.4		32 - 175	57%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	Weston Solutions, Inc.		Date Collected:	04/17/25	
Project:	RFP 911		Date Received:	04/23/25	
Client Sample ID:	P001-SS011-01		SDG No.:	Q1860	
Lab Sample ID:	Q1860-14		Matrix:	SOIL	
Analytical Method:	SW8082A		% Solid:	82.8	Decanted:
Sample Wt/Vol:	30.06	Units: g	Final Vol:	10000	uL
Soil Aliquot Vol:		uL	Test:	PCB	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	SW3541B				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PO110683.D	1	04/24/25 08:32	04/24/25 14:32	PB167721

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	4.80	U	4.80	20.5	ug/kg
11104-28-2	Aroclor-1221	4.90	U	4.90	20.5	ug/kg
11141-16-5	Aroclor-1232	4.50	U	4.50	20.5	ug/kg
53469-21-9	Aroclor-1242	4.80	U	4.80	20.5	ug/kg
12672-29-6	Aroclor-1248	325		7.10	20.5	ug/kg
11097-69-1	Aroclor-1254	282		3.90	20.5	ug/kg
37324-23-5	Aroclor-1262	6.10	U	6.10	20.5	ug/kg
11100-14-4	Aroclor-1268	4.30	U	4.30	20.5	ug/kg
11096-82-5	Aroclor-1260	3.90	U	3.90	20.5	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	16.5		32 - 144	83%	SPK: 20
2051-24-3	Decachlorobiphenyl	14.8		32 - 175	74%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	Weston Solutions, Inc.		Date Collected:	04/17/25	
Project:	RFP 911		Date Received:	04/23/25	
Client Sample ID:	P001-SS012-01		SDG No.:	Q1860	
Lab Sample ID:	Q1860-15		Matrix:	SOIL	
Analytical Method:	SW8082A		% Solid:	81.3	Decanted:
Sample Wt/Vol:	30.03	Units: g	Final Vol:	10000	uL
Soil Aliquot Vol:		uL	Test:	PCB	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	SW3541B				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PO110684.D	1	04/24/25 08:32	04/24/25 14:51	PB167721

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	4.90	U	4.90	20.9	ug/kg
11104-28-2	Aroclor-1221	5.00	U	5.00	20.9	ug/kg
11141-16-5	Aroclor-1232	4.60	U	4.60	20.9	ug/kg
53469-21-9	Aroclor-1242	4.90	U	4.90	20.9	ug/kg
12672-29-6	Aroclor-1248	6500	E	7.30	20.9	ug/kg
11097-69-1	Aroclor-1254	3900	E	3.90	20.9	ug/kg
37324-23-5	Aroclor-1262	6.20	U	6.20	20.9	ug/kg
11100-14-4	Aroclor-1268	4.40	U	4.40	20.9	ug/kg
11096-82-5	Aroclor-1260	4.00	U	4.00	20.9	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	14.8		32 - 144	74%	SPK: 20
2051-24-3	Decachlorobiphenyl	43.8	*	32 - 175	219%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	Weston Solutions, Inc.		Date Collected:	04/17/25	
Project:	RFP 911		Date Received:	04/23/25	
Client Sample ID:	P001-SS012-01DL		SDG No.:	Q1860	
Lab Sample ID:	Q1860-15DL		Matrix:	SOIL	
Analytical Method:	SW8082A		% Solid:	81.3	Decanted:
Sample Wt/Vol:	30.03	Units: g	Final Vol:	10000	uL
Soil Aliquot Vol:		uL	Test:	PCB	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	SW3541B				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PO110710.D	25	04/24/25 08:32	04/25/25 02:20	PB167721

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	121	UD	121	522	ug/kg
11104-28-2	Aroclor-1221	124	UD	124	522	ug/kg
11141-16-5	Aroclor-1232	114	UD	114	522	ug/kg
53469-21-9	Aroclor-1242	123	UD	123	522	ug/kg
12672-29-6	Aroclor-1248	7000	D	182	522	ug/kg
11097-69-1	Aroclor-1254	3900	D	98.6	522	ug/kg
37324-23-5	Aroclor-1262	154	UD	154	522	ug/kg
11100-14-4	Aroclor-1268	111	UD	111	522	ug/kg
11096-82-5	Aroclor-1260	99.2	UD	99.2	522	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	15.3		32 - 144	76%	SPK: 20
2051-24-3	Decachlorobiphenyl	47.5	*	32 - 175	238%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	Weston Solutions, Inc.		Date Collected:	04/17/25	
Project:	RFP 911		Date Received:	04/23/25	
Client Sample ID:	P001-SS013-01		SDG No.:	Q1860	
Lab Sample ID:	Q1860-16		Matrix:	SOIL	
Analytical Method:	SW8082A		% Solid:	85.2	Decanted:
Sample Wt/Vol:	30.05	Units: g	Final Vol:	10000	uL
Soil Aliquot Vol:		uL	Test:	PCB	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	SW3541B				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PO110685.D	1	04/24/25 08:32	04/24/25 15:09	PB167721

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	4.60	U	4.60	19.9	ug/kg
11104-28-2	Aroclor-1221	4.70	U	4.70	19.9	ug/kg
11141-16-5	Aroclor-1232	4.40	U	4.40	19.9	ug/kg
53469-21-9	Aroclor-1242	4.70	U	4.70	19.9	ug/kg
12672-29-6	Aroclor-1248	4300	E	6.90	19.9	ug/kg
11097-69-1	Aroclor-1254	3200	E	3.80	19.9	ug/kg
37324-23-5	Aroclor-1262	5.90	U	5.90	19.9	ug/kg
11100-14-4	Aroclor-1268	4.20	U	4.20	19.9	ug/kg
11096-82-5	Aroclor-1260	3.80	U	3.80	19.9	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	18.3		32 - 144	92%	SPK: 20
2051-24-3	Decachlorobiphenyl	27.1		32 - 175	136%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	Weston Solutions, Inc.		Date Collected:	04/17/25	
Project:	RFP 911		Date Received:	04/23/25	
Client Sample ID:	P001-SS013-01DL		SDG No.:	Q1860	
Lab Sample ID:	Q1860-16DL		Matrix:	SOIL	
Analytical Method:	SW8082A		% Solid:	85.2	Decanted:
Sample Wt/Vol:	30.05	Units: g	Final Vol:	10000	uL
Soil Aliquot Vol:		uL	Test:	PCB	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	SW3541B				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PO110712.D	25	04/24/25 08:32	04/25/25 02:57	PB167721

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	116	UD	116	498	ug/kg
11104-28-2	Aroclor-1221	118	UD	118	498	ug/kg
11141-16-5	Aroclor-1232	109	UD	109	498	ug/kg
53469-21-9	Aroclor-1242	117	UD	117	498	ug/kg
12672-29-6	Aroclor-1248	5100	D	173	498	ug/kg
11097-69-1	Aroclor-1254	3500	D	94.0	498	ug/kg
37324-23-5	Aroclor-1262	147	UD	147	498	ug/kg
11100-14-4	Aroclor-1268	105	UD	105	498	ug/kg
11096-82-5	Aroclor-1260	94.6	UD	94.6	498	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	20.8		32 - 144	104%	SPK: 20
2051-24-3	Decachlorobiphenyl	38.8	*	32 - 175	194%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	Weston Solutions, Inc.		Date Collected:	04/17/25	
Project:	RFP 911		Date Received:	04/23/25	
Client Sample ID:	P001-SS014-01		SDG No.:	Q1860	
Lab Sample ID:	Q1860-17		Matrix:	SOIL	
Analytical Method:	SW8082A		% Solid:	79.8	Decanted:
Sample Wt/Vol:	30.01	Units: g	Final Vol:	10000	uL
Soil Aliquot Vol:		uL	Test:	PCB	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	SW3541B				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PO110686.D	1	04/24/25 08:32	04/24/25 15:27	PB167721

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	4.90	U	4.90	21.3	ug/kg
11104-28-2	Aroclor-1221	5.00	U	5.00	21.3	ug/kg
11141-16-5	Aroclor-1232	4.70	U	4.70	21.3	ug/kg
53469-21-9	Aroclor-1242	5.00	U	5.00	21.3	ug/kg
12672-29-6	Aroclor-1248	398		7.40	21.3	ug/kg
11097-69-1	Aroclor-1254	412		4.00	21.3	ug/kg
37324-23-5	Aroclor-1262	6.30	U	6.30	21.3	ug/kg
11100-14-4	Aroclor-1268	4.50	U	4.50	21.3	ug/kg
11096-82-5	Aroclor-1260	4.00	U	4.00	21.3	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	18.4		32 - 144	92%	SPK: 20
2051-24-3	Decachlorobiphenyl	16.8		32 - 175	84%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	Weston Solutions, Inc.		Date Collected:	04/17/25	
Project:	RFP 911		Date Received:	04/23/25	
Client Sample ID:	P001-SS015-01		SDG No.:	Q1860	
Lab Sample ID:	Q1860-18		Matrix:	SOIL	
Analytical Method:	SW8082A		% Solid:	74.9	Decanted:
Sample Wt/Vol:	30.02	Units: g	Final Vol:	10000	uL
Soil Aliquot Vol:		uL	Test:	PCB	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	SW3541B				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PO110687.D	1	04/24/25 08:32	04/24/25 15:46	PB167721

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	5.30	U	5.30	22.7	ug/kg
11104-28-2	Aroclor-1221	5.40	U	5.40	22.7	ug/kg
11141-16-5	Aroclor-1232	5.00	U	5.00	22.7	ug/kg
53469-21-9	Aroclor-1242	5.40	U	5.40	22.7	ug/kg
12672-29-6	Aroclor-1248	71.7		7.90	22.7	ug/kg
11097-69-1	Aroclor-1254	102		4.30	22.7	ug/kg
37324-23-5	Aroclor-1262	6.70	U	6.70	22.7	ug/kg
11100-14-4	Aroclor-1268	4.80	U	4.80	22.7	ug/kg
11096-82-5	Aroclor-1260	4.30	U	4.30	22.7	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	18.9		32 - 144	95%	SPK: 20
2051-24-3	Decachlorobiphenyl	13.0		32 - 175	65%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	Weston Solutions, Inc.		Date Collected:	04/17/25	
Project:	RFP 911		Date Received:	04/23/25	
Client Sample ID:	P001-SS016-01		SDG No.:	Q1860	
Lab Sample ID:	Q1860-19		Matrix:	SOIL	
Analytical Method:	SW8082A		% Solid:	83.1	Decanted:
Sample Wt/Vol:	30.05	Units: g	Final Vol:	10000	uL
Soil Aliquot Vol:		uL	Test:	PCB	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	SW3541B				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PO110727.D	1	04/25/25 08:20	04/25/25 12:31	PB167737

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	4.70	U	4.70	20.4	ug/kg
11104-28-2	Aroclor-1221	4.80	U	4.80	20.4	ug/kg
11141-16-5	Aroclor-1232	4.50	U	4.50	20.4	ug/kg
53469-21-9	Aroclor-1242	4.80	U	4.80	20.4	ug/kg
12672-29-6	Aroclor-1248	932	E	7.10	20.4	ug/kg
11097-69-1	Aroclor-1254	734	E	3.90	20.4	ug/kg
37324-23-5	Aroclor-1262	6.00	U	6.00	20.4	ug/kg
11100-14-4	Aroclor-1268	4.30	U	4.30	20.4	ug/kg
11096-82-5	Aroclor-1260	3.90	U	3.90	20.4	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	16.5		32 - 144	83%	SPK: 20
2051-24-3	Decachlorobiphenyl	17.1		32 - 175	86%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	Weston Solutions, Inc.		Date Collected:	04/17/25	
Project:	RFP 911		Date Received:	04/23/25	
Client Sample ID:	P001-SS016-01DL		SDG No.:	Q1860	
Lab Sample ID:	Q1860-19DL		Matrix:	SOIL	
Analytical Method:	SW8082A		% Solid:	83.1	Decanted:
Sample Wt/Vol:	30.05	Units: g	Final Vol:	10000	uL
Soil Aliquot Vol:		uL	Test:	PCB	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	SW3541B				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PO110788.D	10	04/25/25 08:20	04/28/25 10:39	PB167737

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	47.5	UD	47.5	204	ug/kg
11104-28-2	Aroclor-1221	48.4	UD	48.4	204	ug/kg
11141-16-5	Aroclor-1232	44.7	UD	44.7	204	ug/kg
53469-21-9	Aroclor-1242	48.2	UD	48.2	204	ug/kg
12672-29-6	Aroclor-1248	986	D	71.1	204	ug/kg
11097-69-1	Aroclor-1254	725	D	38.6	204	ug/kg
37324-23-5	Aroclor-1262	60.3	UD	60.3	204	ug/kg
11100-14-4	Aroclor-1268	43.2	UD	43.2	204	ug/kg
11096-82-5	Aroclor-1260	38.8	UD	38.8	204	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	15.4		32 - 144	77%	SPK: 20
2051-24-3	Decachlorobiphenyl	16.0		32 - 175	80%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	Weston Solutions, Inc.		Date Collected:	04/17/25	
Project:	RFP 911		Date Received:	04/23/25	
Client Sample ID:	P001-SS017-01		SDG No.:	Q1860	
Lab Sample ID:	Q1860-22		Matrix:	SOIL	
Analytical Method:	SW8082A		% Solid:	84.5	Decanted:
Sample Wt/Vol:	30.08	Units: g	Final Vol:	10000	uL
Soil Aliquot Vol:		uL	Test:	PCB	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	SW3541B				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PO110821.D	1	04/24/25 08:32	04/28/25 22:42	PB167721

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	4.70	U	4.70	20.1	ug/kg
11104-28-2	Aroclor-1221	4.80	U	4.80	20.1	ug/kg
11141-16-5	Aroclor-1232	4.40	U	4.40	20.1	ug/kg
53469-21-9	Aroclor-1242	4.70	U	4.70	20.1	ug/kg
12672-29-6	Aroclor-1248	770	E	7.00	20.1	ug/kg
11097-69-1	Aroclor-1254	759	E	3.80	20.1	ug/kg
37324-23-5	Aroclor-1262	5.90	U	5.90	20.1	ug/kg
11100-14-4	Aroclor-1268	4.20	U	4.20	20.1	ug/kg
11096-82-5	Aroclor-1260	3.80	U	3.80	20.1	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	17.3		32 - 144	86%	SPK: 20
2051-24-3	Decachlorobiphenyl	18.4		32 - 175	92%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	Weston Solutions, Inc.		Date Collected:	04/17/25	
Project:	RFP 911		Date Received:	04/23/25	
Client Sample ID:	P001-SS017-01DL		SDG No.:	Q1860	
Lab Sample ID:	Q1860-22DL		Matrix:	SOIL	
Analytical Method:	SW8082A		% Solid:	84.5	Decanted:
Sample Wt/Vol:	30.08	Units: g	Final Vol:	10000	uL
Soil Aliquot Vol:		uL	Test:	PCB	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	SW3541B				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PO110826.D	5	04/24/25 08:32	04/29/25 00:12	PB167721

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	23.3	UD	23.3	100	ug/kg
11104-28-2	Aroclor-1221	23.8	UD	23.8	100	ug/kg
11141-16-5	Aroclor-1232	22.0	UD	22.0	100	ug/kg
53469-21-9	Aroclor-1242	23.7	UD	23.7	100	ug/kg
12672-29-6	Aroclor-1248	749	D	34.9	100	ug/kg
11097-69-1	Aroclor-1254	663	D	18.9	100	ug/kg
37324-23-5	Aroclor-1262	29.6	UD	29.6	100	ug/kg
11100-14-4	Aroclor-1268	21.2	UD	21.2	100	ug/kg
11096-82-5	Aroclor-1260	19.1	UD	19.1	100	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	15.6		32 - 144	78%	SPK: 20
2051-24-3	Decachlorobiphenyl	15.4		32 - 175	77%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	Weston Solutions, Inc.		Date Collected:	04/17/25	
Project:	RFP 911		Date Received:	04/23/25	
Client Sample ID:	P001-SS017-02		SDG No.:	Q1860	
Lab Sample ID:	Q1860-23		Matrix:	SOIL	
Analytical Method:	SW8082A		% Solid:	82.6	Decanted:
Sample Wt/Vol:	30.04	Units: g	Final Vol:	10000	uL
Soil Aliquot Vol:		uL	Test:	PCB	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	SW3541B				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PO110822.D	1	04/24/25 08:32	04/28/25 23:00	PB167721

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	4.80	U	4.80	20.6	ug/kg
11104-28-2	Aroclor-1221	4.90	U	4.90	20.6	ug/kg
11141-16-5	Aroclor-1232	4.50	U	4.50	20.6	ug/kg
53469-21-9	Aroclor-1242	4.80	U	4.80	20.6	ug/kg
12672-29-6	Aroclor-1248	925	E	7.20	20.6	ug/kg
11097-69-1	Aroclor-1254	631	E	3.90	20.6	ug/kg
37324-23-5	Aroclor-1262	6.10	U	6.10	20.6	ug/kg
11100-14-4	Aroclor-1268	4.40	U	4.40	20.6	ug/kg
11096-82-5	Aroclor-1260	3.90	U	3.90	20.6	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	20.2		32 - 144	101%	SPK: 20
2051-24-3	Decachlorobiphenyl	18.5		32 - 175	93%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	Weston Solutions, Inc.		Date Collected:	04/17/25	
Project:	RFP 911		Date Received:	04/23/25	
Client Sample ID:	P001-SS017-02DL		SDG No.:	Q1860	
Lab Sample ID:	Q1860-23DL		Matrix:	SOIL	
Analytical Method:	SW8082A		% Solid:	82.6	Decanted:
Sample Wt/Vol:	30.04	Units: g	Final Vol:	10000	uL
Soil Aliquot Vol:		uL	Test:	PCB	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	SW3541B				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PO110827.D	5	04/24/25 08:32	04/29/25 00:31	PB167721

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	23.9	UD	23.9	103	ug/kg
11104-28-2	Aroclor-1221	24.4	UD	24.4	103	ug/kg
11141-16-5	Aroclor-1232	22.5	UD	22.5	103	ug/kg
53469-21-9	Aroclor-1242	24.2	UD	24.2	103	ug/kg
12672-29-6	Aroclor-1248	878	D	35.8	103	ug/kg
11097-69-1	Aroclor-1254	525	D	19.4	103	ug/kg
37324-23-5	Aroclor-1262	30.3	UD	30.3	103	ug/kg
11100-14-4	Aroclor-1268	21.8	UD	21.8	103	ug/kg
11096-82-5	Aroclor-1260	19.5	UD	19.5	103	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	17.9		32 - 144	90%	SPK: 20
2051-24-3	Decachlorobiphenyl	17.9		32 - 175	90%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	Weston Solutions, Inc.		Date Collected:	04/21/25	
Project:	RFP 911		Date Received:	04/23/25	
Client Sample ID:	P001-SS018-01		SDG No.:	Q1860	
Lab Sample ID:	Q1860-24		Matrix:	SOIL	
Analytical Method:	SW8082A		% Solid:	89.6	Decanted:
Sample Wt/Vol:	30.07	Units: g	Final Vol:	10000	uL
Soil Aliquot Vol:		uL	Test:	PCB	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	SW3541B				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PO110823.D	1	04/24/25 08:32	04/28/25 23:18	PB167721

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	4.40	U	4.40	18.9	ug/kg
11104-28-2	Aroclor-1221	4.50	U	4.50	18.9	ug/kg
11141-16-5	Aroclor-1232	4.10	U	4.10	18.9	ug/kg
53469-21-9	Aroclor-1242	4.50	U	4.50	18.9	ug/kg
12672-29-6	Aroclor-1248	6.60	U	6.60	18.9	ug/kg
11097-69-1	Aroclor-1254	76.3		3.60	18.9	ug/kg
37324-23-5	Aroclor-1262	5.60	U	5.60	18.9	ug/kg
11100-14-4	Aroclor-1268	4.00	U	4.00	18.9	ug/kg
11096-82-5	Aroclor-1260	3.60	U	3.60	18.9	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	20.9		32 - 144	105%	SPK: 20
2051-24-3	Decachlorobiphenyl	16.4		32 - 175	82%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	Weston Solutions, Inc.		Date Collected:	04/21/25	
Project:	RFP 911		Date Received:	04/23/25	
Client Sample ID:	P001-SS019-01		SDG No.:	Q1860	
Lab Sample ID:	Q1860-25		Matrix:	SOIL	
Analytical Method:	SW8082A		% Solid:	88.3	Decanted:
Sample Wt/Vol:	30.06	Units: g	Final Vol:	10000	uL
Soil Aliquot Vol:		uL	Test:	PCB	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	SW3541B				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PO110824.D	1	04/24/25 08:32	04/28/25 23:37	PB167721

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	4.50	U	4.50	19.2	ug/kg
11104-28-2	Aroclor-1221	4.60	U	4.60	19.2	ug/kg
11141-16-5	Aroclor-1232	4.20	U	4.20	19.2	ug/kg
53469-21-9	Aroclor-1242	4.50	U	4.50	19.2	ug/kg
12672-29-6	Aroclor-1248	6.70	U	6.70	19.2	ug/kg
11097-69-1	Aroclor-1254	435	E	3.60	19.2	ug/kg
37324-23-5	Aroclor-1262	5.70	U	5.70	19.2	ug/kg
11100-14-4	Aroclor-1268	4.10	U	4.10	19.2	ug/kg
11096-82-5	Aroclor-1260	297		3.70	19.2	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	21.0		32 - 144	105%	SPK: 20
2051-24-3	Decachlorobiphenyl	18.3		32 - 175	92%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	Weston Solutions, Inc.		Date Collected:	04/21/25	
Project:	RFP 911		Date Received:	04/23/25	
Client Sample ID:	P001-SS019-01DL		SDG No.:	Q1860	
Lab Sample ID:	Q1860-25DL		Matrix:	SOIL	
Analytical Method:	SW8082A		% Solid:	88.3	Decanted:
Sample Wt/Vol:	30.06	Units: g	Final Vol:	10000	uL
Soil Aliquot Vol:		uL	Test:	PCB	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	SW3541B				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PO110837.D	5	04/24/25 08:32	04/29/25 04:57	PB167721

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	22.3	UD	22.3	96.1	ug/kg
11104-28-2	Aroclor-1221	22.8	UD	22.8	96.1	ug/kg
11141-16-5	Aroclor-1232	21.0	UD	21.0	96.1	ug/kg
53469-21-9	Aroclor-1242	22.7	UD	22.7	96.1	ug/kg
12672-29-6	Aroclor-1248	33.5	UD	33.5	96.1	ug/kg
11097-69-1	Aroclor-1254	435	D	18.1	96.1	ug/kg
37324-23-5	Aroclor-1262	28.4	UD	28.4	96.1	ug/kg
11100-14-4	Aroclor-1268	20.3	UD	20.3	96.1	ug/kg
11096-82-5	Aroclor-1260	298	D	18.3	96.1	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	18.6		32 - 144	93%	SPK: 20
2051-24-3	Decachlorobiphenyl	18.6		32 - 175	93%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	Weston Solutions, Inc.		Date Collected:	04/21/25	
Project:	RFP 911		Date Received:	04/23/25	
Client Sample ID:	P001-SS020-01		SDG No.:	Q1860	
Lab Sample ID:	Q1860-26		Matrix:	SOIL	
Analytical Method:	SW8082A		% Solid:	85	Decanted:
Sample Wt/Vol:	30.02	Units: g	Final Vol:	10000	uL
Soil Aliquot Vol:		uL	Test:	PCB	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	SW3541B				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PO110730.D	1	04/25/25 08:20	04/25/25 13:26	PB167737

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	4.60	U	4.60	20.0	ug/kg
11104-28-2	Aroclor-1221	4.70	U	4.70	20.0	ug/kg
11141-16-5	Aroclor-1232	4.40	U	4.40	20.0	ug/kg
53469-21-9	Aroclor-1242	4.70	U	4.70	20.0	ug/kg
12672-29-6	Aroclor-1248	7.00	U	7.00	20.0	ug/kg
11097-69-1	Aroclor-1254	5900	EP	3.80	20.0	ug/kg
37324-23-5	Aroclor-1262	5.90	U	5.90	20.0	ug/kg
11100-14-4	Aroclor-1268	4.20	U	4.20	20.0	ug/kg
11096-82-5	Aroclor-1260	3.80	U	3.80	20.0	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	17.8		32 - 144	89%	SPK: 20
2051-24-3	Decachlorobiphenyl	23.8		32 - 175	119%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	Weston Solutions, Inc.		Date Collected:	04/21/25	
Project:	RFP 911		Date Received:	04/23/25	
Client Sample ID:	P001-SS020-01DL		SDG No.:	Q1860	
Lab Sample ID:	Q1860-26DL		Matrix:	SOIL	
Analytical Method:	SW8082A		% Solid:	85	Decanted:
Sample Wt/Vol:	30.02	Units: g	Final Vol:	10000	uL
Soil Aliquot Vol:		uL	Test:	PCB	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	SW3541B				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PO110789.D	25	04/25/25 08:20	04/28/25 10:57	PB167737

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	116	UD	116	500	ug/kg
11104-28-2	Aroclor-1221	118	UD	118	500	ug/kg
11141-16-5	Aroclor-1232	109	UD	109	500	ug/kg
53469-21-9	Aroclor-1242	118	UD	118	500	ug/kg
12672-29-6	Aroclor-1248	174	UD	174	500	ug/kg
11097-69-1	Aroclor-1254	4500	D	94.3	500	ug/kg
37324-23-5	Aroclor-1262	148	UD	148	500	ug/kg
11100-14-4	Aroclor-1268	106	UD	106	500	ug/kg
11096-82-5	Aroclor-1260	94.9	UD	94.9	500	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	15.8		32 - 144	79%	SPK: 20
2051-24-3	Decachlorobiphenyl	24.0		32 - 175	120%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	Weston Solutions, Inc.		Date Collected:	04/21/25	
Project:	RFP 911		Date Received:	04/23/25	
Client Sample ID:	P001-SS021-01		SDG No.:	Q1860	
Lab Sample ID:	Q1860-27		Matrix:	SOIL	
Analytical Method:	SW8082A		% Solid:	88.5	Decanted:
Sample Wt/Vol:	30.04	Units: g	Final Vol:	10000	uL
Soil Aliquot Vol:		uL	Test:	PCB	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	SW3541B				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PO110731.D	1	04/25/25 08:20	04/25/25 13:44	PB167737

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	4.50	U	4.50	19.2	ug/kg
11104-28-2	Aroclor-1221	4.50	U	4.50	19.2	ug/kg
11141-16-5	Aroclor-1232	4.20	U	4.20	19.2	ug/kg
53469-21-9	Aroclor-1242	4.50	U	4.50	19.2	ug/kg
12672-29-6	Aroclor-1248	6.70	U	6.70	19.2	ug/kg
11097-69-1	Aroclor-1254	381	E	3.60	19.2	ug/kg
37324-23-5	Aroclor-1262	5.70	U	5.70	19.2	ug/kg
11100-14-4	Aroclor-1268	4.10	U	4.10	19.2	ug/kg
11096-82-5	Aroclor-1260	3.60	U	3.60	19.2	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	19.9		32 - 144	99%	SPK: 20
2051-24-3	Decachlorobiphenyl	38.6	*	32 - 175	193%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	Weston Solutions, Inc.		Date Collected:	04/21/25	
Project:	RFP 911		Date Received:	04/23/25	
Client Sample ID:	P001-SS021-01DL		SDG No.:	Q1860	
Lab Sample ID:	Q1860-27DL		Matrix:	SOIL	
Analytical Method:	SW8082A		% Solid:	88.5	Decanted:
Sample Wt/Vol:	30.04	Units: g	Final Vol:	10000	uL
Soil Aliquot Vol:		uL	Test:	PCB	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	SW3541B				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PO110790.D	5	04/25/25 08:20	04/28/25 11:16	PB167737

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	22.3	UD	22.3	95.9	ug/kg
11104-28-2	Aroclor-1221	22.7	UD	22.7	95.9	ug/kg
11141-16-5	Aroclor-1232	21.0	UD	21.0	95.9	ug/kg
53469-21-9	Aroclor-1242	22.6	UD	22.6	95.9	ug/kg
12672-29-6	Aroclor-1248	33.4	UD	33.4	95.9	ug/kg
11097-69-1	Aroclor-1254	400	D	18.1	95.9	ug/kg
37324-23-5	Aroclor-1262	28.3	UD	28.3	95.9	ug/kg
11100-14-4	Aroclor-1268	20.3	UD	20.3	95.9	ug/kg
11096-82-5	Aroclor-1260	18.2	UD	18.2	95.9	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	18.8		32 - 144	94%	SPK: 20
2051-24-3	Decachlorobiphenyl	38.4	*	32 - 175	192%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	Weston Solutions, Inc.		Date Collected:	04/21/25	
Project:	RFP 911		Date Received:	04/23/25	
Client Sample ID:	P001-SS022-01		SDG No.:	Q1860	
Lab Sample ID:	Q1860-28		Matrix:	SOIL	
Analytical Method:	SW8082A		% Solid:	83.5	Decanted:
Sample Wt/Vol:	30.06	Units: g	Final Vol:	10000	uL
Soil Aliquot Vol:		uL	Test:	PCB	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	SW3541B				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PO110732.D	1	04/25/25 08:20	04/25/25 14:03	PB167737

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	4.70	U	4.70	20.3	ug/kg
11104-28-2	Aroclor-1221	4.80	U	4.80	20.3	ug/kg
11141-16-5	Aroclor-1232	4.40	U	4.40	20.3	ug/kg
53469-21-9	Aroclor-1242	4.80	U	4.80	20.3	ug/kg
12672-29-6	Aroclor-1248	7.10	U	7.10	20.3	ug/kg
11097-69-1	Aroclor-1254	203		3.80	20.3	ug/kg
37324-23-5	Aroclor-1262	6.00	U	6.00	20.3	ug/kg
11100-14-4	Aroclor-1268	4.30	U	4.30	20.3	ug/kg
11096-82-5	Aroclor-1260	3.90	U	3.90	20.3	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	18.4		32 - 144	92%	SPK: 20
2051-24-3	Decachlorobiphenyl	30.9		32 - 175	155%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	Weston Solutions, Inc.		Date Collected:	04/21/25	
Project:	RFP 911		Date Received:	04/23/25	
Client Sample ID:	P001-SS023-01		SDG No.:	Q1860	
Lab Sample ID:	Q1860-29		Matrix:	SOIL	
Analytical Method:	SW8082A		% Solid:	84	Decanted:
Sample Wt/Vol:	30.09	Units: g	Final Vol:	10000	uL
Soil Aliquot Vol:		uL	Test:	PCB	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	SW3541B				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PO110733.D	1	04/25/25 08:20	04/25/25 14:21	PB167737

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	4.70	U	4.70	20.2	ug/kg
11104-28-2	Aroclor-1221	4.80	U	4.80	20.2	ug/kg
11141-16-5	Aroclor-1232	4.40	U	4.40	20.2	ug/kg
53469-21-9	Aroclor-1242	4.80	U	4.80	20.2	ug/kg
12672-29-6	Aroclor-1248	7.00	U	7.00	20.2	ug/kg
11097-69-1	Aroclor-1254	410	E	3.80	20.2	ug/kg
37324-23-5	Aroclor-1262	6.00	U	6.00	20.2	ug/kg
11100-14-4	Aroclor-1268	4.30	U	4.30	20.2	ug/kg
11096-82-5	Aroclor-1260	3.80	U	3.80	20.2	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	17.1		32 - 144	86%	SPK: 20
2051-24-3	Decachlorobiphenyl	30.4		32 - 175	152%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	Weston Solutions, Inc.		Date Collected:	04/21/25	
Project:	RFP 911		Date Received:	04/23/25	
Client Sample ID:	P001-SS023-01DL		SDG No.:	Q1860	
Lab Sample ID:	Q1860-29DL		Matrix:	SOIL	
Analytical Method:	SW8082A		% Solid:	84	Decanted:
Sample Wt/Vol:	30.09	Units: g	Final Vol:	10000	uL
Soil Aliquot Vol:		uL	Test:	PCB	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	SW3541B				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PO110791.D	5	04/25/25 08:20	04/28/25 11:34	PB167737

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	23.4	UD	23.4	101	ug/kg
11104-28-2	Aroclor-1221	23.9	UD	23.9	101	ug/kg
11141-16-5	Aroclor-1232	22.1	UD	22.1	101	ug/kg
53469-21-9	Aroclor-1242	23.8	UD	23.8	101	ug/kg
12672-29-6	Aroclor-1248	35.1	UD	35.1	101	ug/kg
11097-69-1	Aroclor-1254	361	D	19.1	101	ug/kg
37324-23-5	Aroclor-1262	29.8	UD	29.8	101	ug/kg
11100-14-4	Aroclor-1268	21.4	UD	21.4	101	ug/kg
11096-82-5	Aroclor-1260	19.2	UD	19.2	101	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	13.1		32 - 144	66%	SPK: 20
2051-24-3	Decachlorobiphenyl	27.8		32 - 175	139%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	Weston Solutions, Inc.		Date Collected:	04/21/25	
Project:	RFP 911		Date Received:	04/23/25	
Client Sample ID:	P001-SS024-01		SDG No.:	Q1860	
Lab Sample ID:	Q1860-30		Matrix:	SOIL	
Analytical Method:	SW8082A		% Solid:	84.6	Decanted:
Sample Wt/Vol:	30.05	Units: g	Final Vol:	10000	uL
Soil Aliquot Vol:		uL	Test:	PCB	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	SW3541B				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PO110734.D	1	04/25/25 08:20	04/25/25 14:39	PB167737

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	4.70	U	4.70	20.1	ug/kg
11104-28-2	Aroclor-1221	4.80	U	4.80	20.1	ug/kg
11141-16-5	Aroclor-1232	4.40	U	4.40	20.1	ug/kg
53469-21-9	Aroclor-1242	4.70	U	4.70	20.1	ug/kg
12672-29-6	Aroclor-1248	7.00	U	7.00	20.1	ug/kg
11097-69-1	Aroclor-1254	347		3.80	20.1	ug/kg
37324-23-5	Aroclor-1262	5.90	U	5.90	20.1	ug/kg
11100-14-4	Aroclor-1268	4.20	U	4.20	20.1	ug/kg
11096-82-5	Aroclor-1260	3.80	U	3.80	20.1	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	19.9		32 - 144	100%	SPK: 20
2051-24-3	Decachlorobiphenyl	14.6		32 - 175	73%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	Weston Solutions, Inc.		Date Collected:	04/21/25	
Project:	RFP 911		Date Received:	04/23/25	
Client Sample ID:	P001-SS025-01		SDG No.:	Q1860	
Lab Sample ID:	Q1860-31		Matrix:	SOIL	
Analytical Method:	SW8082A		% Solid:	86.9	Decanted:
Sample Wt/Vol:	30.03	Units: g	Final Vol:	10000	uL
Soil Aliquot Vol:		uL	Test:	PCB	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	SW3541B				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PO110740.D	1	04/25/25 08:20	04/25/25 18:08	PB167737

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	4.50	U	4.50	19.5	ug/kg
11104-28-2	Aroclor-1221	4.60	U	4.60	19.5	ug/kg
11141-16-5	Aroclor-1232	4.30	U	4.30	19.5	ug/kg
53469-21-9	Aroclor-1242	4.60	U	4.60	19.5	ug/kg
12672-29-6	Aroclor-1248	6.80	U	6.80	19.5	ug/kg
11097-69-1	Aroclor-1254	1200	E	3.70	19.5	ug/kg
37324-23-5	Aroclor-1262	5.80	U	5.80	19.5	ug/kg
11100-14-4	Aroclor-1268	4.10	U	4.10	19.5	ug/kg
11096-82-5	Aroclor-1260	1200	E	3.70	19.5	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	19.8		32 - 144	99%	SPK: 20
2051-24-3	Decachlorobiphenyl	50.0	*	32 - 175	250%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	Weston Solutions, Inc.		Date Collected:	04/21/25	
Project:	RFP 911		Date Received:	04/23/25	
Client Sample ID:	P001-SS025-01DL		SDG No.:	Q1860	
Lab Sample ID:	Q1860-31DL		Matrix:	SOIL	
Analytical Method:	SW8082A		% Solid:	86.9	Decanted:
Sample Wt/Vol:	30.03	Units: g	Final Vol:	10000	uL
Soil Aliquot Vol:		uL	Test:	PCB	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	SW3541B				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PO110792.D	5	04/25/25 08:20	04/28/25 11:53	PB167737

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	22.7	UD	22.7	97.7	ug/kg
11104-28-2	Aroclor-1221	23.2	UD	23.2	97.7	ug/kg
11141-16-5	Aroclor-1232	21.4	UD	21.4	97.7	ug/kg
53469-21-9	Aroclor-1242	23.0	UD	23.0	97.7	ug/kg
12672-29-6	Aroclor-1248	34.0	UD	34.0	97.7	ug/kg
11097-69-1	Aroclor-1254	1200	D	18.5	97.7	ug/kg
37324-23-5	Aroclor-1262	28.9	UD	28.9	97.7	ug/kg
11100-14-4	Aroclor-1268	20.7	UD	20.7	97.7	ug/kg
11096-82-5	Aroclor-1260	1300	D	18.6	97.7	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	20.4		32 - 144	102%	SPK: 20
2051-24-3	Decachlorobiphenyl	48.0	*	32 - 175	240%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	Weston Solutions, Inc.		Date Collected:	04/21/25	
Project:	RFP 911		Date Received:	04/23/25	
Client Sample ID:	P001-SS026-01		SDG No.:	Q1860	
Lab Sample ID:	Q1860-32		Matrix:	SOIL	
Analytical Method:	SW8082A		% Solid:	79.2	Decanted:
Sample Wt/Vol:	30.01	Units: g	Final Vol:	10000	uL
Soil Aliquot Vol:		uL	Test:	PCB	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	SW3541B				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PO110741.D	1	04/25/25 08:20	04/25/25 18:26	PB167737

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	5.00	U	5.00	21.5	ug/kg
11104-28-2	Aroclor-1221	5.10	U	5.10	21.5	ug/kg
11141-16-5	Aroclor-1232	4.70	U	4.70	21.5	ug/kg
53469-21-9	Aroclor-1242	5.10	U	5.10	21.5	ug/kg
12672-29-6	Aroclor-1248	7.50	U	7.50	21.5	ug/kg
11097-69-1	Aroclor-1254	3800	E	4.10	21.5	ug/kg
37324-23-5	Aroclor-1262	6.30	U	6.30	21.5	ug/kg
11100-14-4	Aroclor-1268	4.50	U	4.50	21.5	ug/kg
11096-82-5	Aroclor-1260	4.10	U	4.10	21.5	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	19.5		32 - 144	97%	SPK: 20
2051-24-3	Decachlorobiphenyl	74.5	*	32 - 175	373%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	Weston Solutions, Inc.		Date Collected:	04/21/25	
Project:	RFP 911		Date Received:	04/23/25	
Client Sample ID:	P001-SS026-01DL		SDG No.:	Q1860	
Lab Sample ID:	Q1860-32DL		Matrix:	SOIL	
Analytical Method:	SW8082A		% Solid:	79.2	Decanted:
Sample Wt/Vol:	30.01	Units: g	Final Vol:	10000	uL
Soil Aliquot Vol:		uL	Test:	PCB	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	SW3541B				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PO110794.D	20	04/25/25 08:20	04/28/25 12:29	PB167737

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	99.7	UD	99.7	429	ug/kg
11104-28-2	Aroclor-1221	102	UD	102	429	ug/kg
11141-16-5	Aroclor-1232	93.9	UD	93.9	429	ug/kg
53469-21-9	Aroclor-1242	101	UD	101	429	ug/kg
12672-29-6	Aroclor-1248	149	UD	149	429	ug/kg
11097-69-1	Aroclor-1254	4800	D	81.0	429	ug/kg
37324-23-5	Aroclor-1262	127	UD	127	429	ug/kg
11100-14-4	Aroclor-1268	90.9	UD	90.9	429	ug/kg
11096-82-5	Aroclor-1260	81.5	UD	81.5	429	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	26.2		32 - 144	131%	SPK: 20
2051-24-3	Decachlorobiphenyl	83.6	*	32 - 175	418%	SPK: 20

Comments:

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 P = Indicates >25% difference for detected concentrations between the two GC columns
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.
 () = Laboratory InHouse Limit

Report of Analysis

Client:	Weston Solutions, Inc.		Date Collected:	04/21/25	
Project:	RFP 911		Date Received:	04/23/25	
Client Sample ID:	P001-SS027-01		SDG No.:	Q1860	
Lab Sample ID:	Q1860-33		Matrix:	SOIL	
Analytical Method:	SW8082A		% Solid:	86.5	Decanted:
Sample Wt/Vol:	30.04	Units: g	Final Vol:	10000	uL
Soil Aliquot Vol:		uL	Test:	PCB	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	SW3541B				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PP071536.D	1	04/25/25 08:20	04/28/25 10:43	PB167737

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	4.60	U	4.60	19.6	ug/kg
11104-28-2	Aroclor-1221	4.70	U	4.70	19.6	ug/kg
11141-16-5	Aroclor-1232	4.30	U	4.30	19.6	ug/kg
53469-21-9	Aroclor-1242	4.60	U	4.60	19.6	ug/kg
12672-29-6	Aroclor-1248	114		6.80	19.6	ug/kg
11097-69-1	Aroclor-1254	225		3.70	19.6	ug/kg
37324-23-5	Aroclor-1262	5.80	U	5.80	19.6	ug/kg
11100-14-4	Aroclor-1268	4.20	U	4.20	19.6	ug/kg
11096-82-5	Aroclor-1260	176		3.70	19.6	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	17.3		32 - 144	87%	SPK: 20
2051-24-3	Decachlorobiphenyl	15.7		32 - 175	78%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	Weston Solutions, Inc.		Date Collected:	04/21/25	
Project:	RFP 911		Date Received:	04/23/25	
Client Sample ID:	P001-SS028-01		SDG No.:	Q1860	
Lab Sample ID:	Q1860-34		Matrix:	SOIL	
Analytical Method:	SW8082A		% Solid:	77.7	Decanted:
Sample Wt/Vol:	30.05	Units: g	Final Vol:	10000	uL
Soil Aliquot Vol:		uL	Test:	PCB	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	SW3541B				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PO110743.D	1	04/25/25 08:20	04/25/25 19:03	PB167737

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	5.10	U	5.10	21.8	ug/kg
11104-28-2	Aroclor-1221	5.20	U	5.20	21.8	ug/kg
11141-16-5	Aroclor-1232	4.80	U	4.80	21.8	ug/kg
53469-21-9	Aroclor-1242	5.20	U	5.20	21.8	ug/kg
12672-29-6	Aroclor-1248	1800	E	7.60	21.8	ug/kg
11097-69-1	Aroclor-1254	3000	E	4.10	21.8	ug/kg
37324-23-5	Aroclor-1262	6.50	U	6.50	21.8	ug/kg
11100-14-4	Aroclor-1268	4.60	U	4.60	21.8	ug/kg
11096-82-5	Aroclor-1260	2100	E	4.20	21.8	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	26.6		32 - 144	133%	SPK: 20
2051-24-3	Decachlorobiphenyl	84.5	*	32 - 175	423%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	Weston Solutions, Inc.		Date Collected:	04/21/25	
Project:	RFP 911		Date Received:	04/23/25	
Client Sample ID:	P001-SS028-01DL		SDG No.:	Q1860	
Lab Sample ID:	Q1860-34DL		Matrix:	SOIL	
Analytical Method:	SW8082A		% Solid:	77.7	Decanted:
Sample Wt/Vol:	30.05	Units: g	Final Vol:	10000	uL
Soil Aliquot Vol:		uL	Test:	PCB	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	SW3541B				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PO110855.D	20	04/25/25 08:20	04/29/25 12:11	PB167737

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	102	UD	102	437	ug/kg
11104-28-2	Aroclor-1221	104	UD	104	437	ug/kg
11141-16-5	Aroclor-1232	95.6	UD	95.6	437	ug/kg
53469-21-9	Aroclor-1242	103	UD	103	437	ug/kg
12672-29-6	Aroclor-1248	2100	D	152	437	ug/kg
11097-69-1	Aroclor-1254	2900	D	82.5	437	ug/kg
37324-23-5	Aroclor-1262	129	UD	129	437	ug/kg
11100-14-4	Aroclor-1268	92.5	UD	92.5	437	ug/kg
11096-82-5	Aroclor-1260	2300	D	83.0	437	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	23.8		32 - 144	119%	SPK: 20
2051-24-3	Decachlorobiphenyl	86.8	*	32 - 175	434%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	Weston Solutions, Inc.		Date Collected:	04/21/25	
Project:	RFP 911		Date Received:	04/23/25	
Client Sample ID:	P001-SS029-01		SDG No.:	Q1860	
Lab Sample ID:	Q1860-35		Matrix:	SOIL	
Analytical Method:	SW8082A		% Solid:	80.6	Decanted:
Sample Wt/Vol:	30.02	Units: g	Final Vol:	10000	uL
Soil Aliquot Vol:		uL	Test:	PCB	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	SW3541B				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PP071537.D	1	04/25/25 08:20	04/28/25 10:59	PB167737

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	4.90	U	4.90	21.1	ug/kg
11104-28-2	Aroclor-1221	5.00	U	5.00	21.1	ug/kg
11141-16-5	Aroclor-1232	4.60	U	4.60	21.1	ug/kg
53469-21-9	Aroclor-1242	5.00	U	5.00	21.1	ug/kg
12672-29-6	Aroclor-1248	317		7.30	21.1	ug/kg
11097-69-1	Aroclor-1254	781	E	4.00	21.1	ug/kg
37324-23-5	Aroclor-1262	6.20	U	6.20	21.1	ug/kg
11100-14-4	Aroclor-1268	4.50	U	4.50	21.1	ug/kg
11096-82-5	Aroclor-1260	532	E	4.00	21.1	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	30.2	*	32 - 144	151%	SPK: 20
2051-24-3	Decachlorobiphenyl	32.8		32 - 175	164%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	Weston Solutions, Inc.		Date Collected:	04/21/25	
Project:	RFP 911		Date Received:	04/23/25	
Client Sample ID:	P001-SS029-01DL		SDG No.:	Q1860	
Lab Sample ID:	Q1860-35DL		Matrix:	SOIL	
Analytical Method:	SW8082A		% Solid:	80.6	Decanted:
Sample Wt/Vol:	30.02	Units: g	Final Vol:	10000	uL
Soil Aliquot Vol:		uL	Test:	PCB	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	SW3541B				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PP071604.D	5	04/25/25 08:20	04/29/25 11:56	PB167737

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	24.5	UD	24.5	105	ug/kg
11104-28-2	Aroclor-1221	25.0	UD	25.0	105	ug/kg
11141-16-5	Aroclor-1232	23.1	UD	23.1	105	ug/kg
53469-21-9	Aroclor-1242	24.9	UD	24.9	105	ug/kg
12672-29-6	Aroclor-1248	294	D	36.7	105	ug/kg
11097-69-1	Aroclor-1254	665	D	19.9	105	ug/kg
37324-23-5	Aroclor-1262	31.1	UD	31.1	105	ug/kg
11100-14-4	Aroclor-1268	22.3	UD	22.3	105	ug/kg
11096-82-5	Aroclor-1260	504	D	20.0	105	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	26.7		32 - 144	134%	SPK: 20
2051-24-3	Decachlorobiphenyl	33.0		32 - 175	165%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	Weston Solutions, Inc.		Date Collected:	04/21/25	
Project:	RFP 911		Date Received:	04/23/25	
Client Sample ID:	P001-SS030-01		SDG No.:	Q1860	
Lab Sample ID:	Q1860-36		Matrix:	SOIL	
Analytical Method:	SW8082A		% Solid:	83.7	Decanted:
Sample Wt/Vol:	30.06	Units: g	Final Vol:	10000	uL
Soil Aliquot Vol:		uL	Test:	PCB	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	SW3541B				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PO110745.D	1	04/25/25 08:20	04/25/25 19:39	PB167737

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	4.70	U	4.70	20.3	ug/kg
11104-28-2	Aroclor-1221	4.80	U	4.80	20.3	ug/kg
11141-16-5	Aroclor-1232	4.40	U	4.40	20.3	ug/kg
53469-21-9	Aroclor-1242	4.80	U	4.80	20.3	ug/kg
12672-29-6	Aroclor-1248	2100	E	7.10	20.3	ug/kg
11097-69-1	Aroclor-1254	3300	E	3.80	20.3	ug/kg
37324-23-5	Aroclor-1262	6.00	U	6.00	20.3	ug/kg
11100-14-4	Aroclor-1268	4.30	U	4.30	20.3	ug/kg
11096-82-5	Aroclor-1260	3500	E	3.90	20.3	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	14.7		32 - 144	74%	SPK: 20
2051-24-3	Decachlorobiphenyl	24.3		32 - 175	121%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	Weston Solutions, Inc.		Date Collected:	04/21/25	
Project:	RFP 911		Date Received:	04/23/25	
Client Sample ID:	P001-SS030-01DL		SDG No.:	Q1860	
Lab Sample ID:	Q1860-36DL		Matrix:	SOIL	
Analytical Method:	SW8082A		% Solid:	83.7	Decanted:
Sample Wt/Vol:	30.06	Units: g	Final Vol:	10000	uL
Soil Aliquot Vol:		uL	Test:	PCB	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	SW3541B				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PO110797.D	20	04/25/25 08:20	04/28/25 13:24	PB167737

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	94.2	UD	94.2	405	ug/kg
11104-28-2	Aroclor-1221	96.1	UD	96.1	405	ug/kg
11141-16-5	Aroclor-1232	88.7	UD	88.7	405	ug/kg
53469-21-9	Aroclor-1242	95.6	UD	95.6	405	ug/kg
12672-29-6	Aroclor-1248	2800	D	141	405	ug/kg
11097-69-1	Aroclor-1254	4100	D	76.5	405	ug/kg
37324-23-5	Aroclor-1262	120	UD	120	405	ug/kg
11100-14-4	Aroclor-1268	85.8	UD	85.8	405	ug/kg
11096-82-5	Aroclor-1260	3900	D	77.0	405	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	16.8		32 - 144	84%	SPK: 20
2051-24-3	Decachlorobiphenyl	24.4		32 - 175	122%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	Weston Solutions, Inc.		Date Collected:	04/21/25	
Project:	RFP 911		Date Received:	04/23/25	
Client Sample ID:	P001-SS031-01		SDG No.:	Q1860	
Lab Sample ID:	Q1860-37		Matrix:	SOIL	
Analytical Method:	SW8082A		% Solid:	88.9	Decanted:
Sample Wt/Vol:	30.09	Units: g	Final Vol:	10000	uL
Soil Aliquot Vol:		uL	Test:	PCB	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	SW3541B				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PO110746.D	1	04/25/25 08:20	04/25/25 19:58	PB167737

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	4.40	U	4.40	19.1	ug/kg
11104-28-2	Aroclor-1221	4.50	U	4.50	19.1	ug/kg
11141-16-5	Aroclor-1232	4.20	U	4.20	19.1	ug/kg
53469-21-9	Aroclor-1242	4.50	U	4.50	19.1	ug/kg
12672-29-6	Aroclor-1248	6.60	U	6.60	19.1	ug/kg
11097-69-1	Aroclor-1254	7700	E	3.60	19.1	ug/kg
37324-23-5	Aroclor-1262	5.60	U	5.60	19.1	ug/kg
11100-14-4	Aroclor-1268	4.00	U	4.00	19.1	ug/kg
11096-82-5	Aroclor-1260	3.60	U	3.60	19.1	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	18.9		32 - 144	94%	SPK: 20
2051-24-3	Decachlorobiphenyl	27.3		32 - 175	137%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	Weston Solutions, Inc.		Date Collected:	04/21/25	
Project:	RFP 911		Date Received:	04/23/25	
Client Sample ID:	P001-SS031-01DL		SDG No.:	Q1860	
Lab Sample ID:	Q1860-37DL		Matrix:	SOIL	
Analytical Method:	SW8082A		% Solid:	88.9	Decanted:
Sample Wt/Vol:	30.09	Units: g	Final Vol:	10000	uL
Soil Aliquot Vol:		uL	Test:	PCB	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	SW3541B				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PO110838.D	20	04/25/25 08:20	04/29/25 05:16	PB167737

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	88.6	UD	88.6	381	ug/kg
11104-28-2	Aroclor-1221	90.4	UD	90.4	381	ug/kg
11141-16-5	Aroclor-1232	83.4	UD	83.4	381	ug/kg
53469-21-9	Aroclor-1242	89.9	UD	89.9	381	ug/kg
12672-29-6	Aroclor-1248	133	UD	133	381	ug/kg
11097-69-1	Aroclor-1254	9100	ED	72.0	381	ug/kg
37324-23-5	Aroclor-1262	113	UD	113	381	ug/kg
11100-14-4	Aroclor-1268	80.7	UD	80.7	381	ug/kg
11096-82-5	Aroclor-1260	72.4	UD	72.4	381	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	24.8		32 - 144	124%	SPK: 20
2051-24-3	Decachlorobiphenyl	56.6	*	32 - 175	283%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	Weston Solutions, Inc.		Date Collected:	04/21/25	
Project:	RFP 911		Date Received:	04/23/25	
Client Sample ID:	P001-SS031-01DL2		SDG No.:	Q1860	
Lab Sample ID:	Q1860-37DL2		Matrix:	SOIL	
Analytical Method:	SW8082A		% Solid:	88.9	Decanted:
Sample Wt/Vol:	30.09	Units: g	Final Vol:	10000	uL
Soil Aliquot Vol:		uL	Test:	PCB	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	SW3541B				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PO110839.D	100	04/25/25 08:20	04/29/25 05:34	PB167737

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	443	UD	443	1900	ug/kg
11104-28-2	Aroclor-1221	452	UD	452	1900	ug/kg
11141-16-5	Aroclor-1232	417	UD	417	1900	ug/kg
53469-21-9	Aroclor-1242	450	UD	450	1900	ug/kg
12672-29-6	Aroclor-1248	664	UD	664	1900	ug/kg
11097-69-1	Aroclor-1254	9100	D	360	1900	ug/kg
37324-23-5	Aroclor-1262	563	UD	563	1900	ug/kg
11100-14-4	Aroclor-1268	404	UD	404	1900	ug/kg
11096-82-5	Aroclor-1260	362	UD	362	1900	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	0	*	32 - 144	0%	SPK: 20
2051-24-3	Decachlorobiphenyl	0	*	32 - 175	0%	SPK: 20

Comments:

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 P = Indicates >25% difference for detected concentrations between the two GC columns
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.
 () = Laboratory InHouse Limit

Report of Analysis

Client:	Weston Solutions, Inc.		Date Collected:	04/21/25	
Project:	RFP 911		Date Received:	04/23/25	
Client Sample ID:	P001-SS032-01		SDG No.:	Q1860	
Lab Sample ID:	Q1860-38		Matrix:	SOIL	
Analytical Method:	SW8082A		% Solid:	87	Decanted:
Sample Wt/Vol:	30.02	Units: g	Final Vol:	10000	uL
Soil Aliquot Vol:		uL	Test:	PCB	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	SW3541B				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PO110747.D	1	04/25/25 08:20	04/25/25 20:16	PB167737

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	4.50	U	4.50	19.5	ug/kg
11104-28-2	Aroclor-1221	4.60	U	4.60	19.5	ug/kg
11141-16-5	Aroclor-1232	4.30	U	4.30	19.5	ug/kg
53469-21-9	Aroclor-1242	4.60	U	4.60	19.5	ug/kg
12672-29-6	Aroclor-1248	2100	E	6.80	19.5	ug/kg
11097-69-1	Aroclor-1254	5300	E	3.70	19.5	ug/kg
37324-23-5	Aroclor-1262	5.80	U	5.80	19.5	ug/kg
11100-14-4	Aroclor-1268	4.10	U	4.10	19.5	ug/kg
11096-82-5	Aroclor-1260	3.70	U	3.70	19.5	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	20.9		32 - 144	104%	SPK: 20
2051-24-3	Decachlorobiphenyl	28.0		32 - 175	140%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	Weston Solutions, Inc.		Date Collected:	04/21/25	
Project:	RFP 911		Date Received:	04/23/25	
Client Sample ID:	P001-SS032-01DL		SDG No.:	Q1860	
Lab Sample ID:	Q1860-38DL		Matrix:	SOIL	
Analytical Method:	SW8082A		% Solid:	87	Decanted:
Sample Wt/Vol:	30.02	Units: g	Final Vol:	10000	uL
Soil Aliquot Vol:		uL	Test:	PCB	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	SW3541B				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PO110805.D	10	04/25/25 08:20	04/28/25 16:41	PB167737

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	45.4	UD	45.4	195	ug/kg
11104-28-2	Aroclor-1221	46.3	UD	46.3	195	ug/kg
11141-16-5	Aroclor-1232	42.7	UD	42.7	195	ug/kg
53469-21-9	Aroclor-1242	46.1	UD	46.1	195	ug/kg
12672-29-6	Aroclor-1248	1900	D	68.0	195	ug/kg
11097-69-1	Aroclor-1254	4000	ED	36.9	195	ug/kg
37324-23-5	Aroclor-1262	57.7	UD	57.7	195	ug/kg
11100-14-4	Aroclor-1268	41.4	UD	41.4	195	ug/kg
11096-82-5	Aroclor-1260	37.1	UD	37.1	195	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	17.8		32 - 144	89%	SPK: 20
2051-24-3	Decachlorobiphenyl	32.1		32 - 175	160%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	Weston Solutions, Inc.		Date Collected:	04/21/25	
Project:	RFP 911		Date Received:	04/23/25	
Client Sample ID:	P001-SS032-01DL2		SDG No.:	Q1860	
Lab Sample ID:	Q1860-38DL2		Matrix:	SOIL	
Analytical Method:	SW8082A		% Solid:	87	Decanted:
Sample Wt/Vol:	30.02	Units: g	Final Vol:	10000	uL
Soil Aliquot Vol:		uL	Test:	PCB	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	SW3541B				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PO110806.D	50	04/25/25 08:20	04/28/25 16:59	PB167737

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	227	UD	227	976	ug/kg
11104-28-2	Aroclor-1221	231	UD	231	976	ug/kg
11141-16-5	Aroclor-1232	214	UD	214	976	ug/kg
53469-21-9	Aroclor-1242	230	UD	230	976	ug/kg
12672-29-6	Aroclor-1248	2400	D	340	976	ug/kg
11097-69-1	Aroclor-1254	4200	D	184	976	ug/kg
37324-23-5	Aroclor-1262	288	UD	288	976	ug/kg
11100-14-4	Aroclor-1268	207	UD	207	976	ug/kg
11096-82-5	Aroclor-1260	186	UD	186	976	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	0	*	32 - 144	0%	SPK: 20
2051-24-3	Decachlorobiphenyl	0	*	32 - 175	0%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	Weston Solutions, Inc.		Date Collected:	04/21/25	
Project:	RFP 911		Date Received:	04/23/25	
Client Sample ID:	P001-SS033-01		SDG No.:	Q1860	
Lab Sample ID:	Q1860-39		Matrix:	SOIL	
Analytical Method:	SW8082A		% Solid:	87.7	Decanted:
Sample Wt/Vol:	30.05	Units: g	Final Vol:	10000	uL
Soil Aliquot Vol:		uL	Test:	PCB	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	SW3541B				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PO110748.D	1	04/25/25 08:20	04/25/25 20:34	PB167737

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	4.50	U	4.50	19.4	ug/kg
11104-28-2	Aroclor-1221	4.60	U	4.60	19.4	ug/kg
11141-16-5	Aroclor-1232	4.20	U	4.20	19.4	ug/kg
53469-21-9	Aroclor-1242	4.60	U	4.60	19.4	ug/kg
12672-29-6	Aroclor-1248	3900	E	6.70	19.4	ug/kg
11097-69-1	Aroclor-1254	3600	E	3.70	19.4	ug/kg
37324-23-5	Aroclor-1262	5.70	U	5.70	19.4	ug/kg
11100-14-4	Aroclor-1268	4.10	U	4.10	19.4	ug/kg
11096-82-5	Aroclor-1260	3.70	U	3.70	19.4	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	19.1		32 - 144	95%	SPK: 20
2051-24-3	Decachlorobiphenyl	30.5		32 - 175	153%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	Weston Solutions, Inc.		Date Collected:	04/21/25	
Project:	RFP 911		Date Received:	04/23/25	
Client Sample ID:	P001-SS033-01DL		SDG No.:	Q1860	
Lab Sample ID:	Q1860-39DL		Matrix:	SOIL	
Analytical Method:	SW8082A		% Solid:	87.7	Decanted:
Sample Wt/Vol:	30.05	Units: g	Final Vol:	10000	uL
Soil Aliquot Vol:		uL	Test:	PCB	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	SW3541B				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PO110807.D	20	04/25/25 08:20	04/28/25 17:17	PB167737

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	89.9	UD	89.9	387	ug/kg
11104-28-2	Aroclor-1221	91.8	UD	91.8	387	ug/kg
11141-16-5	Aroclor-1232	84.7	UD	84.7	387	ug/kg
53469-21-9	Aroclor-1242	91.3	UD	91.3	387	ug/kg
12672-29-6	Aroclor-1248	4900	D	135	387	ug/kg
11097-69-1	Aroclor-1254	4000	D	73.1	387	ug/kg
37324-23-5	Aroclor-1262	114	UD	114	387	ug/kg
11100-14-4	Aroclor-1268	82.0	UD	82.0	387	ug/kg
11096-82-5	Aroclor-1260	73.5	UD	73.5	387	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	22.0		32 - 144	110%	SPK: 20
2051-24-3	Decachlorobiphenyl	36.2	*	32 - 175	181%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	Weston Solutions, Inc.		Date Collected:	04/21/25	
Project:	RFP 911		Date Received:	04/23/25	
Client Sample ID:	P001-SS034-01		SDG No.:	Q1860	
Lab Sample ID:	Q1860-40		Matrix:	SOIL	
Analytical Method:	SW8082A		% Solid:	87.9	Decanted:
Sample Wt/Vol:	30.04	Units: g	Final Vol:	10000	uL
Soil Aliquot Vol:		uL	Test:	PCB	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	SW3541B				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PO110749.D	1	04/25/25 08:20	04/25/25 20:53	PB167737

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	4.50	U	4.50	19.3	ug/kg
11104-28-2	Aroclor-1221	4.60	U	4.60	19.3	ug/kg
11141-16-5	Aroclor-1232	4.20	U	4.20	19.3	ug/kg
53469-21-9	Aroclor-1242	4.60	U	4.60	19.3	ug/kg
12672-29-6	Aroclor-1248	1700	E	6.70	19.3	ug/kg
11097-69-1	Aroclor-1254	2700	E	3.60	19.3	ug/kg
37324-23-5	Aroclor-1262	5.70	U	5.70	19.3	ug/kg
11100-14-4	Aroclor-1268	4.10	U	4.10	19.3	ug/kg
11096-82-5	Aroclor-1260	3.70	U	3.70	19.3	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	13.7		32 - 144	69%	SPK: 20
2051-24-3	Decachlorobiphenyl	26.8		32 - 175	134%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	Weston Solutions, Inc.		Date Collected:	04/21/25	
Project:	RFP 911		Date Received:	04/23/25	
Client Sample ID:	P001-SS034-01DL		SDG No.:	Q1860	
Lab Sample ID:	Q1860-40DL		Matrix:	SOIL	
Analytical Method:	SW8082A		% Solid:	87.9	Decanted:
Sample Wt/Vol:	30.04	Units: g	Final Vol:	10000	uL
Soil Aliquot Vol:		uL	Test:	PCB	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	SW3541B				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PO110808.D	20	04/25/25 08:20	04/28/25 17:36	PB167737

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	89.8	UD	89.8	386	ug/kg
11104-28-2	Aroclor-1221	91.6	UD	91.6	386	ug/kg
11141-16-5	Aroclor-1232	84.5	UD	84.5	386	ug/kg
53469-21-9	Aroclor-1242	91.1	UD	91.1	386	ug/kg
12672-29-6	Aroclor-1248	2600	D	135	386	ug/kg
11097-69-1	Aroclor-1254	3600	D	72.9	386	ug/kg
37324-23-5	Aroclor-1262	114	UD	114	386	ug/kg
11100-14-4	Aroclor-1268	81.8	UD	81.8	386	ug/kg
11096-82-5	Aroclor-1260	73.4	UD	73.4	386	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	21.6		32 - 144	108%	SPK: 20
2051-24-3	Decachlorobiphenyl	50.8	*	32 - 175	254%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	Weston Solutions, Inc.		Date Collected:	04/21/25	
Project:	RFP 911		Date Received:	04/23/25	
Client Sample ID:	P001-SS035-01		SDG No.:	Q1860	
Lab Sample ID:	Q1860-41		Matrix:	SOIL	
Analytical Method:	SW8082A		% Solid:	94	Decanted:
Sample Wt/Vol:	30.01	Units: g	Final Vol:	10000	uL
Soil Aliquot Vol:		uL	Test:	PCB	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	SW3541B				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PO110758.D	1	04/25/25 08:20	04/26/25 01:34	PB167737

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	4.20	U	4.20	18.1	ug/kg
11104-28-2	Aroclor-1221	4.30	U	4.30	18.1	ug/kg
11141-16-5	Aroclor-1232	4.00	U	4.00	18.1	ug/kg
53469-21-9	Aroclor-1242	4.30	U	4.30	18.1	ug/kg
12672-29-6	Aroclor-1248	2500	E	6.30	18.1	ug/kg
11097-69-1	Aroclor-1254	3900	E	3.40	18.1	ug/kg
37324-23-5	Aroclor-1262	5.30	U	5.30	18.1	ug/kg
11100-14-4	Aroclor-1268	3.80	U	3.80	18.1	ug/kg
11096-82-5	Aroclor-1260	3.40	U	3.40	18.1	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	21.9		32 - 144	109%	SPK: 20
2051-24-3	Decachlorobiphenyl	36.2	*	32 - 175	181%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	Weston Solutions, Inc.		Date Collected:	04/21/25	
Project:	RFP 911		Date Received:	04/23/25	
Client Sample ID:	P001-SS035-01DL		SDG No.:	Q1860	
Lab Sample ID:	Q1860-41DL		Matrix:	SOIL	
Analytical Method:	SW8082A		% Solid:	94	Decanted:
Sample Wt/Vol:	30.01	Units: g	Final Vol:	10000	uL
Soil Aliquot Vol:		uL	Test:	PCB	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	SW3541B				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PO110809.D	10	04/25/25 08:20	04/28/25 17:53	PB167737

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	42.0	UD	42.0	181	ug/kg
11104-28-2	Aroclor-1221	42.9	UD	42.9	181	ug/kg
11141-16-5	Aroclor-1232	39.6	UD	39.6	181	ug/kg
53469-21-9	Aroclor-1242	42.6	UD	42.6	181	ug/kg
12672-29-6	Aroclor-1248	2900	D	63.0	181	ug/kg
11097-69-1	Aroclor-1254	4100	ED	34.1	181	ug/kg
37324-23-5	Aroclor-1262	53.4	UD	53.4	181	ug/kg
11100-14-4	Aroclor-1268	38.3	UD	38.3	181	ug/kg
11096-82-5	Aroclor-1260	34.4	UD	34.4	181	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	23.2		32 - 144	116%	SPK: 20
2051-24-3	Decachlorobiphenyl	45.8	*	32 - 175	229%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	Weston Solutions, Inc.		Date Collected:	04/21/25	
Project:	RFP 911		Date Received:	04/23/25	
Client Sample ID:	P001-SS035-01DL2		SDG No.:	Q1860	
Lab Sample ID:	Q1860-41DL2		Matrix:	SOIL	
Analytical Method:	SW8082A		% Solid:	94	Decanted:
Sample Wt/Vol:	30.01	Units: g	Final Vol:	10000	uL
Soil Aliquot Vol:		uL	Test:	PCB	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	SW3541B				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PO110810.D	50	04/25/25 08:20	04/28/25 18:12	PB167737

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	210	UD	210	904	ug/kg
11104-28-2	Aroclor-1221	214	UD	214	904	ug/kg
11141-16-5	Aroclor-1232	198	UD	198	904	ug/kg
53469-21-9	Aroclor-1242	213	UD	213	904	ug/kg
12672-29-6	Aroclor-1248	3200	D	315	904	ug/kg
11097-69-1	Aroclor-1254	4200	DP	171	904	ug/kg
37324-23-5	Aroclor-1262	267	UD	267	904	ug/kg
11100-14-4	Aroclor-1268	191	UD	191	904	ug/kg
11096-82-5	Aroclor-1260	172	UD	172	904	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	0	*	32 - 144	0%	SPK: 20
2051-24-3	Decachlorobiphenyl	0	*	32 - 175	0%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Report of Analysis

Client:	Weston Solutions, Inc.		Date Collected:	04/21/25	
Project:	RFP 911		Date Received:	04/23/25	
Client Sample ID:	P001-SS036-01		SDG No.:	Q1860	
Lab Sample ID:	Q1860-42		Matrix:	SOIL	
Analytical Method:	SW8082A		% Solid:	94.4	Decanted:
Sample Wt/Vol:	30.07	Units: g	Final Vol:	10000	uL
Soil Aliquot Vol:		uL	Test:	PCB	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	SW3541B				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PP071539.D	1	04/25/25 08:20	04/28/25 11:32	PB167737

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	4.20	U	4.20	18.0	ug/kg
11104-28-2	Aroclor-1221	4.30	U	4.30	18.0	ug/kg
11141-16-5	Aroclor-1232	3.90	U	3.90	18.0	ug/kg
53469-21-9	Aroclor-1242	4.20	U	4.20	18.0	ug/kg
12672-29-6	Aroclor-1248	6.30	U	6.30	18.0	ug/kg
11097-69-1	Aroclor-1254	19.6	P	3.40	18.0	ug/kg
37324-23-5	Aroclor-1262	5.30	U	5.30	18.0	ug/kg
11100-14-4	Aroclor-1268	3.80	U	3.80	18.0	ug/kg
11096-82-5	Aroclor-1260	3.40	U	3.40	18.0	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	18.2		32 - 144	91%	SPK: 20
2051-24-3	Decachlorobiphenyl	16.4		32 - 175	82%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

LAB CHRONICLE

OrderID:	Q1860	OrderDate:	4/23/2025 10:15:04 AM
Client:	Weston Solutions, Inc.	Project:	RFP 911
Contact:	Smita Sumbaly	Location:	L51

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q1860-01	P001-SS001-01	SOIL	PCB	8082A	04/16/25	04/24/25	04/24/25	04/23/25
Q1860-04	P001-SS002-01	SOIL	PCB	8082A	04/16/25	04/24/25	04/24/25	04/23/25
Q1860-05	P001-SS002-02	SOIL	PCB	8082A	04/16/25	04/24/25	04/24/25	04/23/25
Q1860-06	P001-SS003-01	SOIL	PCB	8082A	04/16/25	04/24/25	04/24/25	04/23/25
Q1860-07	P001-SS004-01	SOIL	PCB	8082A	04/16/25	04/24/25	04/24/25	04/23/25
Q1860-08	P001-SS005-01	SOIL	PCB	8082A	04/16/25	04/24/25	04/24/25	04/23/25
Q1860-09	P001-SS006-01	SOIL	PCB	8082A	04/16/25	04/24/25	04/30/25	04/23/25
Q1860-09DL	P001-SS006-01DL	SOIL	PCB	8082A	04/16/25	04/24/25	04/30/25	04/23/25
Q1860-10	P001-SS007-01	SOIL	PCB	8082A	04/16/25	04/24/25	04/24/25	04/23/25
Q1860-10DL	P001-SS007-01DL	SOIL	PCB	8082A	04/16/25	04/24/25	04/24/25	04/23/25
Q1860-11	P001-SS008-01	SOIL	PCB	8082A	04/16/25	04/24/25	04/24/25	04/23/25
Q1860-11DL	P001-SS008-01DL	SOIL			04/16/25			04/23/25

LAB CHRONICLE

			PCB	8082A		04/24/25	04/24/25	
Q1860-12	P001-SS009-01	SOIL			04/17/25			04/23/25
			PCB	8082A		04/24/25	04/24/25	
Q1860-12DL	P001-SS009-01DL	SOIL			04/17/25			04/23/25
			PCB	8082A		04/24/25	04/25/25	
Q1860-13	P001-SS010-01	SOIL			04/17/25			04/23/25
			PCB	8082A		04/24/25	04/25/25	
Q1860-14	P001-SS011-01	SOIL			04/17/25			04/23/25
			PCB	8082A		04/24/25	04/24/25	
Q1860-15	P001-SS012-01	SOIL			04/17/25			04/23/25
			PCB	8082A		04/24/25	04/24/25	
Q1860-15DL	P001-SS012-01DL	SOIL			04/17/25			04/23/25
			PCB	8082A		04/24/25	04/25/25	
Q1860-16	P001-SS013-01	SOIL			04/17/25			04/23/25
			PCB	8082A		04/24/25	04/24/25	
Q1860-16DL	P001-SS013-01DL	SOIL			04/17/25			04/23/25
			PCB	8082A		04/24/25	04/25/25	
Q1860-17	P001-SS014-01	SOIL			04/17/25			04/23/25
			PCB	8082A		04/24/25	04/24/25	
Q1860-18	P001-SS015-01	SOIL			04/17/25			04/23/25
			PCB	8082A		04/24/25	04/24/25	
Q1860-19	P001-SS016-01	SOIL			04/17/25			04/23/25
			PCB	8082A		04/25/25	04/25/25	
Q1860-19DL	P001-SS016-01DL	SOIL			04/17/25			04/23/25
			PCB	8082A		04/25/25	04/28/25	
Q1860-22	P001-SS017-01	SOIL			04/17/25			04/23/25
			PCB	8082A		04/24/25	04/28/25	
Q1860-22DL	P001-SS017-01DL	SOIL			04/17/25			04/23/25
			PCB	8082A		04/24/25	04/29/25	

LAB CHRONICLE

Q1860-23	P001-SS017-02	SOIL	PCB	8082A	04/17/25	04/24/25	04/28/25	04/23/25
Q1860-23DL	P001-SS017-02DL	SOIL	PCB	8082A	04/17/25	04/24/25	04/29/25	04/23/25
Q1860-24	P001-SS018-01	SOIL	PCB	8082A	04/21/25	04/24/25	04/28/25	04/23/25
Q1860-25	P001-SS019-01	SOIL	PCB	8082A	04/21/25	04/24/25	04/28/25	04/23/25
Q1860-25DL	P001-SS019-01DL	SOIL	PCB	8082A	04/21/25	04/24/25	04/29/25	04/23/25
Q1860-26	P001-SS020-01	SOIL	PCB	8082A	04/21/25	04/25/25	04/25/25	04/23/25
Q1860-26DL	P001-SS020-01DL	SOIL	PCB	8082A	04/21/25	04/25/25	04/28/25	04/23/25
Q1860-27	P001-SS021-01	SOIL	PCB	8082A	04/21/25	04/25/25	04/25/25	04/23/25
Q1860-27DL	P001-SS021-01DL	SOIL	PCB	8082A	04/21/25	04/25/25	04/28/25	04/23/25
Q1860-28	P001-SS022-01	SOIL	PCB	8082A	04/21/25	04/25/25	04/25/25	04/23/25
Q1860-29	P001-SS023-01	SOIL	PCB	8082A	04/21/25	04/25/25	04/25/25	04/23/25
Q1860-29DL	P001-SS023-01DL	SOIL	PCB	8082A	04/21/25	04/25/25	04/28/25	04/23/25
Q1860-30	P001-SS024-01	SOIL	PCB	8082A	04/21/25	04/25/25	04/25/25	04/23/25
Q1860-31	P001-SS025-01	SOIL	PCB	8082A	04/21/25	04/25/25	04/25/25	04/23/25
Q1860-31DL	P001-SS025-01DL	SOIL			04/21/25			04/23/25

LAB CHRONICLE

			PCB	8082A	04/25/25	04/28/25	
Q1860-32	P001-SS026-01	SOIL			04/21/25		04/23/25
			PCB	8082A	04/25/25	04/25/25	
Q1860-32DL	P001-SS026-01DL	SOIL			04/21/25		04/23/25
			PCB	8082A	04/25/25	04/28/25	
Q1860-33	P001-SS027-01	SOIL			04/21/25		04/23/25
			PCB	8082A	04/25/25	04/28/25	
Q1860-34	P001-SS028-01	SOIL			04/21/25		04/23/25
			PCB	8082A	04/25/25	04/25/25	
Q1860-34DL	P001-SS028-01DL	SOIL			04/21/25		04/23/25
			PCB	8082A	04/25/25	04/29/25	
Q1860-35	P001-SS029-01	SOIL			04/21/25		04/23/25
			PCB	8082A	04/25/25	04/28/25	
Q1860-35DL	P001-SS029-01DL	SOIL			04/21/25		04/23/25
			PCB	8082A	04/25/25	04/29/25	
Q1860-36	P001-SS030-01	SOIL			04/21/25		04/23/25
			PCB	8082A	04/25/25	04/25/25	
Q1860-36DL	P001-SS030-01DL	SOIL			04/21/25		04/23/25
			PCB	8082A	04/25/25	04/28/25	
Q1860-37	P001-SS031-01	SOIL			04/21/25		04/23/25
			PCB	8082A	04/25/25	04/25/25	
Q1860-37DL	P001-SS031-01DL	SOIL			04/21/25		04/23/25
			PCB	8082A	04/25/25	04/29/25	
Q1860-37DL 2	P001-SS031-01DL2	SOIL			04/21/25		04/23/25
			PCB	8082A	04/25/25	04/29/25	
Q1860-38	P001-SS032-01	SOIL			04/21/25		04/23/25
			PCB	8082A	04/25/25	04/25/25	
Q1860-38DL	P001-SS032-01DL	SOIL			04/21/25		04/23/25

LAB CHRONICLE

			PCB	8082A	04/25/25	04/28/25	
Q1860-38DL	P001-SS032-01DL2	SOIL			04/21/25		04/23/25
2							
			PCB	8082A	04/25/25	04/28/25	
Q1860-39	P001-SS033-01	SOIL			04/21/25		04/23/25
			PCB	8082A	04/25/25	04/25/25	
Q1860-39DL	P001-SS033-01DL	SOIL			04/21/25		04/23/25
			PCB	8082A	04/25/25	04/28/25	
Q1860-40	P001-SS034-01	SOIL			04/21/25		04/23/25
			PCB	8082A	04/25/25	04/25/25	
Q1860-40DL	P001-SS034-01DL	SOIL			04/21/25		04/23/25
			PCB	8082A	04/25/25	04/28/25	
Q1860-41	P001-SS035-01	SOIL			04/21/25		04/23/25
			PCB	8082A	04/25/25	04/26/25	
Q1860-41DL	P001-SS035-01DL	SOIL			04/21/25		04/23/25
			PCB	8082A	04/25/25	04/28/25	
Q1860-41DL	P001-SS035-01DL2	SOIL			04/21/25		04/23/25
2							
			PCB	8082A	04/25/25	04/28/25	
Q1860-42	P001-SS036-01	SOIL			04/21/25		04/23/25
			PCB	8082A	04/25/25	04/28/25	