

284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CASE NARRATIVE

Tully Environmental, Inc

Project Name: Transfer Station-SPDES

Project # N/A

Chemtech Project # Q1863 Test Name: Metals Group 10

A. Number of Samples and Date of Receipt:

4 Water samples were received on 04/23/2025.

B. Parameters:

According to the Chain of Custody document, the following analyses were requested: Metals Group 10 and TSS. This data package contains results for Metals Group 10.

C. Analytical Techniques:

The analysis and digestion of Metals Group 10 was based on method 200.7.

D. QA/ QC Samples:

The Holding Times were met for all analysis.

The Blank Spike met requirements for all samples.

The Duplicate analysis met criteria for all samples.

The Matrix Spike analysis met criteria for all samples.

The Matrix Spike Duplicate analysis met criteria for all samples.

The Blank analysis did not indicate the presence of lab contamination.

The Calibration met the requirements.

The Serial Dilution met criteria for all samples.

E. Additional Comments:

I certify that the data package is in compliance with the terms and conditions of the
contract, both technically and for completeness, for other than the conditions detailed
above. The laboratory manager or his designee, as verified by the following signature has
authorized release of the data contained in this hard copy data package.

a •			
Signature			
<i>-</i>	 	 	