

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tully Environmental, Inc **SDG No.:** Q1863
Contract: TULL01 **Lab Code:** CHEM **Case No.:** Q1863 **SAS No.:** Q1863
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Copper	489	510	96	95 - 105	P	04/23/2025	15:29	LB135538
	Iron	10500	10000	105	95 - 105	P	04/23/2025	15:29	LB135538
	Lead	972	1000	97	95 - 105	P	04/23/2025	15:29	LB135538

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tully Environmental, Inc **SDG No.:** Q1863
Contract: TULL01 **Lab Code:** CHEM **Case No.:** Q1863 **SAS No.:** Q1863
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Copper	22.3	20.0	111	80 - 120	P	04/23/2025	15:44	LB135538
	Iron	106	100	106	80 - 120	P	04/23/2025	15:44	LB135538
	Lead	12.5	12.0	104	80 - 120	P	04/23/2025	15:44	LB135538

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tully Environmental, Inc **SDG No.:** Q1863
Contract: TULL01 **Lab Code:** CHEM **Case No.:** Q1863 **SAS No.:** Q1863
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Copper	1260	1250	100	90 - 110	P	04/23/2025	16:15	LB135538
	Iron	5160	5000	103	90 - 110	P	04/23/2025	16:15	LB135538
	Lead	4970	5000	100	90 - 110	P	04/23/2025	16:15	LB135538
CCV02	Copper	1250	1250	100	90 - 110	P	04/23/2025	17:05	LB135538
	Iron	4950	5000	99	90 - 110	P	04/23/2025	17:05	LB135538
	Lead	4930	5000	99	90 - 110	P	04/23/2025	17:05	LB135538
CCV03	Copper	1260	1250	101	90 - 110	P	04/23/2025	17:55	LB135538
	Iron	5200	5000	104	90 - 110	P	04/23/2025	17:55	LB135538
	Lead	5020	5000	100	90 - 110	P	04/23/2025	17:55	LB135538
CCV04	Copper	1250	1250	100	90 - 110	P	04/23/2025	18:38	LB135538
	Iron	5070	5000	102	90 - 110	P	04/23/2025	18:38	LB135538
	Lead	4960	5000	99	90 - 110	P	04/23/2025	18:38	LB135538