

Metals
- 4 -
INTERFERENCE CHECK SAMPLE

Client: Tully Environmental, Inc **SDG No.:** Q1863
Contract: TULL01 **Lab Code:** CHEM **Case No.:** Q1863 **SAS No.:** Q1863
ICS Source: EPA **Instrument ID:** P4

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Copper	5.71	2.0	286	-18	22	04/23/2025	15:58	LB135538
	Iron	94900	101000	94	85600	116500	04/23/2025	15:58	LB135538
	Lead	-1.66			-12	12	04/23/2025	15:58	LB135538
ICSAB01	Copper	480	511	94	434	588	04/23/2025	16:02	LB135538
	Iron	97700	99300	98	84400	114500	04/23/2025	16:02	LB135538
	Lead	45.3	49.0	92	37	61	04/23/2025	16:02	LB135538
ICSA	Copper	507	2.0	25350	-18	22	04/23/2025	16:06	LB135538
	Iron	98900	101000	98	85600	116500	04/23/2025	16:06	LB135538
	Lead	29.5			-12	12	04/23/2025	16:06	LB135538
ICSAB	Copper	10.4	511	2	434	588	04/23/2025	16:10	LB135538
	Iron	103000	99300	104	84400	114500	04/23/2025	16:10	LB135538
	Lead	-29.0	49.0	59	37	61	04/23/2025	16:10	LB135538