



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CASE NARRATIVE

Aramark Uniforms

Project Name: Monthly 2025

Project # N/A

Chemtech Project # Q1864

Test Name: TPH,BOD5,TSS

A. Number of Samples and Date of Receipt:

2 Water samples were received on 04/23/2025.

B. Parameters:

According to the Chain of Custody document, the following analyses were requested: BOD5, TPH and TSS. This data package contains results for TPH,BOD5,TSS.

C. Analytical Techniques:

The analysis of TPH was based on method 1664A, The analysis of TSS was based on method SM2540 D and The analysis of BOD5 was based on method SM5210 B.

D. QA/ QC Samples:

The Holding Times were met for all analysis.

The Blank Spike met requirements for all samples.

The Duplicate analysis met criteria for all samples.

The Blank analysis did not indicate the presence of lab contamination.

The Calibration met the requirements.

E. Additional Comments:

As per method 1664A, MS/MSD is required to be performed with the sample analysis. However, Lab did not receive sufficient volume to perform the MS/MSD therefore MS/MSD were not performed for this project.

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature_____