

Instrument ID: **BNA_M**

Daily Analysis Runlog For Sequence/QC Batch ID # BM042825

Review By	Rahul	Review On	4/29/2025 8:56:26 AM		
Supervise By	Jagrut	Supervise On	4/29/2025 11:57:30 AM		
SubDirectory	BM042825	HP Acquire Method	BNA_M	HP Processing Method	BM042825
STD. NAME		STD REF.#			
Tune/Reschk		SP6757			
Initial Calibration Stds		SP6722,SP6723,SP6724,SP6725,SP6726,SP6727,SP6728,SP6729			
CCC		SP6725			
Internal Standard/PEM		S12661,10ul/1000ul sample			
ICV/I.BLK		SP6770			
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	DFTPP	BM050023.D	28 Apr 2025 11:46	RC/JU	Ok
2	SSTDICC2.5	BM050024.D	28 Apr 2025 12:30	RC/JU	Ok
3	SSTDICC005	BM050025.D	28 Apr 2025 13:09	RC/JU	Ok,M
4	SSTDICC010	BM050026.D	28 Apr 2025 13:48	RC/JU	Ok,M
5	SSTDICC020	BM050027.D	28 Apr 2025 14:27	RC/JU	Ok,M
6	SSTDICCC040	BM050028.D	28 Apr 2025 15:06	RC/JU	Ok
7	SSTDICC050	BM050029.D	28 Apr 2025 15:45	RC/JU	Ok
8	SSTDICC060	BM050030.D	28 Apr 2025 16:24	RC/JU	Ok
9	SSTDICC080	BM050031.D	28 Apr 2025 17:04	RC/JU	Ok
10	SSTDICV040	BM050032.D	28 Apr 2025 17:43	RC/JU	Ok,M
11	PB167739BL	BM050033.D	28 Apr 2025 19:40	RC/JU	Not Ok

M : Manual Integration

Instrument ID: **BNA_M**

Daily Analysis Runlog For Sequence/QC Batch ID # BM042925

Review By	Rahul	Review On	4/30/2025 11:38:15 AM		
Supervise By	Jagrut	Supervise On	4/30/2025 12:40:09 PM		
SubDirectory	BM042925	HP Acquire Method	BNA_M	HP Processing Method	BM042825
STD. NAME		STD REF.#			
Tune/Reschk		SP6757			
Initial Calibration Stds		SP6722,SP6723,SP6724,SP6725,SP6726,SP6727,SP6728,SP6729			
CCC		SP6725			
Internal Standard/PEM		S12661,10ul/1000ul sample			
ICV/I.BLK		SP6770			
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	DFTPP	BM050034.D	29 Apr 2025 08:34	RC/JU	Ok
2	SSTDCCC040	BM050035.D	29 Apr 2025 09:13	RC/JU	Ok
3	PB167729BL	BM050036.D	29 Apr 2025 09:52	RC/JU	Ok
4	Q1502-06DL	BM050037.D	29 Apr 2025 11:00	RC/JU	Ok
5	Q1870-03	BM050038.D	29 Apr 2025 11:39	RC/JU	Ok
6	Q1870-03DL	BM050039.D	29 Apr 2025 12:19	RC/JU	Not Ok
7	PB167729BS	BM050040.D	29 Apr 2025 12:58	RC/JU	Ok

M : Manual Integration